

Stratton VA - Albany, New York
Renovate for Sterile Processing Services
VA Project #: 528A8-08-847
EQUIPMENT LIST

Item	Description	Furnished By	Installed By	Relocated	New	Remarks
A	Computer & Monitor	VA	VA			
B	Computer Work Station Cart	VA	VA			
C	3-IN-1 Copier	VA	VA			
D	Desktop Printer	VA	VA			
E	Barcode Reader	VA	VA			
F	Desk	VA	VA			
G	Chair/Stool	VA	VA			
H	Not Used	-	-			
J	Table and Chairs	VA	VA			
K	Lateral File Cabinet	VA	VA			
L	Label Printer	VA	VA			
M	Waste Receptacle	VA	VA			
N	Coffee Maker	VA	VA			
P	Microwave	VA	VA			
Q	Full Size Refrigerator	VA	VA			
R	Fabric Wrapped Tack Panel	GC	GC		3	
S	Flat Panel Display (Smartboard)	VA	VA			
T	Coat Rack/Hooks	GC	GC		1	
U	Wall Clock	VA	VA			
V	Lockers	GC	GC		X	see drawings for quantity/locations
W	Locker Bench	GC	GC		5	
X	Adjustable Height Stainless Steel Triple Clean Up Sink	VA	GC		1	
Y	Stainless Steel Work Table	VA	VA			
Z	Washer / Disinfector - Getinge 88 Turbo	VA	GC		2	GC hire Manuf. Authorized installer
AA	Washer Loading Dolly	VA	VA			
BB	Washer Receiving Tray	VA	VA			
CC	Cart Washer / Disinfector - Getinge 9128	VA	GC		1	GC hire Manuf. Authorized installer
DD	Access Ramp	VA	GC		2	GC hire Manuf. Authorized installer
EE	Ultrasonic Cleaner - Getinge 2460-UC	VA	GC	1		
FF	Ultrasonic Cleaner - Branson B5510-DTH or equal	VA	VA			
GG	Dental Instrument Oiler - Assistina 301 or equal	VA	GC	3		
HH	Nitrogen Tank	VA	VA			
JJ	Getinge Model 8001 Pass Through Window / Counter	VA	GC		2	
KK	Personal Protective Equipment Storage Rack	VA	VA			
LL	Endoscope Pre-Processing Sink - 83-S6 or equal	VA	GC		3	
MM	Sonic Irrigator - PCF System or equal	VA	GC	1		GC hire Manuf. Authorized installer
NN	Vapor Control System (Wall Mtd.) - PCI G14FH or equal	VA	VA	1	1	
PP	Scope Reprocessor/Medivator	VA	GC	1	2	GC hire Manuf. Authorized installer
QQ	Scope Storage Cabinet (Wall Mounted)	VA	GC		1	
RR	Emergency Shower	GC	GC		1	
SS	Portable Emergency Eyewash Station	VA	GC		5	
TT	Soiled Linen Container	VA	VA			
UU	Infectious Waste Disposal	VA	VA			
VV	Automatic Detergent Dispensing System	VA	Vendor			GC hire Manuf. Authorized installer
WW	R.O. Water System	VA	Vendor	1		GC hire Manuf. Authorized installer
XX	Cardboard Recycling Cart	VA	VA			

Item	Description	Furnished By	Installed By	Relocated	New	Remarks
YY	Pallet	VA	VA			
ZZ	Wire Storage Rack	VA	VA			
A1	Wire Storage Rack on Wheels	VA	VA			
A2	Steam Sterilizer - Getinge 733HC or equal	VA	GC	2	1	GC hire Manuf. Authorized installer
A3	Steam Sterilizer - Getinge 533HC or equal	VA	GC	1		GC hire Manuf. Authorized installer
A4	Sterilizer Load / Unload Carriage	VA	VA			
A5	Compact Sterilizer - Sterrad NX or equal	VA	GC	2	1	
A6	Sterilizer - Sterrad 100NX or equal	VA	GC	1		GC hire Manuf. Authorized installer to inspect
A7	Sterile-Dryer - Olympic Model 44 or equal	VA	VA	2		
A8	Adjustable Height Instrument Handling Work Station	GC	GC		6	
A9	Instrument Storage Cabinet (5 Drawer Nom.)	VA	VA			
A10	Instrument Storage Cabinet (10 Drawer Nom.)	VA	VA			
A11	Wrap Rack (Wheeled)	VA	VA			
A12	Sterile / Prep Sealer Equipment	VA	VA			
A13	Auto-Reader	VA	VA			
A14	Incubator - Sterrad or equal	VA	VA			
A15	Gus Soak Station (Wall Mtd.) - PCI G14KA or equal	VA	GC		1	
A16	Gus Soak Station (Wall Mtd.) - PCI G14ENT-3 or equal	VA	GC		1	
A17	Cleanascope Transport Cart or equal	VA	VA			
A18	Scrubex Machine or equal	VA	VA			
A19	Glove Dispensing Box	VA	GC		X	see drawings for quantity/locations
A20	Trophon (Probe HLD) or equal	VA	GC		2	
A21	Binder Storage Cabinet	VA	VA			
A22	Markerboard	VA	VA			
A23	TBJ Inc. Model PT. 44-38 Pass Through Window	GC	GC		1	
A24	Case Carts	VA	VA			
A25	Code Carts	VA	VA			
A26	Suction Pump Carts	VA	VA			
A27	Medi-Therm Carts	VA	VA			

Notes:

All items noted as being relocated are currently in the existing basement SPS Department area except those noted with (*). Items noted with (*) are to be relocated from existing Scope Processing Rooms on higher floors. Relocations to be coordinated with the VA and with phasing schedules to maintain operations in the SPS.

Contractor to retain an authorized manufacturer's representative to relocate item WW in accordance with a phasing plan to keep existing space operational. Authorized installer to supply a new storage tank and pump unit to permit existing system to remain operational.

Contractor to retain an authorized installer for items A2, A3 to be relocated in accordance with a phasing plan to keep existing space operational. Work includes temporary walls and maintaining sterile environment in existing area to permit existing system to remain operational. A total of two sterilizers must remain operational at all times.

An authorized manufacturer's representative and installer for the above noted equipment is Mr. Jeff Carrig, Getinge Albany Office, 518-281-0435, jeff.carrig@getingeusa.com. A second contact is Tom Wescott, 716-863-5909, tom.wescott@getinge.com.

An authorized manufacturer's representative and installer for item MM is Medisafe, Service Group, 877-444-6334, Roy Gonzales or Edgar Zayas, email - support@medisafeamerica.com with Roy Gonzales in subject line.

An authorized manufacturer's representative and installer for item PP is Medivator, Ray Wiseman 612-554-9384, rwiseman@medivators.com

Manufacturer's representative before energizing to avoid serious damage to the machine. Retain Manufacturer's inspection service by contacting Chris Sica, 518-879-5105 for inspection.