

621-B30012 - XR RAD, VAMC MTN HOME, TN, REDACTED BENCHMARK

VISN 9 Standardized Radiology Equipment Project

Basic Digital Radiography Room Generic Specs

Technical Requirements:

- Vendor Imaging Approved List
- Fully automatic tube and detector Digital system
- Minimum Table weight of 450lbs with all movements in any table position
- 80 KW High Frequency generator
- KV 40 - 150
- upright and horizontal detector orientations
- wireless detectors with extra batteries and recharging system (at least one wireless and one fixed configuration or dual wireless) specify sizes of detectors in quote
- tethered cables available for backup of wireless detector system
- imaging table- three axis
- Operators console Flat Panel monitors
- PC acquisition station
- General purpose system
- Overhead tube suspension system
- Patient Hand Grips for Upright system
- Repeat and Reject Analysis
- Protective weight bearing wireless detector cover (quote should describe size & wt limit)
- DICOM 3.0 or higher compatibility
- Interface with PACS
- Patient Radiation dose captured as a downloadable data element
- Uninterruptible Power Supply
- Stitching including hardware and software (optional)
- Subtraction Chest Double Exposure or software (Lung and Bone Windows on Chest exams) Optional
- Tomography (Optional)
- Wireless detector damage insurance (optional)

Training

- All-inclusive off site training for multiple techs for minimum 1 week each from each facility (optional - please quote unit price for each tech)
- All-inclusive off site training for minimum of 1 or 2 Biomed Tech from each facility including any necessary prerequisites (optional- please quote for 1 and for 2)
- Additional 2 weeks on site training, separated by 1 month (mandatory)
- Physician Training (optional)

Service

- 24/7/365 access to Vendor Help Center (mandatory)
- Remote Service Capability (preferred)
- Full access to service diagnostic software for service, maintenance, trouble-shooting (preferred)
- Monthly in person site visits to every VISN facility by a Field Engineer (optional)
- 30 minute Field Engineer response via telephone (mandatory)
- 1-2 hour on-site presence by Field Engineer (highly desirable)
- Two (2) copies of operator's instruction manuals
- Two (2) copies of complete technical service manuals including detailed troubleshooting guides, necessary diagnostic software, service keys, schematic diagrams, and parts lists
- Two (2) copies of a system manager's manual outlining back-up procedures, managing privilege group limits, routine tasks, etc.

Price/Installation

- Additional discount below VA IDIQ price if entire VISN quantity purchased from single Vendor (optional)
- Standard Warranty (mandatory)
- 2 Service manuals per site (mandatory)

Unique Quote/Estim for each Site

- Trade in for existing RF equipment, including De-installation and Removal (optional)
- Turn-key minor construction site prep (optional)

Note: "Optional" listed by a line item indicates a desirable Option – it is not mandatory this be included as an Option on all units. Must be listed as a separate Line Item on every Vendor's bid, even if they do not offer/bid this option.