

Line #	Description	Qty
1	Allura Xper FD20 C Rel. 8.1	1

Allura Xper FD20 monoplane system is a state of art X-ray imaging system that can be customized to support a wide range of applications including peripheral, abdominal, cerebral, thoracic, cardiac and non-vascular interventional and diagnostic procedures.

The Allura Xper FD20 system uses an integrated single-host concept. The system is comprised of five functional building blocks: Geometry, X-ray Generation, User Interface, Image Detection, and Viewing. Each functional building block is explained in further detail.

GEOMETRY

The Allura Xper FD20 Stand

The Allura stand consists of a ceiling-mounted C-arm. The stand has the following capability:

- The L-arm can be rotated and can be moved in longitudinal direction allowing a three-sided patient approach and total body coverage.
- L-arm rotation around the patient table: +90, 0, -90 degrees.
- L-arm longitudinal movement: 300 cm
- This movement features auto-stops at the parking position, cardio/neuro position and lower peripheral position.

The Allura stand allows a very wide range of projections, including PA and AP imaging.

- In the head position (0 degrees position, L-arm parallel to patient table):
 - C-arm rotation range (degrees): 120 LAO to 185 RAO
 - C-arm angulation range (degrees): 90 CA to 90 CR
 - (Full angulation capability determined by patient position)
- In the side position (+90 / -90 degrees position, L-arm perpendicular to patient table):
 - C-arm rotation range (degrees): 90 LAO to 90 RAO
 - C-arm angulation range (degrees): 185 CA to 120 CR or 120 CA to 185 CR
 - (Full angulation capability determined by patient position)
- The stand provides fully motorized fast movements with variable and configurable maximum speed.
 - Variable C-arm rotation speed, up to 25 degrees per second
 - Variable C-arm angulation speed, up to 18 degrees per second
- L-arm rotation and longitudinal movement: motorized and manual
- C-arm depth is 90 cm
- The FD20 Dynamic Flat Detector features Xper Access which allows the flat detector to be positioned in either portrait or landscape imaging modes in 3 seconds.
- The variable source image distance between focus and Dynamic Flat Detector input screen is motorized from 86.5 to 123 cm.

- The stand features BodyGuard a capacitive sensing collision avoidance system for patient protection.

Patient support

The Xper Table

Patient support with flat carbon fiber tabletop

- Table top length of 319 cm, width 50 cm
- Metal-free overhang 125 cm
- Floating table-top movement of 120 cm longitudinal and 35 cm transversal range.
- Motorized height adjustment from 79 to 107 cm
- Maximum cantilever of 223 cm , for full patient coverage
- Maximum patient weight 250 kg with 25 kg of accessories plus 500 N for CPR in any longitudinal position of the table top
- Xper Geometry and Imaging Modules for exam room controls.
 - The operating modules can be attached to either side of the table.

Patient Support Accessories set

- One cerebral filter
- Three rail accessory clamps
- One IV stand
- One slow recovery foam mattress
- One Set of Arm Supports (FCV0248)
- One Set of Patient Straps (FCV0250)
- One Head Support (FCV0251)
- One Arm Support (FCV0258)
- One Table-mounted Radiation Shield
- One anti-fatigue mat with Philips logo

X-ray Generation

The Allura Xper FD20 comprises an integrated dedicated X-ray system, micro-processor controlled Velara CFD generator based on high frequency converter technique. The user interface control of this X-ray Generator is incorporated in the Xper module, Xper Desktop Viewing Console, and the Xper on-screen displays. The Velara CFD generator comprises:

- X-ray generator 100 kW
- Voltage range is 40 - 125 kV
- Maximum current 1250 mA at 80 kV
- Program selection
 - Pulsed X-ray for pulsed fluoroscopy; 3.75, 7.5, 15 and 30 frames/s
 - Pulsed X-ray for (subtracted) acquisition up to 6 frames/s for vascular applications
 - Minimum exposure time of 1 ms
 - Automatic kV and mA control for optimal image quality prior to run to save dose
 - An X-ray depth collimator with two semi-transparent wedged filters with manual and automatic positioning

- SpectraBeam filtering of low energy radiation to optimize image quality and dose efficiency with MRC-GS 0407 X-ray tube.
- Grid switching at dynamic pulsed fluoroscopy
- Xper Beam Shaping, positioning of both shutters and wedges on the Last image Hold without the need for X-ray radiation

Fluoroscopy

- Three programmable fluoroscopy modes
 - Each mode can be set to different composition of dose rate, pulse speed, filter setting, and image processing (noise reduction, adaptive contour enhancement, and adaptive harmonization).
- Roadmap Pro
 - Roadmap Pro can be selected from the Xper imaging module and/or Xper module.
 - A vessel map is created and superimposed with (un)subtracted live fluoroscopy. Acquisition runs can be done during Roadmap without losing the vessel map. Roadmap Pro features Smart Settings in special clinical modes that are optimized to visualize special materials such as coils and glue. Live processing of the vessel map, the device map and the landmark map can be done on the Xper Module. Xres for vascular procedures is standard part of Roadmap Pro.
 - **Disclaimer:** AMC only corrects movement artifacts in two dimensions. Three dimensional movements such as swallowing or rotation of the head cannot be corrected.
 - In Roadmap Pro R2 "Automatic Motion Compensation" (AMC) is added to the roadmap functionality. During roadmap, small movements of the patient can lead to subtraction artifacts. These artifacts might conceal important clinical information. "Automatic Motion Compensation" compensates for rigid, uniform (skeletal/table) translations and is therefore very effective in interventional (neurology) applications where subtraction imaging is applied.

§ Disclaimer: AMC only corrects movement artifacts in 2 dimensions. 3 dimensional movements like swallowing or rotation of the head cannot be corrected.

- Xper Fluoro Storage, a grab function allows storage and archiving of both a fluoro image and the last 20 seconds of Fluoroscopy, called Xper Fluoro Storage. These fluoro images or fluoro runs can be archived as a regular exposure run.

X-ray tube

The Allura Xper FD20 has the Maximus ROTALIX Ceramic grid switch tube assembly MRC 200 GS 0407 integrated in the C-arc. This MRC tube has an anode heat storage capacity of 2.4 MHU and 0.4/0.7 mm. nominal focal spot values. The tube has a maximal loading of 30 and 67 kW.

Dynamic pulsed fluoroscopy uses grid switching technology to eliminate soft radiation and improve image quality. SpectraBeam allows for filtration of the x-ray beam with (a combination of) 0.2, 0.5 or 1 mm CU-equivalent filters.

Tube housing ROT-GS 1004 is for oil-cooling and has a build-in thermal safety switch. A rotor control unit is build-in for continuous rotation of the anode disk. The heat exchanger CU 3101 is for direct and continuous forced cooling with oil.

IMAGE DETECTION

The Allura Xper FD20 comprises the following image detection chain:

- A 30 cm by 40 cm FD20 Dynamic Flat Detector with eight imaging modes.
 - 30 x 38, 30 x 30, 26 x 26, 22 x 22, 19 x 19, 16 x 16, 13.5 x 13.5, and 11 x 11 cm
- The digital output of the FD20 flat detector is 2k*2.5k image matrix at 14 bits depth for the largest mode
- The flat detector subsystem features Xper Access, the detector can be rotated over 90 degrees, it moves from portrait to landscape back & forth
- DQE (Detective Quantum Efficiency) >73 %
- The pixel pitch: 154 x 154 microns

Viewing

The Allura Xper FD20 comprises the following components in order to display the clinical images in the control and examination room:

Displays

Examination Room

Two 18-inch monochrome LCD monitors designed for medical applications. The first display is used for viewing live images. The second display is the reference monitor.

- 18-inch monochrome TFT-LCD display with a 160 degree viewing angle.
- Native format 1280x1024 SXGA
- 10-bit gray-scale resolution with gray-scale correction

These monitors are not delivered when FlexVision XL, EP Cockpit or EP Cockpit XL is selected.

The monitor ceiling suspension in the exam room can be configured to accommodate 3, 4, 6 or 8 LCD monitors and includes motorized height adjustment. The height-adjust feature is dependent on the room ceiling height. When FlexVision XL, EP Cockpit or EP Cockpit XL is selected the monitor ceiling suspension is configured for one of those options.

- Of the two medical monochrome LCD monitors included in the MCS, one is used for viewing of live images and the other serves as the first reference display. Reference images or runs are controlled by infra-red remote-control Xper ViewPad.
- The On-Screen Display provides status information on stand rotation, angulation, display of system messages, X-ray tube load status, selected fluoroscopy mode, selected detector Field of View, and both the rate and accumulation of the dose area product and skin dose. For cardiac applications, the system also monitors and displays body zone specific Air Kerma data (10 zones).

Control Room

One 19-inch color LCD monitor used as a data monitor.

- 19-inch color TFT-LCD display
- Native format 1280x1024 SXGA

One 18-inch monochrome LCD monitor (Xper review monitor) designed for medical applications.

- 18-inch monochrome TFT-LCD display

- Native format 1280x1024 SXGA
- 10-bit gray-scale resolution with gray-scale correction

These control room monitors are not delivered when EP Cockpit or EP Cockpit XL is selected.

The Graphical User Interface on the monochrome monitor has the following features and functions:

- Step through file, run, or images
- File, and run overview
- Contrast, brightness, and edge enhancement settings
- Flagging of runs or images for transfer
- Applying text annotation in images
- Optional DICOM printing
- Executing Quantitative Analysis Packages if available
- Subtraction functionality
- Zoom/pan functionality
- Electronic shutters
- Video invert
- View trace, stacking of images
- Landmarking

Acquisition

The acquisition segment coordinates the parameters for automatic exposure control. The program is selected via the Xper module or Xper Desktop Console.

Exposure techniques:

- Serial imaging for DA and DSA with automatic exposure setting
- Single shot mode
- Acquisition frame rates: 0.5 to 6 images/s at 2048 x 2048, 12-bit matrix

The Allura Xper FD20 offers a storage capacity of:

- 50,000 images at matrix size of 1024 x 1024
- 12,500 images at matrix size of 2048 x 2048
- Maximum number of examinations is 999, with no limit to the maximum number of images per examination

USER INTERFACE

Xper is comprised of three elements: 1) Xper Settings, which customizes the system to each user preferred settings. 2) Xper User Interface 3) Xper Integration, which makes advanced integration functionality available such as DICOM Query / Retrieve, background archiving, and Xper Fluoro Storage.

The Xper User Interface uses User Interface modules in the Examination Room with On-Screen Display.

The On-Screen Display is positioned on the left side of the reference monitor. The following system information is displayed

- X-ray indicator and X-ray tube temperature condition

- Gantry position in rotation, angulation, and Source Image Distance
- Detector field size display
- General System messages
- Selected Frame speed
- Fluoroscopy mode
- Integrated fluoroscopy time
- Skin Dose and Dose Area Product
- Stopwatch

The Xper ViewPad contains the preprogrammed function settings. The system is provided with two Xper Viewpads. The following functions are provided:

- Run and image selection
- File and run cycle
- File overview
- Store to Reference image file
- Copy image to photo file
- Digital (fixed) zoom and panning
- Recall reference images
- Laser pointer, intended to point at regions of interest on the imaging monitors
 - LED indication of laser pointer on/off and battery low
- Subtraction on/off
- Remasking
- Landmarking

Remote Intercom

The separate intercom which is connected independently from the system that allows separate placement of the intercom at the preferred working position in the control room and examination room.

Table Side Modules

Two Xper Modules are provided for use. The first Xper Module is mounted tableside. The Second Xper Module (NCVA778) is located in the control room. These modules use a touch screen, which can be operated when draped with sterile covers. The Xper Module contains the following functionality:

- Acquisition settings
- Selection of Xper Setting allows the user to set frame rates and X-ray generation settings applicable for the type of the preferred intervention
- Image Processing

The Xper Geometry module can be positioned on all sides of the patient table, while keeping the button operation intuitive. The Xper Geometry module provides the following functionality:

- Tabletop float and table height position
- Source Image Distance selection
- Longitudinal movement of the Gantry along the ceiling
- Gantry rotation in an axis perpendicular to the ceiling

- Store and recall of two scratch gantry positions including SID
- Emergency stop button

The Xper Imaging module can also be positioned on three sides of the patient table, while keeping the button operation intuitive. The Xper Imaging module provides the following functionality:

- Fluoroscopy Flavor selection defined per Xper Setting
- Shutters and Wedge positioning
- Xper Fluoro Storage and Grab
- Selection of the Detector field size
- Shutter positioning
- Reset of the fluoroscopy buzzer

Pan Handle (NCVA081)

- The Pan Handle is an extension of the control facility for floating movements of the tabletop.

Control Room

The control room comprises a Xper Review Module, Xper Desktop Module, a keyboard, and a mouse. The Xper Review Module offers the basic functions for review. The Xper Review Module contains the following functionality:

- Power on/off
- Tagarno wheel to control the review of a patient file
- File and run cycle
- Contrast, Brightness, and Edge enhancement settings
- File, Run, Image stepping and run and file overview
- Delete run
- Image invert and digital zoom
- Reset fluoroscopy timer and enable/disable X-ray

System information is displayed on the bottom of the data monitor:

- Stopwatch and Time
- System guidance information
- Dose Area Product (DAP) and Skin Dose, and accumulative dose
- Frame speed settings, fluoroscopy mode, and accumulated fluoroscopy time
- Exposure and fluoroscopy settings as Voltage (kV), Current (mA) and pulse time (ms)
- Geometry information as rotation, angulation, and SID

Scheduling

The patients can be listed and selected per date, physician, and intervention type. Previous DICOM patient studies can be uploaded with the DICOM Query Retrieve function in the Allura system.

Patient management protocols are flexible and allow for multiple studies to be selected under one patient identification number. This means that new studies can be appended to an earlier patient file. Furthermore, each study can contain multiple examinations to allow for split administrative purposes. Each examination contains multiple files, like acquisition file, reference file, and QA results file.

Preparation

The preparation page provides the information of the room and patient preparation of each individual physician. The preparation page is customizable per Xper Setting and allows each physician to provide his own room protocols. This preparation page makes hard copies of the protocol instructions redundant.

Acquisition

The acquisition page contains information on the current selected patient.

Review

The review page allows for reviewing of patients:

- Previous examination cases
- Review of other DICOM XA or DICOM SC studies

Radiation Dose Structured Report

Collection of dose relevant parameters and settings and export to a DICOM database (e.g. PACS, RIS), according IEC60601-2-43, 2nd Edition.

The reported data can be used for, for example:

- Quality improvement: evaluating trends in X-ray dose performance per facility, system and operator.
- RDSR enables analysis of average dose levels & variance for routinely performed exams and procedures.
- Typical system usage can be extracted from the data.

Archive

Continuous Autopush (NCVA090)

Continuous Autopush is an archive accelerator, which ensures that background archiving continues with minimal disruptions.

Clinical studies can be archived to a CD or a PACS. The archive process can be completely automated and customized with Xper Settings. Parameters like multiple destinations, archive formats can be selected to the individual needs and wishes for programming under the Xper Settings,

The Xper DICOM Image Interface enables the export of clinical images to PACS. The export formats are based on DICOM 3.0 protocols. The system exports clinical studies in Cardiac DICOM XA Multi-Frame or DICOM Secondary Capture formats.

- The export format is configurable in 512x512, 1024x1024 2048 x 2048 (unprocessed) matrix.
- The examination can be sent to multiple destinations for archiving and reviewing purposes.
- The Xper DICOM Image Interface provides DICOM Storage and DICOM Storage Commitment Services.
- The DICOM Query/Retrieve function allows older DICOM XA MF and DICOM SC studies to be uploaded in the system. Furthermore, additional information can be appended to a study, while keeping the patient identification the same.

Remote Service

Access to the system from a Remote location is possible via network or modem connection. Remote access to a system can shorten the time needed for e.g. changing system settings or problem diagnosis.

Clinical Education Program for the Allura Xper System

Essentials OffSite Education: Philips will provide up to two (2) Cardiovascular Technologists, Registered Technologists Registered Nurses, or other system operator as selected by customer, with in-depth didactic, tutorial, and hands-on training covering basic functionality and work-flow of the cardiovascular imaging system. In order to provide trainees with the ability to apply all fundamental functioning on their system, and to achieve maximum effectiveness, this class should be attended no earlier than two weeks prior to system installation. This twenty-eight (28) hour class is located in Cleveland, Ohio, and is scheduled based on your equipment configuration and availability. Due to program updates, the number of class hours is subject to change without notice. Customer will be notified of current, total class hours at the time of registration. This class is a prerequisite to your equipment handover OnSite Education. CEU credits may be available for each participant that meets the guidelines provided by Philips. Please refer to guidelines for more information. **Travel and lodging are not included, but may be purchased through Philips. It is highly recommended that 989801292102 (CV Full Travel Pkg OffSite) is purchased with all OffSite courses**

Handover OnSite Education: Philips Education Specialists will provide twenty-eight (28) hours of education for up to four (4) students, selected by customer, including technologists from night/weekend shifts if necessary. Students should attend all 28 hours, and must include the two OffSite education attendees. CEU credits may be available for each participant that meets the guidelines provided by Philips. Please refer to guidelines for more information. Note: Site must be patient-ready. Philips personnel are not responsible for actual patient contact or operation of equipment during education sessions except to demonstrate proper equipment operation. **It is highly recommended for systems that are fully loaded or for customers with a large number of staff members to also purchase 989801292099 (CV Add OnSite Clin Educ 24h).**

The above education entitlements expire one (1) year from equipment delivery date. Ref# 106107318-091207

2

Isolated Wall Connection Box

5

Isolated Wall Connection Box

This Isolated Wall connection Box facilitates connection of the video source via standard DVI cable/connector and lossless transfer of the video signal over the approximate 30 m cable distance. It can be mounted in the exam room or in the control room, depending on the location of the video source.

The quantity of the VWCB's has to be calculated as follows:
For each video signal via MultiVision: 1 VWCB (max = 4)

For each video signal to FlexVision XL on Cardio System: 1 VWCB (max = 9)
 For each video signal to FlexVision XL on Vascular System: 1 VWCB (max = 8)
 For each 3rd party video signal directly connected to an LCD in the MCS: 1x VWCB.

Note:

No VWCB is required in case a video signal is connected directly to a dedicated LCD from the following sources:

- 1) Xper Live/ref Slaving
- 2) Interventional HW (XtraVision), ViewForum, Xcelera (only if workstations are powered by Allura Xper)
- 3) Xper IM

3

MultiVision 4x 1

1

MultiVision is the integrated video switch for high quality Progressive Display video sources. It can switch either B&W (RGB based) or color signals, with up to 4 inputs to one output.

MultiVision allows to share a Reference Display- or extra color monitor in the ceiling suspension in the exam room between the Xray system and other sources, such as a Xcelera viewstation, an Ultrasound system, a Interventional Tools station (like StentBoost, 3D CA) etc. These sources can be

allocated in the exam room or in the control room of the cath lab.

The switch is controlled by a button on the Xper Module. MultiVision provides a black image when a not operational input is selected.

Each external video source requires a Wall Connection box (not included in the MultiVision package) for the connection to a MultiVision input: only the Xper 2nd Ref Display as possible source does not require such box. The wall connection box also provides 230 V Power Supply for the connected video sources; however, in total maximum 1400 W can be supplied to the sources all together.

Comprising:

- video switch unit with cabling for max 4 B&W- or Color sources
- soft key button implemented on the Xper Module

Compatible with:

- Allura Xper series Rel 3 (monoplane versions) resp Rel2 (biplane versions) onwards.
- external video sources that comply with following requirements:
 - qualified medical electrical equipment [IEC 60601-1], or IEC 950 equipment combined with a multiple socket outlet [IEC 60601-1]
 - can be connected to the same earth as the Philips Protective Conductor Bar (PPCB).
 - provide video-output that matches the display range of the XB monitor or Colour

Monitor that is used for display

- provide a slave monitor output

Power requirements: refer to system configuration

4

Legacy Video Convertor

4

Legacy Video Convertor

The Legacy Video Convertor enables conversion from VGA towards DVI.

The Legacy Video Convertor enables conversion from VGA towards DVI for supported input resolutions, as listed in the table below.

Signal type	Native resolution	Image Aspect Ratio
-------------	-------------------	--------------------

VGA	640x480	4:3
-----	---------	-----

SVGA	800x600	4:3
------	---------	-----

XGA	1024x768	4:3
-----	----------	-----

SXGA	1280x1024	5:4
------	-----------	-----

SXGA+	1400x1050	4:3
-------	-----------	-----

UXGA	1600x1200	4:3
------	-----------	-----

WXGA	1280x800	16:10 (8:5)
------	----------	-------------

WSXGA	1440x900	16:10 (8:5)
-------	----------	-------------

WSXGA+	1680x1050	16:10 (8:5)
--------	-----------	-------------

WUXGA	1920x1200	16:10 (8:5)
-------	-----------	-------------

2K	2048x1080	19:10
----	-----------	-------

TV1080I/P	1920x1080	16:9
-----------	-----------	------

TV 480I	720x480	4:3
---------	---------	-----

TV 480P	704x480	4:3
---------	---------	-----

TV 576I	720x576	4:3
---------	---------	-----

TV 576P	704x576	4:3
---------	---------	-----

TV 720P	1280x720	16:9
---------	----------	------

5

RIS / CIS DICOM interface

1

This package allows communication of the Allura Xper system with a local information system (CIS or RIS). The interface uses the DICOM Worklist Management (DICOM WLM) and Modality Performed Procedure Step (DICOM MPPS) standards.

If a hospital has an Allura Xper system and an information system it can receive patient and examination request information from the information system and report examination results in order to:

- Eliminate the need for retyping patient information on the Allura Xper
- Prevent errors in typing patient names and registration numbers (ensuring consistency with IS information to prevent problems in archive clusters or to search for a name in case of later retrieval)
- Inform the IS about the acquired images and radiation dose

Upon request from the Allura Xper system the complete worklist with all relevant patient and examination data is returned from the IS to the Allura Xper system. For each patient the following information will be shown on the Allura Xper after it has been retrieved from the IS:

Patient Identification:

- Patient name
- Patient ID
- Birth date
- Sex

Examination/Request Information:

- Accession number
- Scheduled procedure step start time

- Scheduled performing physician's name

It is possible at all times to enter patient demographics information manually within the Allura Xper system in case of an emergency or in case the local Information System connection is down.

On request of the clinical user the Allura Xper will report the following information about the selected patient to the IS:

Patient Identification:

- Patient name
- Patient ID
- Birth date
- Sex

Examination/Request Information:

- Accession number
- Performed procedure step status start/end date and time
- Performing physician's name
- Referenced image sequence

Radiation dose:

- Total time of fluoroscopy
- Accumulated fluoroscopy dose
- Accumulated exposure dose
- Total dose
- Total number of exposures
- Total number of frames

Further detailed information can be found in the Allura Xper DICOM Conformance Statement.

The interface requires an EasyLink (hardware and software) if the IS is not compliant with DICOM Work List Management and Modality Performed Procedure Step.

6

Lab Reporting

1

Lab Reporting allows the user to generate and print simple reports in modality stand-alone situations. The user is able to incorporate free text and clinical images. The reporting functionality is suited for local printing and email. Part of the report is generated automatically from administrative data (e.g. patient/exam data hospital name) and required data (e.g. run-log dose information and event-log).

7

Automatic Position Control (APC)

1

The Automatic Position Controller (APC) for Integris Allura Flat Detector systems provides two modes of operation:

- Preset Position Sequence; the sequence of projections is determined per Xper Settings. Each set contains a maximum of 10 positions. Positions can be recalled in sequence or directly. The projection sequence comprises rotation, angulation, and SID settings, related to the selected reference image.
- Reference driven positioning. The projections on the reference monitors can be recalled with the push of a button. The reference driven positioning recollects the rotation, angulation, and SID.

8

Table APC

1

The Automatic Position Controller (APC) for the table provides two modes of operation:

- Auto positioning. The tabletop position and table height will be adjusted automatically to the pre-defined default point of interest. This to save time and x-ray dose at the start of an exam or for setting up the system for rotation scans.
- Store/recall of a position of the table top. This includes the height-, longitudinal- and lateral position of the table top.

9

Subtracted Bolus Chase

1

For visualization of vessel structures when the blood flow is difficult to estimate, in particular in the lower peripherals.

Bolus Chase solves the problem of cumbersome step movements, the mismatch between blood flow and selected program, and lack of real-time image information.

During digital acquisition in non-subtracted mode with uninterrupted real-time image display, the contrast bolus is followed (chased) interactively by a motorized table scan movement using a hand-held speedcontroller to adapt the speed of the table scan to the contrast flow. The framespeed can be adapted as well.

The bolus run is followed with a mask run while using the same speedcurve and framespeed as generated during the bolus run. Viewing is possible in the subtracted and non-subtracted mode. If subtracted viewing is not required, the mask run can be skipped.

Subtracted Bolus Chase gives fast, accurate results for increased patient throughput and improved patient management. Automated exposure control and precise speed control assure a high quality images and excellent subtraction studies.

Comprising:

- automatic exposure control
- tabletop motordrive and hand-held speed controller (tableside)
- technique selection using Xper module, available both tableside and in control room (Xper FD20, FD20/10)

10

CO2 View Trace Software

1

Software package which enables tracing (stacking) of images acquired with CO2 injections. This function can be used during postprocessing next to view trace of images acquired with iodine injection.

100215 Allura Xper FD20

Line #	Description	Qty
--------	-------------	-----

11	FULL AUTOCAL	1
----	---------------------	---

The AutoCal option is a software package to be used in conjunction with quantitative analysis software packages. It provides an auto calibration procedure for an object to be analyzed that is placed in the iso-center. When the object to be analyzed (e.g. Left Ventricle Vessel Segment) is placed in the iso-center AutoCal avoids the need to:

- acquire an additional image series containing a sphere or grid for calibration purposes
- calibrate manually on a calibration object (e.g. catheter) displayed in the image or image series to be analyzed

12	Vascular Quant.Sw pkg(Xper)	1
----	------------------------------------	---

Functions:

- vessel diameter / stenotic index
- automated vessel analysis
- calibration routines

In addition the package allows manual measurements of line lengths (absolute and ratio's) and angulations. Multiple measurements in one image are possible.

Compatible with:

- Allura Xper FD10 Rel 3 and FD10/10 Rel 2 onwards
- Allura Xper FD20 Rel 2 and FD20/10 Rel 2 onwards
- Allura CV20 R1 onwards

13	Peripheral X-ray Filter	1
----	--------------------------------	---

Set of flexible x-ray filters to provide an uniform density in angiographic examinations of the lower peripheral area.

Comprising:

- one central filter, at the top edge provided with sizing markers at every 5 cm, length : 1 m
- two side filters, length: 1 m

14	Pivot for table base.	1
----	------------------------------	---

For angiographic- and interventional procedures of the upper peripherals.

Provides improved table access for patient transfer.

Allows pivoting of the table base around its vertical axes.

Pivot range from -90 degrees to + 180 degrees (or -180 to +90 degrees) with locked positions on 0, -13/+13 (facilitating arm-angiography) and -90/+90 and 180 degrees.

Comprising:

- pivot device with graduated scale to be mounted on the universal floor plate of the table.

Compatible with Xper Table

15	Add.op-rail with cable ext.kit	1
----	---------------------------------------	---

100215 Allura Xper FD20

Line #	Description	Qty
	The content of the additional OP-Rail kit is:	
	<ul style="list-style-type: none"> [A] One additional OP-Rail (mechanical) [B] Cable Extension for OP-Rail <ul style="list-style-type: none"> One Extension cable for Geo Module One Extension cable for Imaging Module One connection box (wherein the extension cables are coupled with the UI-Module cables. 	
	<p>[A]</p> <ul style="list-style-type: none"> An extension for the table op-rail (30cm). The additional op-rail can be mounted at the both sides of the tabletop part where no op-rails are mounted. The additional op-rail is compatible with AD5 and XperTable (cardio and neuro) patient-tabletops. The op-rail has the same profile /dimensions as the current standard op-rail The maximum load (downwards) on the additional op-Rail is 100 N (F=100N) <ul style="list-style-type: none"> (this is limited by the tabletop of the Patient Table) The maximum mechanical moment on the additional op-Rail is 40Nm downwards and 20Nm upwards <ul style="list-style-type: none"> (this is limited by the tabletop of the Patient Table) 	
	<p>[B]</p> <ul style="list-style-type: none"> The cable extension consists out of two cables with a length of 1.3 m; one for the Geo and one for the Imaging module, and an interface box were the coupling to the Geo and Imaging module cables can be made. 	

16 **19" Color LCD monitor in 1**

Exam Room

19" Color LCD monitor in Exam Room

19" flat panel color monitor. This LCD monitor is intended for viewing in the examination room and is designed for medical applications.

The main characteristics are:

- 19 inch Color TFT-LCD display
- Native format: 1280x1024 SXGA
- Wide viewing angle (approx 170 degr)
- operated Brightness level 200 Cd/m2
- On Screen Display of control functions operated via touch buttons on front
- Internal power supply (90-264 VAC)

Compatible with:

Standard PC format (RGBHV)

DVI interface standard

UL60601-1

Allura Cardio/Vascular systems

Mains connection: 110 - 240 V

Dimensions : 425(W)x375(H)x97(D) mm

Weight: 7 kg.

Colour: mushroom, front ultra dark grey

100215 Allura Xper FD20

Line #	Description	Qty
--------	-------------	-----

17	18 INCH LCD MONITOR IN THE EXAM ROOM	1
----	---	---

18" Monochrome LCD monitor. The LCD monitor is intended for viewing in the examination room and is designed for medical applications.

The main characteristics are:

- 18 inch monochrome TFT-LCD display
- Native format 1280x1024 SXGA
- 10 bit gray-scale resolution with gray-scale correction
- Wide viewing angle (approx 160 degr)
- High brightness (max 600 Cd/m2 default 500 Cd/m2) with ambient light dependent brightness control
- Push buttons for control functions on front
- User programmable and standard reference setting
- On Screen Display
- Internal selectable lookup table for gray-scale transfer function
- Internal power supply (110-240 VAC)
- Including LCD protection screen

18	Two rows of 2 (4M)	1
----	---------------------------	---

19	Rad Shield w/ Arm (Contoured) 61X76	1
----	--	---

Contoured Rad Shield with Arm rest. 61X76

20	PIVOTING TABLE-MOUNTED RADIATION SHIELD	1
----	--	---

Table-mounted radiation shield for additional protection of physician and staff against scatter radiation. The shield consists of two protective parts: a lower shield and an upper shield. The shield is specially designed for use with the AD5 patient table.

The table mounted radiation shield provides the following features:

- Mounting to either the right or left table accessory rails;
- Pivoting into the required working position;
- Pivoting into the parking underneath the tabletop facilitating patient preparation;
- The upper shield can be positioned upright providing optimal protection or can be folded down for free access to the patient.

The table mounted radiation shield includes:

- Lower shield measuring 70 cm high 80 cm wide 0.5 mm Pbequivalence;
- Upper shield measuring 40 cm high 50 cm wide 0.5 mm Pbequivalence;
- Mounting clamp;

Docking device for wall mounting.

21	Exam Lamp 220v	1
----	-----------------------	---

Spring arm mounted examination light for cardiovascular applications

100215 Allura Xper FD20

Line #	Description	Qty
--------	-------------	-----

22	Portegra 2 360 Ceiling Column	1
----	--------------------------------------	---

Portegra 2 360 Column w/ trolley and ceiling track

23	Medrad Mark 7 Arterion Pedestal	1
----	--	---

The Arterion Mark 7 Pedestal contrast medium injector can be positioned anywhere at the patient positioning table on a mobile unit, for direct operation of all functions in the examination room.

The injector system includes:

- A mobile pedestal stand with electronics unit and a connection cable to the manual release.
- A support arm with injector head and a control lever for moving the injector head.
- A user control console with large touch screen and corresponding additional monitoring display on the injector head.

Functions

Pressure limitation:

- for 150 ml syringes 689 to 8273 kPa, corresponds to 100 to 1200 psi. .

Flow rates for 150 ml syringes:

- 0.1 to 45 ml/s in increments of 0.1 ml/s
- 0.1 to 59.9 ml/min in increments of 0.1 ml/min
- rise/fall: 0 to 9.9 s in increments of 0.1 seconds

Release delay for injection or radiation:

- 0 to 99.9 s in increments of 0.1 s.

Adjustable volume for 150 ml syringes:

- 1 ml to the max. syringe capacity in increments of 1 ml.

Fill rate:

- Variable syringe filling speed 1-20ml/s.

Injection protocols:

- Up to 40 injection protocols possible.

Parameters currently displayed on the touch screen display and on the head display:

- Injection speed
- Injection volume
- Remaining volume
- Injection duration
- Applied pressure

Contrast medium heating:

- Nominal 35°C (95°F)+-5°C (9°F)

Injection data memory

- Up to 50 injection data items stored

Included in the scope of delivery

- Injector standard configuration 150 ml
- Philips interface cable
- Operator Manual
- Service manual (English).

Power supply

100-240 VAC 50/60 Hz 1000VA.

100215 Allura Xper FD20

OPTIONS

SELECTION OF ANY OPTION WILL INCREASE THE CONTRACT PRICE BY THE AMOUNT SHOWN IN THE PRICE COLUMN. OPTIONAL EQUIPMENT PRICING VALID ONLY IF PURCHASED IN CONJUNCTION WITH EQUIPMENT QUOTED.

Line #	Description	Qty
1	<p>3D-RA R.6</p> <p>Allura 3D-RA assists physicians in decision making for treatment strategy in endovascular procedures, neuro or vascular surgery or even radiotherapy.</p> <p>Allura 3D-RA reduces the number of DSA acquisitions and fluoroscopy time needed to perform an examination. This means less X-Ray dose for the patient and the medical staff and a reduced quantity of dye, leading to reduced procedure costs.</p> <p>Allura 3D-RA provides a unique assessment after treatment due to the use of non-subtracted images that allows to shows devices stents, coils, clips and provide the optimal stand projection for endovascular treatment.</p> <p>Allura 3D-RA provides a wide range of communication facilities to export 3D images.</p> <p>1 Image Acquisition</p> <p>Image acquisition is performed with the Rotational Angiography feature of the Allura Xper FD series with the flexibility to position the C-arm in either head or side position.</p> <p>C-arm in Head position: the Rotational Angiography run is performed over a scan range of 240 degrees with a rotation speed up to 55 degrees/sec.</p> <p>C-arm in Side position: the Rotational Angiography run is performed over a scan range of 180 degrees with a rotation speed up to 30 degrees/sec.</p> <p>2 3D Vessel Reconstruction</p> <p>The rotational run is automatically transferred and displayed as a 3D vessel model: with the Real-Time digital link (option) 120 images are reconstructed into a 3 dimensional model within seconds. Additional reconstructions, using the Reconstructive Zooming Technique, can be performed as well.</p> <p>3 Workflow:</p> <p>Allura 3D-RA in combination with the Allura Xper FD series will provide an optimal workflow via the following workflow enhancers:</p> <p>Complete automated 3D-RA process from 3D acquisition to 3D Viewing: no user interaction needed.</p> <p>3D at Xper Module (option); With the Xper module the physician has all required 3D functionality at tableside. At the touch screen module functionality like rotating, panning, zooming, AVA, virtual stenting, 3D-APC and 3D Follow C-arc can be performed. With the mouse tablet all other functions can be performed so that there is no need for the Physician to leave the examination room.</p> <p>3D Automatic Position Control (3D-APC); When the optimal working position has been chosen via the Allura 3D-RA interventional tool, the C-arc will automatically steer to this position.</p> <p>3D Follow C-arc; When the position of the C-arc (not using any X-ray) is changed, the 3D volume will automatically follow the position of the C-arc. This means the position of the C-arc (and therefore the 2D projection) and the 3D volume are always aligned. As last seen; when the user leaves the patient in the model and later selects that patient again, the Allura 3D-RA interventional tool will return to the image last used by the user.</p> <p>Mouse over: When moving the mouse cursor over a button the mouse over text will show up to explain the function of that specific button.</p> <p>4 Calibration</p> <p>Allura 3D-RA calibrations are performed by Philips Healthcare Customer Support. Allura 3D-RA calibration data are stable over at least 6 months time.</p> <p>5 Viewing</p> <p>A Real Time user interface is available with 3D-RA, providing 3D object viewing in any space</p>	1

100215 Allura Xper FD20

OPTIONS

SELECTION OF ANY OPTION WILL INCREASE THE CONTRACT PRICE BY THE AMOUNT SHOWN IN THE PRICE COLUMN. OPTIONAL EQUIPMENT PRICING VALID ONLY IF PURCHASED IN CONJUNCTION WITH EQUIPMENT QUOTED.

Line #	Description	Qty
	<p>direction. A graphical display of (C-arm) stand position including angulation/rotation for any projection.</p> <p>Philips' CRM (Contrast Resolution Management) Technology for a considerable increase in contrast resolution in all volumes.</p> <p>Various Image Rendering possibilities: Volume/Surface Rendering, MIP, Endoscopy, SUM (pseudo x-ray image) Gradient rendering; the possibility to display the vessel structure transparently.</p> <p>Cut-plane function to get a precise insight of the shape of the pathology</p> <p>Orthoviewer providing a multi-planar visualization of objects using the different Image Rendering possibilities.</p> <p>MPR (Multi-Planar Reformatting): enables visualization of the volume in all three standard projections (coronal, sagittal and axial) Especially useful for optimal viewing of spine procedures (e.g. Vertebroplasty)</p> <p>SpineView: special acquisition protocol for optimal viewing of the spine, especially osteoporotic vertebrae</p> <p>CalciView: allows visualization of Hyper dense plaque in 3D, separately or in relation to the lumen.</p> <p>5 different distance measurements calculated in the same volume, including "Quick measurement" feature</p> <p>Volume calculation</p> <p>Automated Vessel Analysis (AVA), provides information on vessel segment diameter, area and length with only three mouse-clicks. Endoscopic and cross sectional views are available.</p> <p>Computer Assisted Aneurysm Analysis (CAAA), providing information on Aneurysms, like volume, neck size etc..</p> <p>Catheter tip shape simulation, providing information on how to shape the catheter tip.</p> <p>Virtual stenting; Ability to simulate a stent placement in a selected vessel segment for proper stent sizing. All relevant data of the simulated stent are displayed</p> <p>Annotation: text can be added to a volume to capture comments.</p> <p>Interpolative Zoom</p> <p>Reconstructive Zooming Technique, 2 additional user defined reconstructions focused on the Volume Of Interest (VOI) using different cube size and voxel resolution.</p> <p>Subtraction of reconstructed volumes, allowing to visualize vessels without embolization devices (stents, coils, clips,...) to assess the outcomes of treatment</p> <p>Automatic Voxelshift: compensates for movement when rendering subtracted or superimposed volumes</p> <p>Set the grey values WW/WL</p> <p>Store/Recall of user defined projections.</p>	
6	Archiving:	
	Transfer to:	
	Optional Hard Copy unit (DICOM Print)	
	Any optional DICOM compatible device (e.g. PACS/ViewForum/Xcelera), supported are DICOM	
	XA, DICOM SC, DICOM CT and DICOM 3D	
	Any PC in a standard PC compatible format (JPEG,AVI)	
	One or multiple DVD's, CD-ROM(s) for easy archiving	
	Store a subset of exportable objects (snapshots and AVI Movies) to a USB removable memory device.	

100215 Allura Xper FD20

OPTIONS

SELECTION OF ANY OPTION WILL INCREASE THE CONTRACT PRICE BY THE AMOUNT SHOWN IN THE PRICE COLUMN. OPTIONAL EQUIPMENT PRICING VALID ONLY IF PURCHASED IN CONJUNCTION WITH EQUIPMENT QUOTED.

Line #	Description	Qty
	Clinical Education Specialists will provide sixteen (16) hours of tailored CV 3DRA OnSite Education for up to four (4) students, selected by customer, including technologists from night/weekend shifts if necessary. CEUs are not available in all cases. Please read Guidelines for more information, which will be provided to you during the scheduling process. Education Hours: Mon – Fri 8:00am to 5:00pm, except Monday and Friday are half-days to allow for trainer's travel. Note: Philips personnel are not responsible for actual patient contact or operation of equipment during education sessions except to demonstrate proper equipment operation. Education expires one (1) year from equipment delivery date (or purchase date if not sold with equipment).	
2	Xper Live/Ref Slaving Xper Live/Ref Slaving The Xper Live/Ref Slaving will enable the option to slave the Live or Ref video source from the Allura Xper. The total amount of Xper Live/Ref Slaving that can be selected is max 4. Xper Live/Ref Slaving is possible: - In Control Room icw FCV0011(B/W monitor in Control Room) - In Philips MCS (additional monitor excluded from this option) - Icw FCV0519 1 or 2 MCS from Skytron/Steris	1
3	FD Rotational Angio Rotational angiography provides real-time 3D impressions of complex vasculature and coronary artery tree. It acquires multiple projections with just one contrast injection via a fast rotational scan of the region of interest. Rotational Angiography can be used during screening procedures to quickly determine the optimal projection for the study as the angle (rotation/angulation) of the projection is indicated on each image. Compared with traditional angiography, Rotational Angiography can save considerable time, dose and contrast, while providing image detail required for diagnostic and therapeutic decisions. A rotational scan is possible both with the Allura Xper systems in the side position (ceiling mounted systems) and in the head position, providing the flexibility to perform procedures virtually from head to toe. C-arm in side position: <ul style="list-style-type: none">• Max. rotation Speed: 30 degrees/s• Max. rotation Angle: 180 degrees C-arm in head position: <ul style="list-style-type: none">• Max. rotation Speed: 55 degrees/s• Max. rotation Angle: 305 degrees	1

Max. Frame speeds are given by the framespeed specifications of the system configuration.

100215 Allura Xper FD20

OPTIONS

SELECTION OF ANY OPTION WILL INCREASE THE CONTRACT PRICE BY THE AMOUNT SHOWN IN THE PRICE COLUMN. OPTIONAL EQUIPMENT PRICING VALID ONLY IF PURCHASED IN CONJUNCTION WITH EQUIPMENT QUOTED.

Line #	Description	Qty
	The speed and range of rotation are the highest available (see table). The very high speed allows using less contrast, whereas the very wide rotation range provides a complete evaluation of the anatomy.	

A contrast run can be followed up with a mask run, to allow image/run subtraction.

The stand is designed for a very high mechanical stability. It offers precise positioning and high reproducibility, assuring you of high quality images and excellent subtraction studies.

Operation of Rotational Angiography is extremely easy. The procedure is selected, set up and executed virtually in a matter of seconds, supporting the highest patient throughput.

A set of dedicated acquisition programs is available on the Xper Module and can be selected at the touch of a button. The rotation end- and start-positions are easily selected. The procedure is controlled from the exposure hand- or footswitch.

4	FD Dual Fluoro	1
---	-----------------------	---

Dual Fluoro for Flat detector systems

The Dual Fluoroscopy mode allows digitally processed fluoroscopy in parallel with trace subtract fluoroscopy, providing a non subtracted reference fluoro image for complex interventions.

This option provides an additional fluoro channel in parallel to the default fluoro channel. The Dual fluoroscopy mode is selected via the Xper module.

The trace subtracted fluoro image will be displayed on the exam monitor, the non-subtracted fluoro image is displayed on the reference monitor.

In Dual Fluoro mode, The fluoroscopy image on the exam monitor can be zoomed digitally with a factor 2, providing a larger view of the region of interest for complex interventions. The fluoro zoom function is controlled via the Xper module.

5	FD SmartMask	1
---	---------------------	---

SmartMask simplifies roadmapping procedures by overlaying a selected reference image with fluoroscopy on the live monitor in the exam room.

The reference image can be faded in/out with variable intensity, controlled from tableside.

SmartMask uses the reference image displayed on the reference monitor.

Any previously acquired image can be used as reference.

SmartMask facilitates pre- and post- intervention comparisons to assess treatment results

6	Xper Table Tilt	1
---	------------------------	---

This innovating SyncraTilt enhances the accuracy and efficiency of gravity-oriented procedures. It is available as an option for the Xper table in Allura Xper series systems.

SyncraTilt is ideal for interventional, myelography, phlebography and head down procedures because it provides more precise imaging of contrast medium, blood, or objects in the body.

With SyncraTilt, the isocentre is automatically located at the isocentre of rotation and angulation of the stand. If the longitudinal position of the stand changes, the tilt isocentre is changed to match with the new stand position. As a result, the region of interest is always centred

100215 Allura Xper FD20

OPTIONS

SELECTION OF ANY OPTION WILL INCREASE THE CONTRACT PRICE BY THE AMOUNT SHOWN IN THE PRICE COLUMN. OPTIONAL EQUIPMENT PRICING VALID ONLY IF PURCHASED IN CONJUNCTION WITH EQUIPMENT QUOTED.

Line #	Description	Qty
	<p>As the table tilts, the X-ray beam automatically coordinates to the movement.</p> <p>The table floats even when tilted, and the region of interest can be followed by panning the tabletop.</p> <p>When combined with the Bolus Chase option, SyncraTilt enables phlebography to be performed with a head-up tilted patient.</p> <p>The option provides:</p> <ul style="list-style-type: none">• maximum tilt range:• 17 degrees (head down) to +17 degrees (head up).• tilt speed: 2 degrees/sec• automatic safeguarding system with manual override• panning range in tilted plane: equal to the standard• tabletop specifications (longitudinal 120cm, lateral 35cm)• easy to use controls Comprising:<ul style="list-style-type: none">• Tilt drive with user controls <p>Compatible with:</p> <ul style="list-style-type: none">• Xper table in Allura Xper FD series Rel 3 onwards (monoplane versions) and Rel 2 onwards (biplane versions)• Bolus Chase• Pivot for table base• swivel for table base	
7	Interventional Tools Hardware	1
	<p>The interventional tools hardware is the computer that enables the 3D interventional tools, it allows to import and view DICOM compatible data from other imaging modalities The interventional Hardware comprises at least a Harddisk containing operating system and application software.</p>	
8	Real time image link	1
	<p>Real Time digital image link to an off-line Allura Interventional Hardware station. This applies on the applications 3D-RA, StentBoost and 3D-CA on the Interventional Hardware. This dedicated digital link sends raw or processed image data (depending on the application) real time during monoplane exposures to the connected Interventional Hardware station, to allow instant results of the applicable reconstruction after the exposure run.</p> <p>In biplane systems, this digital link is available for the frontal channel only.</p>	
9	3D RA Control for Xper Module	1

100215 Allura Xper FD20

OPTIONS

SELECTION OF ANY OPTION WILL INCREASE THE CONTRACT PRICE BY THE AMOUNT SHOWN IN THE PRICE COLUMN. OPTIONAL EQUIPMENT PRICING VALID ONLY IF PURCHASED IN CONJUNCTION WITH EQUIPMENT QUOTED.

Line #	Description	Qty
	Table Side Module functionality for Allura Xper FD20 used with Integris 3D-RA Release 4.2. For further improvement of interventional procedures efficiency the following workflow enhancers are made available in the examination room: With the Xper touchscreen module the physician has all 3D functionality needed at tableside. Functionality like rotating panning zooming AVA Virtual stenting 3 and 3D Follow C-arc can be performed. No need for the Physician to leave the examination room. 3D Automatic Position Control (3D-APC); when the optimal working position has been chosen via the Integris 3D-RA interventional tool the C-arc will automatically steer to this position. 3D Follow C-arc: When the position of the C-arc (not using any X-ray) is changed the 3D volume will automatically follow the position of the C-arc. This means the position of the C-arc (and therefore the 2D projection) and the 3D volume are always aligned.	

10	3D Roadmap	1
	3D Roadmap extends the capabilities of the integrated 3D product by providing a sustainable 3D roadmap to support interventional procedures. The 3D Roadmap option matches the real-time 2D fluoroscopy images with the 3D-RA reconstruction of the vessel tree. It provides a 3D real time insight of the advancement of the guide wire, catheter and coils through complex vessel structures. 3D roadmap has automatic motion compensation for the neuro runs. When the automatic motion compensation function is active, this functionality will constantly correct the motion artifacts which can be present in the 3D Roadmap image.	

Image Acquisition

The 3D Roadmap is based on the visualization of the vessel tree out of 3D-RA. The 3D Roadmap is activated with one button touch at tableside (Xper Module). Select the 3D Roadmap function on the touch screen module, activate fluoroscopy and the 3D Roadmap is activated. The "live" 2D fluoroscopy image is overlaid with the 3D volume of the vessel tree and is automatically displayed on the 3D roadmap monitor in both the examination and control room.

Intuitive, fully controlled from tableside:

The bidirectional link between the X-ray system and the 3D Roadmap allows the user to select the optimal stand position for the procedure in two ways. 3D Automatic Position Control allows the gantry to automatically move to the best interventional projection as shown on the 3D Roadmap monitor. 3D Follow C-arc allows the 3D Roadmap to remain in sync with the 2D projection, automatically adjusting viewpoint as the gantry is repositioned.

- Landmarking to adjust the intensity of the anatomical reference surrounding the vessels;
- 3D blending to fade in/out the 3D view;
- WW/WL settings to control the contrast/brightness;
- Store and review runs for reporting and archive purposes;
- Store snapshots and movies.

3D Roadmaps can be sent to:

Any optional DICOM compatible device (e.g. PACS/ViewForum/Xcelera), supported are DICOM XA, DICOM SC, DICOM CT and DICOM 3D
Any PC in a standard PC compatible format (JPEG,AVI)

And stored/archieved on

A PACS systems as DICOM Secondary Capture images or movies
USB removable memory device
One or multiple DVD's, CD-ROM(s) for easy archiving
Hard copy via the (DICOM Print) protocol

OPTIONS

SELECTION OF ANY OPTION WILL INCREASE THE CONTRACT PRICE BY THE AMOUNT SHOWN IN THE PRICE COLUMN. OPTIONAL EQUIPMENT PRICING VALID ONLY IF PURCHASED IN CONJUNCTION WITH EQUIPMENT QUOTED.

Line #	Description	Qty
--------	-------------	-----

11	MR/CT Roadmap	1
----	---------------	---

MR/CT Roadmap extends the capabilities of the integrated 3D product by providing a sustainable 3D roadmap based on previous acquired CT or MR scans to support interventional procedures. The MR/CT Roadmap option matches the real-time 2D fluoroscopy images with the 3D volume of CT or MR.

The CT or MR data can visualize in either 3D (e.g vascular structure) or with 2D slice in the same orientation as the 2D fluoro image. It provides a 3D real time insight of the advancement of the guide wire, catheter and coils through complex vessel and anatomical structures

Image Acquisition

A previously acquired CT or MR scan can be imported into the system and matched with a low dose 3D-RA or XperCT scan. The MR/CT Roadmap is activated with one button touch at tableside (Xper Module). Select the MR/CT Roadmap function on the touch screen module, activate fluoroscopy and the MR/CT Roadmap is activated. The "live" 2D fluoroscopy image is overlaid with the MR/CT volume presented in 2D or 3D and is automatically displayed on the roadmap monitor in both the examination and control room.

Intuitive, fully controlled from tableside:

The bidirectional link between the X-ray system and the MR/CT Roadmap allows the user to select the optimal stand position for the procedure in two ways. 3D Automatic Position Control allows the gantry to automatically move to the best interventional projection as shown on the MR/CT Roadmap monitor. 3D Follow C-arc allows the MR/CT Roadmap to remain in sync with the 2D projection, automatically adjusting viewpoint as the gantry is repositioned.

- Easy 2 step registration of the MR/ CT volumes
- Landmarking to adjust the intensity of the anatomical reference surrounding the vessels and tissue
- 2D and 3D blending to fade in/out the 2D or 3D view;
- WW/WL settings to control the contrast/brightness;
- Store and review runs for reporting and archive purposes;
- Store snapshots and movies.

MR/CT Roadmaps can be sent to:

- Any optional DICOM compatible device (e.g. PACS/ViewForum/Xcelera), supported are DICOM XA, DICOM SC, DICOM CT and DICOM 3D.
- Any PC in a standard PC compatible format (JPEG,AVI).

And stored/archived on

- APACS systems as DICOM Secondary Capture images or movies.
- USB removable memory device.
- One or multiple DVD's, CD-ROM(s) for easy archiving.
- Hard copy via the (DICOM Print) protocol.

12	Xper CT R2	1
----	------------	---

XperCT extends the capabilities of the angio system offering CT like imaging. The XperCT acquisition scan acquires up to 620 images. The 620 images ensure a high quality reconstruction of a CT-like volume to visualise soft tissue.

XperCT includes frame rate extension to increase the system acquisition speed up to 60 frames per second. The high frame rates are beneficial for the dedicated abdomen protocols: fast acquisitions times in 5 or 10 seconds.

100215 Allura Xper FD20

OPTIONS

SELECTION OF ANY OPTION WILL INCREASE THE CONTRACT PRICE BY THE AMOUNT SHOWN IN THE PRICE COLUMN. OPTIONAL EQUIPMENT PRICING VALID ONLY IF PURCHASED IN CONJUNCTION WITH EQUIPMENT QUOTED.

Line #	Description	Qty
	<p>The XperCT imaging process is fully automated in the Xper system. The XperCT 3D volume is displayed automatically within 1 minute (from acquisition to display); no user interaction required. Especially in critical cases it is important to obtain a fast overview.</p> <p>The 3D volume can be viewed in the control room and in the examination room. The slice view is performed by scrolling through the volume. Slice thickness and ww/wl can be varied upon user need. XperCT can be controlled via the Xper 3D module at tableside.</p> <p>In addition the XperCT volume can be matched with Allura 3D-RA. This view combines soft tissue information with high-resolution vessel information. The optimal view can be chosen with the orientation of the 3D volume: the C-arc follows automatically.</p> <p>Pre-requisite:</p> <ul style="list-style-type: none"> • Interventional HardWare • Real Time Link • FD Rotational Angio • Frame rate extension 	

Clinical Education Program for XperCT

CV XperCT Handover OnSite Education: Philips Education Specialists will provide eight (08) hours of education for up to four (4) students, selected by customer, including technologists from night/weekend shifts if necessary. CEU credits may be available for each participant that meets the guidelines provided by Philips. Please refer to guidelines for more information. Note: Site must be patient-ready. Philips personnel are not responsible for actual patient contact or operation of equipment during education sessions except to demonstrate proper equipment operation. Education expires one (1) year from equipment installation date (or purchase date if sold separately). Ref# 335-100615

13

XperGuide Rel 2 SW

1

XperGuide enables real-time needle guidance in the angio suite. Virtual needle paths are created on an XperCT dataset or on the previous acquired CT or MR dataset. XperGuide option matches the real-time 2D fluoroscopy images with the 3D volume of XperCT, CT or MR; to visualize the actual needle path versus the virtual path previously planned.

This volumetric dataset can be viewed in any slice direction. A wide range of gantry projections can be used to define the needle path.

Path planning can be done:

- By drawing a virtual needle path on an XperCT, MR or CT slice
- By defining entry and target points on different XperCT, MR or CT slices
- By defining a help line on a 3D volume XperGuide automatically calculates the optimal gantry projections for the path and transfers them to the planning to draw the needle path.

The calculated virtual needle paths can be viewed on the XperCT, MR or CT slices, to verify if this path is feasible. XperGuide supports planning of multiple needle trajectories. During the needle procedure, XperGuide is fully controlled at tableside. When XperGuide is active, guidance is

OPTIONS

SELECTION OF ANY OPTION WILL INCREASE THE CONTRACT PRICE BY THE AMOUNT SHOWN IN THE PRICE COLUMN. OPTIONAL EQUIPMENT PRICING VALID ONLY IF PURCHASED IN CONJUNCTION WITH EQUIPMENT QUOTED.

Line #	Description	Qty
	<p>automatically active when the fluoro pedal is pressed. The live 2D image is projected over the XperCT, MR or CT volume. The gantry can be positioned in the calculated gantry positions or controlled manually. The XperGuide images (live 2D fluoro projected over the XperCT, MR or CT volume) will follow the gantry projections.</p> <p>At table side, XperGuide adapts in real-time to the following parameters:</p>	

- Changes in the angulation of the C-arm
- Changes in the rotation of the C-arm
- Changes in the field of view
- Changes in the source image distance

XperGuide run are in the same patient file as all other patient related data. All this data can be reviewed at any time.

XperGuide runs are stored together with the XperGuide movies and snapshots can be sent to:

- Any optional DICOM compatible device (e.g. PACS/ViewForum/Xcelera), supported are DICOM XA, DICOM SC, DICOM CT and DICOM 3D
- Any PC in a standard PC compatible format (JPEG,AVI)

And stored/archieved on:

- A PACS systems as DICOM Secondary Capture images or movies
- USB removable memory device
- One or multiple DVD's, CD-ROM(s) for easy archiving
- Hard copy via the (DICOM Print) protocol

Clinical Education Program for XperGuide

CV XperGuide Handover OnSite Education: Philips Education Specialists will provide eight (08) hours of education for up to four (4) students, selected by customer, including technologists from night/weekend shifts if necessary. CEU credits may be available for each participant that meets the guidelines provided by Philips. Please refer to guidelines for more information. Note: Site must be patient-ready. Philips personnel are not responsible for actual patient contact or operation of equipment during education sessions except to demonstrate proper equipment operation.

Education expires one (1) year from equipment installation date (or purchase date if sold separately). Ref # 336-100316