



**VA**

U.S. Department  
of Veterans Affairs

Department of Veterans Affairs  
Office of Construction & Facilities Management

Manhattan Veteran's Affairs Medical Center  
New York, New York

Manhattan VAMC Bldg. 1-Ward Renovations-8N/17N

Medical Equipment Activation List

ISSUED FOR CONSTRUCTION May 21, 2014

REVISION: ADDENDUM 1 - 8N, 17N, 17W

6/20/14

**CANNONDESIGN**

*THIS PAGE INTENTIONALLY LEFT BLANK*

NOTE: THE SUMMARY OF EQUIPMENT ON THIS LIST IS COMPILED FROM USER GROUP MEETINGS, THE EQUIPMENT INVENTORY LIST PROVIDED BY THE VA MANHATTAN MEDICAL CENTER, THE SURVEY OF MEDICAL EQUIPMENT BY THE MEDICAL EQUIPMENT PLANNER, THE PROGRAM FOR DESIGN (PFD), AND THE PROJECT ROOM CONTENTS (PRC). DECISIONS ON RELOCATION OF EQUIPMENT ARE TO BE DETERMINED BY THE USERS. EQUIPMENT ITEMS LISTED FOR MULTIPLE ROOMS, E.G. TOILET, PUBLIC (QTY 2), ARE INTENDED TO BE DIVIDED BETWEEN THE QUANTITY OF ROOMS LISTED. PLEASE REVIEW THE MANUFACTURERS' TECHNICAL DATA SHEETS FOR ADDITIONAL SPECIFICATION INFORMATION; THOSE DATA SHEETS ARE PROVIDED AS A BASIS OF DESIGN ONLY.

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8N MED SURG															
STORAGE, EQUIPMENT															
8109C	M2055.5	SHELVING, WIRE	1	VV	5 WIRE CHROME SHELVES;MOBILE;SWIVEL& BRAKE CASTERS	60	24	69	112	FLOOR				YES	YES
8109C	M4250	PUMP, SYRINGE, INFUSION	4	VV	THE INFUSION SYRINGE PUMP ENSURES HIGHLY ACCURATEVOLUME DELIVERY AND CONSISTENT FLOW FOR SMALLVOLUMES (<50 ML) OF PHARMACOLOGIC AGENTS OR THICKFEEDING SOLUTIONS.IT SHALL BE SMALL , LIGHTWEIGHTCONSTRUCTION,MAKING IT TRANSPORTABLE.SHALL HAVEMENU-DRIVEN PROGRAMMING CAPABLE OF FLOW RATES(E.G. 0.1 OR 1.0 ML/HR) THAT ARE INTENDED FORLONG-TERM BEDSIDE USE AND/OR CRITICAL CARE PATIENTTRANSPORT, PLUNGER POSITIONING SENSOR, LCD DISPLAYFOR EASY VIEWING, VOLUME LIMIT PROGRAMMING TO SERVE AS A CONVENIENT CUE OF VOLUME OR DOSEDELIVERY COMPLETION AND MULTIPLE DELIVERY MODESFOR ALL APPLICATIONS REQUIRING PRECISELYCONTROLLED INFUSION RATES.THE INFUSION PUMP SHALLHAVE AUTOMATIC SYRINGE SIZE SENSING WHICH WILLGIVE THE FLEXIBILITY TO ACCEPT A WIDE RANGE OFSYRINGE SIZES (UP TO 60 ML) FROM DIFFERENTMANUFACTURERS.SHALL BE BATTERY POWERED/AC ADAPTER.VENDOR 1: SMITHS MEDICALCAGE 1: 0N0T8MODEL 1: MEDFUSION 3500VENDOR 2: B. BRAUN MEDICAL INCORPORATEDCAGE 2: 0Y7P7MODEL 2: PERFUSOR SPACE (8713030U)VENDOR 3: BAXTER HEALTHCARE CORP., IV SYSTEMS/GOVERNMENT SALESCAGE 3: 04687MODEL 3: SIGMA SPECTRUM 35700BAX	11	7	9	4	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8109C	M4250	PUMP, SYRINGE, INFUSION	1	VV	THE INFUSION SYRINGE PUMP ENSURES HIGHLY ACCURATEVOLUME DELIVERY AND CONSISTENT FLOW FOR SMALLVOLUMES (<50 ML) OF PHARMACOLOGIC AGENTS OR THICKFEEDING SOLUTIONS.IT SHALL BE SMALL , LIGHTWEIGHTCONSTRUCTION,MAKING IT TRANSPORTABLE.SHALL HAVEMENU-DRIVEN PROGRAMMING CAPABLE OF FLOW RATES(E.G. 0.1 OR 1.0 ML/HR) THAT ARE INTENDED FORLONG-TERM BEDSIDE USE AND/OR CRITICAL CARE PATIENTTRANSPORT, PLUNGER POSITIONING SENSOR, LCD DISPLAYFOR EASY VIEWING, VOLUME LIMIT PROGRAMMING TO SERVE AS A CONVENIENT CUE OF VOLUME OR DOSEDELIVERY COMPLETION AND MULTIPLE DELIVERY MODESFOR ALL APPLICATIONS REQUIRING PRECISELYCONTROLLED INFUSION RATES.THE INFUSION PUMP SHALLHAVE AUTOMATIC SYRINGE SIZE SENSING WHICH WILLGIVE THE FLEXIBILITY TO ACCEPT A WIDE RANGE OFSYRINGE SIZES (UP TO 60 ML) FROM DIFFERENTMANUFACTURERS.SHALL BE BATTERY POWERED/AC ADAPTER.VENDOR 1: SMITHS MEDICALCAGE 1: 0N0T8MODEL 1: MEDFUSION 3500VENDOR 2: B. BRAUN MEDICAL INCORPORATEDCAGE 2: 0Y7P7MODEL 2: PERFUSOR SPACE (8713030U)VENDOR 3: BAXTER HEALTHCARE CORP., IV SYSTEMS/GOVERNMENT SALESCAGE 3: 04687MODEL 3: SIGMA SPECTRUM 35700BAX	11	7	9	4	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8109C	M4655.2	STRETCHER, CHAIR	1	VV	400LB. CAPACITY;27.5"W PATIENT SURFACE;2" MATTRESSTHERMOFORMED ABS HOOD W/INTEGRATED STORAGE TRAYAND O2 BOTTLE HOLDER;HANDS FREE TRENDELENBURG;DUAL SIDE-MTD FOOT CONTROL HYDRAULICS;6"CASTERS;4-WHEEL RING BRAKE SYSTEM;SWING-DOWN SIDERAILS4 IV RECEPTACLES	30	76	34	0	FLOOR				YES	NO
8109C	M4801M	CHAIR, COMMODOE/SHOWER	2	VV	TRANSFER TUB BENCH;PADDED OPEN SEAT AND BACKALUMINUM CONSTRUCTION;ADJUSTABLE BACK RESTVARIABLE HEIGHT LEGS;LARGE SUCTION CUPS ON INSIDELEGS;SIDE ARM	30	17	38	20	FLOOR				YES	NO
8109C	U0005	DISPENSER	1	CC	Plastic bag dispenser to cover clean equipment. 12x12x42 (For covering clean equipment)				0	WALL				YES	NO
8109C	U0008.1	LIFTER, PATIENT	1	VV	500LB WORKING LOAD;BATTERY OPERATED;SLINGBATTERY INDICATOR;SERVICE METER;AUTOMATIC SAFETYCUT-OUT;LOW FRICTION CASTERS;REAR CASTERS W/BRAKESTO INCLUDE:1 EA MFA1000M-L LARGE CLIP BOX1 EA MFA1000M-M MEDIUM CLIP BOX1 EA MFA1000M-XL EXTRA LARGE CLIP BOX1 EA MFA1000M-XXL EXTRA EXTRA LARGE CLIP BOX1 EA MAA2090M-L LARGE AMPUTEE SLING1 EA MAA2090M-XL EXTRA LARGE AMPUTEE SLING 600lb CAPACITY for Bariatric usage.	46	32	61	154	FLOOR	24V, 5.5AH SEALED LEAD ACID BATTERY			YES	YES

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
STAFF LOUNGE															
8112C	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
8112C	A1066	MIRROR, FLOAT GLASS, WITH SS FRAME	1	CC	A HIGH QUALITY 1/4" POLISHED FLOAT GLASS MIRROR36X18, FRAMED IN A ONE-PIECE, BRIGHT POLISHED,STAINLESS STEEL CHANNEL FRAME WITH 90° MITEREDCORNERS.ALL EDGES OF THE MIRROR ARE PROTECTED BYABSORBING FILLER STRIPS.MIRROR HAS A GALVANIZEDSTEEL BACK WITH INTEGRAL HORIZONTAL HANGINGBRACKETS AND WALL HANGER FOR CONCEALEDMOUNTING.FOR MOUNTING ABOVE SINGLE WALL MOUNTEDLAVATORIES LOCATED IN TOILET AREAS, DOCTORSEXAMINATION OFFICES, ETC.MAY ALSO BE USED ABOVEDOUBLE LAVATORIES, EITHER WALL OR COUNTERTOPMOUNTED, FOUND IN RESTROOM AREAS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-165 1836VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 781-1836VENDOR 3: MEEK MANUFACTURING COMPANYCAGE 3: SR301MODEL 3: 1500 (18X36)	18	1	36	22	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8112C	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
8112C	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8112C	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8112C	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WV40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8112C	F2017.2	WASTE RECEPTACLE, STEP-ON	1	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
8112C	K1550	COFFEE BREWER	1	VV	12 cup drip coffee maker 9x10x13				0	COUNTER				YES	NO
8112C	K4665	OVEN, MICROWAVE, CONSUMER	1	VV	COUNTER MOUNTED MICROWAVE OVEN FOR AVERAGE DUTYUSE.THE EXTERIOR CABINET CAN BE METAL OR HEAVYDUTY IMPACT RESISTANT PLASTIC.THE OVEN DELIVERSINSTANT ENERGY FOR RAPID HEATING, DEFROSTING ORPRIME COOKING.THE OVEN HAS TOUCH PAD CONTROLS.DIGITAL TIMER, POWER LEVEL SELECTOR ANDPREPROGRAMMED SELECTORS FOR COMMONLY COOKEDITEMS.THIS OVEN IS COMMONLY FOUND IN STAFFLOUNGES.VENDOR 1: KITCHENAIDCAGE 1: AM179MODEL 1: KMS1555RWHVENDOR 2: SEARS, COMMERCIAL SALESCAGE 2: SR050MODEL 2: 02074062000VENDOR 3: GENERAL ELECTRIC APPLIANCESCAGE 3: 0G0L9MODEL 3: JE2160WF	24	20	14	44	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 15AMPS 2: 0WATTS 1: 1800WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8112C	R7250.1	REFRIGERATOR/FREEZER, UPRIGHT	1	VV	TOP FREEZER;WHITE;21 CUFT;NO ICE;REVERSIBLE DOOR	30	32	70	205	FLOOR	115V;60HZ;15A			YES	NO
DAY ROOM/FAMILY LOUNGE															
8113C	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
8113C	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8113C	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8113C	A5080.1	DISPENSER, PAPER TOWEL	1	CC	A recessed, C-fold, paper towel dispenser. Dispenser features: tumble lock; front hinged at bottom; and refill indicator slot. Minimum capacity 400 towels. For use in Behavior Health only.				0	WALL				YES	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8113C	A5210.1	TELEVISION ACCESSORY, BRACKET, WALL MOUNTED	1	CC	Bracket, Television, Wall Mounted, w/Adjust Arm 40W X 4D X 13H				0	WALL				YES	NO
8113C	F2017.2	WASTE RECEPTACLE, STEP-ON	1	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
8113C	K1550	COFFEE BREWER	1	VV	12 cup drip coffee maker 9x10x13				0	COUNTER				YES	NO
8113C	M0506.2	TELEVISION, FLAT PANEL LCD	1	VV	Color television with HD.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance. 42" wall mounted unit. 120V;60HZ				0	WALL				YES	NO
8113C	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	2	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THESYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMDVDCOMBO; 1,44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RE5MODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
PUBLIC TLT															
8115C	A1066	TABLE, DIAPER CHANGING, WALL MOUNTED	1	CC	WALL MOUNTED DIAPER CHANGING TABLE FOR USE IN PUBLIC RESTROOMS.UNIT IS A FOLD DOWN SYSTEM THAT FOLDS UPWARD WHEN NOT IN USE FOR MINIMAL INTRUSION INTO THE ROOM.CHANGING SURFACE IS CONTOURED AND HAS SAFETY BELT TO INSURE THAT AN INFANT WILL NOT FALL FROM THE TABLE.TABLE IS DESIGNED TO HOLD NO LESS THAN 250 POUNDS STATIC WEIGHT.CLOSED DIMENSIONS ARE APPROXIMATELY 21" H X 36" W X 5" D.OPEN DIMENSIONS VARY ACCORDING TO STYLE.ONE SYSTEMS OPEN DIMENSIONS ARE 15 " HIGH, 32" WIDE	35	5	22	26	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8115C	A1066	DISPOSAL, SANITARY NAPKIN, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL, SANITARY NAPKIN DISPOSAL.DISPOSAL FEATURES A FLIP-UP COVER, SECURED TO THE CONTAINER BY A HEAVY DUTY STAINLESS STEEL PIANO-HINGE.DISPOSAL MAY BE SECURED TO WALL OR TOILET PARTITION.FOR GENERAL PURPOSE USE IN FEMALE TOILET STALLS OR ROOMS AND UNI-SEX TOILET ROOMS.	9	5	12	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8115C	A1066	MIRROR, FLOAT GLASS, WITH SS FRAME	1	CC	A HIGH QUALITY 1/4" POLISHED FLOAT GLASS MIRROR36X18, FRAMED IN A ONE-PIECE, BRIGHT POLISHED,STAINLESS STEEL CHANNEL FRAME WITH 90° MITEREDCORNERS.ALL EDGES OF THE MIRROR ARE PROTECTED BYABSORBING FILLER STRIPS.MIRROR HAS A GALVANIZEDSTEEL BACK WITH INTEGRAL HORIZONTAL HANGINGBRACKETS AND WALL HANGER FOR CONCEALEDMOUNTING.FOR MOUNTING ABOVE SINGLE WALL MOUNTEDLAVATORIES LOCATED IN TOILET AREAS. DOCTORSEXAMINATION OFFICES, ETC.MAY ALSO BE USED ABOVEDOUBLE LAVATORIES, EITHER WALL OR COUNTERTOPMOUNTED, FOUND IN RESTROOM AREAS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-165 1836VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 781-1836VENDOR 3: MEEK MANUFACTURING COMPANYCAGE 3: SR301MODEL 3: 1500 (18X36)	18	1	36	22	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8115C	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZB9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8115C	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WPA40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8115C	A5109	GRAB BAR, 1-1/4" DIA., SS, 2 WALL, W/C ACCESSIBLE	1	CC	A 1-1/4" DIAMETER, SATIN FINISH STAINLESS STEEL,PEENED GRIPPING SURFACE, 2 WALL TOILET STALL/ROOM, GRAB BAR WITH CONCEALED MOUNTINGFLANGES.SNAP OVER FLANGES ARE PROVIDED TO CONCEALMOUNTING SCREWS.A SELECTION OF MOUNTING KITS ANDCONCEALED ANCHOR DEVICES ARE AVAILABLE FROM THEMANUFACTURERS FOR DIFFERENT TYPES OFINSTALLATIONS.GRAB BAR SHALL COMPLY WITHBARRIER-FREE ACCESSIBILITY GUIDELINES FORSTRUCTURAL STRENGTH. FOR TYPICAL WATER CLOSETAPPLICATIONS IN TOILET STALLS AND ROOMS WHERE ADA(AMERICAN'S WITH DISABILITIES ACT) REQUIREMENTS MUST BE MET.THESE MANUFACTURERS PROVIDE NUMEROUSCONFIGURATIONS OF THESE GRAB BARS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-5837.99VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 8322-059VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: SERIES 3700 TYPE 56 (3756-P)	57	3	3	7	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8115C	A5145.2	HOOK, GARMENT	1	CC	Fabricate hook units either of chromium plated brass with a satin finish, or stainless steel, using 6 mm (1/4 inch) minimum thick stock, with edges and corners rounded smooth to the thickness of the metal, or 3 mm (1/8 inch) minimum radius. Fabricate each unit as a double hook on a single shaft, integral with or permanently fastened to the wall flange, provided with concealed fastenings.				0	WALL				NO	NO
8115C	A5200	DISPENSER, TOILET TISSUE, SS, 2-ROLL, SURFACE MNTD	1	CC	A CONCEALED SURFACE MOUNTED, DOUBLE ROLL, SATINFINISH STAINLESS STEEL, TOILET TISSUEDISPENSER.UNIT ACCOMMODATES TWO STANDARD-CORETOILET TISSUE ROLLS THROUGH 5" DIAMETER.SPINDLES ARE CHROME PLATED PLASTIC WITH AHEAVY-DUTY INTERNAL SPRING AND TURN FREELY FORNON-CONTROLLED DELIVERY.FOR GENERAL PURPOSE USE INRESTROOMS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-27460VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 5234VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 1CF35	15	7	13	6	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8115C	F2017.2	WASTE RECEPTACLE, STEP-ON	1	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
BATH TLT (QTY 16)															
8116AN	A1066	MIRROR, FLOAT GLASS, WITH SS FRAME	16	CC	A HIGH QUALITY 1/4" POLISHED FLOAT GLASS MIRROR36X18, FRAMED IN A ONE-PIECE, BRIGHT POLISHED,STAINLESS STEEL CHANNEL FRAME WITH 90° MITEREDCORNERS.ALL EDGES OF THE MIRROR ARE PROTECTED BYABSORBING FILLER STRIPS.MIRROR HAS A GALVANIZEDSTEEL BACK WITH INTEGRAL HORIZONTAL HANGINGBRACKETS AND WALL HANGER FOR CONCEALEDMOUNTING.FOR MOUNTING ABOVE SINGLE WALL MOUNTEDLAVATORIES LOCATED IN TOILET AREAS, DOCTORSEXAMINATION OFFICES, ETC.MAY ALSO BE USED ABOVEDOUBLE LAVATORIES, EITHER WALL OR COUNTERTOPMOUNTED, FOUND IN RESTROOM AREAS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-165 1836VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 781-1836VENDOR 3: MEEK MANUFACTURING COMPANYCAGE 3: SR301MODEL 3: 1500 (18X36)	18	1	36	22	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8116AN	A5109.5	GRAB BAR	16	CC	Retractable Grab Bar for 2 1/2" and 3 1/2" (65 and 90mm) or deeper hollow walls with finished wall surfaces 1/4" to 1 1/2" (6 to 38mm) thick. Corrosion resistant stainless steel. Fastener will support 300 lbs (136 kg) load. Wall Mounted NOTE: Located at Water Closet				0	WALL				YES	NO
8116AN	A5075	DISPENSER, SOAP, DISPOSABLE	16	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREFILLABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU8MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8116AN	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	16	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8116AN	A5109.2	GRAB BAR	16	CC	36" GRAB BAR. A 1-1/4" DIAMETER, SATIN FINISH STAINLESS STEEL, PEENED GRIPPING SURFACE, GRAB BAR WITH CONCEALED MOUNTING FLANGES.SNAP OVER FLANGES ARE PROVIDED TO CONCEAL MOUNTING SCREWS.A SELECTION OF MOUNTING KITS AND CONCEALED ANCHOR DEVICES ARE AVAILABLE FROM THE MANUFACTURERS FOR DIFFERENT TYPES OF INSTALLATIONS.GRAB BAR SHALL COMPLY WITH BARRIER-FREE ACCESSIBILITY GUIDELINES FOR STRUCTURAL STRENGTH. FOR TYPICAL WATER CLOSET APPLICATIONS IN TOILET STALLS AND ROOMS WHERE ADA (AMERICAN'S WITH DISABILITIES ACT) REQUIREMENTS MUST BE MET. NOTE: LOCATED AT SHOWER				0	WALL				YES	NO
8116AN	A5109.6	GRAB BAR	16	CC	18" GRAB BAR. A 1-1/4" DIAMETER, SATIN FINISH STAINLESS STEEL, PEENED GRIPPING SURFACE, GRAB BAR WITH CONCEALED MOUNTING FLANGES.SNAP OVER FLANGES ARE PROVIDED TO CONCEAL MOUNTING SCREWS.A SELECTION OF MOUNTING KITS AND CONCEALED ANCHOR DEVICES ARE AVAILABLE FROM THE MANUFACTURERS FOR DIFFERENT TYPES OF INSTALLATIONS.GRAB BAR SHALL COMPLY WITH BARRIER-FREE ACCESSIBILITY GUIDELINES FOR STRUCTURAL STRENGTH. FOR TYPICAL WATER CLOSET APPLICATIONS IN TOILET STALLS AND ROOMS WHERE ADA (AMERICAN'S WITH DISABILITIES ACT) REQUIREMENTS MUST BE MET. NOTE: LOCATED AT SHOWER				0	WALL				YES	NO
8116AN	A5145.2	HOOK, GARMENT	16	CC	Fabricate hook units either of chromium plated brass with a satin finish, or stainless steel, using 6 mm (1/4 inch) minimum thick stock, with edges and corners rounded smooth to the thickness of the metal, or 3 mm (1/8 inch) minimum radius. Fabricate each unit as a double hook on a single shaft, integral with or permanently fastened to the wall flange, provided with concealed fastenings.				0	WALL				NO	NO
8116AN	A5165.1	SHELF	16	CC	RECESSED SHELF; 17W X 4D X 7H				0	WALL				YES	NO
8116AN	A5200	DISPENSER, TOILET TISSUE, SS, 2-ROLL, SURFACE MNTD	16	CC	A CONCEALED SURFACE MOUNTED, DOUBLE ROLL, SATINFINISH STAINLESS STEEL, TOILET TISSUEDISPENSER UNIT ACCOMMODATES TWO STANDARD-CORETOILET TISSUE ROLLS THROUGH 5" DIAMETER.SPINDLES ARE CHROME PLATED PLASTIC WITH AHEAVY-DUTY INTERNAL SPRING AND TURN FREELY FORNON-CONTROLLED DELIVERY.FOR GENERAL PURPOSE USE INRESTROOMS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-27460VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: B-234VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 1CF35	15	7	13	6	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8116AN	F2017.2	WASTE RECEPTACLE, STEP-ON	16	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
STORAGE, EQUIPMENT															
8117N	M4710	WHEELCHAIR, PATIENT, BARIATRIC, TRANSPORT, FOLDING	3	VV	FOLDING BARIATRIC PATIENT TRANSPORT WHEELCHAIRWITH A CAPACITY OF AT LEAST 500 LBS.CONSTRUCTEDWITH A HEAVY DUTY STEEL FRAME AND DOUBLE CROSSBRACE, WITH DETACHABLE ARMREST AND SWING AWAYFOOTRESTS.UNIT HAS HEAVY DUTY	36	49	39	70	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N)	BTU/HR: 0	YES	NO
8117N	U0005	DISPENSER	1	CC	Plastic bag dispenser to cover clean equipment. 12x12x42 (For covering clean equipment)				0	WALL				YES	NO
PCTC															
8120N	A1011.1	TELEPHONE	1	VV	This JSN is a place holder to indicate the placement of an instrument provided by a contractor as part of a total telephone system.				0	SPACE				NO	NO
8120N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
8120N	A5145.2	HOOK, GARMENT	1	CC	Fabricate hook units either of chromium plated brass with a satin finish, or stainless steel, using 6 mm (1/4 inch) minimum thick stock, with edges and corners rounded smooth to the thickness of the metal, or 3 mm (1/8 inch) minimum radius. Fabricate each unit as a double hook on a single shaft, integral with or permanently fastened to the wall flange, provided with concealed fastenings.				0	WALL				NO	NO
8120N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	1	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THESYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMDVDCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RESMODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8120N	M1801.1	COPY MACHINE	1	VV	Desktop copier.Unit features automatic paper size selection, automatic document feeder and sorter.The system also has zoom capabilities and automatic two-sided copying.For use where medium volume reproduction is required in the range of 30 to 40 copies per minute.				0	COUNTER				NO	NO
RESIDENT WORKROOM															
8121N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
SOILED UTILITY															
8122N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GQJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8122N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITs ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8122N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8122N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	1	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH AWHITE BAKED ENAMEL FINISH, PLASTIC ORACRYLIC.PROVIDED WITH WALL BRACKET TO FACILITATEMOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10.VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M644MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8122N	A5108.1	DISPOSAL CONTAINER, SHARPS W/BRACKET	1	VV	Waste Disposal Unit, Sharps 16W X 8D X 17H				0	WALL				YES	NO
8122N	F2017.1	WASTE RECEPTACLE, STEP-ON	1	VV	23 GAL.:MOBILE;LEAK-PROOF BASE;TIGHT FITTING LIDCALIFORNIA FIRE MARSHALL APPROVED; ROUNDED CORNERS; DEODORANT BLOCK HOLDER; COLOR:BEIGEWITH LINERS	20	17	33	18	FLOOR				YES	NO
8122N	F2020	CAN, TRASH, 44 GALLON	1	VV	FORTY FOUR (44) GALLON TRASH CAN, 32" HIGH X 24"DIAMETER, WITH LID.USED TO COLLECT AND TRANSPORTREFUSE FROM A POINT OF ORIGIN TO POINT OF DISPOSAL(EXAMPLE: FROM SOILED UTILITY OR A NURSING UNIT TOTHE TRASH COMPACTOR AT HOUSEKEEPING).VENDOR 1: MSC INDUSTRIAL SUPPLY CO.CAGE 1: 4J007MODEL 1: 66737016 W/09312661/09312745VENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 3W224 W/4W016 & 5W007VENDOR 3: GLOBAL INDUSTRIAL EQUIPMENTCAGE 3: 0JSZ8MODEL 3: WB239007GY W/WB239008/WB239003	24	24	39	22	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8122N	M3070.1	HAMPER, LINEN	4	VV	Hamper, Linen, Mobile, w/Lid 21W X 22D X 38H				0	FLOOR				YES	NO
8122N	U0006	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL	1	VV	CHEMOTHERAPY;PLASTIC;18 GAL;YELLOW;SLIDING LIDSOLD 5 TO A CASE	18	13	26	8	FLOOR				YES	NO
8122N	U0012	WASTE MANAGEMENT SYSTEM	1	VV	FOR DISPOSAL OF VERNACARE DISPOSABLE PRODUCTS Macerator will dispose a maximum of four disposable products such as bedpan liners, bottles etc, in a single cycle.	18	27	54	250	FLOOR	110V,12A,60HZ	4 GAL/MIN SUPPLY,25 LITERS/CYCLE,2-INCH DRAIN		YES	NO
8122N	U0018	WASTE RECEPTACLE, STEP-ON	1	VV	24 GAL;RED;FOOT PEDAL;RETAINER BAND;QUIET OPERATION;LINERS	15	15	30	36	FLOOR				YES	NO
HAC															
8123N	A5135.1	RACK, MOP/BROOM	1	CC	Shelf, Utility W/ Mop/Broom Holders, SS, Surf Mntd. DIMENSIONS 34W X 8D X 13H				0	WALL				YES	NO



ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8123N	F0500	CART, JANITOR'S	1	VV	JANITORS CART, 38" HIGH X 66" WIDE X 20" DEEP WITH25 GALLON HEAVY DUTY COLLECTION BAG.ACCOMMODATIONS FOR LOCKING VAULT AND JANITORIALSUPPORT ITEMS (BROOM, DUST PAN, CHEMICALS, MOP ANDMOP BUCKET; NOT INCLUDED).PRICE DIFFERENCE VERIFIED.VENDOR 1: RUBBERMAID COMMERCIAL PRODUCTS LLC.CAGE 1: 2H789MODEL 1: 9T72VENDOR 2: GLOBAL INDUSTRIAL EQUIPMENT.CAGE 2: 0JSZ8MODEL 2: WBB54619VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 2U670	22	49	54	77	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
STAFF TLT															
8124N	A1066	MIRROR, FLOAT GLASS, WITH SS FRAME	1	CC	A HIGH QUALITY 1/4" POLISHED FLOAT GLASS MIRROR36X18, FRAMED IN A ONE-PIECE, BRIGHT POLISHED,STAINLESS STEEL CHANNEL FRAME WITH 90° MITEREDCORNERS.ALL EDGES OF THE MIRROR ARE PROTECTED BYABSORBING FILLER STRIPS.MIRROR HAS A GALVANIZEDSTEEL BACK WITH INTEGRAL HORIZONTAL HANGINGBRACKETS AND WALL HANGER FOR CONCEALEDMOUNTING. FOR MOUNTING ABOVE SINGLE WALL MOUNTEDLAVATORIES LOCATED IN TOILET AREAS. DOCTORSEXAMINATION OFFICES, ETC. MAY ALSO BE USED ABOVEDOUBLE LAVATORIES, EITHER WALL OR COUNTERTOPMOUNTED, FOUND IN RESTROOM AREAS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-165 1836VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 781-1836VENDOR 3: MEEK MANUFACTURING COMPANYCAGE 3: SR301MODEL 3: 1500 (18X36)	18	1	36	22	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8124N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8124N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK: FRONT HINGED AT BOTTOM:AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8124N	A5109	GRAB BAR, 1-1/4" DIA., SS, 2 WALL, W/C ACCESSIBLE	1	CC	A 1-1/4" DIAMETER, SATIN FINISH STAINLESS STEEL,PEENED GRIPPING SURFACE, 2 WALL TOILET STALL/ROOM, GRAB BAR WITH CONCEALED MOUNTINGFLANGES.SNAP OVER FLANGES ARE PROVIDED TO CONCEALMOUNTING SCREWS.A SELECTION OF MOUNTING KITS ANDCONCEALED ANCHOR DEVICES ARE AVAILABLE FROM THEMANUFACTURERS FOR DIFFERENT TYPES OFINSTALLATIONS.GRAB BAR SHALL COMPLY WITHBARRIER-FREE ACCESSIBILITY GUIDELINES FORSTRUCTURAL STRENGTH. FOR TYPICAL WATER CLOSETAPPLICATIONS IN TOILET STALLS AND ROOMS WHERE ADA(AMERICAN'S WITH DISABILITIES ACT) REQUIREMENTS MUST BE MET.THESE MANUFACTURERS PROVIDE NUMEROUSCONFIGURATIONS OF THESE GRAB BARS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-5837.99VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 8322-059VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: SERIES 3700 TYPE 56 (3756-P)	57	3	3	7	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8124N	A5145.2	HOOK, GARMENT	1	CC	Fabricate hook units either of chromium plated brass with a satin finish, or stainless steel, using 6 mm (1/4 inch) minimum thick stock, with edges and corners rounded smooth to the thickness of the metal, or 3 mm (1/8 inch) minimum radius. Fabricate each unit as a double hook on a single shaft, integral with or permanently fastened to the wall flange, provided with concealed fastenings.				0	WALL				NO	NO
8124N	A5200	DISPENSER, TOILET TISSUE, SS, 2-ROLL, SURFACE MNTD	1	CC	A CONCEALED SURFACE MOUNTED, DOUBLE ROLL, SATINFINISH STAINLESS STEEL, TOILET TISSUEDISPENSER.UNIT ACCOMMODATES TWO STANDARD-CORETOILET TISSUE ROLLS THROUGH 5" INDIAMETER.SPINDLES ARE CHROME PLATED PLASTIC WITH AHEAVY-DUTY INTERNAL SPRING AND TURN FREELY FORNON-CONTROLLED DELIVERY.FOR GENERAL PURPOSE USE INRESTROOMS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-27460VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 8234VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 1CF35	15	7	13	6	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8124N	F2017.2	WASTE RECEPTACLE, STEP-ON	1	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
CLEAN LINEN															

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8131N	F0515.1	CART, LINEN	1	VV	LINEN/SUPPLY;WIRE;STEM CASTERS;ENCLOSURES;COVERTO INCLUDE THE FOLLOWING:1 EA N566BC;4 WIRE SHELVES;4 POSTS;PLASTIC SPLIT SLEEVES;DONUT BUMPERS;2 EA 5M STEM CASTERS;2 EA 5MB CASTERS WITH BRAKES, CHROME5 EA EP56C-24X63 WIRE ENCLOSURE PANEL FOR SIDES AND BACK1 EA 24X60X62UC UNCOATED CART COVER, ZIPPER,WHITE	60	24	68	0	FLOOR				YES	NO
8131N	F0515.1	CART, LINEN	1	VV	LINEN/SUPPLY;WIRE;STEM CASTERS;ENCLOSURES;COVERTO INCLUDE THE FOLLOWING:1 EA N566BC;4 WIRE SHELVES;4 POSTS;PLASTIC SPLIT SLEEVES;DONUT BUMPERS;2 EA 5M STEM CASTERS;2 EA 5MB CASTERS WITH BRAKES, CHROME5 EA EP56C-24X63 WIRE ENCLOSURE PANEL FOR SIDES AND BACK1 EA 24X60X62UC UNCOATED CART COVER, ZIPPER,WHITE	60	24	68	0	FLOOR				YES	YES
ANTE ROOM (QTY 2)															
8133AN	A5075	DISPENSER, SOAP, DISPOSABLE	2	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8133AN	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	2	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8133AN	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	2	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL SINGLE-FOLD, PAPER TOWEL DISPENSERDISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8133AN	E0963	CART, ISOLATION	2	VV	YELLOW;4 DRAWER;W/TOP MAT/DRAWERLABEL HOLDERSTO INCLUDE:-1 EA 12093 CLEAR TOP MAT-DRAWER LABEL HOLDERSPRE-DRILLED ACCESSORY POSITIONS #2 AND #4 ON BOTHEND PANELS	31	24	43	0	FLOOR				YES	NO
8133AN	F2017.2	WASTE RECEPTACLE, STEP-ON	2	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
1 BED MED SURG ISO															
8133N	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
8133N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
8133N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8133N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8133N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL SINGLE-FOLD, PAPER TOWEL DISPENSERDISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8133N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	1	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH AWHITE BAKED ENAMEL FINISH, PLASTIC ORACRYLIC.PROVIDED WITH WALL BRACKET TO FACILITATEMOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10.VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M644MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8133N	A5108.1	DISPOSAL CONTAINER, SHARPS W/BACKET	1	VV	Waste Disposal Unit, Sharps 16W X 8D X 17H				0	WALL				YES	NO
8133N	A5210.1	TELEVISION ACCESSORY, BRACKET, WALL MOUNTED	1	CC	Bracket, Television, Wall Mounted, w/Adjust Arm 40W X 4D X 13H				0	WALL				YES	NO
8133N	F2017.1	WASTE RECEPTACLE, STEP-ON	1	VV	23 GAL.;MOBILE;LEAK-PROOF BASE;TIGHT FITTING LIDCALIFORNIA FIRE MARSHALL APPROVED; ROUNDED CORNERS; DEODORANT BLOCK HOLDER; COLOR:BEIGEWITH LINERS	20	17	33	18	FLOOR				YES	NO
8133N	F2700	READER, BAR CODE, HAND HELD, WITH INTERFACE	1	VV	HAND HELD LASER BAR CODE READER WITH COMPUTERINTERFACE.USED FOR AUTOMATED INVENTORY, USING BARCODE STICKERS / LABELS.CONVENIENCE OUTLET REQUIREDAT POINT OF USE.THESE UNITS ARE WIRELESS BAR CODE SCANNERS ANDHAND HELD COMPUTERS.MANY DIFFERENT MODELS ANDAPPLICATIONS ARE AVAILABLE FROM THESE VENDORS.VENDOR 1: UNITECH AMERICA, INC.CAGE 1: 4QR03MODEL 1: PT600 (PT600HD-0L00B)VENDOR 2: SCAN SOURCE, INC.CAGE 2: AM282MODEL 2: SF51 W/87-221-001/235-195-002VENDOR 3: BARCODES, INC.CAGE 3: AM283MODEL 3: DOLPHIN 9500 9500LWP431C30KIT	4	10	2	2	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8133N	M0506.2	TELEVISION, FLAT PANEL LCD	1	VV	Color television with HD.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance. 42" wall mounted unit. 120V/60HZ				0	WALL				YES	NO
8133N	M0750	FLOWMETER, AIR, CONNECT W/50 PSISUPPLY	2	VV	AIR FLOWMETER.UNIT HAS A STAINLESS STEEL NEEDLEVALVE WITH CLEAR FLOWTUBE FOR CONNECTION TO 50 PSIAIR OUTLET FROM CENTRAL PIPELINE SYSTEM.REQUIRESTHE APPROPRIATE ADAPTER FOR CONNECTION TO THE WALLOUTLET AND FITTING TO CONNECT TO TUBING.DATABASEPRICES REFLECT FITTINGS WITH AN ATTACHED DISSPOWER OUTLET.OTHER OUTLET AND ADAPTERCONFIGURATIONS ARE AVAILABLE.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 7702-1266-907VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 2: 11978MODEL 2: 15002-03PVENDOR 3: WESTERN MEDICACAGE 3: SR258MODEL 3: FM104	4	4	10	3	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (K) VACUUM AND LP AIR (30 PSIG) MAXIMUMUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8133N	M0755	FLOWMETER, OXYGEN, LOW FLOW	2	VV	OXYGEN FLOWMETER.CONSISTS OF A CLEAR CRYSTALFLOWTUBE CALIBRATED TO 3.5 OR 8 LPM DEPENDING ONMANUFACTURER.FOR OXYGEN REGULATION IN HOSPITALSETTINGS.DATABASE PRICING INCLUDES DISS FITTINGAND DISS POWER OUTLET AND WALL ADAPTER.OTHERFITTING AND ADAPTER CONFIGURATIONS ARE AVAILABLE.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 6702/1306/921VENDOR 2: PRECISION MEDICALCAGE 2: 0UK20MODEL 2: 1MFA4001/1MFA4004VENDOR 3: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 3: 11978MODEL 3: 08008-03	4	4	10	1	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (A) OXYGENUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8133N	M0765	REGULATOR, VACUUM	4	VV	VACUUM PRESSURE REGULATOR FOR CONNECTION TOCENTRAL PIPED VACUUM SYSTEM.STANDARD DISPLAY SCALEIS GRADUATED AT LEAST FROM 0 TO 200 MM HG OFVACUUM.DISPLAYS ON SPECIALIZED REGULATORS MAYCOVER OTHER VACUUM RANGES.REGULATOR TYPE(CONTINUOUS, INTERMITTENT,CONTINUOUS/INTERMITTENT, SURGICAL, PEDIATRIC,THORACIC, ETC.) AS REQUIRED.TO BE USED INDELIVERY, NEONATAL, PEDIATRICS OR ANY AREA WHERESECTION IS REQUIRED.DATABASE PRICING REFLECTSCONTINUOUS REGULATORS GRADUATED TO 200 MM HG WITHA FULL LINE VACUUM SELECTION SWITCH AND DISSCONFIGURED INLETS AND OUTLETS.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 6702-1224-907VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 2: 11978MODEL 2: VACUTRON #22-12-1111VENDOR 3: PRECISION MEDICALCAGE 3: 0UK20MODEL 3: PM3114	3	4	6	1	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (B) VACUUMUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8133N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	1	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THE SYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROM/DVDCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RE5MODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8133N	M4116	MONITOR, VITAL SIGNS	1	VV	ELECTRONIC SPHYGMOMANOMETER.LCD DISPLAYSNON-INVASIVE BLOOD PRESSURE, PULSE RATE ANDTEMPERATURE.USED IN HOSPITALS AND CLINICS.INCLUDESAN OPTIONAL MOBILE STAND.DIMENSIONS ARE FOR MONITOR ONLY.VENDOR 1: WELCH-ALLYNCAGE 1: 05CJ3MODEL 1: 53070-E1G & 008-0835-01VENDOR 2: DATASCOPE CORPORATIONCAGE 2: 1L753MODEL 2: DUO W/TEMP PROBE & STANDVENDOR 3: CARDINAL HEALTH PRODUCTS GROUP, CARDINAL HEALTHCAGE 3: 07TA6MODEL 3: 420TB-E1-M	10	6	10	6	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8133N	M4250	PUMP, SYRINGE, INFUSION	1	VV	THE INFUSION SYRINGE PUMP ENSURES HIGHLY ACCURATE VOLUME DELIVERY AND CONSISTENT FLOW FOR SMALL VOLUMES (<50 ML) OF PHARMACOLOGIC AGENTS OR THICK FEEDING SOLUTIONS. IT SHALL BE SMALL, LIGHTWEIGHT CONSTRUCTION, MAKING IT TRANSPORTABLE. SHALL HAVE MENU-DRIVEN PROGRAMMING CAPABLE OF FLOW RATES (E.G. 0.1 OR 1.0 ML/HR) THAT ARE INTENDED FOR LONG-TERM BEDSIDE USE AND/OR CRITICAL CARE PATIENT TRANSPORT. PLUNGER POSITIONING SENSOR, LCD DISPLAY FOR EASY VIEWING, VOLUME LIMIT PROGRAMMING TO SERVE AS A CONVENIENT CUE OF VOLUME OR DOSE DELIVERY COMPLETION, AND MULTIPLE DELIVERY MODES FOR ALL APPLICATIONS REQUIRING PRECISELY CONTROLLED INFUSION RATES. THE INFUSION PUMP SHALL HAVE AUTOMATIC SYRINGE SIZE SENSING WHICH WILL GIVE THE FLEXIBILITY TO ACCEPT A WIDE RANGE OF SYRINGE SIZES (UP TO 60 ML) FROM DIFFERENT MANUFACTURERS. SHALL BE BATTERY POWERED/AC ADAPTER. VENDOR 1: SMITHS MEDICAL CAGE 1: 0N0T8MODEL 1: MEDFUSION 3500VENDOR 2: B. BRAUN MEDICAL INCORPORATEDCAGE 2: 0Y7P7MODEL 2: PERFUSOR SPACE (8713030U)VENDOR 3: BAXTER HEALTHCARE CORP., IV SYSTEMS/GOVERNMENT SALESCAGE 3: 04687MODEL 3: SIGMA SPECTRUM 35700BAX	11	7	9	4	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8133N	M4255	STAND, IV, ADJUSTABLE	1	VV	ADJUSTABLE IV STAND WITH 4-HOOK ARRANGEMENT. STAND HAS STAINLESS STEEL CONSTRUCTION WITH HEAVY WEIGHT BASE. IT ADJUSTS FROM 66 INCHES TO 100 INCHES AND IS MOUNTED ON CONDUCTIVE RUBBER, BALL BEARING, SWIVEL CASTERS. STAND IS USED FOR ADMINISTERING INTRAVENOUS SOLUTIONS. VENDOR 1: BLICKMAN, INCORPORATEDCAGE 1: 4M773MODEL 1: 7793SS-4VENDOR 2: CLINTON INDUSTRIES, INC.CAGE 2: 0UKA5MODEL 2: IVS-354VENDOR 3: PEDIGO PRODUCTS, INC.CAGE 3: 13180MODEL 3: P-1576-4	24	24	108	35	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8133N	M5030.1	STOOL, REVOLVING	1	VV	5 CASTER STOOL PNEUMATIC; 14" ROUND CUSHIONED SEAT ONE TOUCH ADJUSTMENT FROM 18"-23"; 2.5" CASTERS	14	14	27	25	FLOOR				YES	NO
8133N	M7010	BED, PATIENT, ELECTRIC	1	VV	ELECTRIC PATIENT BED INCLUDES SAFETY SIDE RAILS, IV ROD INSERTION SLOTS AND PATIENT CONTROL PANEL FOR A FULL RANGE OF POSITIONS. FOUR WHEEL LOCKING BRAKES. OVERALL BED HEIGHT ADJUSTABLE BETWEEN 16.5 AND 31 INCHES. PRICES CAN VARY BASED ON OPTIONS SELECTED. NO OTHERS. SOURCE(S) IDENTIFIED WITH THIS DESCRIPTION/SPECIFICATION. VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 1: 45044MODEL 1: STRYKER S3 (#3002S3)VENDOR 2: HILL-ROM CO., INC., A HILLENBRAND INDUSTRY CAGE 2: 28713MODEL 2: VERSACARE BED VC200VENDOR 3: CAGE 3: MODEL 3:	42	95	37	335	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 10AMPS 2: 0WATTS 1: 1200WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8133N	M7040	TABLE, OVERBED	1	VV	OVERBED TABLE. ADJUSTABLE HEIGHT TABLE CONSTRUCTED OF HEAVY GAUGE STEEL MOUNTED ON 2" DIAMETER TWIN SWIVEL CASTERS WITH BUMPER CAPS. TABLE TOP IS CONSTRUCTED WITH A HIGH PRESSURE PLASTIC LAMINATED SURFACE THAT RESISTS CHIPPING, SCRATCHING, AND STAINING. IT INCLUDES A VANITY TRAY AND A MIRROR. TABLE IS DESIGNED FOR USE OVER BED, WHEELCHAIR OR LARGE CHAIR. VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 1: 45044MODEL 1: SINGLE TOP W/VANITYVENDOR 2: HILL-ROM CO., INC., A HILLENBRAND INDUSTRY CAGE 2: 28713MODEL 2: OBT220 W/LOW BASEVENDOR 3: GF HEALTH PRODUCTS, INC.CAGE 3: 1JY01MODEL 3: A797002 W/VANITY MIRROR	18	36	46	58	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8133N	U0008	LIFTER, PATIENT	1	VC	An overhead rail system specifically designed for patient lifting and movement for a single bed patient room. The system will consist of recessed or ceiling mounted primary and secondary rails, lift motor with rolling carriage, patient harness or seat, and a hand controller or control box with charger. System will facilitate lifting and movement of patient to and from bed to gurney, chair or other requirement. Minimum lift capability is 550 pounds. Custom design of track layout by manufacturer is essential to meet individual facility requirements.				0	CEILING				YES	NO
8133N	U0008.3	LIFTER, PATIENT, ACCESSORY, SLING	1	VV	Variant 4 non-optional Lifter Sling 47X44X56				0	SPACE				NO	NO
8133N	U0030	WALL CHANNEL	1	VV	19"; Aluminum; seismic wall channel with hardware				0	WALL				YES	NO
8133N	U0031	COMPUTER ACCESSORY	1	VV	Mounting assembly and accessories. Includes CPU holder, (1) variable height, adjustable monitor arm supporting up to 40lbs. 180 degree stop mechanism to prevent wall damage, monitor extends 26.25 inches from wall.				0	WALL				YES	NO
8133N	U0032	COMPUTER ACCESSORY	1	VV	Undercounter Mounting assembly for computer keyboard and mouse; ergonomic keyboard holder with tilt and height adjustable; powder coated steel.				0	COUNTER				YES	NO
8133N	U0034	COMPUTER ACCESSORY, BRACKET, WALL	1	VV	27" track to support computer accessories including monitor, keyboard arm, CPU holder. Designed to be mounted to a wall stud or mounting board. Includes channel for cables and covers for sealed cable management.				0	WALL				YES	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8133N	U0036	BRACKET	2	VV	Rail mount bracket mount for regulators, wall bottle bracket, disposable bottle brackets, and blenders 5X1X6				0	WALL				YES	NO
8133N	U0039	HEADWALL, PATIENT, ACCESSORY	1	VV	Basket; Mounts on Headwall Accessory Rail; 6 1/2" W X 4"D X 5"H; Functional loading to 5 lbs.; (Meant to hold Bar Code Reader)				0	SPACE				YES	NO
1 BED ISOLATION, BARIATRIC															
8134N	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
8134N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
8134N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER. ONE-HANDED DISPENSING OPERATION. DESIGNED TO ACCOMMODATE DISPOSABLE SOAP CARTRIDGE AND VALVE. BASED ON INFECTION CONTROL STANDARDS ONLY SOAP DISPENSERS REFILLED BY MEANS OF A CARTRIDGE OR REPLACEABLE BAG ARE TO BE USED. VENDOR 1: STAPLES, INC. CAGE 1: 0ZBU9 MODEL 1: 603696 VENDOR 2: JACOBS GARDNER CAGE 2: 7R579 MODEL 2: GOJ515006 VENDOR 3: W. W. GRAINGER, INC. CAGE 3: 25795 MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8134N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZER DISPENSER. FOR USE THROUGHOUT A HEALTHCARE FACILITY. UNIT DOES NOT INCLUDE THE SANITIZING LIQUID. UNITS ARE BATTERY OPERATED. VENDOR 1: STAPLES, INC. CAGE 1: 0ZBU9 MODEL 1: TFX TOUCH FREE DISPENSER VENDOR 2: W. W. GRAINGER, INC. CAGE 2: 25795 MODEL 2: 1PKP9 VENDOR 3: CAGE 3: AM490 MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8134N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL, SINGLE-FOLD, PAPER TOWEL DISPENSER. DISPENSER FEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM; AND REFILL INDICATOR SLOT. MINIMUM CAPACITY 400 SINGLE-FOLD PAPER TOWELS. FOR GENERAL PURPOSE USE THROUGHOUT THE FACILITY. VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATED CAGE 1: 89488 MODEL 1: B-263 VENDOR 2: BRADLEY CORPORATION CAGE 2: 01055 MODEL 2: 251-15 VENDOR 3: AMERICAN SPECIALTIES, INC. CAGE 3: 0WP40 MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8134N	A5107	DISPENSER, GLOVE, SURGICAL/EXAMINATION, WALL MNTD	1	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING. FABRICATED OF EITHER COLD ROLLED STEEL WITH A WHITE BAKED ENAMEL FINISH, PLASTIC OR ACRYLIC. PROVIDED WITH WALL BRACKET TO FACILITATE MOUNTING AND DEMOUNTING. MODEL 2 IS NORMALLY SOLD BY A CASE OF 10. VENDOR 1: LAB SAFETY SUPPLY, INCORPORATED CAGE 1: 6M644 MODEL 1: 83818 VENDOR 2: S-CURVE TECHNOLOGIES CAGE 2: AM044 MODEL 2: EBD-1000 VENDOR 3: SURGICAL SUPPLY SERVICE CAGE 3: AM045 MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8134N	A5108.1	DISPOSAL CONTAINER, SHARPS W/BRACKET	1	VV	Waste Disposal Unit, Sharps 16W X 8D X 17H				0	WALL				YES	NO
8134N	A5210.1	TELEVISION ACCESSORY, BRACKET, WALL MOUNTED	1	CC	Bracket, Television, Wall Mounted, w/Adjust Arm 40W X 4D X 13H				0	WALL				YES	NO
8134N	F2017.1	WASTE RECEPTACLE, STEP-ON	1	VV	23 GAL.; MOBILE; LEAK-PROOF BASE; TIGHT FITTING LID CALIFORNIA FIRE MARSHALL APPROVED; ROUNDED CORNERS; DEODORANT BLOCK HOLDER; COLOR: BEIGE WITH LINERS	20	17	33	18	FLOOR				YES	NO
8134N	F2700	READER, BAR CODE, HAND HELD, WITH INTERFACE	1	VV	HAND HELD LASER BAR CODE READER WITH COMPUTER INTERFACE. USED FOR AUTOMATED INVENTORY, USING BARCODE STICKERS / LABELS. CONVENIENCE OUTLET REQUIRED AT POINT OF USE. THESE UNITS ARE WIRELESS BAR CODE SCANNERS AND HAND HELD COMPUTERS. MANY DIFFERENT MODELS AND APPLICATIONS ARE AVAILABLE FROM THESE VENDORS. VENDOR 1: UNITECH AMERICA, INC. CAGE 1: 4QR03 MODEL 1: PT600 (PT600HD-0L00B) VENDOR 2: SCAN SOURCE, INC. CAGE 2: AM282 MODEL 2: SF51 W/87-221-001/235-195-002 VENDOR 3: BARCODES, INC. CAGE 3: AM283 MODEL 3: DOLPHIN 9500 9500LWP431C30KIT	4	10	2	2	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8134N	M0506.2	TELEVISION, FLAT PANEL LCD	1	VV	Color television with HD. Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance. 42" wall mounted unit. 120V/60HZ				0	WALL				YES	NO
8134N	M0750	FLOWMETER, AIR, CONNECT W/50 PSISUPPLY	2	VV	AIR FLOWMETER. UNIT HAS A STAINLESS STEEL NEEDLE VALVE WITH CLEAR FLOW TUBE FOR CONNECTION TO 50 PSIAIR OUTLET FROM CENTRAL PIPELINE SYSTEM. REQUIRES THE APPROPRIATE ADAPTER FOR CONNECTION TO THE WALL OUTLET AND FITTING TO CONNECT TO TUBING. DATABASE PRICES REFLECT FITTINGS WITH AN ATTACHED DISPOPOWER OUTLET OTHER OUTLET AND ADAPTER CONFIGURATIONS ARE AVAILABLE. VENDOR 1: OHIO MEDICAL CORPORATION CAGE 1: 45874 MODEL 1: 7702-1266-907 VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC. CAGE 2: 11978 MODEL 2: 15002-03P VENDOR 3: WESTERN MEDICA CAGE 3: SR258 MODEL 3: FM104	4	4	10	3	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (K) VACUUM AND LP AIR (30 PSIG) MAXIMUM UTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8134N	M0755	FLOWMETER, OXYGEN, LOW FLOW	2	VV	OXYGEN FLOWMETER.CONSISTS OF A CLEAR CRYSTALFLOWTUBE CALIBRATED TO 3.5 OR 8 LPM DEPENDING ONMANUFACTURER.FOR OXYGEN REGULATION IN HOSPITALSETTINGS.DATABASE PRICING INCLUDES DISS FITTINGAND DISS POWER OUTLET AND WALL ADAPTER.OTHERFITTING AND ADAPTER CONFIGURATIONS ARE AVAILABLE.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 6702/1306/921VENDOR 2: PRECISION MEDICALCAGE 2: 0UK20MODEL 2: 1MFA4001/1MFA4004VENDOR 3: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 3: 11978MODEL 3: 08008-03	4	4	10	1	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (A) OXYGENUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8134N	M0765	REGULATOR, VACUUM	4	VV	VACUUM PRESSURE REGULATOR FOR CONNECTION TOCENTRAL PIPED VACUUM SYSTEM.STANDARD DISPLAY SCALEIS GRADUATED AT LEAST FROM 0 TO 200 MM HG OFVACUUM.DISPLAYS ON SPECIALIZED REGULATORS MAYCOVER OTHER VACUUM RANGES.REGULATOR TYPE(CONTINUOUS, INTERMITTENT,CONTINUOUS/INTERMITTENT, SURGICAL, PEDIATRIC,THORACIC, ETC.) AS REQUIRED.TO BE USED INDELIVERY, NEONATAL, PEDIATRICS OR ANY AREA WHERESUCTION IS REQUIRED.DATABASE PRICING REFLECTSCONTINUOUS REGULATORS GRADUATED TO 200 MM HG WITHA FULL LINE VACUUM SELECTION SWITCH AND DISSCONFIGURED INLETS AND OUTLETS.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 6702-1224-907VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 2: 11978MODEL 2: VACUTRON #22-12-1111VENDOR 3: PRECISION MEDICALCAGE 3: 0UK20MODEL 3: PM3114	3	4	6	1	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (B) VACUUMUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8134N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	1	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THE SYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMDVDCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RESMODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8134N	M4116	MONITOR, VITAL SIGNS	1	VV	ELECTRONIC SPHYGMOMANOMETER.LCD DISPLAYSNON-INVASIVE BLOOD PRESSURE, PULSE RATE ANDTEMPERATURE.USED IN HOSPITALS AND CLINICS.INCLUDESAN OPTIONAL MOBILE STAND.DIMENSIONS ARE FOR MONITOR ONLY.VENDOR 1: WELCH-ALLYNLCAGE 1: 05CJ3MODEL 1: 530T0-E1G & 008-0835-01VENDOR 2: DATASCOPE CORPORATIONCAGE 2: 1L753MODEL 2: DUO W/TEMP PROBE & STANDVENDOR 3: CARDINAL HEALTH PRODUCTS GROUP, CARDINAL HEALTHCAGE 3: 07TA6MODEL 3: 420TB-E1-M	10	6	10	6	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8134N	M4250	PUMP, SYRINGE, INFUSION	1	VV	THE INFUSION SYRINGE PUMP ENSURES HIGHLY ACCURATEVOLUME DELIVERY AND CONSISTENT FLOW FOR SMALLVOLUMES (<50 ML) OF PHARMACOLOGIC AGENTS OR THICKFEEDING SOLUTIONS.IT SHALL BE SMALL, LIGHTWEIGHTCONSTRUCTION,MAKING IT TRANSPORTABLE.SHALL HAVEMENU-DRIVEN PROGRAMMING CAPABLE OF FLOW RATES(E.G. 0.1 OR 1.0 ML/HR) THAT ARE INTENDED FORLONG-TERM BEDSIDE USE AND/OR CRITICAL CARE PATIENTTRANSPORT, PLUNGER POSITIONING SENSOR, LCD DISPLAYFOR EASY VIEWING, VOLUME LIMIT PROGRAMMING TOSERVE AS A CONVENIENT CUE OF VOLUME OR DOSEDELIVERY COMPLETION AND MULTIPLE DELIVERY MODESFOR ALL APPLICATIONS REQUIRING PRECISELYCONTROLLED INFUSION RATES.THE INFUSION PUMP SHALLHAVE AUTOMATIC SYRINGE SIZE SENSING WHICH WILLGIVE THE FLEXIBILITY TO ACCEPT A WIDE RANGE OFSYRINGE SIZES (UP TO 60 ML) FROM DIFFERENTMANUFACTURERS.SHALL BE BATTERY POWERED/AC ADAPTER.VENDOR 1: SMITHS MEDICALCAGE 1: 0N0T8MODEL 1: MEDFUSION 3500VENDOR 2: B. BRAUN MEDICAL INCORPORATEDCAGE 2: 0Y7P7MODEL 2: PERFUSOR SPACE (8713030U)VENDOR 3: BAXTER HEALTHCARE CORP., IV SYSTEMS/GOVERNMENT SALESCAGE 3: 04687MODEL 3: SIGMA SPECTRUM 35700BAX	11	7	9	4	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8134N	M4255	STAND, IV, ADJUSTABLE	1	VV	ADJUSTABLE IV STAND WITH 4-HOOK ARRANGEMENT.STANDHAS STAINLESS STEEL CONSTRUCTION WITH HEAVY WEIGHTBASE.IT ADJUSTS FROM 66 INCHES TO 100 INCHES ANDIS MOUNTED ON CONDUCTIVE RUBBER, BALL BEARING,SWIVEL CASTERS.STAND IS USED FOR ADMINISTERINGINTRAVENOUS SOLUTIONS.VENDOR 1: BLICKMAN, INCORPORATEDCAGE 1: 4M773MODEL 1: 7793SS-4VENDOR 2: CLINTON INDUSTRIES, INC.CAGE 2: 0UKA5MODEL 2: IVS-354VENDOR 3: PEDIGO PRODUCTS, INC.CAGE 3: 13180MODEL 3: P-1576-4	24	24	108	35	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8134N	M5030.1	STOOL, REVOLVING	1	VV	5 CASTER STOOL PNEUMATIC;14" ROUND CUSHIONED SEATONE TOUCH ADJUSTMENT FROM 18"-23";2.5"CASTERS	14	14	27	25	FLOOR				YES	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8134N	M7009	BED, PATIENT, ELECTRIC, BARIATRIC	1	VV	ELECTRIC BARIATRIC PATIENT BED WITH A WEIGHTCAPACITY OF 1,000 POUNDS.THE BED WILL INCLUDE ANELECTRONIC BED SCALE TO ALLOW PATIENT TO BEWEIGHED WHILE IN BED.BED INCLUDES MATTRESS(FOAMOR AIR), SAFETY SIDE RAILS, IV ROD INSERTION SLOTSAND PATIENT CONTROL PANEL FOR A FULL RANGE OFPOSITIONS ALSO INCLUDES FOUR CASTERS WITH LOCKINGBRAKES.PRICE MAY VARY BASED ON OPTIONS AND ACCESSORIESSELECTED.MODEL 2 IS PROVIDED AS A RENTAL AT THEPRESENT TIME.VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 1: 45044MODEL 1: BARI 10-A 2239 W/2236-000-000VENDOR 2: HILL-ROM CO., INC. A HILLENBRAND INDUSTRYCAGE 2: 28713MODEL 2: EXCEL CARE ES BEDVENDOR 3:CAGE 3: 24285MODEL 3: ELITE 5154	54	98	34	700	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 6AMPS 2: 0WATTS 1: 720WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8134N	M7040	TABLE, OVERBED	1	VV	OVERBED TABLE.ADJUSTABLE HEIGHT TABLE CONSTRUCTEDOF HEAVY GAUGE STEEL.MOUNTED ON 2" DIAMETER TWINSWIVEL CASTERS WITH BUMPER CAPS.TABLE TOP ISCONSTRUCTED WITH A HIGH PRESSURE PLASTIC LAMINATEDSURFACE THAT RESISTS CHIPPING, SCRATCHING, ANDSTAINING.IT INCLUDES A VANITY TRAY AND A MIRROR.TABLE IS DESIGNED FOR USE OVER BED. WHEELCHAIR ORLARGE CHAIR.VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 1: 45044MODEL 1: SINGLE TOP W/VANITYVENDOR 2: HILL-ROM CO., INC. A HILLENBRAND INDUSTRYCAGE 2: 28713MODEL 2: OBT220 W/LOW BASEVENDOR 3: GF HEALTH PRODUCTS, INC.CAGE 3: 1JY01MODEL 3: A797002 W/VANITY &MIRROR	18	36	46	58	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8134N	U0008	LIFTER, PATIENT	1	VC	An overhead rail system specifically designed for patient lifting and movement for a single bed patient room. The system will consist of recessed or ceiling mounted primary and secondary rails, lift motor with rolling carriage,patient harness or seat, and a hand controller or control box with charger.System will facilitate lifting and movement of patient to and from bed to gurney, chair or other requirement.Minimum lift capability is 550 pounds.Custom design of track layout by manufacturer is essential to meet individual facility requirements.				0	CEILING				YES	NO
8134N	U0008.3	LIFTER, PATIENT, ACCESSORY, SLING		VV	Bariatric 4 point patient lifts sling 47x44X56				0	SPACE				NO	NO
8134N	U0030	WALL CHANNEL	1	VV	19"; Aluminum; seismic wall channel with hardware				0	WALL				YES	NO
8134N	U0031	COMPUTER ACCESSORY	1	VV	Mounting assembly and accessories. Includes CPU holder, (1) variable height, adjustable monitor arm supporting up to 40lbs. 180 degree stop mechanism to prevent wall damage, monitor extends 26.25 inches from wall.				0	WALL				YES	NO
8134N	U0032	COMPUTER ACCESSORY	1	VV	Undercounter Mounting assembly for computer keyboard and mouse; ergonomic keyboard holder with tilt and height adjustable; powder coated steel.				0	COUNTER				YES	NO
8134N	U0034	COMPUTER ACCESSORY, BRACKET, WALL	1	VV	27" track to support computer accessories including monitor, keyboard arm, CPU holder. Designed to be mounted to a wall stud or mounting board. Includes channel for cables and covers for sealed cable management.				0	WALL				YES	NO
8134N	U0036	BRACKET	2	VV	Rail mount bracket mount for regulators, wall bottle bracket, disposable bottle brackets, and blenders 5X1X6				0	WALL				YES	NO
8134N	U0039	HEADWALL, PATIENT, ACCESSORY	1	VV	Basket; Mounts on Headwall Accessory Rail; 6 1/2" W X 4"D X 5"H; Functional loading to 5 lbs.; (Meant to hold Bar Code Reader)				0	SPACE				YES	NO
EQUIPMENT ROOM															
8135N	F0470.1	CART, AUDIO/VISUAL	1	VV	3 OPEN ADJUSTABLE SHELVES;WITH ELECTRICAL ASSEMBLYTHREE SHELVES STEEL CONSTRUCTION; HEIGHT ADJUSTABLE 26 TO 42 IN. IN 4 IN. INCREMENTS;RUBBER MAT ON TOP SHELF; 4 IN. (10.16 CM)CASTERS; 2 WITH BRAKES; 15 FT CORD,THREE OUTLETS	24	18	42	44	FLOOR				NO	YES
8135N	M2055	SHELVING, STORAGE, WIRE, CRS, W/ADJUSTABLE SHELVES	1	VV	STATIONARY, WIRE, SHELVING UNIT. UNIT HAS FULLYADJUSTABLE SHELVES CONSTRUCTED OF STAINLESSSTEEL.FOR USE IN GENERAL PURPOSE STORAGEAREAS.SHELVING IS PROVIDED IN VARIOUS SIZES ANDCONFIGURATIONS.PRICE PROVIDED IS FOR A UNITAPPROXIMATELY 74"H X 18"D X 48"W WITH FOURSHELVES.PRICE DIFFERENCE VERIFIED.VENDOR 1: W. W. GRAINGER, INC.CAGE 1: 25795MODEL 1: 2KPV1VENDOR 2: GILLIS ASSOCIATED INDUSTRIES, INC.CAGE 2: 0URP9MODEL 2: A1848SS72S-4VENDOR 3: GLOBAL INDUSTRIAL EQUIPMENTCAGE 3: 0JSZ8MODEL 3: WB189410	48	18	74	80	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8135N	M4250	PUMP, SYRINGE, INFUSION	1	VV	THE INFUSION SYRINGE PUMP ENSURES HIGHLY ACCURATEVOLUME DELIVERY AND CONSISTENT FLOW FOR SMALLVOLUMES (<50 ML) OF PHARMACOLOGIC AGENTS OR THICKFEEDING SOLUTIONS.IT SHALL BE SMALL, LIGHTWEIGHTCONSTRUCTION,MAKING IT TRANSPORTABLE.SHALL HAVEMENU-DRIVEN PROGRAMMING CAPABLE OF FLOW RATES(E.G. 0.1 OR 1.0 ML/HR) THAT ARE INTENDED FORLONG-TERM BEDSIDE USE AND/OR CRITICAL CARE PATIENTTRANSPORT, PLUNGER POSITIONING SENSOR, LCD DISPLAYFOR EASY VIEWING, VOLUME LIMIT PROGRAMMING TOSERVE AS A CONVENIENT CUE OF VOLUME OR DOSEDELIVERY COMPLETION AND MULTIPLE DELIVERY MODESFOR ALL APPLICATIONS REQUIRING PRECISELYCONTROLLED INFUSION RATES.THE INFUSION PUMP SHALLHAVE AUTOMATIC SYRINGE SIZE SENSING WHICH WILLGIVE THE FLEXIBILITY TO ACCEPT A WIDE RANGE OFSYRINGE SIZES (UP TO 60 ML) FROM DIFFERENTMANUFACTURERS.SHALL BE BATTERY POWERED/AC ADAPTER .VENDOR 1: SMITHS MEDICALCAGE 1: 0N0T8MODEL 1: MEDFUSION 3500VENDOR 2: B. BRAUN MEDICAL INCORPORATEDCAGE 2: 0Y77MODEL 2: PERFUSOR SPACE (8713030)VENDOR 3: BAXTER HEALTHCARE CORP., IV SYSTEMS/GOVERNMENT SALES CAGE 3: 04687MODEL 3: SIGMA SPECTRUM 35700BAX	11	7	9	4	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8135N	M4270	PUMP, ENTERAL FEEDING	2	VV	ENTERAL FEEDING PUMP THAT IS SELF-REGULATING, WITHAUTOMATIC SENSOR, AND ADJUSTABLERATE EQUIPPED WITHVISUAL AND AUDIO ALARMS.USED FOR FEEDING PATIENTSVIA A NASO GASTRIC TUBE .VENDOR 1: COVIDIENCAGE 1: 4YQT2MODEL 1: KANGAROO EPUMP (#382400)VENDOR 2:CAGE 2: AM468MODEL 2: SENTINELVENDOR 3: ARIA MEDICALCAGE 3: AM375MODEL 3: COMPAT NOVARTIS #19923500	8	6	8	6	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8135N	M4710.1	WHEELCHAIR, PATIENT, BARIATRIC, TRANSPORT, FOLDING	4	VV	FOLDING BARIATRIC PATIENT TRANSPORT WHEELCHAIRWITH A CAPACITY OF AT LEAST 500 LBS.CONSTRUCTEDWITH A HEAVY DUTY STEEL FRAME AND DOUBLE CROSSBRACE, WITH DETACHABLE ARMREST AND SWING AWAYFOOTRESTS.UNIT HAS HEAVY DUTY FRONT CASTERS ANDREAR WHEELS WITH PRECISION WHEEL BEARINGS IN FRONTAND REAR ASSEMBLY.UNIT IS USED FOR PATIENTMOVEMENT.PRICE DIFFERENCE VERIFIED.VENDOR 1: PREFERRED HEALTHCARECAGE 1: AM245MODEL 1: BCW 600VENDOR 2:CAGE 2: AM498MODEL 2: STD20DDA-SFVENDOR 3: GRAHAM-FIELD HEALTH PRODUCTS, INC.CAGE 3: 12671MODEL 3: PARAMOUNT XD #5PX10620	36	49	39	70	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8135N	M7710	ELECTROCARDIOGRAPH, 12 LEAD, PORTABLE	1	VV	USED TO DETECT THE ELECTRICAL SIGNALS ASSOCIATEDWITH CARDIAC ACTIVITY, DIAGNOSE CARDIACABNORMALITIES, DETERMINE A PATIENTS RESPONSE TODRUG THERAPY AND REVEAL TRENDS OR CHANGES IN HEARTFUNCTION.CAPABLE OF RECORDING TWO OR MORE LEADSSIMULTANEOUSLY, RECORDING AN ENTIRE 12 LEAD ECG INABOUT 10 SECONDS.INCLUDES OF A 3.5 INCH, HIGHDENSITY, FLOPPY DISK DRIVE FOR TESTSTORAGE. PORTABLE .VENDOR 1: PHILIPS MEDICAL SYSTEMS, N.A.CAGE 1: 1CS84MODEL 1: PAGEWRITER TC 70 (#860315)VENDOR 2: CARDIAC SCIENCE CORPORATIONCAGE 2: 1CDW6MODEL 2: ECLIPSE PREMIER (#E10-1AE17)VENDOR 3: GE HEALTHCARECAGE 3: 50542MODEL 3: MAC 5500 W/TROLLEY	21	28	52	87	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8135N	M8825	TABLE, INSTRUMENT/DRESSING, CRS, APPROX. 36X20X34	1	VV	INSTRUMENT AND DRESSING TABLE.MADE OF CORROSIONRESISTANT STAINLESS STEEL WITH A SOUND DEADENEDTOP.INCLUDES GUARD RAIL, SHELF AND TWOSIDE-BY-SIDE DRAWERS. THE TABLE IS MOUNTED ONSWIVEL, BALL-BEARING CASTERS.VENDOR 1: SUBURBAN SURGICAL, INCORPORATEDCAGE 1: 7V465MODEL 1: 108520-00VENDOR 2: UNITED METAL FABRICATORS, INC.CAGE 2: 13127MODEL 2: SS8120VENDOR 3: BLICKMAN, INCORPORATEDCAGE 3: 4M773MODEL 3: 7856SS "FINGER STICK CART" FOR POINT OF CARE TESTING EQUIPMENT	37	21	36	90	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8135N	U0004	CART, CYLINDER, MEDICAL GAS	1	VV	SINGLE D & E:8"SEMI PNEU WHEELS:GURNEY HOOK:GREEN	12	10	41	8	FLOOR				YES	YES
8135N	U0013	WALKER, NON-FOLDING	2	VV	8"WHEELS:BRAKES:FLIP-UP SEAT:440 LBS CAPFORWARD SWEPT HANDLES	28	29	39	22	FLOOR				YES	YES
8135N	U0017	PURIFIER, AIR	1	VV	Wall mounted air purifier. 24X24X24				0	WALL				NO	YES
2 BED MED SURG (QTY 2)															
8136N	A1011.1	TELEPHONE	4	VV	WALL OR DESK TBD				0	WALL				NO	NO
8136N	A4015.1	CLOCK, WALL MOUNT	4	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
8136N	A5075	DISPENSER, SOAP, DISPOSABLE	2	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603690VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 1HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO



ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8136N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	6	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8136N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	2	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8136N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	4	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH AWHITE BAKED ENAMEL FINISH. PLASTIC ORACRYLIC PROVIDED WITH WALL BRACKET TO FACILITATEMOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10.VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M644MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8136N	A5108.1	DISPOSAL CONTAINER, SHARPS W/BRACKET	4	VV	Waste Disposal Unit, Sharps 16W X 8D X 17H				0	WALL				YES	NO
8136N	A5215.1	TELEVISION ACCESSORY, BRACKET, CEILING MOUNTED	4	CC	CEILING MOUNTED TELEVISION BRACKET FOR FLAT PANEL LCD AND PLASMA SCREENS.THE BRACKET SHALL BE A UNIVERSAL STYLE MOUNT WITH ADJUSTMENTS OF 0-15 DEGREE TILT. ACCOMMODATING VARIOUS SIZE UNITS. MAY REQUIRE ADDITIONAL MOUNTING HARDWARE DUE TO CEILING VARIATIONS.  DIMENSIONS: 7W X 14D X 10H				0	CEILING				YES	NO
8136N	F2017.1	WASTE RECEPTACLE, STEP-ON	4	VV	23 GAL.,MOBILE,LEAK-PROOF BASE,TIGHT FITTING LIDCALIFORNIA FIRE MARSHALL APPROVED; ROUNDED CORNERS; DEODORANT BLOCK HOLDER; COLOR:BEIGEWITH LINERS	20	17	33	18	FLOOR				YES	NO
8136N	F2700	READER, BAR CODE, HAND HELD, WITH INTERFACE	4	VV	HAND HELD LASER BAR CODE READER WITH COMPUTERINTERFACE.USED FOR AUTOMATED INVENTORY, USING BARCODE STICKERS / LABELS.CONVENIENCE OUTLET REQUIREDAT POINT OF USE.THESE UNITS ARE WIRELESS BAR CODE SCANNERS ANDHAND HELD COMPUTERS.MANY DIFFERENT MODELS ANDAPPLICATIONS ARE AVAILABLE FROM THESE VENDORS.VENDOR 1: UNITECH AMERICA, INC.CAGE 1: 4QR03MODEL 1: PT600 (PT600HD-0L00B)VENDOR 2: SCAN SOURCE, INC.CAGE 2: AM282MODEL 2: SF51 W/87-221-001/235-195-002VENDOR 3: BARCODES, INC.CAGE 3: AM283MODEL 3: DOLPHIN 9500 9500LWP431C30KIT	4	10	2	2	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8136N	M0506.1	TELEVISION, FLAT PANEL LCD	4	VV	Color television.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance.Unit is HD with a minimum of 720p resolution. 32" ceiling mounted unit. 120V/60HZ				0	CEILING				YES	NO
8136N	M0750	FLOWMETER, AIR, CONNECT W/50 PSIISUPPLY	8	VV	AIR FLOWMETER.UNIT HAS A STAINLESS STEEL NEEDLEVALVE WITH CLEAR FLOWTUBE FOR CONNECTION TO 50 PSIAIR OUTLET FROM CENTRAL PIPELINE SYSTEM.REQUIRESTHE APPROPRIATE ADAPTER FOR CONNECTION TO THE WALLOUTLET AND FITTING TO CONNECT TO TUBING.DATABASEPRICES REFLECT FITTINGS WITH AN ATTACHED DISSPOWER OUTLET.OTHER OUTLET AND ADAPTERCONFIGURATIONS ARE AVAILABLE.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 7702-1266-907VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 2: 11978MODEL 2: 15002-03PVENDOR 3: WESTERN MEDICACAGE 3: SR258MODEL 3: FM104	4	4	10	3	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (K) VACUUM AND LP AIR (30 PSIG) MAXIMUMUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8136N	M0755	FLOWMETER, OXYGEN, LOW FLOW	8	VV	OXYGEN FLOWMETER.CONSISTS OF A CLEAR CRYSTALFLOWTUBE CALIBRATED TO 3.5 OR 8 LPM DEPENDING ONMANUFACTURER.FOR OXYGEN REGULATION IN HOSPITALSETTINGS.DATABASE PRICING INCLUDES DISS FITTINGAND DISS POWER OUTLET AND WALL ADAPTER.OTHERFITTING AND ADAPTER CONFIGURATIONS ARE AVAILABLE.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 6702/1306/521VENDOR 2: PRECISION MEDICALCAGE 2: 0UK20MODEL 2: 1MFA4001/1MFA4004VENDOR 3: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 3: 11978MODEL 3: 08008-03	4	4	10	1	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (A) OXYGENUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8136N	M0765	REGULATOR, VACUUM	16	VV	VACUUM PRESSURE REGULATOR FOR CONNECTION TO CENTRAL PIPED VACUUM SYSTEM. STANDARD DISPLAY SCALE IS GRADUATED AT LEAST FROM 0 TO 200 MM HG OF VACUUM. DISPLAYS ON SPECIALIZED REGULATORS MAY COVER OTHER VACUUM RANGES. REGULATOR TYPE (CONTINUOUS, INTERMITTENT, CONTINUOUS/INTERMITTENT, SURGICAL, PEDIATRIC, THORACIC, ETC.) AS REQUIRED. TO BE USED IN DELIVERY, NEONATAL, PEDIATRIC OR ANY AREA WHERE SUCCTION IS REQUIRED. DATABASE PRICING REFLECTS CONTINUOUS REGULATORS GRADUATED TO 200 MM HG WITH A FULL LINE VACUUM SELECTION SWITCH AND DISCONNECTED INLETS AND OUTLETS. VENDOR 1: OHIO MEDICAL CORPORATION CAGE 1: 45B74 MODEL 1: 6702-1224-907 VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC. CAGE 2: 11978 MODEL 2: VACUTRON #22-12-1111 VENDOR 3: PRECISION MEDICAL CAGE 3: 0UK20 MODEL 3: PM3114	3	4	6	1	WALL	VOLTS 1: 0 VOLTS 2: 0 HERTZ 1: 0 HERTZ DEP: 0 HERTZ SWIT: 0 PHASE 1: 0 PHASE 2: 0 AMPS 1: 0 AMPS 2: 0 WATTS 1: 0 WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/A UTILITY2: Electrical (N) N/A UTILITY3: Medical Gas (B) N/A UTILITY4: Miscellaneous Gas (N) N/A UTILITY5: Non-Medical Gas (N) N/A UTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8136N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	4	VV	DESK TOP MICROPROCESSING COMPUTER. THE UNIT SHALL CONSIST OF A CENTRAL PROCESSING MINI TOWER, FLAT PANEL MONITOR, KEYBOARD, MOUSE AND SPEAKERS. THE SYSTEM SHALL HAVE THE FOLLOWING MINIMUM CHARACTERISTICS: A 2.8 GHZ PENTIUM PROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROM/DVD COMBO; 1.44MB NETWORK INTERFACE CARD; VIDEO 32MB NVIDIA; A 18 INCH FLAT PANEL MONITOR. THE COMPUTER IS USED THROUGHOUT THE FACILITY TO INPUT, MANIPULATE AND RETRIEVE INFORMATION. PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIME OF PURCHASE. VENDOR 1: DELL COMPUTER CORP. CAGE 1: 04RES MODEL 1: OPTIPLEX 780 W/19" MONITOR VENDOR 2: GATEWAY 2000 CAGE 2: 03ZB4 MODEL 2: ZX6810-01 VENDOR 3: CAGE 3: 1U0E6 MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120 VOLTS 2: 0 HERTZ 1: 60 HERTZ DEP: 0 HERTZ SWIT: 1 PHASE 1: 1 PHASE 2: 0 AMPS 1: 5 AMPS 2: 0 WATTS 1: 600 WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/A UTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLET UTILITY3: Medical Gas (N) N/A UTILITY4: Miscellaneous Gas (N) N/A UTILITY5: Non-Medical Gas (N) N/A UTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8136N	M4116	MONITOR, VITAL SIGNS	4	VV	ELECTRONIC SPHYGMOMANOMETER. LCD DISPLAYS NON-INVASIVE BLOOD PRESSURE, PULSE RATE AND TEMPERATURE. USED IN HOSPITALS AND CLINICS. INCLUDES AN OPTIONAL MOBILE STAND. DIMENSIONS ARE FOR MONITOR ONLY. VENDOR 1: WELCH-LIN CAGE 1: 05CJ3 MODEL 1: 53070-E1G & 008-0835-01 VENDOR 2: DATASCOPE CORPORATION CAGE 2: 1L753 MODEL 2: DUO W/TEMP PROBE & STAND VENDOR 3: CARDINAL HEALTH PRODUCTS GROUP, CARDINAL HEALTH CAGE 3: 07TA6 MODEL 3: 420TB-E1-M	10	6	10	6	SPACE	VOLTS 1: 120 VOLTS 2: 0 HERTZ 1: 60 HERTZ DEP: 0 HERTZ SWIT: 0 PHASE 1: 1 PHASE 2: 0 AMPS 1: 2 AMPS 2: 0 WATTS 1: 240 WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/A UTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLET UTILITY3: Medical Gas (N) N/A UTILITY4: Miscellaneous Gas (N) N/A UTILITY5: Non-Medical Gas (N) N/A UTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8136N	M4250	PUMP, SYRINGE, INFUSION	4	VV	THE INFUSION SYRINGE PUMP ENSURES HIGHLY ACCURATE VOLUME DELIVERY AND CONSISTENT FLOW FOR SMALL VOLUMES (<50 ML) OF PHARMACOLOGIC AGENTS OR THICK FEEDING SOLUTIONS. IT SHALL BE SMALL, LIGHTWEIGHT CONSTRUCTION, MAKING IT TRANSPORTABLE. SHALL HAVE MENU-DRIVEN PROGRAMMING CAPABLE OF FLOW RATES (E.G. 0.1 OR 1.0 ML/HR) THAT ARE INTENDED FOR LONG-TERM BEDSIDE USE AND/OR CRITICAL CARE PATIENT TRANSPORT, PLUNGER POSITIONING SENSOR, LCD DISPLAY FOR EASY VIEWING, VOLUME LIMIT PROGRAMMING TO SERVE AS A CONVENIENT CUE OF VOLUME OR DOSE DELIVERY COMPLETION AND MULTIPLE DELIVERY MODES FOR ALL APPLICATIONS REQUIRING PRECISELY CONTROLLED INFUSION RATES. THE INFUSION PUMP SHALL HAVE AUTOMATIC SYRINGE SIZE SENSING WHICH WILL GIVE THE FLEXIBILITY TO ACCEPT A WIDE RANGE OF SYRINGE SIZES (UP TO 60 ML) FROM DIFFERENT MANUFACTURERS. SHALL BE BATTERY POWERED/AC ADAPTER. VENDOR 1: SMITHS MEDICAL CAGE 1: 0N0T8 MODEL 1: MEDFUSION 3500 VENDOR 2: B. BRAUN MEDICAL INCORPORATED CAGE 2: 0Y7P7 MODEL 2: PERFUSOR SPACE (8713030U) VENDOR 3: BAXTER HEALTHCARE CORP., IV SYSTEMS/GOVERNMENT SALES CAGE 3: 04687 MODEL 3: SIGMA SPECTRUM 35700BAX	11	7	9	4	SPACE	VOLTS 1: 120 VOLTS 2: 0 HERTZ 1: 60 HERTZ DEP: 1 HERTZ SWIT: 0 PHASE 1: 1 PHASE 2: 0 AMPS 1: 1 AMPS 2: 0 WATTS 1: 120 WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/A UTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLET UTILITY3: Medical Gas (N) N/A UTILITY4: Miscellaneous Gas (N) N/A UTILITY5: Non-Medical Gas (N) N/A UTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8136N	M4255	STAND, IV, ADJUSTABLE	4	VV	ADJUSTABLE IV STAND WITH 4-HOOK ARRANGEMENT. STAND HAS STAINLESS STEEL CONSTRUCTION WITH HEAVY WEIGHT BASE. IT ADJUSTS FROM 68 INCHES TO 100 INCHES AND IS MOUNTED ON CONDUCTIVE RUBBER BALL BEARING. SWIVEL CASTERS STAND IS USED FOR ADMINISTERING INTRAVENOUS SOLUTIONS. VENDOR 1: BLICKMAN, INCORPORATED CAGE 1: 4M773 MODEL 1: 7793SS-4 VENDOR 2: CLINTON INDUSTRIES, INC. CAGE 2: 0UKA5 MODEL 2: IVS-354 VENDOR 3: PEDIGO PRODUCTS, INC. CAGE 3: 13180 MODEL 3: P-1576-4	24	24	108	35	FLOOR	VOLTS 1: 0 VOLTS 2: 0 HERTZ 1: 0 HERTZ DEP: 0 HERTZ SWIT: 0 PHASE 1: 0 PHASE 2: 0 AMPS 1: 0 AMPS 2: 0 WATTS 1: 0 WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/A UTILITY2: Electrical (N) N/A UTILITY3: Medical Gas (N) N/A UTILITY4: Miscellaneous Gas (N) N/A UTILITY5: Non-Medical Gas (N) N/A UTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8136N	M5030.1	STOOL, REVOLVING	4	VV	5 CASTER STOOL. PNEUMATIC; 14" ROUND CUSHIONED SEAT ONE TOUCH ADJUSTMENT FROM 18"-23"; 2.5" CASTERS	14	14	27	25	FLOOR				YES	NO
8136N	M7010	BED, PATIENT, ELECTRIC	4	VV	ELECTRIC PATIENT BED. INCLUDES SAFETY SIDE RAILS, IV ROD INSERTION SLOTS AND PATIENT CONTROL PANEL FOR A FULL RANGE OF POSITIONS. FOUR WHEEL LOCKING BRAKES. OVERALL BED HEIGHT ADJUSTABLE BETWEEN 16.5 AND 31 INCHES. PRICES CAN VARY BASED ON OPTIONS SELECTED. NO OTHER SOURCE(S) IDENTIFIED WITH THIS DESCRIPTION. SPECIFICATION VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV. CAGE 1: 45044 MODEL 1: STRYKER S3 (#3002S3) VENDOR 2: HILL-ROM CO., INC., A HILLENBRAND INDUSTRY CAGE 2: 28713 MODEL 2: VERSACARE BED VC200 VENDOR 3: CAGE 3: MODEL 3:	42	95	37	335	FLOOR	VOLTS 1: 120 VOLTS 2: 0 HERTZ 1: 60 HERTZ DEP: 1 HERTZ SWIT: 0 PHASE 1: 1 PHASE 2: 0 AMPS 1: 10 AMPS 2: 0 WATTS 1: 1200 WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/A UTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLET UTILITY3: Medical Gas (N) N/A UTILITY4: Miscellaneous Gas (N) N/A UTILITY5: Non-Medical Gas (N) N/A UTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8136N	M7040	TABLE, OVERBED	4	VV	OVERBED TABLE.ADJUSTABLE HEIGHT TABLE CONSTRUCTEDOF HEAVY GAUGE STEEL.MOUNTED ON 2" DIAMETER TWINSWIVEL CASTERS WITH BUMPER CAPS.TABLE TOP ISCONSTRUCTED WITH A HIGH PRESSURE PLASTIC LAMINATEDSURFACE THAT RESISTS CHIPPING, SCRATCHING, ANDSTAINING.IT INCLUDES A VANITY TRAY AND A MIRROR.TABLE IS DESIGNED FOR USE OVER BED. WHEELCHAIR ORLARGE CHAIR.VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 1: 45044MODEL 1: SINGLE TOP W/VANITYVENDOR 2: HILL-ROM CO., INC. A HILLENBRAND INDUSTRY.CAGE 2: 28713MODEL 2: OBT220 W/LOW BASEVENDOR 3: GF HEALTH PRODUCTS, INC.CAGE 3: 1JY01MODEL 3: A797002 W/VANITY &MIRROR	18	36	46	58	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8136N	U0008	LIFTER, PATIENT	4	VC	An overhead rail system specifically designed for patient lifting and movement for a single bed patient room. The system will consist of recessed or ceiling mounted primary and secondary rails, lift motor with rolling carriage,patient harness or seat, and a hand controller or control box with charger.System will facilitate lifting and movement of patient to and from bed to gurney, chair or other requirement.Minimum lift capability is 550 pounds.Custom design of track layout by manufacturer is essential to meet individual facility requirements.				0	CEILING				YES	NO
8136N	U0008.3	LIFTER, PATIENT, ACCESSORY, SLING	4	VV	Bariatric 4 joint patient lifter sling. 47X44X56				0	SPACE				NO	NO
8136N	U0030	WALL CHANNEL	4	VV	19"; Aluminum; seismic wall channel with hardware				0	WALL				YES	NO
8136N	U0031	COMPUTER ACCESSORY	4	VV	Mounting assembly and accessories. Includes CPU holder, (1) variable height, adjustable monitor arm supporting up to 40lbs. 180 degree stop mechanism to prevent wall damage, monitor extends 26.25 inches from wall.				0	WALL				YES	NO
8136N	U0032	COMPUTER ACCESSORY	4	VV	Undercounter Mounting assembly for computer keyboard and mouse; ergonomic keyboard holder with tilt and height adjustable; powder coated steel.				0	COUNTER				YES	NO
8136N	U0034	COMPUTER ACCESSORY, BRACKET, WALL	4	VV	27" track to support computer accessories including monitor, keyboard arm, CPU holder. Designed to be mounted to a wall stud or mounting board. Includes channel for cables and covers for sealed cable management.				0	WALL				YES	NO
8136N	U0036	BRACKET	8	VV	Rail mount bracket mount for regulators, wall bottle bracket, disposable bottle brackets, and blenders 5X1X6				0	WALL				YES	NO
8136N	U0039	HEADWALL, PATIENT, ACCESSORY	4	VV	Basket; Mounts on Headwall Accessory Rail; 6 1/2" W X 4"D X 5"H; Functional loading to 5 lbs.; (Meant to hold Bar Code Reader)				0	SPACE				YES	NO
1 BED MED SURG (QTY 11)															
8140N	A1011.1	TELEPHONE	11	VV	WALL OR DESK TBD				0	WALL				NO	NO
8140N	A4015.1	CLOCK, WALL MOUNT	11	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
8140N	A5075	DISPENSER, SOAP, DISPOSABLE	11	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8140N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	11	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITs ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8140N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	11	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8140N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	11	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH AWHITE BAKED ENAMEL FINISH, PLASTIC ORACRYLIC.PROVIDED WITH WALL BRACKET TO FACILITATEMOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10.VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M644MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8140N	A5108.1	DISPOSAL CONTAINER, SHARPS W/BRACKET	11	VV	Waste Disposal Unit, Sharps 16W X 8D X 17H				0	WALL				YES	NO
8140N	A5210.1	TELEVISION ACCESSORY, BRACKET, WALL MOUNTED	7	CC	Bracket, Television, Wall Mounted, w/Adjust Arm 40W X 4D X 13H				0	WALL				YES	NO
8140N	A5215.1	TELEVISION ACCESSORY, BRACKET, CEILING MOUNTED	4	CC	CEILING MOUNTED TELEVISION BRACKET FOR FLAT PANEL LCD AND PLASMA SCREENS.THE BRACKET SHALL BE A UNIVERSAL STYLE MOUNT WITH ADJUSTMENTS OF 0-15 DEGREE TILT. ACCOMMODATING VARIOUS SIZE UNITS. MAY REQUIRE ADDITIONAL MOUNTING HARDWARE DUE TO CEILING VARIATIONS.  DIMENSIONS: 7W X 14D X 10H				0	CEILING				YES	NO
8140N	F2017.1	WASTE RECEPTACLE, STEP-ON	11	VV	23 GAL.,MOBILE,LEAK-PROOF BASE,TIGHT FITTING LIDCALIFORNIA FIRE MARSHALL APPROVED, ROUNDED CORNERS; DEODORANT BLOCK HOLDER. COLOR:BEIGEWITH LINERS	20	17	33	18	FLOOR				YES	NO
8140N	F2700	READER, BAR CODE, HAND HELD, WITH INTERFACE	11	VV	HAND HELD LASER BAR CODE READER WITH COMPUTERINTERFACE.USED FOR AUTOMATED INVENTORY, USING BARCODE STICKERS / LABELS.CONVENIENCE OUTLET REQUIREDAT POINT OF USE.THESE UNITS ARE WIRELESS BAR CODE SCANNERS ANDHAND HELD COMPUTERS.MANY DIFFERENT MODELS ANDAPPLICATIONS ARE AVAILABLE FROM THESE VENDORS.VENDOR 1: UNITECH AMERICA, INC.CAGE 1: 4QR03MODEL 1: PT600 (PT600HD-0L00B)VENDOR 2: SCAN SOURCE, INC.CAGE 2: AM282MODEL 2: SF51 W/87-221-001/235-195-002VENDOR 3: BARCODES, INC.CAGE 3: AM283MODEL 3: DOLPHIN 9500 9500LWP431C30KIT	4	10	2	2	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8140N	M0506.1	TELEVISION, FLAT PANEL LCD	4	VV	Color television.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance.Unit is HD with a minimum of 720p resolution. 32" ceiling mounted unit. 120V;60HZ				0	CEILING				YES	NO
8140N	M0506.2	TELEVISION, FLAT PANEL LCD	7	VV	Color television with HD.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance. 42" wall mounted unit. 120V;60HZ				0	WALL				YES	NO
8140N	M0750	FLOWMETER, AIR, CONNECT W/50 PSISUPPLY	22	VV	AIR FLOWMETER.UNIT HAS A STAINLESS STEEL NEEDLEVALVE WITH CLEAR FLOWTUBE FOR CONNECTION TO 50 PSIAIR OUTLET FROM CENTRAL PIPELINE SYSTEM.REQUIRESTHE APPROPRIATE ADAPTER FOR CONNECTION TO THE WALLOUTLET AND FITTING TO CONNECT TO TUBING.DATABASEPRICES REFLECT FITTINGS WITH AN ATTACHED DISSPOWER OUTLET.OTHER OUTLET AND ADAPTERCONFIGURATIONS ARE AVAILABLE.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 7702-1266-907VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 2: 11978MODEL 2: 15002-03PVENDOR 3: WESTERN MEDICACAGE 3: SR258MODEL 3: FM104	4	4	10	3	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (K) VACUUM AND LP AIR (30 PSIG) MAXIMUMUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8140N	M0755	FLOWMETER, OXYGEN, LOW FLOW	22	VV	OXYGEN FLOWMETER.CONSISTS OF A CLEAR CRYSTALFLOWTUBE CALIBRATED TO 3.5 OR 8 LPM DEPENDING ONMANUFACTURER.FOR OXYGEN REGULATION IN HOSPITALSETTINGS.DATABASE PRICING INCLUDES DISS FITTINGAND DISS POWER OUTLET AND WALL ADAPTER.OTHER FITTING AND ADAPTER CONFIGURATIONS ARE AVAILABLE.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 6702/1306/921VENDOR 2: PRECISION MEDICALCAGE 2: 0UK20MODEL 2: 1MFA4001/1MFA4004VENDOR 3: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 3: 11978MODEL 3: 08008-03	4	4	10	1	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (A) OXYGENUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8140N	M0765	REGULATOR, VACUUM	44	VV	VACUUM PRESSURE REGULATOR FOR CONNECTION TOCENTRAL PIPED VACUUM SYSTEM.STANDARD DISPLAY SCALEIS GRADUATED AT LEAST FROM 0 TO 200 MM HG OFVACUUM.DISPLAYS ON SPECIALIZED REGULATORS MAYCOVER OTHER VACUUM RANGES.REGULATOR TYPE(CONTINUOUS, INTERMITTENT,CONTINUOUS/INTERMITTENT, SURGICAL, PEDIATRIC,THORACIC, ETC.) AS REQUIRED.TO BE USED INDELIVERY, NEONATAL, PEDIATRICS OR ANY AREA WHERESUCTION IS REQUIRED.DATABASE PRICING REFLECTSCONTINUOUS REGULATORS GRADUATED TO 200 MM HG WITHA FULL LINE VACUUM SELECTION SWITCH AND DISSCONFIGURED INLETS AND OUTLETS.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 6702-1224-907VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 2: 11978MODEL 2: VACUTRON #22-12-1111VENDOR 3: PRECISION MEDICALCAGE 3: 0UR20MODEL 3: PM3114	3	4	6	1	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (B) VACUUMUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8140N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	11	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THESYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMDVDCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RESMODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8140N	M4116	MONITOR, VITAL SIGNS	11	VV	ELECTRONIC SPHYGMOMANOMETER.LCD DISPLAYSNON-INVASIVE BLOOD PRESSURE, PULSE RATE ANDTEMPERATURE.USED IN HOSPITALS AND CLINICS.INCLUDESAN OPTIONAL MOBILE STAND.DIMENSIONS ARE FOR MONITOR ONLY.VENDOR 1: WELCH-ALLYN.CAGE 1: 05CJ38MODEL 1: 50070-E1G & 00835-01VENDOR 2: DATASCOPE CORPORATIONCAGE 2: IL753MODEL 2: DUO W/TEMP PROBE & STANDVENDOR 3: CARDINAL HEALTH PRODUCTS GROUP, CARDINAL HEALTHCAGE 3: 07TA6MODEL 3: 420TB-E1-M	10	6	10	6	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8140N	M4250	PUMP, SYRINGE, INFUSION	11	VV	THE INFUSION SYRINGE PUMP ENSURES HIGHLY ACCURATEVOLUME DELIVERY AND CONSISTENT FLOW FOR SMALLVOLUMES (<50 ML) OF PHARMACOLOGIC AGENTS OR THICKFEEDING SOLUTIONS.IT SHALL BE SMALL, LIGHTWEIGHTCONSTRUCTION,MAKING IT TRANSPORTABLE.SHALL HAVEMENU-DRIVEN PROGRAMMING CAPABLE OF FLOW RATES(E.G. 0.1 OR 1.0 ML/HR) THAT ARE INTENDED FORLONG-TERM BEDSIDE USE AND/OR CRITICAL CARE PATIENTTRANSPORT, PLUNGER POSITIONING SENSOR, LCD DISPLAYFOR EASY VIEWING, VOLUME LIMIT PROGRAMMING TOSERVE AS A CONVENIENT CUE OF VOLUME OR DOSEDELIVERY COMPLETION AND MULTIPLE DELIVERY MODESFOR ALL APPLICATIONS REQUIRING PRECISELYCONTROLLED INFUSION RATES.THE INFUSION PUMP SHALLHAVE AUTOMATIC SYRINGE SIZE SENSING WHICH WILLGIVE THE FLEXIBILITY TO ACCEPT A WIDE RANGE OFSYRINGE SIZES (UP TO 60 ML) FROM DIFFERENTMANUFACTURERS.SHALL BE BATTERY POWERED/AC ADAPTER.VENDOR 1: SMITHS MEDICALCAGE 1: 0N0T8MODEL 1: MEDFUSION 3500VENDOR 2: B. BRAUN MEDICAL INCORPORATEDCAGE 2: 0Y7P7MODEL 2: PERFUSOR SPACE (8713030U)VENDOR 3: BAXTER HEALTHCARE CORP., IV SYSTEMS/GOVERNMENT SALESCAGE 3: 04687MODEL 3: SIGMA SPECTRUM 35700BAX	11	7	9	4	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8140N	M4255	STAND, IV, ADJUSTABLE	11	VV	ADJUSTABLE IV STAND WITH 4-HOOK ARRANGEMENT.STANDHAS STAINLESS STEEL CONSTRUCTION WITH HEAVY WEIGHTBASE.IT ADJUSTS FROM 66 INCHES TO 100 INCHES ANDIS MOUNTED ON CONDUCTIVE RUBBER BALL BEARING.SWIVEL CASTERS STAND IS USED FOR ADMINISTERINGINTRAVENOUS SOLUTIONS.VENDOR 1: BLICKMAN, INCORPORATEDCAGE 1: 4M773MODEL 1: 7793SS-4VENDOR 2: CLINTON INDUSTRIES, INC.CAGE 2: 0UKA5MODEL 2: IVS-354VENDOR 3: PEDIGO PRODUCTS, INC.CAGE 3: 13180MODEL 3: P-1576-4	24	24	108	35	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8140N	M5030.1	STOOL, REVOLVING	11	VV	5 CASTER STOOL.PNEUMATIC;14" ROUND CUSHIONED SEATONE TOUCH ADJUSTMENT FROM 18"-23";2.5"CASTERS	14	14	27	25	FLOOR				YES	NO
8140N	M7010	BED, PATIENT, ELECTRIC	11	VV	ELECTRIC PATIENT BED.INCLUDES SAFETY SIDE RAILS,IV ROD INSERTION SLOTS AND PATIENT CONTROL PANELFOR A FULL RANGE OF POSITIONS.FOUR WHEEL LOCKINGBRAKES.OVERALL BED HEIGHT ADJUSTABLE BETWEEN 16.5AND 31 INCHES.PRICES CAN VARY BASED ON OPTIONS SELECTED.NO OTHERSOURCE(S) IDENTIFIED WITH THISDESCRIPTIONSPECIFICATION.VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 1: 45044MODEL 1: STRYKER S3 (#3002S3)VENDOR 2: HILL-ROM CO., INC., A HILLENBRAND INDUSTRYCAGE 2: 28713MODEL 2: VERSACARE BED VC200VENDOR 3:CAGE 3:MODEL 3:	42	95	37	335	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 10AMPS 2: 0WATTS 1: 1200WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8140N	M7040	TABLE, OVERBED	11	VV	OVERBED TABLE.ADJUSTABLE HEIGHT TABLE CONSTRUCTEDOF HEAVY GAUGE STEEL.MOUNTED ON 2" DIAMETER TWINSWIVEL CASTERS WITH BUMPER CAPS.TABLE TOP ISCONSTRUCTED WITH A HIGH PRESSURE PLASTIC LAMINATEDSURFACE THAT RESISTS CHIPPING, SCRATCHING, ANDSTAINING.IT INCLUDES A VANITY TRAY AND A MIRROR.TABLE IS DESIGNED FOR USE OVER BED, WHEELCHAIR ORLARGE CHAIR.VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 1: 45044MODEL 1: SINGLE TOP W/VANITYVENDOR 2: HILL-ROM CO., INC., A HILLENBRAND INDUSTRYCAGE 2: 28713MODEL 2: OBT220 W/LOW BASEVENDOR 3: GF HEALTH PRODUCTS, INC.CAGE 3: 1JY01MODEL 3: A797002 W/VANITY &MIRROR	18	36	46	58	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO

ROOM #	JSN	NAME	QTY	ACQ/INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8140N	U0008	LIFTER, PATIENT	11	VC	An overhead rail system specifically designed for patient lifting and movement for a single bed patient room. The system will consist of recessed or ceiling mounted primary and secondary rails, lift motor with rolling carriage,patient harness or seat, and a hand controller or control box with charger.System will facilitate lifting and movement of patient to and from bed to gurney, chair or other requirement.Minimum lift capability is 550 pounds.Custom design of track layout by manufacturer is essential to meet individual facility requirements.				0	CEILING				YES	NO
8140N	U0008.3	LIFTER, PATIENT, ACCESSORY, SLING	11	VV	Baketric 4-point Patient Lifter sling. 47X44X56				0	SPACE				NO	NO
8140N	U0030	WALL CHANNEL	11	VV	19"; Aluminum; seismic wall channel with hardware				0	WALL				YES	NO
8140N	U0031	COMPUTER ACCESSORY	11	VV	Mounting assembly and accessories. Includes CPU holder, (1) variable height, adjustable monitor arm supporting up to 40lbs. 180 degree stop mechanism to prevent wall damage, monitor extends 26.25 inches from wall.				0	WALL				YES	NO
8140N	U0032	COMPUTER ACCESSORY	11	VV	Undercounter Mounting assembly for computer keyboard and mouse; ergonomic keyboard holder with tilt and height adjustable; powder coated steel.				0	COUNTER				YES	NO
8140N	U0034	COMPUTER ACCESSORY, BRACKET, WALL	11	VV	27" track to support computer accessories including monitor, keyboard arm, CPU holder. Designed to be mounted to a wall stud or mounting board. Includes channel for cables and covers for sealed cable management.				0	WALL				YES	NO
8140N	U0036	BRACKET	22	VV	Rail mount bracket mount for regulators, wall bottle bracket, disposable bottle brackets, and blenders 5X1X6				0	WALL				YES	NO
8140N	U0039	HEADWALL, PATIENT, ACCESSORY	11	VV	Basket; Mounts on Headwall Accessory Rail; 6 1/2" W X 4"D X 5"H; Functional loading to 5 lbs.; (Meant to hold Bar Code Reader)				0	SPACE				YES	NO
8140N		NOTE:	1	VV	32" CEILING MOUNTED TV LOCATED IN ROOMS: 8129N, 8125N, 8127N, 8126N				0	CEILING				NO	NO
CLEAN UTILITY															
8142N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8142N	M2055	SHELVING, STORAGE, WIRE, CRS, W/ADJUSTABLE SHELVES	1	VV	STATIONARY, WIRE, SHELVING UNIT. UNIT HAS FULLYADJUSTABLE SHELVES CONSTRUCTED OF STAINLESSSTEEL.FOR USE IN GENERAL PURPOSE STORAGEAREAS.SHELVING IS PROVIDED IN VARIOUS SIZES ANDCONFIGURATIONS.PRICE PROVIDED IS FOR A UNITAPPROXIMATELY 74"H X 18"D X 48"W WITH FOURSHELVES.PRICE DIFFERENCE VERIFIED.VENDOR 1: W. W. GRAINGER, INC.CAGE 1: 25795MODEL 1: 2KPV1VENDOR 2: GILLIS ASSOCIATED INDUSTRIES, INC.CAGE 2: 0URP9MODEL 2: A1848SS72S-4VENDOR 3: GLOBAL INDUSTRIAL EQUIPMENTCAGE 3: 0JSZ8MODEL 3: WB189410	48	18	74	80	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8142N	M2055.1	SHELVING, WIRE	1	VV	5 SHELF;CHROME;FREESTANDINGTO CONSIST OF:4 EA. 74P: 1" O.D. X 16GSX73 3/4L. CHR POST4 EA. 9993Z: FOOT PLATE5 EA. 2472NC: WIRE CHROME SHELF	72	24	74	0	FLOOR				YES	YES
8142N	M2055.2	SHELVING, WIRE	2	VV	6 SHELF, FREE STANDING, STAINLESS STEELTO CONSIST OF: 6 EA 1836NS SHELF, WIRE 18 X 36 IN 4 EA 74PS POST, 74-5/8 IN	36	18	75	73	FLOOR				YES	YES
8142N	M3105	CABINET, WARMING	1	VV	FOR USE WITH COMFORT PERSONAL CLEANSING BRAND ACCOMMODATES 14 INDIVIDUAL PACKAGES;ONE PER SLOT; 125 DEGREES F TARGET PRODUCT TEMPERATURE;POWDER COATED STEEL EXTERIOR;ALUMINUM INTERIOR TO INCLUDE: -1 EA. 7920 CART 33WX18DX35H TOTAL DIMENSIONS FOR CABINET AND CART: 33X18X53				0	FLOOR				YES	NO
MEDS															
8143N	A5065.1	BOX, RETURN, DRUGS	1	VV	Security Drop box. Contents are deposited through the 12" front opening. A mailbox type baffle deters the removal of contents. Made of 18 gauge steel. Two cam locks keyed differently. DIMENSIONS: 14.5W X 17D X 20H				0	WALL				YES	YES
8143N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8143N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8143N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8143N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	1	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH AWHITE BAKED ENAMEL FINISH, PLASTIC ORACRYLIC PROVIDED WITH WALL BRACKET TO FACILITATE MOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10 VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M644MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8143N	A5108.1	DISPOSAL CONTAINER, SHARPS W/BRACKET	1	VV	Waste Disposal Unit, Sharps 16W X 8D X 17H				0	WALL				YES	NO
8143N	E0957.1	CART, MEDICATION	3	VV	DOCUMENTATION CART W/BIN STORAGE:HEIGHT ADJ.KEYBOARD TRAY;COMMERCIAL GRADE POWER CORD;CORD HANGER;MOTORIZED LIFT MECHANISM;MOUSE SURFACEWORK SURFACE;NOTEBOOK/PC COMPARTMENT	18	20	46	0	FLOOR	120V;7.8 AMPS			YES	YES
8143N	F2017.1	WASTE RECEPTACLE, STEP-ON	1	VV	23 GAL.,MOBILE;LEAK-PROOF BASE;TIGHT FITTING LIDCALIFORNIA FIRE MARSHALL APPROVED; ROUNDED CORNERS; DEODORANT BLOCK HOLDER; COLOR:BEIGEWITH LINERS	20	17	33	18	FLOOR				YES	NO
8143N	M3150	DISTRIBUTION SYSTEM, MEDICATION,AUTOMATIC	1	VV	AN AUTOMATED DISPENSING SYSTEM THAT PROVIDESCONTROLLED DISPENSING, INVENTORY AND SECURITY.SIZEAND COST WILL VARY DEPENDENT ON NUMBER OF MODULESSELECTED.PRICES ARE ESTIMATES ONLY AND ARE FOR A ONESTATION CABINET.THE PRICE CAN VARY SIGNIFICANTLYDEPENDING ON THE SPECIFIC REQUIREMENTS OF EACHFACILITY.MODEL 1 IS A 6-DRAWER MAIN UNIT WITHTOUCH SCREEN MONITOR AND COMPUTER. MODEL 2 PRICEIS FOR THE LARGE MAIN CABINET.MODEL 3 PRICE IS FORTHE BASE 72 DRAWER DISPENSING UNIT AND A TOUCHSCREEN AND COMPUTER.IT CAN BE INTERFACED WITH ANYHL7-COMPLIANT INFORMATION SYSTEM.THERE ARE SEVERALCONFIGURATIONS WITH VARIOUS SIZED CABINETS WITHOPTIONS/ACCESSORIES AVAILABLE INCLUDING VARIOUSSOFTWARE PACKAGES, BARCODE READER AND FINGER PRINTSCANNER.VENDOR 1: PYXIS CORPORATION, A CARDINALHEALTH COMPANYCAGE 1: SR350MODEL 1: MEDSTATION 4000 6-DRAWERVENDOR 2:CAGE 2: 1F1E9MODEL 2: ACUDOSE-RXVENDOR 3: INTERMETRO INDUSTRIES CORPORATIONCAGE 3: 40038MODEL 3: MEDDISPENSE BASE 72 (MDB-72)	33	29	57	659	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 10AMPS 2: 0WATTS 1: 1200WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (B) COLD WATER AND DRAINUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8143N	M3150M	DISPENSER, MEDICATION, AUTOMATED	2	VV	AUXILIARY;7 DRAWER	23	28	47	133	FLOOR	PLUGS INTO MAIN AND PYXIS CONTROL CABLE120V;100W;0.9A;EMERGENCY POWER		368 BTU/H	YES	NO
8143N	R6200.1	REFRIGERATOR, UNDERCOUNTER	1	VV	S.S.;MEDICAL GRADE;4 CU FT;MAGNETIC DOOR GASKETEXT. TEMP LOCKING DOOR;FIELD REVERSIBLE DOOR;2 WIRE SHELVES	24	28	32	186	FLOOR	115V;60HZ;1PH;8A MAX.;FUSE 15 AMPS;NEMA 5-15 PLUG;(1.6M)HOSPITAL-GRADE CORD		717 BTU/HR	YES	NO
NURSE STATION															
8144N	A1011.1	TELEPHONE	4	VV	WALL OR DESK TBD				0	WALL				NO	NO
8144N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
8144N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8144N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8144N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK, FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8144N	F2017.2	WASTE RECEPTACLE, STEP-ON	1	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR		UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A		YES	NO
8144N	F2026.1	CONTAINER, RECYCLING, LARGE	1	VV	RECEPTACLE FOR PAPER PRODUCTS SLATED FOR COLLECTION, SHREDDING, RECYCLING DIMENSIONS: 19W X 19D X 37H	21	21	34	25	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8144N	L1185.1	LAB ANALYZER, CHEMISTRY, MANUAL	2	VV	POC;W/DOWNLOADER/RECHARGER & BATTERY;WIRELESSUTILIZES TEST CARTRIDGE OR GLUCOSE STRIPMEASURES:BLOOD GASES,ELECTROLYTES,CHEMISTRIES,COAGULATION,HEMATOLOGY,G LUCOSE & CARDIAC MARKERSTO INCLUDE:1 EA 03P75-01 ISTAT 1 WIRELESS ANALYZER1 EA 04P73-04 ISTAT 1 DOWNLOADER RECHARGER KIT1 EA 06F23-55 RECHARGEABLE POWER PACK1 EA 06F11-01 ISTAT ELECTRONIC STIMULATOR	3	3	10	0	COUNTER	REQUIRES 2 9V LITHIUM OR NICKEL CADMIUM BATTERIES120V AC ADAPTER/CHARGER			YES	YES
8144N	M0501	MONITOR, VIDEO	1	VV	Flat screen monitor with approximately 19" diagonal screen size; for staff communications; wall mounted				0	WALL	120 V;60 HZ			NO	NO
8144N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	4	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THESYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMDVDCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RE5MODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8144N	M1801.1	COPY MACHINE	1	VV	Desktop copier.Unit features automatic paper size selection, automatic document feeder and sorter.The system also has zoom capabilities and automatic two-sided copying.For use where medium volume reproduction is required in the range of 30 to 40 copies per minute.				0	COUNTER				NO	NO
8144N	M1825	PRINTER, COMPUTER	1	VV	HIGH RESOLUTION COMPUTER PRINTER WITH A VARIETY OFTYPE STYLES AND SHEET/ENVELOPE FEEDERTRAYS.DATABASE INFORMATION REFLECTS NETWORK READY,MEDIUM DUTY OFFICE STYLE LASER PRINTERS.OTHERTYPES OF PRINTERS (BUBBLE JET, DOT MATRIX, LINE ORPLOTTER) AS WELL AS LIGHT OR HEAVY USECAPABILITIES ARE AVAILABLE.VENDOR 1: LEXMARK INTERNATIONAL, INCORPORATEDCAGE 1: 0SD37MODEL 1: C736NVENDOR 2: HEWLETT PACKARD COMPANYCAGE 2: 28480MODEL 2: CP3524DN (CC470A)VENDOR 3: BROTHER INDUSTRIES, LTD.CAGE 3: 65533MODEL 3: HL-4570CDWT	21	20	18	73	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 6AMPS 2: 0WATTS 1: 720WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8144N	M1855	FACSIMILE MACHINE	1	VV	PLAIN PAPER FAXSIMILE UNIT USED TO TRANSMITPRINTED MATTER OR IMAGES OVER STANDARD TELEPHONELINES.THE UNIT OPERATES AT A MODEM SPEED OF 33.6KBPS AND AUTOMATICALLY RECOGNIZES AND COMMUNICATESWITH A VARIETY OF STANDARD FACSIMILEMACHINES.COMMONLY REFERRED TO AS "FAX" UNIT.VENDOR 1: HEWLETT PACKARD COMPANYCAGE 1: 28480MODEL 1: HP 3180 FAXVENDOR 2: CANON USA, INC.CAGE 2: 1BA63MODEL 2: FAXPHONE L170VENDOR 3: BROTHER INTERNATIONAL CORP.CAGE 3: SR335MODEL 3: INTELLIFAX 2920	21	23	14	25	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 1HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8144N	M7780	SCANNER, DOCUMENT	1	VV	SCANNER FOR RECORDING PRINTED MATERIAL IN ANELECTRONIC FORMAT.SOME MODELS PROCESS BOTH COLORAND BLACK-AND-WHITE IMAGES.OPTIONS INCLUDEAUTOMATIC PAPER FEEDERS AND TRANSPARENCYADAPTERS.MOST MODELS REQUIRE CONNECTION TO AMICROCOMPUTER FOR INDEPENDENT OPERATION OR USE ASPART OF A NETWORK.LARGER MODELS FOR HIGH USEAPPLICATIONS WHICH CONNECT DIRECTLY TO A NETWORKWITHOUT PASSING INFORMATION THROUGH AMICROCOMPUTER ARE AVAILABLE.VENDOR 1: HEWLETT PACKARD COMPANYCAGE 1: 28480MODEL 1: SCANJET 7000VENDOR 2: UMAX TECHNOLOGIES, INC.CAGE 2: SR324MODEL 2: AV830CVENDOR 3: OFFICE DEPOTCAGE 3: SR203MODEL 3: KODAK I1200 PLUS	23	14	10	22	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8144N	U0007	I.D. SYSTEM, EMBOSSEER ACCESSORY, COMPUTER CONTROL	1	VV	Computer controlled ID System Embosser. Approximately 20Wx20Dx16H.				0	COUNTER				NO	NO
8144N	U0021	SAFE	1	CC					0	SPACE				NO	NO
8144N	U0030	WALL CHANNEL	1	VV	19"; Aluminum; seismic wall channel with hardware				0	WALL				YES	NO
8144N	U0041	MONITOR ACCESSORY, MOUNTING ASSEMBLY	1	VV	M-SERIES:FLUSH MOUNT FOR FLAT PANEL DISPLAY W/ 75/100MM VESA;TILT/SWIVEL;30# MAX LOAD				0	WALL				YES	NO



ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
NOURISH															
8145N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8145N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8145N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8145N	F2017.2	WASTE RECEPTACLE, STEP-ON	1	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
8145N	K4665	OVEN, MICROWAVE, CONSUMER	1	VV	COUNTER MOUNTED MICROWAVE OVEN FOR AVERAGE DUTYUSE.THE EXTERIOR CABINET CAN BE METAL OR HEAVYDUTY IMPACT RESISTANT PLASTIC.THE OVEN DELIVERSINSTANT ENERGY FOR RAPID HEATING, DEFROSTING ORPRIME COOKING.THE OVEN HAS TOUCH PAD CONTROLS.DIGITAL TIMER, POWER LEVEL SELECTOR ANDPREPROGRAMMED SELECTORS FOR COMMONLY COOKEDITEMS.THIS OVEN IS COMMONLY FOUND IN STAFFLOUNGES.VENDOR 1: KITCHENAIDCAGE 1: AM179MODEL 1: KCMS1555RWHVENDOR 2: SEARS, COMMERCIAL SALESCAGE 2: SR050MODEL 2: 02074062000VENDOR 3: GENERAL ELECTRIC APPLIANCESCAGE 3: 0G0L9MODEL 3: JE2160WF	24	20	14	44	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 15AMPS 2: 0WATTS 1: 1800WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8145N	R4650.1	ICE MAKER	1	VV	12# STORAGE;453#/24HRS;COUNTERTOP;TOUCH FREEW/FILTER;FLAKED;AIR COOLED;S.S.;BIN STORAGEICE PRODUCTION;WATER/ICE SELECTION SWITCHTO INCLUDE: 1EA SSM1-P SINGLE SYSTEM WATER FILTER 1EA SC10-A 10" COARSE PREFILTER	14	24	35	155	COUNTER	115V;60HZ;14.9A;1PH;1/2HP;LINE CORD NOT PROVIDED	FLARE INLET: 1/4";I.D. TUBE DRAIN: 5/8" I.D. SINK DRAIN: 7/8"SSM1-P. INLET - 3/8"FNPT;OUTLET - 3/8" FNPTSC10A: INLET - 3/4" FNPT;OUTLET - 3/4" FNPT20-80 PSIG WATER PRESSURE;WATER TEMP 40-100 F	6 IN. (15.24 CM) CLEARANCE REQUIRED ON BOTH SIDESFOR VENTILATION	YES	NO
8145N	R5135.1	FREEZER, UNDERCOUNTER	1	VV	3.5CF;MED GRADE;-25 DEG-TO -23 DEG C W/ALARMAUTO DEFROST;EXT. TEMP. DISPLAY	24	26	32	175	FLOOR	115V;60HZ;1PH;8A;MAX FUSE-15A;HG PLUG & CORD			YES	NO
8145N	R6200.1	REFRIGERATOR, UNDERCOUNTER	1	VV	S.S.;MEDICAL GRADE;4 CU FT;MAGNETIC DOOR GASKETEXT. TEMP LOCKING DOOR;FIELD REVERSIBLE DOOR;2 WIRE SHELVES	24	28	32	186	FLOOR	115V;60HZ;1PH;8A MAX.;FUSE 15 AMPS;NEMA 5-15 PLUG(1.8M)HOSPITAL-GRADE CORD		717 BTU/HR	YES	NO
1 BED, BARIATRIC															
8148N	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
8148N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
8148N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8148N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8148N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8148N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	1	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH AWHITE BAKED ENAMEL FINISH, PLASTIC ORACRYLIC.PROVIDED WITH WALL BRACKET TO FACILITATEMOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10.VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M644MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8148N	A5108.1	DISPOSAL CONTAINER, SHARPS W/BACKET	1	VV	Waste Disposal Unit, Sharps 16W X 8D X 17H				0	WALL				YES	NO
8148N	A5215.1	TELEVISION ACCESSORY, BRACKET, CEILING MOUNTED	1	CC	CEILING MOUNTED TELEVISION BRACKET FOR FLAT PANEL LCD AND PLASMA SCREENS.THE BRACKET SHALL BE A UNIVERSAL STYLE MOUNT WITH ADJUSTMENTS OF 0-15 DEGREE TILT. ACCOMMODATING VARIOUS SIZE UNITS. MAY REQUIRE ADDITIONAL MOUNTING HARDWARE DUE TO CEILING VARIATIONS.  DIMENSIONS: 7W X 14D X 10H				0	CEILING				YES	NO
8148N	F2017.1	WASTE RECEPTACLE, STEP-ON	1	VV	23 GAL.;MOBILE;LEAK-PROOF BASE;TIGHT FITTING LIDCALIFORNIA FIRE MARSHALL APPROVED; ROUNDED CORNERS; DEODORANT BLOCK HOLDER. COLOR:BEIGEWITH LINERS	20	17	33	18	FLOOR				YES	NO
8148N	F2700	READER, BAR CODE, HAND HELD, WITH INTERFACE	1	VV	HAND HELD LASER BAR CODE READER WITH COMPUTERINTERFACE.USED FOR AUTOMATED INVENTORY, USING BARCODE STICKERS / LABELS.CONVENIENCE OUTLET REQUIREDAT POINT OF USE.THESE UNITS ARE WIRELESS BAR CODE SCANNERS ANDHAND HELD COMPUTERS.MANY DIFFERENT MODELS ANDAPPLICATIONS ARE AVAILABLE FROM THESE VENDORS.VENDOR 1: UNITECH AMERICA, INC.CAGE 1: 4QR03MODEL 1: PT600 (PT600HD-0L00B)VENDOR 2: SCAN SOURCE, INC.CAGE 2: AM282MODEL 2: SF51 W/87-221-001/235-195-002VENDOR 3: BARCODES, INC.CAGE 3: AM283MODEL 3: DOLPHIN 9500 9500LWP431C30KIT	4	10	2	2	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8148N	M0506.1	TELEVISION, FLAT PANEL LCD	1	VV	Color television.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance.Unit is HD with a minimum of 720p resolution. 32" ceiling mounted unit. 120V;60HZ				0	CEILING				YES	NO
8148N	M0750	FLOWMETER, AIR, CONNECT W/50 PSISUPPLY	2	VV	AIR FLOWMETER.UNIT HAS A STAINLESS STEEL NEEDLEVALVE WITH CLEAR FLOWTUBE FOR CONNECTION TO 50 PSIAIR OUTLET FROM CENTRAL PIPELINE SYSTEM.REQUIRESTHE APPROPRIATE ADAPTER FOR CONNECTION TO THE WALLOUTLET AND FITTING TO CONNECT TO TUBING.DATABASEPRICES REFLECT FITTINGS WITH AN ATTACHED DISSPOWER OUTLET OTHER OUTLET AND ADAPTERCONFIGURATIONS ARE AVAILABLE.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 7702-1266-907VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 2: 11978MODEL 2: 15002-03PVENDOR 3: WESTERN MEDICACAGE 3: SR258MODEL 3: FM104	4	4	10	3	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (K) VACUUM AND LP AIR (30 PSIG) MAXIMUMUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8148N	M0755	FLOWMETER, OXYGEN, LOW FLOW	2	VV	OXYGEN FLOWMETER.CONSISTS OF A CLEAR CRYSTALFLOWTUBE CALIBRATED TO 3.5 OR 8 LPM DEPENDING ONMANUFACTURER.FOR OXYGEN REGULATION IN HOSPITALSETTINGS.DATABASE PRICING INCLUDES DISS FITTINGAND DISS POWER OUTLET AND WALL ADAPTER.OTHERFITTING AND ADAPTER CONFIGURATIONS ARE AVAILABLE.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 6702/1306/921VENDOR 2: PRECISION MEDICALCAGE 2: 0UK20MODEL 2: 1MFA4001/1MFA4004VENDOR 3: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 3: 11978MODEL 3: 08008-03	4	4	10	1	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (A) OXYGENUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8148N	M0765	REGULATOR, VACUUM	4	VV	VACUUM PRESSURE REGULATOR FOR CONNECTION TOCENTRAL PIPED VACUUM SYSTEM.STANDARD DISPLAY SCALEIS GRADUATED AT LEAST FROM 0 TO 200 MM HG OFVACUUM.DISPLAYS ON SPECIALIZED REGULATORS MAYCOVER OTHER VACUUM RANGES.REGULATOR TYPE(CONTINUOUS, INTERMITTENT,CONTINUOUS/INTERMITTENT, SURGICAL, PEDIATRIC,THORACIC, ETC.) AS REQUIRED.TO BE USED INDELIVERY, NEONATAL, PEDIATRICS OR ANY AREA WHERESUCTION IS REQUIRED.DATABASE PRICING REFLECTSCONTINUOUS REGULATORS GRADUATED TO 200 MM HG WITHA FULL LINE VACUUM SELECTION SWITCH AND DISSCONFIGURED INLETS AND OUTLETS.VENDOR 1: OHIO MEDICAL CORPORATIONCAGE 1: 45B74MODEL 1: 6702-1224-907VENDOR 2: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 2: 11978MODEL 2: VACUTRON #22-12-1111VENDOR 3: PRECISION MEDICALCAGE 3: 0UK20MODEL 3: PM3114	3	4	6	1	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (B) VACUUMUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8148N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	1	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THESYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMDVDCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RE5MODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8148N	M4116	MONITOR, VITAL SIGNS	1	VV	ELECTRONIC SPHYGMOMANOMETER.LCD DISPLAYSNON-INVASIVE BLOOD PRESSURE, PULSE RATE ANDTEMPERATURE.USED IN HOSPITALS AND CLINICS.INCLUDESAN OPTIONAL MOBILE STAND.DIMENSIONS ARE FOR MONITOR ONLY.VENDOR 1: WELCH-ALLYNCAGE 1: 05CJ3MODEL 1: 530T0-E1G & 008-0835-01VENDOR 2: DATASCOPE CORPORATIONCAGE 2: 1L753MODEL 2: DUO W/TEMP PROBE & STANDVENDOR 3: CARDINAL HEALTH PRODUCTS GROUP, CARDINAL HEALTHCAGE 3: 07TA6MODEL 3: 420TB-E1-M	10	6	10	6	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8148N	M4250	PUMP, SYRINGE, INFUSION	1	VV	THE INFUSION SYRINGE PUMP ENSURES HIGHLY ACCURATEVOLUME DELIVERY AND CONSISTENT FLOW FOR SMALLVOLUMES (<50 ML) OF PHARMACOLOGIC AGENTS OR THICKFEEDING SOLUTIONS.IT SHALL BE SMALL, LIGHTWEIGHTCONSTRUCTION,MAKING IT TRANSPORTABLE.SHALL HAVEMENU-DRIVEN PROGRAMMING CAPABLE OF FLOW RATES(E.G. 0.1 OR 1.0 ML/HR) THAT ARE INTENDED FORLONG-TERM BEDSIDE USE AND/OR CRITICAL CARE PATIENTTRANSPORT, PLUNGER POSITIONING SENSOR, LCD DISPLAYFOR EASY VIEWING, VOLUME LIMIT PROGRAMMING TOSERVE AS A CONVENIENT CUE OF VOLUME OR DOSEDELIVERY COMPLETION AND MULTIPLE DELIVERY MODESFOR ALL APPLICATIONS REQUIRING PRECISELYCONTROLLED INFUSION RATES.THE INFUSION PUMP SHALLHAVE AUTOMATIC SYRINGE SIZE SENSING WHICH WILLGIVE THE FLEXIBILITY TO ACCEPT A WIDE RANGE OFSYRINGE SIZES (UP TO 60 ML) FROM DIFFERENTMANUFACTURERS.SHALL BE BATTERY POWERED/AC ADAPTER.VENDOR 1: SMITHS MEDICALCAGE 1: 0N0T8MODEL 1: MEDFUSION 3500VENDOR 2: B. BRAUN MEDICAL INCORPORATEDCAGE 2: 0Y7P7MODEL 2: PERFUSOR SPACE (8713030U)VENDOR 3: BAXTER HEALTHCARE CORP., IV SYSTEMS/GOVERNMENT SALESCAGE 3: 04687MODEL 3: SIGMA SPECTRUM 35700BAX	11	7	9	4	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 120WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8148N	M4255	STAND, IV, ADJUSTABLE	1	VV	ADJUSTABLE IV STAND WITH 4-HOOK ARRANGEMENT.STANDHAS STAINLESS STEEL CONSTRUCTION WITH HEAVY WEIGHTBASE.IT ADJUSTS FROM 68 INCHES TO 100 INCHES ANDIS MOUNTED ON CONDUCTIVE RUBBER BALL BEARING SWIVEL CASTERS STAND IS USED FOR ADMINISTERINGINTRAVENOUS SOLUTIONS.VENDOR 1: BUCKMAN, INCORPORATEDCAGE 1: 4M773MODEL 1: 7793SS-4VENDOR 2: CLINTON INDUSTRIES, INC.CAGE 2: 0UKA5MODEL 2: IVS-354VENDOR 3: PEDIGO PRODUCTS, INC.CAGE 3: 13180MODEL 3: P-1576-4	24	24	108	35	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 1AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8148N	M5030.1	STOOL, REVOLVING	1	VV	5 CASTER STOOL PNEUMATIC;14" ROUND CUSHIONED SEATONE TOUCH ADJUSTMENT FROM 18"-23";2.5"CASTERS	14	14	27	25	FLOOR				YES	NO
8148N	M7009	BED, PATIENT, ELECTRIC, BARIATRIC	1	VV	ELECTRIC BARIATRIC PATIENT BED WITH A WEIGHTCAPACITY OF 1,000 POUNDS.THE BED WILL INCLUDE ANELECTRONIC BED SCALE TO ALLOW PATIENT TO BEWEIGHED WHILE IN BED.BED INCLUDES MATTRESS(FOAMOR AIR), SAFETY SIDE RAILS, IV ROD INSERTION SLOTSAND PATIENT CONTROL PANEL FOR A FULL RANGE OFPOSITIONS.ALSO INCLUDES FOUR CASTERS WITH LOCKINGBRAKES.PRICE MAY VARY BASED ON OPTIONS AND ACCESSORIESSELECTED.MODEL 2 IS PROVIDED AS A RENTAL AT THEPRESENT TIME.VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 1: 45044MODEL 1: BARI 10-A 2239 W/2236-000-000VENDOR 2: HILL-ROM CO., INC.,A HILLENBRAND INDUSTRYCAGE 2: 28713MODEL 2: EXCEL CARE ES BEDVENDOR 3:CAGE 3: 24285MODEL 3: ELITE 5154	54	98	34	700	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 6AMPS 2: 0WATTS 1: 720WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
8148N	M7040	TABLE, OVERBED	1	VV	OVERBED TABLE.ADJUSTABLE HEIGHT TABLE CONSTRUCTEDOF HEAVY GAUGE STEEL.MOUNTED ON 2" DIAMETER TWINSWIVEL CASTERS WITH BUMPER CAPS.TABLE TOP ISCONSTRUCTED WITH A HIGH PRESSURE PLASTIC LAMINATEDSURFACE THAT RESISTS CHIPPING, SCRATCHING, ANDSTAINING.IT INCLUDES A VANITY TRAY AND A MIRROR.TABLE IS DESIGNED FOR USE OVER BED, WHEELCHAIR ORLARGE CHAIR.VENDOR 1: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 1: 45044MODEL 1: SINGLE TOP W/VANITYVENDOR 2: HILL-ROM CO., INC.,A HILLENBRAND INDUSTRYCAGE 2: 28713MODEL 2: 0BT220 W/LOW BASEVENDOR 3: GF HEALTH PRODUCTS, INC.CAGE 3: 1JY01MODEL 3: A797002 W/VANITY &MIRROR	18	36	46	58	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO

ROOM #	JSN	NAME	QTY	ACQ/INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8148N	U0008	LIFTER, PATIENT	1	VC	An overhead rail system specifically designed for patient lifting and movement for a single bed patient room. The system will consist of recessed or ceiling mounted primary and secondary rails, lift motor with rolling carriage,patient harness or seat, and a hand controller or control box with charger.System will facilitate lifting and movement of patient to and from bed to gurney, chair or other requirement.Minimum lift capability is 550 pounds.Custom design of track layout by manufacturer is essential to meet individual facility requirements.				0	CEILING				YES	NO
8148N	U0008.3	LIFTER, PATIENT, ACCESSORY, SLING	1	VV	Bariatric 4 point patient lifter sling 47x44X56				0	SPACE				NO	NO
8148N	U0030	WALL CHANNEL	1	VV	19"; Aluminum; seismic wall channel with hardware				0	WALL				YES	NO
8148N	U0031	COMPUTER ACCESSORY	1	VV	Mounting assembly and accessories. Includes CPU holder, (1) variable height, adjustable monitor arm supporting up to 40lbs. 180 degree stop mechanism to prevent wall damage, monitor extends 26.25 inches from wall.				0	WALL				YES	NO
8148N	U0032	COMPUTER ACCESSORY	1	VV	Undercounter Mounting assembly for computer keyboard and mouse; ergonomic keyboard holder with tilt and height adjustable; powder coated steel.				0	COUNTER				YES	NO
8148N	U0034	COMPUTER ACCESSORY, BRACKET, WALL	1	VV	27" track to support computer accessories including monitor, keyboard arm, CPU holder. Designed to be mounted to a wall stud or mounting board. Includes channel for cables and covers for sealed cable management.				0	WALL				YES	NO
8148N	U0036	BRACKET	2	VV	Rail mount bracket mount for regulators, wall bottle bracket, disposable bottle brackets, and blenders 5X1X6				0	WALL				YES	NO
8148N	U0039	HEADWALL, PATIENT, ACCESSORY	1	VV	Basket; Mounts on Headwall Accessory Rail; 6 1/2" W X 4"D X 5"H; Functional loading to 5 lbs.. (Meant to hold Bar Code Reader)				0	SPACE				YES	NO
ALCOVE, FOOD CART															
8204AN	K1285	CART, BUFFET, MEDIUM	1	CC	MEDIUM SIZE BUFFET CART.THE CART CAN SUPPORT SELF-SERVICE ITEMS SUCH AS A LARGE SOUP TUREEN OR COFFEE SERVICE. THE UNIT HAS STAINLESS STEEL CONSTRUCTION AND HEAVY DUTY CASTERS.THE CABINET TOP CAN BE STAINLESS STEEL OR WOOD LAMINATE.TWO CASTERS ARE FIXED AND TWO CASTERS SWIVEL AND HAVE BRAKES.VARYING MODELS HAVE LAMINATED SIDES OVER THE STAINLESS STEEL AND STORAGE AREAS BENEATH THE CART TOP.VENDOR 1: LAKESIDE MANUFACTURING, INC.CAGE 1: 16063MODEL 1: 675VENDOR 2: PIPER PRODUCTS, INC.CAGE 2: SR256MODEL 2: 2ST (500)VENDOR 3: DUKE MANUFACTURING COMPANYCAGE 3: 18759MODEL 3: TST-32PG	32	41	39	225	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
ALCOVE, CRASH CART															
8204BN	E0954.1	CART, RESUSCITATION, CARDIAC	1	VV	CART, EMERGENCY, MOBILE - 42W X 30D X 73H "THIS TYPICAL INCLUDES: 1 Cart body, style-A narrow, w/raised edge top 1 Accessory rail, side 1 Accessory rail, back 1 Defibrillator tray 1 IV pole 1 Breakaway bar 1 Flip-up shelf 1 Wastebasket 1 Oxygen tank holder 1 Electrical box-4 outlet 1 Cord wrap 4 Drawer, 3"H 3 Drawer, 6"H Drawer organizer bins."  Herman Miller for Healthcare , Inc. - Model 0BAV4				0	FLOOR				YES	YES

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS															
ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
8204BN	M7650	DEFIBRILLATOR/MONITOR, ACUTE CARE	1	VV	PORTABLE DEFIBRILLATOR/MONITOR FOR ACUTE CAREINCLUDES BIPHASIC DEFIBRILLATOR, PACING, SPO2,INTERPRETIVE 12-LEAD, NIBP MONITORING, ETCO2MONITORING, INVASIVE PRESSURE MONITORING, VITALSIGN MONITORING, TEMPERATURE PROBE, FAXTRANSMISSION, PCMCIA DATA CARDS, PADDLEACCESSORIES, AND A COLOR LCD.MODEL 3 PRICE INCLUDES OPTIONS.MODEL 3 HASRECEIVED DOD APPROVAL AS MEETING MILITARYAIRWORTHINESS STANDARDS FOR USE IN ROTARY ANDFIXED-WING AIRCRAFT.VENDOR 1: ZOLL MEDICAL CORPORATIONCAGE 1: AM088MODEL 1: M SERIES CCTVENDOR 2: PHYSIO-CONTROL, INC.CAGE 2: 28494MODEL 2: LIFEPAK 15 W/OPTIONSVENDOR 3: PHILIPS MEDICAL SYSTEMS, N.A.CAGE 3: 1CS84MODEL 3: HEARTSTART MRX ALX (#M3536A)	16	10	14	20	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (D) 120 AND 208/220 VOLTUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
8204BN	M8755.1	ASPIRATOR, LOW PRESSURE/LOW VOLUME, SURGICAL	1	VV	LOW PRESSURE, LOW VOLUME SURGICAL ASPIRATOR.STANDMOUNTED STAINLESS STEEL UNIT.1/8 HP MOTOR WITHVARIABLE SUCTION INTENSITY.IT INCLUDES 600 MLGRADUATED COLLECTION BOTTLE, HYDROPHOBIC BACTERIAFILTER, A PRECISION SUCTION REGULATING VALVE,TIMER AND A FOOT REGULATOR.FOR POST OPERATIVESURGICAL WOUND DRAINAGE AND GENERAL SUCTION USE.VENDOR 1: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 1: 11978MODEL 1: 3020VENDOR 2: OHIO MEDICAL CORPORATIONCAGE 2: 45B74MODEL 2: MOBLVAC III (#756000)VENDOR 3: MEDICAL DEVICE RESOURCE CORPORATIONCAGE 3: AM300MODEL 3: LS2	18	17	33	110	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
MACERATOR ROOM															
8204cN	U0012	WASTE MANAGEMENT SYSTEM	1	VV	FOR DISPOSAL OF VERNACARE DISPOSABLE PRODUCTS Macerator will dispose a maximum of four disposable products such as bedpan liners, bottles etc, in a single cycle.	18	27	54	250	FLOOR	110V,12A,60HZ	4 GAL/MIN SUPPLY,25 LITERS/CYCLE,2-INCH DRAIN		YES	NO
ALCOVE, NURSE/TD (QTY 2)															
8204DN	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	2	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THESYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMDVDCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RE5MODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
8204DN	U0033	COMPUTER ACCESSORY	2	VV	Mounting assembly and accessories. Includes CPU holder, (1) variable height, adjustable monitor arm supporting up to 40lbs. 180 degree stop mechanism to prevent wall damage, monitor extends 26.25 inches from wall. (1) variable height, adjustable Keyboard arm and tilting keyboard tray with mousepad; keyboard (including platform) extends up to 36.75 inches form wall; folds to a maximum depth fo 5 inches from wall. 26X37X30				0	WALL				YES	NO
8204DN	U0034	COMPUTER ACCESSORY, BRACKET, WALL	2	VV					0	WALL				YES	NO
ALCOVE (QTY 3)															
8204EN	E0963	CART, ISOLATION	3	VV	YELLOW;4 DRAWER;W/TOP MAT/DRAWERLABEL HOLDERSTO INCLUDE:-1 EA 12093 CLEAR TOP MAT-DRAWER LABEL HOLDERSPRE-DRILLED ACCESSORY POSITIONS #2 AND #4 ON BOTHEND PANELS	31	24	43	0	FLOOR				YES	NO
17N BEHAVIORAL HEALTH															
STAFF LOUNGE															

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17038C	A1066	MIRROR, FLOAT GLASS, WITH SS FRAME	1	CC	A HIGH QUALITY 1/4" POLISHED FLOAT GLASS MIRROR36X18, FRAMED IN A ONE-PIECE, BRIGHT POLISHED,STAINLESS STEEL CHANNEL FRAME WITH 90° MITERED CORNERS.ALL EDGES OF THE MIRROR ARE PROTECTED BYABSORBING FILLER STRIPS.MIRROR HAS A GALVANIZEDSTEEL BACK WITH INTEGRAL HORIZONTAL HANGINGBRACKETS AND WALL HANGER FOR CONCEALEDMOUNTING.FOR MOUNTING ABOVE SINGLE WALL MOUNTEDLAVATORIES LOCATED IN TOILET AREAS. DOCTORSEXAMINATION OFFICES, ETC.MAY ALSO BE USED ABOVEDOUBLE LAVATORIES, EITHER WALL OR COUNTERTOPMOUNTED, FOUND IN RESTROOM AREAS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-165 1836VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 781-1836VENDOR 3: MEEK MANUFACTURING COMPANYCAGE 3: SR301MODEL 3: 1500 (18X36)	18	1	36	22	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17038C	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
17038C	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOU515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17038C	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17038C	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17038C	F2017.1	WASTE RECEPTACLE, STEP-ON	1	VV	23 GAL.;MOBILE;LEAK-PROOF BASE;TIGHT FITTING LIDCALIFORNIA FIRE MARSHALL APPROVED; ROUNDED CORNERS; DEODORANT BLOCK HOLDER; COLOR:BEIGEWITH LINERS	20	17	33	18	FLOOR				YES	NO
17038C	K1550	COFFEE BREWER	1	VV	12 cup drip coffee maker 9x10x13				0	COUNTER				YES	YES
17038C	K4665	OVEN, MICROWAVE, CONSUMER	1	VV	COUNTER MOUNTED MICROWAVE OVEN FOR AVERAGE DUTYUSE.THE EXTERIOR CABINET CAN BE METAL OR HEAVYDUTY IMPACT RESISTANT PLASTIC.THE OVEN DELIVERSINSTANT ENERGY FOR RAPID HEATING, DEFROSTING ORPRIME COOKING.THE OVEN HAS TOUCH PAD CONTROLS,DIGITAL TIMER, POWER LEVEL SELECTOR ANDPREPROGRAMMED SELECTORS FOR COMMONLY COOKEDITEMS.THIS OVEN IS COMMONLY FOUND IN STAFFLOUNGES.VENDOR 1: KITCHENAIDCAGE 1: AM179MODEL 1: KCM51555RWHVENDOR 2: SEARS, COMMERCIAL SALESCAGE 2: SR050MODEL 2: 02074062000VENDOR 3: GENERAL ELECTRIC APPLIANCESCAGE 3: 0G0L9MODEL 3: JE2160WF	24	20	14	44	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 15AMPS 2: 0WATTS 1: 1800WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
17038C	R7250.1	REFRIGERATOR/FREEZER, UPRIGHT	1	VV	TOP FREEZER;WHITE;21 CUFT;NO ICE;REVERSIBLE DOOR	30	32	70	205	FLOOR	115V;60HZ;15A			YES	NO
STORAGE/ LAUNDRY															
17043C	M2325	WASHER/DRYER, CLOTHES, STACKABLE	1	VV	AN AUTOMATIC, STACKABLE CLOTHES WASHER/DRYER.THEUNIT WILL HAVE, AS A MINIMUM, A 2.5 CU.FT., 8CYCLE WASHER AND 5.9 CU.FT. ELECTRIC DRYER ANDWILL BE STACKED ONE ON TOP OF THE OTHER.MODEL 3 IS A 220V WASHER/DRYER.VENDOR 1: WHIRLPOOL CORPORATIONCAGE 1: 0FVB1MODEL 1: WET33000X0VENDOR 2: GENERAL ELECTRIC COMPANY, APPLIANCE DIVISIONCAGE 2: 9A849MODEL 2: GUTP270EMWWVENDOR 3: SEARS, COMMERCIAL SALESCAGE 3: SR050MODEL 3: 02608362000	28	32	76	312	FLOOR	VOLTS 1: 120VOLTS 2: 220HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 1AMPS 1: 30AMPS 2: 30WATTS 1: 3600WATTS 2: 6600	UTILITY1: Plumbing (Water and Drainage) (D) COLD AND HOT WATER AND DRAINUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
RESIDENTS															
17044C	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
17044C	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17044C	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZER.DISPENSER.FOR USE THROUGHOUT A HEALTHCARE FACILITY.UNIT DOES NOT INCLUDE THE SANITIZING LIQUID. UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSER.VENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17044C	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	6	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALL CONSIST OF A CENTRAL PROCESSING MINI TOWER, FLAT PANEL MONITOR, KEYBOARD, MOUSE AND SPEAKERS.THE SYSTEM SHALL HAVE THE FOLLOWING MINIMUM CHARACTERISTICS: A 2.8 GHZ PENTIUM PROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROM/DVD COMBO; 1.44MB NETWORK INTERFACE CARD; VIDEO 32MB NVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTER IS USED THROUGHOUT THE FACILITY TO INPUT, MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIME OF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RESMODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17044C	M1825	PRINTER, COMPUTER	1	VV	HIGH RESOLUTION COMPUTER PRINTER WITH A VARIETY OF TYPE STYLES AND SHEET/ENVELOPE FEEDER TRAYS.DATABASE INFORMATION REFLECTS NETWORK READY,MEDIUM DUTY OFFICE STYLE LASER PRINTERS.OTHER TYPES OF PRINTERS (BUBBLE JET, DOT MATRIX, LINE OR PLOTTER) AS WELL AS LIGHT OR HEAVY USE CAPABILITIES ARE AVAILABLE.VENDOR 1: LEXMARK INTERNATIONAL, INC. INCORPORATEDCAGE 1: 0SD37MODEL 1: C736NVENDOR 2: HEWLETT PACKARD COMPANYCAGE 2: 28480MODEL 2: CP3524DN (CC470A)VENDOR 3: BROTHER INDUSTRIES, LTD.CAGE 3: 65533MODEL 3: HL-4570CDWT	21	20	18	73	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 6AMPS 2: 0WATTS 1: 720WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
DINING/ACTIVITY/REC															
17083C	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
17083C	A5075.1	DISPENSER, LIQUID, SOAP/DISINFECTANT	1	CC	Tamper proof soap dispenser. One-handed dispensing operation. Designed to accommodate disposable soap cartridge and valve.				0	WALL				NO	NO
17083C	A5080.1	DISPENSER, PAPER TOWEL	1	CC	A recessed, C-fold, paper towel dispenser. Dispenser features: tumbler lock; front hinged at bottom; and refill indicator slot. Minimum capacity 400 towels. For use in Behavior Health only.				0	WALL				YES	NO
SERVERY															
17084C	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
17084C	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER. ONE-HANDED DISPENSING OPERATION. DESIGNED TO ACCOMMODATE DISPOSABLE SOAP CARTRIDGE AND VALVE. BASED ON INFECTION CONTROL STANDARDS ONLY SOAP DISPENSERS REFILLED BY MEANS OF A CARTRIDGE OR REPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 1HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17084C	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL, SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSER FEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM; AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400 SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USE THROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INC. INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 1HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17084C	K1550.1	COFFEE BREWER	1	VV	COUNTERTOP; DISPENSER PROVED ON A SERVICE; PLUMBED UNIT	13	21	30	0	COUNTER	120V;60HZ	1/4" WATER LINE; PRESSURE 10-120 PSI		YES	NO
17084C	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	1	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALL CONSIST OF A CENTRAL PROCESSING MINI TOWER, FLAT PANEL MONITOR, KEYBOARD, MOUSE AND SPEAKERS.THE SYSTEM SHALL HAVE THE FOLLOWING MINIMUM CHARACTERISTICS: A 2.8 GHZ PENTIUM PROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROM/DVD COMBO; 1.44MB NETWORK INTERFACE CARD; VIDEO 32MB NVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTER IS USED THROUGHOUT THE FACILITY TO INPUT, MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIME OF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RESMODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17084C	M1825	PRINTER, COMPUTER	1	VV	HIGH RESOLUTION COMPUTER PRINTER WITH A VARIETY OFTYPE STYLES AND SHEET/ENVELOPE FEEDERTRAYS.DATABASE INFORMATION REFLECTS NETWORK READY.MEDIUM DUTY OFFICE STYLE LASER PRINTERS.OTHERTYPES OF PRINTERS (BUBBLE JET, DOT MATRIX, LINE ORPLOTTER) AS WELL AS LIGHT OR HEAVY USECAPABILITIES ARE AVAILABLE.VENDOR 1: LEXMARK INTERNATIONAL, INCORPORATEDCAGE 1: 0SD37MODEL 1: C736NVENDOR 2: HEWLETT PACKARD COMPANYCAGE 2: 28480MODEL 2: CP3524DN (CC470A)VENDOR 3: BROTHER INDUSTRIES, LTD.CAGE 3: 65533MODEL 3: HL-4570CDWT	21	20	18	73	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 6AMPS 2: 0WATTS 1: 720WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17084C	R7250.1	REFRIGERATOR/FREEZER, UPRIGHT	1	VV	TOP FREEZER,WHITE;21 CUFT:NO ICE;REVERSIBLE DOOR	30	32	70	205	FLOOR	115V,60HZ;15A			YES	NO
17084C	U0019	TABLE	1	VV	STEAM TABLE				0	FLOOR				NO	NO
17084C	U0020	EQUIPMENT	1	CC	STARTER STATION				0	FLOOR				NO	NO
STORAGE															
17086C	M2055.5	SHELVING, WIRE	1	VV	5 WIRE CHROME SHELVES;MOBILE;SWIVEL& BRAKE CASTERS	60	24	69	112	FLOOR				YES	NO
17086C	M4655	STRETCHER, MOBILE, CRS, 9 POSITION	1	VV	MOBILE STRETCHER.ALL CORROSION RESISTANT STAINLESSSTEEL CONSTRUCTION.IT CONSISTS OF A TUBULAR FRAMEWITH SIDE RAILS, A 9-POSITION HYDRAULIC BASE WITHPNEUMATIC FOWLER ADJUSTMENT, AND A 2" PAD.UNIT ISMOUNTED ON 8" CONDUCTIVE CASTERS.DESIGNED FORPATIENT TRANSPORT AS WELL AS FOR MINOR SURGICALPROCEDURES.PRICE MAY VARY BASED ON OPTIONS OR ACCESSORIESSELECTED.PRICES SHOWN ARE WITH STANDARD FEATURESONLY.VENDOR 1: HILL-ROM CO., INC.,A HILLENBRAND INDUSTRYCAGE 1: 28713MODEL 1: P8005VENDOR 2: STRYKER CORPORATION, MEDICAL PRODUCTS DIV.CAGE 2: 45044MODEL 2: 738VENDOR 3: STERIS CORPORATIONCAGE 3: 02TU3MODEL 3: 461-EMC-ST	34	83	36	228	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
GROUP ROOM															
17087N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
17087N	A5075.1	DISPENSER, LIQUID, SOAP/DISINFECTANT	1	CC	Tamper proof soap dispenser. One-handed dispensing operation. Designed to accommodate disposable soap cartridge and valve.				0	WALL				NO	NO
17087N	A5080.1	DISPENSER, PAPER TOWEL	1	CC	A recessed, C-fold, paper towel dispenser. Dispenser features: tumbler lock; front hinged at bottom; and refill indicator slot. Minimum capacity 400 towels. For use in Behavior Health only.				0	WALL				YES	NO
17087N	A5210.1	TELEVISION ACCESSORY, BRACKET, WALL MOUNTED	1	CC	Bracket, Television, Wall Mounted, w/Adjust Arm 40W X 4D X 13H				0	WALL				YES	NO
17087N	M0430	RECORDER/ PLAYER, DIGITAL, VIDEO, DVD	1	VV	PROFESSIONAL MEDICAL GRADE DVD VIDEORECORDER/PLAYER.APPROXIMATELY 5" HIGH X 10" WIDE X14" DEEP. MEETSNITSC AND PAL STANDARDS AND EMPLOYSMPEG2 VIDEO COMPRESSION.CAPABLE OF HIGH SPEEDPICTURE SEARCH, DIRECT RECORDING INTO AVAILABLEDISC SPACE AND RECORDING AND PLAYBACK ON VARIOUSDVD FORMATS.AUDIO USES DOLBY DIGITALFORMAT.CONVENIENCE OUTLET REQUIRED AT POINT OFUSE.MODELS 1 AND 2 ALSO RECORD AND PLAY BLU-RAY DISCS.VENDOR 1:CAGE 1: 50356MODEL 1: BD-R2000VENDOR 2: JVC AMERICACAGE 2: 3NPW4MODEL 2: SR-HD1500VENDOR 3: PANASONIC USA, GOVERNMENT MARKETINGCAGE 3: 5Y727MODEL 3: LQ-MD800	19	15	5	12	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17087N	M0506.2	TELEVISION, FLAT PANEL LCD	1	VV	Color television with HD.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance. 42" wall mounted unit. 120V,60HZ				0	WALL				YES	NO
INTERVIEW															
17088N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
PCTC															
17089N	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
17089N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
17089N	A5145.2	HOOK, GARMENT	1	CC	Fabricate hook units either of chromium plated brass with a satin finish, or stainless steel, using 6 mm (1/4 inch) minimum thick stock, with edges and corners rounded smooth to the thickness of the metal, or 3 mm (1/8 inch) minimum radius. Fabricate each unit as a double hook on a single shaft.				0	WALL				NO	NO
17089N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	1	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDNSPEAKERS.THE SYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4:	BTU/HR: 0	NO	NO
17089N	M1825	PRINTER, COMPUTER	1	VV	HIGH RESOLUTION COMPUTER PRINTER WITH A VARIETY OFTYPE STYLES AND SHEET/ENVELOPE FEEDERTRAYS.DATABASE INFORMATION REFLECTS NETWORK READY.MEDIUM DUTY OFFICE STYLE LASER PRINTERS.OTHERTYPES OF PRINTERS (BUBBLE JET,	21	20	18	73	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 6AMPS 2: 0WATTS 1: 720WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4:	BTU/HR: 0	NO	NO



ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
2 BED BEHAV HEALTH (QTY 2)															
17091N	F2000	BASKET, WASTEPAPER, ROUND, METAL	2	VV	ROUND WASTEPAPER BASKET, APPROXIMATELY 18" HIGH X16" DIAMETER.THIS METAL UNIT IS USED TO COLLECTAND TEMPORARILY STORE SMALL QUANTITIES OF PAPERREFUSE IN PATIENT ROOMS, ADMINISTRATIVE AREAS ANDNURSING STATIONS.VENDOR 1: JACOBS	16	16	18	6	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N)	BTU/HR: 0	YES	NO
SOILED UTILITY															
17093N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17093N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17093N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17093N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	1	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH AWHITE BAKED ENAMEL FINISH, PLASTIC ORACRYLIC.PROVIDED WITH WALL BRACKET TO FACILITATEMOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10.VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M644MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
17093N	F2017.2	WASTE RECEPTACLE, STEP-ON	1	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
17093N	M3070.1	HAMPER, LINEN	1	VV	Hamper, Linen, Mobile, w/Lid 21W X 22D X 38H				0	FLOOR				YES	NO
17093N	U0006	DISPOSAL CONTAINER, WASTE, PHARMACEUTICAL	1	VV	Plastic 18 Gallon, Chemotherapy waste disposal container with sliding lid.	18	13	26	8	FLOOR				YES	NO
17093N	U0012	WASTE MANAGEMENT SYSTEM	1	VV	FOR DISPOSAL OF VERNACARE DISPOSABLE PRODUCTS Macerator will dispose a maximum of four disposable products such as bedpan liners, bottles etc. in a single cycle.	18	27	54	250	FLOOR	110V,12A,60HZ	4 GAL/MIN SUPPLY,25 LITERS/CYCLE,2-INCH DRAIN		YES	NO
17093N	U0018	WASTE RECEPTACLE, STEP-ON	1	VV	24 GAL;RED;FOOT PEDAL;RETAINER BAND;QUIET OPERATION;LINERS Red Biohazard. Rectangular waste receptacle with step-on lid and 24 gallon capacity	15	15	30	36	FLOOR				YES	NO
INTERVIEW/LOUNGE/VISITING															
17094N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS. 17W X 2D X 11H				0	WALL				YES	NO
17094N	A5210.1	TELEVISION ACCESSORY, BRACKET, WALL MOUNTED	1	CC	Bracket, Television, Wall Mounted, w/Adjust Arm 40W X 4D X 13H				0	WALL				YES	NO
17094N	M0506.2	TELEVISION, FLAT PANEL LCD	1	VV	Color television with HD.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance. 42" wall mounted unit. 120V;60HZ				0	WALL				YES	NO
BATH TLT (QTY 17)															
17095AN	A1066.1	MIRROR	17	CC	Polycarbonate mirror. Approximate size 18W x 36H. Must be non-shatterproof and non-breakable.				0	WALL				NO	NO
17095AN	A5075.1	DISPENSER, LIQUID, SOAP/DISINFECTANT	17	CC	Tamper proof soap dispenser. One-handed dispensing operation. Designed to accommodate disposable soap cartridge and valve.				0	WALL				NO	NO
17095AN	A5080.1	DISPENSER, PAPER TOWEL	17	CC	A recessed, C-fold, paper towel dispenser. Dispenser features: tumbler lock; front hinged at bottom; and refill indicator slot. Minimum capacity 400 towels. For use in Behavior Health only.				0	WALL				YES	NO
17095AN	A5109.3	GRAB BAR	51	CC	42" long SECURITY grab bar. A 1-1/4" diameter. Grab bar shall comply with barrier-free accessibility guidelines for structural strength. For typical water closet applications in toilet stalls and rooms where ADA (American's With Disabilities Act) requirements must be met.				0	WALL				YES	NO

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17095AN	A5109.7	GRAB BAR	17	CC	18" long SECURITY grab bar. A 1-1/4" diameter. Grab bar shall comply with barrier-free accessibility guidelines for structural strength. For typical water closet applications in toilet stalls and rooms where ADA (American's With Disabilities Act) requirements must be met.				0	WALL				YES	NO
17095AN	A5109.8	GRAB BAR	34	CC	36" long SECURITY grab bar. A 1-1/4" diameter. Grab bar shall comply with barrier-free accessibility guidelines for structural strength. For typical water closet applications in toilet stalls and rooms where ADA (American's With Disabilities Act) requirements must be met.				0	WALL				YES	NO
17095AN	A5145.1	HOOK, GARMENT	17	CC	NYS OMH Approved Collapsible, anti-ligature type; stainless steel hook snaps down with weight in excess of 20 to 25 pounds is applied.				0	WALL				NO	NO
17095AN	A5165.1	SHELF	17	CC	RECESSED SHELF: 17W X 4D X 7H				0	WALL				YES	NO
17095AN	A5200.1	DISPENSER, TOILET TISSUE	17	CC	STAINLESS STEEL, SINGLE ROLL, RECESSED				0	WALL				NO	NO
17095AN	F2017.2	WASTE RECEPTACLE, STEP-ON	17	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
1 BED BEHAV HEALTH (QTY 14)															
17095N	F2000	BASKET, WASTEPAPER, ROUND, METAL	14	VV	ROUND WASTEPAPER BASKET, APPROXIMATELY 18" HIGH X16" DIAMETER.THIS METAL UNIT IS USED TO COLLECTAND TEMPORARILY STORE SMALL QUANTITIES OF PAPERREFUSE IN PATIENT ROOMS, ADMINISTRATIVE AREAS ANDNURSING STATIONS.VENDOR 1: JACOBS GARDNERCAGE 1: 7R579MODEL 1: SAF96048LVENDOR 2: BRODHEAD-GARRETT COMPANYCAGE 2: 89791MODEL 2: 068127VENDOR 3: RUBBERMAID COMMERCIAL PRODUCTS LLC CAGE 3: 2H789MODEL 3: WB44	16	16	18	6	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
STORAGE															
17097N	M4801	COMMODE CHAIR	1	VV	COMMODE CHAIR TO BE USED IN URODYNAMICSTESTING.ALUMINUM FRAME WITH PLASTIC SEAT ANDFUNNEL ATTACHED.CHAIR TO BE USED IN URODYNAMICSTESTING.FREE-STANDING ADJUSTABLE HEIGHTFAPPROXIMATELY 27" TO 31".ALUMINUM FRAME WITH ARMRESTS, PLASTIC SEAT AND FUNNEL.NO OTHER SOURCE(S) IDENTIFIED WITH THISDESCRIPTION/SPECIFICATION.VENDOR 1: SRS MEDICALCAGE 1: SR351MODEL 1: 84411206VENDOR 2: LIFE-TECH, INC.CAGE 2: 63965MODEL 2: SA0591AVENDOR 3:CAGE 3:MODEL 3:	21	20	33	16	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
17097N	X2126	SCANNER, ULTRASOUND, PORTABLE, URINARY RETENTION	1	VV	A PORTABLE 3D ULTRASOUND SCANNER FOR MEASURING THEURINARY BLADDER VOLUME AND POST-VOID RESIDUAL(PVR) VOLUME IN ADULTS AND CHILDREN.MEASURESBLADDER VOLUME NONINVASIVELY ON A WIDE RANGE OFPATIENTS.UNIT IS BATTERY OPERATED AND COMES WITH ACHARGING UNIT.A MOBILE STAND IS AVAILABLE FROM BOTH VENDORS.VENDOR 1:CAGE 1: 0T27MODEL 1: BLADDER SCAN BVI 9400VENDOR 2: MEDIWATCH, USACAGE 2: 4RD27MODEL 2: PORTASCAN+ U-0007VENDOR 3:CAGE 3:MODEL 3:	14	10	14	7	SPACE	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
17099N	M2055.5	SHELVING, WIRE	1	VV	5 WIRE CHROME SHELVES:MOBILE;SWIVEL& BRAKE CASTERS	60	24	69	112	FLOOR				YES	NO
17099N	M4801	COMMODE CHAIR	1	VV	COMMODE CHAIR TO BE USED IN URODYNAMICSTESTING ALUMINUM FRAME WITH PLASTIC SEAT ANDFUNNEL ATTACHED CHAIR TO BE USED IN URODYNAMICSTESTING.FREE-STANDING ADJUSTABLE HEIGHTFAPPROXIMATELY 27" TO 31".ALUMINUM FRAME WITH ARMRESTS, PLASTIC SEAT AND FUNNEL.NO OTHER SOURCE(S) IDENTIFIED WITH THISDESCRIPTION/SPECIFICATION.VENDOR 1: SRS MEDICALCAGE 1: SR351MODEL 1: 84411206VENDOR 2: LIFE-TECH, INC.CAGE 2: 63965MODEL 2: SA0591AVENDOR 3:CAGE 3:MODEL 3:	21	20	33	16	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
HAC															
17101N	A5135.1	RACK, MOP/BROOM	1	CC	Shelf, Utility W/ Mop/Broom Holders, SS, Surf Mntd. DIMENSIONS 34W X 8D X 13H				0	WALL				YES	NO
17101N	F0500	CART, JANITOR'S	1	VV	JANITORS CART, 38" HIGH X 66" WIDE X 20" DEEP WITH25 GALLON HEAVY DUTY COLLECTION BAG.ACCOMMODATIONS FOR LOCKING VAULT AND JANITORIAL SUPPORT ITEMS (BROOM, DUST PAN, CHEMICALS, MOP ANDMOP BUCKET; NOT INCLUDED) PRICE DIFFERENCE VERIFIED.VENDOR 1: RUBBERMAID COMMERCIAL PRODUCTS LLC CAGE 1: 2H789MODEL 1: 9T72VENDOR 2: GLOBAL INDUSTRIAL EQUIPMENTCAGE 2: 0JSZ8MODEL 2: WBB54619VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 2U670	22	49	54	77	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
CLEAN SUPPLY															
17104N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNIT'S ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17104N	U0002	BIN, ACCESSORY, PANEL, LOUVERED	1	CC	Wall Hanging system, fits Akrobins for off the floor storage. 2 panel; 36W X 1D X 73H				0	WALL				YES	NO
17104N	U0002.1	BIN, ACCESSORY, PANEL, LOUVERED	1	CC	Wall Hanging system, fits Akrobins for off the floor storage. 2 panel; 36W X 1D X 55H				0	WALL				YES	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17104N	U0002.2	BIN	1	VV	Akrobins for storage on louvered panels. 13x9x6				0	SPACE				NO	NO
LAUNDRY															
17105N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSING OPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAP CARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAP DISPENSERS REFILLED BY MEANS OF A CARTRIDGE OR REPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17105N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZER DISPENSER.FOR USE THROUGHOUT A HEALTHCARE FACILITY.UNIT DOES NOT INCLUDE THE SANITIZING LIQUID. UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17105N	A5080.1	DISPENSER, PAPER TOWEL	1	CC	A recessed, C-fold, paper towel dispenser. Dispenser features: tumbler lock; front hinged at bottom; and refill indicator slot. Minimum capacity 400 towels. For use in Behavior Health only.				0	WALL				YES	NO
17105N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	1	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH A WHITE BAKED ENAMEL FINISH, PLASTIC OR ACRYLIC.PROVIDED WITH WALL BRACKET TO FACILITATE MOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10.VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M644MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
17105N	M2325	WASHER/DRYER, CLOTHES, STACKABLE	2	VV	AN AUTOMATIC, STACKABLE CLOTHES WASHER/DRYER.THE UNIT WILL HAVE, AS A MINIMUM, A 2.5 CU.FT., 8CYCLE WASHER AND 5.9 CU.FT.	28	32	76	312	FLOOR	VOLTS 1: 120VOLTS 2: 220HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT:	UTILITY1: Plumbing (Water and Drainage) (D) COLD AND HOT WATER AND	BTU/HR: 0	YES	NO
QUIT/CONF															
17106N	U0042	MONITOR, VIDEO	1	VC	SKY FACTORY 46" RECESSED MONITOR					WALL				YES	NO
CLEAN LINEN															
17107N	F0515.1	CART, LINEN	1	VV	LINEN/SUPPLY,WIRE,STEM CASTERS,ENCLOSURES,COVERTO INCLUDE THE FOLLOWING:1 EA N66BC,4 WIRE SHELVES,4 POSTS,PLASTIC SPLIT SLEEVES,DONUT BUMPERS,2 EA 6M STEM CASTERS,2 EA 6MB CASTERS WITH BRAKES, CHROME5 EA EP56C-24X63 WIRE ENCLOSURE PANEL FOR SIDES AND BACK1 EA 24X60X62UC UNCOATED CART COVER, ZIPPER,WHITE	60	24	68	0	FLOOR				YES	YES
17107N	F0515M	CART, SUPPLY	1	VV	DOUBLE WIDE,SR 24,RAISED EDGE TOP-FULL SPINE,42HCOLOR:SOFT WHITETO INCLUDE:2 EA CT511.03 DRAWER,3H8 EA CT511.06 DRAWER,6H	59	22	42	0	FLOOR				YES	YES
VESTIBULE															
17111BN	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
RESTRAINT															
17111N	M7005	BED, HI-LOW, MANUAL	1	VV	MANUAL HI-LOW BED.HEAVILY CONSTRUCTED STEEL FRAME WITH CRANK HANDLES TO LOWER THE BED OR OBTAIN A FULL RANGE OF POSITIONS.HANDLE GRIP FOLDS DOWN WHEN NOT IN USE.IT ALSO FEATURES SIX 1/4" SOCKET AND REMOVABLE HEADBOARD MOUNTED ON FOUR 3" BALL BEARING SWIVEL CASTERS.MAY INCLUDE BED RAILS TO BE USED IN HOSPITALS, CLINICS, NURSING HOMES, AND OTHER APPROPRIATE AREAS.PRICE DIFFERENCE VERIFIED.VENDOR 1: PREFERRED HEALTHCARECAGE 1: AM245MODEL 1: IVC 5307VENDOR 2: GRAHAM-FIELD HEALTH PRODUCTS, INC.CAGE 2: 12671MODEL 2: SD0100NVENDOR 3: MED1ONLINECAGE 3: AM359MODEL 3: 449-1081	37	89	33	220	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
TELEPHONE ROOM															
17116N	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
TEAM CHARTING															
17117N	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
17117N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17117N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17117N	A5210.1	TELEVISION ACCESSORY, BRACKET, WALL MOUNTED	1	CC	Bracket, Television, Wall Mounted, w/Adjust Arm 40W X 4D X 13H				0	WALL				YES	NO
17117N	M0506.2	TELEVISION, FLAT PANEL LCD	1	VV	Color television with HD.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo sound and automatic color balance. 42" wall mounted unit. 120V;60HZ				0	WALL				YES	NO
17117N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	2	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THESYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMDVDCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RESMODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
TREATMENT/EXAM															
17118N	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
17118N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
17118N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17118N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17118N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WP40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17118N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	1	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH AWHITE BAKED ENAMEL FINISH. PLASTIC ORACRYLIC.PROVIDED WITH WALL BRACKET TO FACILITATEMOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10.VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M644MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
17118N	A5108.1	DISPOSAL CONTAINER, SHARPS W/BRACKET	1	VV	Waste Disposal Unit, Sharps 16W X 8D X 17H				0	WALL				YES	NO
17118N	A5145.1	HOOK, GARMENT	1	CC	NYS OMH Approved Collapsible, anti-ligature type; stainless steel hook snaps down with weight in excess of 20 to 25 pounds is applied.				0	WALL				NO	NO

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17118N	E0954.1	CART, RESUSCITATION, CARDIAC	1	VV	CART, EMERGENCY, MOBILE - 42W X 30D X 73H "THIS TYPICAL INCLUDES: 1 Cart body, style-A narrow, w/raised edge top 1 Accessory rail, side 1 Accessory rail, back 1 Defibrillator tray 1 IV pole 1 Breakaway bar 1 Flip-up shelf 1 Wastebasket 1 Oxygen tank holder 1 Electrical box-4 outlet 1 Cord wrap 4 Drawer, 3"H 3 Drawer, 6"H Drawer organizer bins."  Herman Miller for Healthcare , Inc. - Model 0BAV4				0	FLOOR				YES	YES
17118N	F2017.2	WASTE RECEPTACLE, STEP-ON	1	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
17118N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	1	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THESYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMDVDCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA; A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RE5MODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17118N	M4020	SCALE, PERSON WEIGHING, HIGH CAPACITY	1	VV	HIGH CAPACITY, PERSON WEIGHING SCALE.LARGE DIGITALREADOUT DISPLAYS WEIGHT IN POUNDS, ORKILOGRAMS.HAND RAILS.SHALL BE SUPPORTED BY FIXEDHEAVY-DUTY LEGS IN STATIONARY POSITION.CAPACITY UPTO 800 POUNDS.OPTIONAL EQUIPMENT:BATTERY PACK WITHRECHARGEABLE BATTERIES AND BUILT-INRECHARGER.DESIGNED TO WEIGH LARGE AMBULATORYPATIENTS.ALL THREE MODELS ARE BATTERY OPERATED.VENDOR 1: SCALE-TRONIX, INC.CAGE 1: 8S293MODEL 1: 5002VENDOR 2: DETECTO, A DIVISION OF CARDINAL SCALE MFG. CO.CAGE 2: 009P6MODEL 2: 6868VENDOR 3: ALGEN SCALE CORPORATIONCAGE 3: AM021MODEL 3: SR555 W/SR3070	32	34	36	166	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
17118N	M5030	STOOL, SURGEON, REVOLVING	1	VV	REVOLVINGSTOOL.CONSISTS OF A PADDED UPHOLSTEREDSEAT WITH HEIGHT ADJUSTMENT.UNIT ROTATES AND ISMOUNTED ON BALL BEARING SWIVEL CASTERS.DESIGNEDFOR USE IN EXAMINATIONS, TREATMENT, AND SURGICALPROCEDURES.VENDOR 1: BOYD INDUSTRIES, INC.CAGE 1: 8V232MODEL 1: BOS-46 (200-4610)VENDOR 2: PEDIGO PRODUCTS, INC.CAGE 2: 13180MODEL 2: P-536-GSVENDOR 3: MIDMARK CORPORATIONCAGE 3: 90768MODEL 3: 425	24	24	25	23	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
17118N	M7401	LIGHT, EXAM, MOBILE, SPOTLIGHT, MOBILE STAND	1	VV	THE EXAM LIGHT SHALL BE A MOBILE FLOOR UNIT.THELIGHT WILL BE A HALOGEN BULB OR LED THATCANPRODUCE A CONTINUOUS AND HOMOGENEOUS SPOT OF LIGHTADJUSTABLE FROM 5 TO 9 INCHES IN DIAMETER FROM ASET DISTANCE.THE LIGHT INTENSITY SHALL BE AMINIMUM OF 750 FOOT-CANDLES AT A DISTANCE OF 16INCHES AND HAVE A COLOR TEMPERATURE OF 3,200DEGREES KELVIN. THE UNIT WILL CONSIST OF AN ARM ORSLEEVE OF APPROXIMATELY 45 INCHES IN LENGTH TOALLOW FOR EASY ARM ROTATION AND ARM MOVEMENTUP ANDDOWN. THE UNIT SHALL BE MOUNTED ON A CASTER BASEFOR EASY MOVEMENT.VENDOR 1: WELCH-ALLYNCAGE 1: 05CJ3MODEL 1: GS EXAM LIGHT (#48810 W/48950)VENDOR 2: MIDMARK CORPORATIONCAGE 2: 90768MODEL 2: 253-006VENDOR 3: BURTON MEDICAL PRODUCTS, INC.CAGE 3: 8T151MODEL 3: SB40FL	21	21	83	30	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 50 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
17118N	M7650	DEFIBRILLATOR/MONITOR, ACUTE CARE	1	VV	PORTABLE DEFIBRILLATOR/MONITOR FOR ACUTE CAREINCLUDES BIPHASIC DEFIBRILLATOR, PACING, SPO2,INTERPRETIVE 12-LEAD, NIBP MONITORING, ETCO2MONITORING, INVASIVE PRESSURE MONITORING, VITALSIGN MONITORING, TEMPERATURE PROBE, FAXTRANSMISSION, PCMCIA DATA CARDS, PADDLEACCESSORIES, AND A COLOR LCD.MODEL 3 PRICE INCLUDES OPTIONS.MODEL 3 HASRECEIVED DOD APPROVAL AS MEETING MILITARYAIRWORTHINESS STANDARDS FOR USE IN ROTARY ANDFIXED-WING AIRCRAFT.VENDOR 1: ZOLL MEDICAL CORPORATIONCAGE 1: AM088MODEL 1: M SERIES CCTVENDOR 2: PHYSIO-CONTROL, INC.CAGE 2: 28494MODEL 2: LIFEPAK 15 W/OPTIONSVENDOR 3: PHILIPS MEDICAL SYSTEMS, N.A.CAGE 3: 1CS84MODEL 3: HEARTSTART MRX ALX (#M3536A)	16	10	14	20	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (D) 120 AND 208/220 VOLTUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17118N	M7710	ELECTROCARDIOGRAPH, 12 LEAD, PORTABLE	1	VV	USED TO DETECT THE ELECTRICAL SIGNALS ASSOCIATEDWITH CARDIAC ACTIVITY, DIAGNOSE CARDIACABNORMALITIES, DETERMINE A PATIENTS RESPONSE TODRUG THERAPY AND REVEAL TRENDS OR CHANGES IN HEARTFUNCTION.CAPABLE OF RECORDING TWO OR MORE LEADSSIMULTANEOUSLY, RECORDING AN ENTIRE 12 LEAD ECG INABOUT 10 SECONDS.INCLUDES OF A 3.5 INCH, HIGHDENSITY, FLOPPY DISK DRIVE FOR TESTSTORAGE.PORTABLE.VENDOR 1: PHILIPS MEDICAL SYSTEMS, N.A.CAGE 1: 1CS84MODEL 1: PAGEWRITER TC 70 (#860315)VENDOR 2: CARDIAC SCIENCE CORPORATIONCAGE 2: 1CDW6MODEL 2: ECLIPSE PREMIER (#E10-1AE17)VENDOR 3: GE HEALTHCARECAGE 3: 50542MODEL 3: MAC 5500 W/TROLLEY	21	28	52	87	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
17118N	M8755.1	ASPIRATOR, LOW PRESSURE/LOW VOLUME, SURGICAL	1	VV	LOW PRESSURE, LOW VOLUME SURGICAL ASPIRATOR.STANDMOUNTED STAINLESS STEEL UNIT.1/8 HP MOTOR WITHVARIABLE SUCTION INTENSITY.IT INCLUDES 600 MLGRADUATED COLLECTION BOTTLE, HYDROPHOBIC BACTERIAFILTER, A PRECISION SUCTION REGULATING VALVE.TIMER AND A FOOT REGULATOR.FOR POST OPERATIVESURGICAL WOUND DRAINAGE AND GENERAL SUCTION USE.VENDOR 1: ALLIED HEALTHCARE PRODUCTS, INC.CAGE 1: 11978MODEL 1: 3020VENDOR 2: OHIO MEDICAL CORPORATIONCAGE 2: 45B74MODEL 2: MOBLVAC III (#756000)VENDOR 3: MEDICAL DEVICE RESOURCE CORPORATIONCAGE 3: AM300MODEL 3: LS2	18	17	33	110	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	YES
17118N	M9025	TABLE, EXAMINATION/TREATMENT, WITH CABINET	1	VV	TABLE, EXAMINATION/TREATMENT WITH CABINET.CABINETTYPE EXAM TABLE WITH THREE SECTION UPHOLSTERED TOPADJUSTABLE TO A FULL CHAIR POSITION.CONCEALED HEELSTIRRUPS, DUAL OUTLETS, DRAWERS, AND STORAGE SPACEWITH OPTIONAL ARMBBOARD.UNIT IS DESIGNED FORGENERAL USE EXAMINATION/TREATMENT ROOMS.VENDOR 1: MIDMARK CORPORATIONCAGE 1: 90768MODEL 1: 604-001VENDOR 2: UNITED METAL FABRICATORS, INC.CAGE 2: 13127MODEL 2: 5140VENDOR 3: HAUSMANN INDUSTRIES, INCORPORATEDCAGE 3: 90694MODEL 3: 4420	27	78	34	330	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
17118N	U0003	BOARD, PATIENT TRANSFER	1	VV	A high-pressure laminated base board for performing side-to-side sliding exercises.				0	WALL				YES	NO
17118N	U0004	CART, CYLINDER, MEDICAL GAS	1	VV	SINGLE D & E,8"SEMI PNEU WHEELS;GURNEY HOOK;GREEN	12	10	41	8	FLOOR				YES	YES
17118N	U0018	WASTE RECEPTACLE, STEP-ON	1	VV	24 GAL;RED;FOOT PEDAL;RETAINER BAND;QUIET OPERATION;LINERS Red Biohazard. Rectangular waste receptacle with step-on lid and 24 gallon capacity	15	15	30	36	FLOOR				YES	NO
17118N	U0037	DIAGNOSTIC SET	1	VV	WALL TRANSFORMER;COAXIAL OPTHALMOSCOPE;MACROVIEWOTOSCOPE;KLEENSPEC DISPENSER;WALL ANERIODTO INCLUDE:1 EA 77710 GS77 WALL TRANSFORMER1 EA 11720 COAXIAL OPTHALMOSCOPE1 EA 23810 MACROVIEW OTOSCOPE1 EA 7670-01 WALL ANEROID1 EA 52400-PF KEELSPEC DISPENSER1 EA 01690-300 SURETEMP PLUS THERMOMETER  Wall Mounted Integrated system on wall board. To include 1 ea. Ophthalmoscope, 1 ea. Diagnostic Macroview Otoscope, 1 ea. Kleenspec Dispenser; aneroid sphygmanometer; thermometer	34	7	12	0	WALL	100-240V;50/60HZ;.,18A MAXTRANSFORMER: 5V;1A			YES	NO
DAY ROOM															
17119N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZB9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17119N	A5080.1	DISPENSER, PAPER TOWEL	1	CC	A recessed, C-fold, paper towel dispenser. Dispenser features: tumbler lock; front hinged at bottom; and refill indicator slot. Minimum capacity 400 towels. For use in Behavior Health only.				0	WALL				YES	NO
17119N	A5210.1	TELEVISION ACCESSORY, BRACKET, WALL MOUNTED	1	CC	Bracket, Television, Wall Mounted, w/Adjust Arm 40W X 4D X 13H				0	WALL				YES	NO
17119N	K4665	OVEN, MICROWAVE, CONSUMER	1	VV	COUNTER MOUNTED MICROWAVE OVEN FOR AVERAGE DUTYUSE.THE EXTERIOR CABINET CAN BE METAL OR HEAVYDUTY IMPACT RESISTANT PLASTIC.THE OVEN DELIVERSINSTANT ENERGY FOR RAPID HEATING. DEFROSTING ORPRIME COOKING.THE OVEN HAS TOUCH PAD CONTROLS.DIGITAL TIMER, POWER LEVEL SELECTOR ANDPREPROGRAMMED SELECTORS FOR COMMONLY COOKEDITEMS.THIS OVEN IS COMMONLY FOUND IN STAFFLOUNGES.VENDOR 1: KITCHENAIDCAGE 1: AM179MODEL 1: KCMS1555RWHVENDOR 2: SEARS, COMMERCIAL SALESCAGE 2: SR050MODEL 2: 02074062000VENDOR 3: GENERAL ELECTRIC APPLIANCESCAGE 3: 0G0L9MODEL 3: JE2160WF	24	20	14	44	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 15AMPS 2: 0WATTS 1: 1800WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17119N	M0430	RECORDER/ PLAYER, DIGITAL, VIDEO, DVD	1	VV	PROFESSIONAL MEDICAL GRADE DVD VIDEORECORDER/PLAYER,APPROXIMATELY 5" HIGH X 10" WIDE X14" DEEP. MEETSNTSC AND PAL STANDARDS AND EMPLOYSMPEG2 VIDEO COMPRESSION,CAPABLE OF HIGH SPEEDPICTURE SEARCH, DIRECT RECORDING INTO AVAILABLEDISC SPACE AND RECORDING AND PLAYBACK ON VARIOUSDVD FORMATS.AUDIO USES DOLBY DIGITALFORMAT.CONVENIENCE OUTLET REQUIRED AT POINT OFUSE.MODEL5 1 AND 2 ALSO RECORD AND PLAY BLU-RAY DISCS.VENDOR 1:CAGE 1: 50356MODEL 1: BD-R2000VENDOR 2: JVC AMERICACAGE 2: 3NPW4MODEL 2: SR-HD1500VENDOR 3: PANASONIC USA, GOVERNMENT MARKETINGCAGE 3: 5Y727MODEL 3: LQ-MD800	19	15	5	12	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 2AMPS 2: 0WATTS 1: 240WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17119N	M0506.2	TELEVISION, FLAT PANEL LCD	1	VV	Color television with HD.Unit consists of a color receiver/monitor, with remote control, full off-air channel capability, automatic light sensor, broadcast stereo				0	WALL				YES	NO
17119N	R4650.1	ICE MAKER	1	VV	12# STORAGE;453#/24HRS;COUNTERTOP;TOUCH FREEW/FILTER;FLAKED;AIR COOLED;S.S.;BIN STORAGEICE PRODUCTION;WATER/ICE SELECTION SWITCHTO INCLUDE: 1EA SSM1-P SINGLE SYSTEM WATER FILTER 1EA SC10-A 10" COARSE PREFILTER	14	24	35	155	COUNTER	115V;60HZ;14.9A;1PH;1/2HP;LINE CORD NOT PROVIDED	FLARE INLET: 1/4";I.D. TUBE DRAIN: 5/8";I.D. SINK DRAIN: 7/8"SSM1-P. INLET - 3/8"FNPT;OUTLET - 3/8" FNPTSC10A: INLET - 3/4" FNPT;OUTLET - 3/4" FNPT20-60 PSIG WATER PRESSURE;WATER TEMP 40-100 F	6 IN. (15.24 CM) CLEARANCE REQUIRED ON BOTH SIDESFOR VENTILATION	YES	NO
17119N	R7250.2	REFRIGERATOR/FREEZER, UPRIGHT	1	VV	20.5CU FT;WHITE;FROST FREE;REVERSIBLE DOOR;LOCKING2 ADJ. GLASS SHELVES;2 CRISPERS;GALLON STORAGE IN DOOR;REMOVABLE FREEZER SHELF;CONTOURED DOOR	30	32	69	250	FLOOR	115V;60HZ			YES	NO
NURSE STATION															
17120N	A1011.1	TELEPHONE	4	VV	WALL OR DESK TBD				0	WALL				NO	NO
17120N	F2026.1	CONTAINER, RECYCLING, LARGE	1	VV	RECEPTACLE FOR PAPER PRODUCTS SLATED FOR COLLECTION, SHREDDING, RECYCLING DIMENSIONS: 19W X 19D X 37H	21	21	34	25	FLOOR	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
17120N	M1801	COMPUTER, MICROPROCESSING, W/FLAT PANEL MONITOR	4	VV	DESK TOP MICROPROCESSING COMPUTER.THE UNIT SHALLCONSIST OF A CENTRAL PROCESSING MINI TOWER, FLATPANEL MONITOR, KEYBOARD, MOUSE ANDSPEAKERS.THESYSTEM SHALL HAVE THE FOLLOWINGMINIMUMCHARACTERISTICS: A 2.8 GHZ PENTIUMPROCESSOR; 512MB MEMORY; 80GB HARD DRIVE; 32/48X CD-ROMVDCCOMBO; 1.44MBNETWORK INTERFACE CARD; VIDEO 32MBNVIDIA: A 18 INCH FLAT PANEL MONITOR.THE COMPUTERIS USED THROUGHOUT THE FACILITY TO INPUT,MANIPULATE AND RETRIEVE INFORMATION.PRICES WILL VARY BASED ON OPTIONS SELECTED AT TIMEOF PURCHASE.VENDOR 1: DELL COMPUTER CORP.CAGE 1: 04RE5MODEL 1: OPTIPLEX 780 W/19" MONITORVENDOR 2: GATEWAY 2000CAGE 2: 03ZB4MODEL 2: ZX6810-01VENDOR 3:CAGE 3: 1U0E6MODEL 3: THINKCENTRE M58 #7373	24	19	23	45	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 0HERTZ SWIT: 1PHASE 1: 1PHASE 2: 0AMPS 1: 5AMPS 2: 0WATTS 1: 600WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17120N	M1825	PRINTER, COMPUTER	1	VV	HIGH RESOLUTION COMPUTER PRINTER WITH A VARIETY OFTYPE STYLES AND SHEET/ENVELOPE FEEDERTRAYS.DATABASE INFORMATION REFLECTS NETWORK READY,MEDIUM DUTY OFFICE STYLE LASER PRINTERS.OTHERTYPES OF PRINTERS (BUBBLE JET, DOT MATRIX, LINE ORPLOTTER) AS WELL AS LIGHT OR HEAVY USECAPABILITIES ARE AVAILABLE.VENDOR 1:LEXMARK INTERNATIONAL, INCORPORATEDCAGE 1: OSD37MODEL 1: C736NVENDOR 2: HEWLETT PACKARD COMPANYCAGE 2: 28480MODEL 2: CP3524DN (CC470A)VENDOR 3: BROTHER INDUSTRIES, LTD.CAGE 3: 65533MODEL 3: HL-4570CDWT	21	20	18	73	COUNTER	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 6AMPS 2: 0WATTS 1: 720WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17120N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
MEDS															
17122N	A1011.1	TELEPHONE	1	VV	WALL OR DESK TBD				0	WALL				NO	NO
17122N	A4015.1	CLOCK, WALL MOUNT	1	VV	CLOCK, DIGITAL, BATTERY OPERATED; WALL MOUNT; APPROXIMATE DIMENSIONS: 17W X 2D X 11H				0	WALL				YES	NO
17122N	A5065.1	BOX, RETURN, DRUGS	1	VV	Security Drop box. Contents are deposited through the 12" front opening. A mailbox type baffie deters the removal of contents. Made of 18 gauge steel. Two cam locks keyed differently. DIMENSIONS: 14.5W X 17D X 20H				0	WALL				YES	YES
17122N	A5075	DISPENSER, SOAP, DISPOSABLE	1	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOU515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17122N	A5077	DISPENSER, HAND SANITIZER, HANDS-FREE	1	VV	A TOUCH FREE WALL-MOUNTED HAND SANITIZERDISPENSER.FOR USE THROUGHOUT A HEALTHCAREFACILITY.UNIT DOES NOT INCLUDE THE SANITIZINGLIQUID.UNITS ARE BATTERY OPERATED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: TFX TOUCH FREE DISPENSERVENDOR 2: W. W. GRAINGER, INC.CAGE 2: 25795MODEL 2: 1PKP9VENDOR 3:CAGE 3: AM490MODEL 3: SJS-1450	7	6	15	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO

ACTIVATION EQUIPMENT LIST VA NY HARBOR HEALTHCARE SYSTEM MANHATTAN VAMC BUILDING 1 PKG 1 FLOOR RENOVATIONS

ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17122N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	1	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL,SINGLE-FOLD, PAPER TOWEL DISPENSER.DISPENSERFEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM;AND REFILL INDICATOR SLOT.MINIMUM CAPACITY 400SINGLE-FOLD PAPER TOWELS.FOR GENERAL PURPOSE USETHROUGHOUT THE FACILITY.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-263VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 251-15VENDOR 3: AMERICAN SPECIALTIES, INC.CAGE 3: 0WPA40MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17122N	A5107	DISPENSER, GLOVE,SURGICAL/EXAMINATION, WALL MNTD	1	VV	EXAMINATION GLOVE DISPENSER BOX FOR WALL MOUNTING.FABRICATED OF EITHER COLD ROLLED STEEL WITH AWHITE BAKED ENAMEL FINISH, PLASTIC ORACRYLIC.PROVIDED WITH WALL BRACKET TO FACILITATEMOUNTING AND DEMOUNTING.MODEL 2 IS NORMALLY SOLD BY A CASE OF 10.VENDOR 1: LAB SAFETY SUPPLY, INCORPORATEDCAGE 1: 6M844MODEL 1: 83818VENDOR 2: S-CURVE TECHNOLOGIESCAGE 2: AM044MODEL 2: EBD-1000VENDOR 3: SURGICAL SUPPLY SERVICECAGE 3: AM045MODEL 3: 900990	11	4	6	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: 0 HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
17122N	A5108.1	DISPOSAL CONTAINER, SHARPS W/BRACKET	1	VV	Waste Disposal Unit, Sharps 16W X 8D X 17H				0	WALL				YES	NO
17122N	E0957.1	CART, MEDICATION	1	VV	DOCUMENTATION CART W/BIN STORAGE;HEIGHT ADJ.KEYBOARD TRAY;COMMERCIAL GRADE POWER CORD;CORD HANGER;MOTORIZED LIFT MECHANISM;MOUSE SURFACEWORK SURFACE;NOTEBOOK/PC COMPARTMENT	18	20	46	0	FLOOR	120V;7.8 AMPS			YES	YES
17122N	E0957.1	CART, MEDICATION	1	VV	DOCUMENTATION CART W/BIN STORAGE;HEIGHT ADJ.KEYBOARD TRAY;COMMERCIAL GRADE POWER CORD;CORD HANGER;MOTORIZED LIFT MECHANISM;MOUSE SURFACEWORK SURFACE;NOTEBOOK/PC COMPARTMENT	18	20	46	0	FLOOR	120V;7.8 AMPS			YES	NO
17122N	F2017.2	WASTE RECEPTACLE, STEP-ON	1	VV	8 GALLON;PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO
17122N	M3150	DISTRIBUTION SYSTEM, MEDICATION,AUTOMATIC	1	VV	AN AUTOMATED DISPENSING SYSTEM THAT PROVIDESCONTROLLED DISPENSING, INVENTORY AND SECURITY.SIZEAND COST WILL VARY DEPENDENT ON NUMBER OF MODULESSELECTED.PRICES ARE ESTIMATES ONLY AND ARE FOR A ONESTATION CABINET.THE PRICE CAN VARY SIGNIFICANTLYDEPENDING ON THE SPECIFIC REQUIREMENTS OF EACHFACILITY.MODEL 1 IS A 6-DRAWER MAIN UNIT WITHTOUCH SCREEN MONITOR AND COMPUTER. MODEL 2 PRICES FOR THE LARGE MAIN CABINET.MODEL 3 PRICE IS FORTHE BASE 72 DRAWER DISPENSING UNIT AND A TOUCHSCREEN AND COMPUTER.IT CAN BE INTERFACED WITH ANYHL7-COMPLIANT INFORMATION SYSTEM.THERE ARE SEVERALCONFIGURATIONS WITH VARIOUS SIZED CABINETS WITHOPTIONS/ACCESSORIES AVAILABLE INCLUDING VARIOUSSOFTWARE PACKAGES, BARCODE READER AND FINGER PRINTSCANNER.VENDOR 1: PYXIS CORPORATION, A CARDINALHEALTH COMPANYCAGE 1: SR350MODEL 1: MEDSTATION 4000 6-DRAWERVENDOR 2:CAGE 2: 1F1E9MODEL 2: ACUDOSE-RXVENDOR 3: INTERMETRO INDUSTRIES CORPORATIONCAGE 3: 40038MODEL 3: MEDDISPENSE BASE 72 (MDB-72)	33	29	57	659	FLOOR	VOLTS 1: 120VOLTS 2: 0HERTZ 1: 60 HERTZ DEP: 1HERTZ SWIT: 0PHASE 1: 1PHASE 2: 0AMPS 1: 10AMPS 2: 0WATTS 1: 1200WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (B) COLD WATER AND DRAINAUTILITY2: Electrical (A) 120 VOLT, CONVENTIONAL OUTLETUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	YES	NO
17122N	M3150M	DISPENSER, MEDICATION, AUTOMATED	1	VV	AUXILIARY;7 DRAWER	23	28	47	133	FLOOR	PLUGS INTO MAIN AND PYXIS CONTROL CABLE120V;100W;0.9A;EMERGENCY POWER		368 BTU/H	YES	NO
17122N	R6200.1	REFRIGERATOR, UNDERCOUNTER	1	VV	S.S.;MEDICAL GRADE;4 CU FT;MAGNETIC DOOR GASKETEXT. TEMP LOCKING DOOR;FIELD REVERSIBLE DOOR;2 WIRE SHELVES	24	28	32	186	FLOOR	115V;60HZ;1PH;8A MAX.;FUSE 15 AMPS;NEMA 5-15 PLUG(1.8M)HOSPITAL-GRADE CORD		717 BTU/HR	YES	NO
STAFF TLT (QTY 2)															
17127N	A1066	MIRROR, FLOAT GLASS, WITH SS FRAME	2	CC	A HIGH QUALITY 1/4" POLISHED FLOAT GLASS MIRROR36X18, FRAMED IN A ONE-PIECE, BRIGHT POLISHED,STAINLESS STEEL CHANNEL FRAME WITH 90° MITEREDCORNERS.ALL EDGES OF THE MIRROR ARE PROTECTED BYABSORBING FILLER STRIPS.MIRROR HAS A GALVANIZEDSTEEL BACK WITH INTEGRAL HORIZONTAL HANGINGBRACKETS AND WALL HANGER FOR CONCEALEDMOUNTING. FOR MOUNTING ABOVE SINGLE WALL MOUNTEDLAVATORIES LOCATED IN TOILET AREAS. DOCTORSEXAMINATION OFFICES, ETC.MAY ALSO BE USED ABOVEDOUBLE LAVATORIES, EITHER WALL OR COUNTERTOPMOUNTED, FOUND IN RESTROOM AREAS.VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATEDCAGE 1: 89488MODEL 1: B-165 1836VENDOR 2: BRADLEY CORPORATIONCAGE 2: 01055MODEL 2: 781-1836VENDOR 3: MEEK MANUFACTURING COMPANYCAGE 3: SR301MODEL 3: 1500 (18X36)	18	1	36	22	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17127N	A5075	DISPENSER, SOAP, DISPOSABLE	2	VV	DISPOSABLE SOAP DISPENSER.ONE-HANDED DISPENSINGOPERATION.DESIGNED TO ACCOMMODATE DISPOSABLE SOAPCARTRIDGE AND VALVE.BASED ON INFECTION CONTROL STANDARDS ONLY SOAPDISPENSERS REFILLED BY MEANS OF A CARTRIDGE ORREPLACEABLE BAG ARE TO BE USED.VENDOR 1: STAPLES, INC.CAGE 1: 0ZBU9MODEL 1: 603696VENDOR 2: JACOBS GARDNERCAGE 2: 7R579MODEL 2: GOJ515006VENDOR 3: W. W. GRAINGER, INC.CAGE 3: 25795MODEL 3: 5VN37	5	4	11	2	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO



ROOM #	JSN	NAME	QTY	ACQ/ INS	DESCRIPTION	W	D	H	LBS	LOCATION	ELECT	HVAC	PLUMB	CUTSHEET	RELOCATION
17127N	A5080	DISPENSER, PAPER TOWEL, SS, SURFACE MOUNTED	2	CC	A SURFACE MOUNTED, SATIN FINISH STAINLESS STEEL, SINGLE-FOLD, PAPER TOWEL DISPENSER. DISPENSER FEATURES: TUMBLER LOCK; FRONT HINGED AT BOTTOM; AND REFILL INDICATOR SLOT. MINIMUM CAPACITY 400. SINGLE-FOLD PAPER TOWELS. FOR GENERAL PURPOSE USE THROUGHOUT THE FACILITY. VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATED. CAGE 1: 89488. MODEL 1: B-263. VENDOR 2: BRADLEY CORPORATION. CAGE 2: 01055. MODEL 2: 251-15. VENDOR 3: AMERICAN SPECIALTIES, INC. CAGE 3: 0WP40. MODEL 3: 0245-SS	13	7	9	4	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17127N	A5109	GRAB BAR, 1-1/4" DIA., SS, 2 WALL, W/C ACCESSIBLE	2	CC	A 1-1/4" DIAMETER, SATIN FINISH STAINLESS STEEL, PEENED GRIPPING SURFACE, 2 WALL TOILET STALL/ROOM, GRAB BAR WITH CONCEALED MOUNTING FLANGES. SNAP OVER FLANGES ARE PROVIDED TO CONCEAL MOUNTING SCREWS. A SELECTION OF MOUNTING KITS AND CONCEALED ANCHOR DEVICES ARE AVAILABLE FROM THE MANUFACTURERS FOR DIFFERENT TYPES OF INSTALLATIONS. GRAB BAR SHALL COMPLY WITH BARRIER-FREE ACCESSIBILITY GUIDELINES FOR STRUCTURAL STRENGTH. FOR TYPICAL WATER CLOSET APPLICATIONS IN TOILET STALLS AND ROOMS WHERE ADA (AMERICAN'S WITH DISABILITIES ACT) REQUIREMENTS MUST BE MET. THESE MANUFACTURERS PROVIDE NUMEROUS CONFIGURATIONS OF THESE GRAB BARS. VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATED. CAGE 1: 89488. MODEL 1: B-5837.99. VENDOR 2: BRADLEY CORPORATION. CAGE 2: 01055. MODEL 2: 8322-059. VENDOR 3: AMERICAN SPECIALTIES, INC. CAGE 3: 0WP40. MODEL 3: SERIES 3700 TYPE 56 (3756-P)	57	3	3	7	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17127N	A5145.2	HOOK, GARMENT	2	CC	Fabricate hook units either of chromium plated brass with a satin finish, or stainless steel, using 6 mm (1/4 inch) minimum thick stock, with edges and corners rounded smooth to the thickness of the metal, or 3 mm (1/8 inch) minimum radius. Fabricate each unit as a double hook on a single shaft, integral with or permanently fastened to the wall flange, provided with concealed fastenings.				0	WALL				NO	NO
17127N	A5200	DISPENSER, TOILET TISSUE, SS, 2-ROLL, SURFACE MNTD	2	CC	A CONCEALED SURFACE MOUNTED, DOUBLE ROLL, SATIN FINISH STAINLESS STEEL, TOILET TISSUE DISPENSER. UNIT ACCOMMODATES TWO STANDARD-CORE TOILET TISSUE ROLLS THROUGH 5" DIAMETER. SPINDLES ARE CHROME PLATED PLASTIC WITH A HEAVY-DUTY INTERNAL SPRING AND TURN FREELY FOR NON-CONTROLLED DELIVERY. FOR GENERAL PURPOSE USE IN RESTROOMS. VENDOR 1: BOBRICK WASHROOM EQUIPMENT, INCORPORATED. CAGE 1: 89488. MODEL 1: B-27460. VENDOR 2: BRADLEY CORPORATION. CAGE 2: 01055. MODEL 2: 5234. VENDOR 3: W. W. GRAINGER, INC. CAGE 3: 25795. MODEL 3: 1CF35	15	7	13	6	WALL	VOLTS 1: 0VOLTS 2: 0HERTZ 1: HERTZ DEP: 0HERTZ SWIT: 0PHASE 1: 0PHASE 2: 0AMPS 1: 0AMPS 2: 0WATTS 1: 0WATTS 2: 0	UTILITY1: Plumbing (Water and Drainage) (N) N/AUTILITY2: Electrical (N) N/AUTILITY3: Medical Gas (N) N/AUTILITY4: Miscellaneous Gas (N) N/AUTILITY5: Non-Medical Gas (N) N/AUTILITY6: Miscellaneous (N) N/A	BTU/HR: 0	NO	NO
17127N	F2017.2	WASTE RECEPTACLE, STEP-ON	2	VV	8 GALLON; PLASTIC; WITH 6243 RIGID LINER	16	17	18	11	FLOOR				YES	NO