

VA AMARILLO, TX
PO# 504-B47003

Line #	Description	Qty
1	CX50 2D xMATRIX	1

Interface:

- 15.0 inch high resolution display with wide viewing angle
- Quick Keys and Active Mode
- Laptop style Alphanumeric QWERTY keyboard
- 8 TGCs and 2 LGCs
- Ergonomic carrying handle
- Includes AC adapter , power cord and system battery pack
- 2 USB flash drives on system
- 80 GB hard drive
- Internal DVD RW drive

Architecture:

All-digital compact broadband beamformer, Microfine 2D focusing with Dynamic Focal Tuning that includes Advanced X-Res signal processing, 170 dB full time input dynamic range 18,432 digitally-processed channels, Continuously variable steering in 2D, color and Doppler modes 2D Opt signal processing with 4X multi-line parallel processing and frequency compounding.

Intelligent Controls:

The CX50 has been designed to make portable exams easy and efficient. With a single button, iSCAN technology automatically samples data for a new level of 2D and Doppler optimization iSCAN one-touch Intelligent Optimization, iSCAN one-touch Intelligent Color Optimization, iSCAN Doppler one-touch optimization.

Transducers:

Supports Compact family of transducers featuring PureWave imaging technology in the S5-1, CX7-2t, C5-1, D5CWC. Also supports the high resolution S12-4, S8-3, C8-5 and L12-3 transducers. All transducers provide breakthrough frequency bandwidths and array configurations. These transducers also have ergonomically designed lightweight flexible cables and compact connectors.

Modes:

- 2D
- M-mode
- Anatomical M-mode
- Color M-mode
- Pulsed Wave Doppler
- Color Power Angio (CPA)
- Continuous Wave Doppler
- Invert and Color Invert
- Color compare mode

- Dual mode
- Duplex for simultaneous 2D and Doppler
- 2D Optimization Signal Processing
- Live Compare
- Tissue Harmonic Imaging (THI)
- Reconstructed zoom with pan (read zoom)
- Write zoom
- Pulse Inversion Harmonic imaging
- Adaptive Doppler
- Adaptive Color Doppler
- Color Tissue Doppler imaging
- Pulsed Wave Tissue Doppler imaging
- Active Native Data - manipulation of image data
- DICOM Networking
 - Ethernet @100Mb/second Includes DICOM wired and wireless "G" and "N".
 - Provides DICOM 3.0 network print and store and storage commitment
 - Performed Procedure Step (pps)
 - Modality Worklist
- DICOM Structured Reporting

Cineloop review

- On-board workstation-class data management with thumbnail previews and storage of images, loops, and reports. Retrospective and prospective clip capture to internal drive or removable media
Integrated DVD/CD burning capability for storage of images or export in DICOM, JPEG and .avi for PC compatibility. Philips DICOM viewer option to imbed in media transfer for easy viewing of study on most PCs.
- Maintenance and Serviceability
- Remote Access for Expedient Clinical and Technical Support
- Flexible Service Agreements
- Clinical Application and Educational Support

Cinical Education

1 Day PAS Onsite - Ultrasound system or upgrade onsite training provided by a PAS (Product Applications Specialist) for specific system applications or upgrades; not per modality. *Education is provided Monday - Friday during normal business hours.* Note: Philips Healthcare personnel are not responsible for actual patient contact or operation of equipment during education sessions except to demonstrate proper equipment operation. The training sessions should be attended by the appropriate healthcare professional as identified by the department director. *Repeat training for staff non-attendance will not be accepted.* Site must be patient-ready to meet training expectations. All onsite training day expires within 90 days from system or upgrade installation date. Exceptions are for 3D Stress onsite training (which expires 9 months from system or upgrade installation date) and Fusion & Needle Navigation onsite training (which expires 180 days from system or upgrade installation date).

THE NUMBER OF ONSITE TUITIONS YOU RECIEVE MAY VARY BASED ON PURCHASED OPTIONS. PLEASE CONSULT YOUR SALES REPRESENTIVE FOR FURTHER DETAILS

All Tuitions must be registered prior to the expiration date. The course chosen must be taken within 90 days of expiration.

- | | | |
|----------|--|----------|
| 2 | DICOM Package | 1 |
| | Dicom
Provides DICOM 3.0 network Print and Store, Performed Procedure Step (PPS), and Modality Worklist functionality. Networking capability supportable in both wired and wireless environments | |
| | Dicom Structured Reporting
Provides adult echo and vascular DICOM Structured Reporting. (Requires appropriate adult echo or vascular clinical options. | |
| 3 | Vascular Clinical Option | 1 |
| | This clinical option includes Tissue Specific Imaging software and SonoCT for Cerebrovascular, Peripheral vascular, abdominal vascular and Transcrainial applications. This clinical option also includes in depth analysis and reporting packages for vascular applications. Freehand 3D is also provide within this clinical option. Allows operation for vascular applications of the C5-1, S5-1, L12-3, L15,7io, L12-5 50 and D5cwc transducers. | |
| 4 | Adult Echo Clinical Option | 1 |
| | Tissue Specific Imaging software for adult cardiac ultrasound applications. Display optimization software with Tissue Specific presets for adult cardiac imaging and Doppler applications. Analysis software package includes cardiac imaging protocol measurements and configurable reports and finding codes. Active native data for post-process optimization and advanced XRES adaptive image processing for improved tissue conspicuity. iSCAN intelligent one-button optimization for adaptive gain compensation in 2D, Doppler, Tissue Doppler Imaging and LVO contrast functions. Includes Live compare mode, cardiac High-Q Automatic Doppler Analysis and respiration waveform from chest impedance. Allows operation of S5-1, CX X7-2t and D2cwc transducers. | |
| 5 | Cart with Multi-port Adapter | 1 |
| | Highly mobile cart that features hardware module to support transducer switching among up to three imaging transducers at the touch of a button. Includes: 4 swivel wheels with 2 locking casters, rear handle, micro-positioning grips, quick-connect tray, storage shelf, footrest, internal isolation transformer, B&W printer brackets, integrated transducer connector holder, gel holders and cable management. Includes USB hub for additional connectivity. | |
| 6 | USA Power Cord | 1 |
| 7 | D2cwc Static Transducer | 1 |
| | Non-imaging 2 MHz PW/CW Doppler transducer for cardiac applications | |
| 8 | S5-1 Transducer | 1 |
| | Features compact connector designed for reliability and improved ergonomics. Compatible with both EPIQ and CX50 systems. Manufactured in accordance with the European Union's Restriction of Hazardous Substances (RoHS) directive. | |

PureWave crystal Sector array transducer with 5 to 1 MHz extended operating frequency range for adult cardiology, abdominal, vascular, TCD and Acute Care.

9

X7-2t Transducer

1

Features compact connector designed for reliability and improved ergonomics. Compatible with both EPIQ and CX50 systems. Manufactured in accordance with the European Union's Restriction of Hazardous Substances (RoHS) directive.

X7-2t PureWave matrix array TEE Transducer for superior 2D quality. 7 to 2 MHz extended operating frequency range. Includes M-Mode, PW doppler, CW doppler, harmonics, true electrocautery suppression, and adaptive autocool.

Clinical Education

X7-2T Clinical Education; ***1 day of Implementation Onsite Training (expires 90 days after install, provided Mon-Fri during normal business hours), a 2 Day offsite TEE University (expires 365 days after install) and one subscription to E-Echocardiography.com (must be activated within 90 days of code notification). All offsite training includes travel, see travel disclaimer**

**TRAVEL Disclaimer: Travel & Accommodations for registered attendees. Each tuition includes one (1) participant's airfare from a North American customer location to a Philips North America Ultrasound Clinical Education training location with modest lodging, ground transportation and meal expenses for the course duration. Breakfast/dinner are provided by the hotel and lunch/breaks are catered by Philips Healthcare. All other expenses will be the responsibility of the attendee (ie. Baggage fees, meals while traveling, transportation to and from customer's home airport). Details are provided during the scheduling process. Note: 21 day Cancellation/Rescheduling policy is strictly enforced.

***Note: Philips Healthcare personnel are not responsible for actual patient contact or operation of equipment during education sessions except to demonstrate proper equipment operation. The training sessions should be attended by the appropriate healthcare professional as identified by the department director. Repeat training for staff non-attendance will not be accepted. Site must be patient-ready to meet training expectations.

All Tuitions must be registered prior to the expiration date. The course chosen must be taken within 90 days of expiration.

10

L12-3 Transducer

1

Features compact connector designed for reliability and improved ergonomics. Compatible with both EPIQ and CX50 systems. Manufactured in accordance with the European Union's Restriction of Hazardous Substances (RoHS) directive.

L12-3 fine pitched, high resolution linear array with 12 to 3 Mhz extended operating frequency range for vascular, small parts, breast, musculoskeletal, contrast regional anesthesia and acute care applications

11

English Manual

1

Operation Manual

12

US2791 Bio CX50 CTC3

1

This course provides technical instruction for biomedical engineers (hospital engineers) on the CX50 Ultrasound System. Students receive instruction on system components, theory of operation, disassembly, reassembly, preventative maintenance and safety checks. Hands-on labs train the students to verify proper equipment operation and diagnostic troubleshooting techniques. Philips support philosophy is explained to facilitate working successfully with our support professionals.

Upon completion of this course, the student will be able to:

- Operate the system to determine basic functionality
- Collect and deliver system status and failure information (logs and settings) to Philips support organization
- Disassemble and reassemble the major parts of the system
- Perform Preventative Maintenance and Mechanical adjustments
- Back-up and restore user presets
- Install and upgrade system software
- Set-up the system for DICOM communication
- Configure/enable remote connections

Key topics:

- Features
- Theory of Operation
- Disassembly and Reassembly
- System Administration

Prerequisites: Basic computer knowledge, Basic Ultrasound and Transducer knowledge, Networking and DICOM familiarity, ESD training, Electrical Safety Testing.

Accreditation: None.

Location: CTC; Cleveland, OH, USA.

Class Length: 3 days (excludes Saturdays, Sundays, and Philips holidays)

Materials:

- Instructor-presented power point presentations and demonstration videos
- Field Service Manual
- Test is on Philips Learning Center

*** PHILIPS PROPRIETARY MATERIALS SUCH AS DIAGNOSTIC SOFTWARE AND SERVICE DOCUMENTATION ARE NOT INCLUDED IN THE TRAINING AND WILL NOT BE AVAILABLE FOR USE OUTSIDE OF THE TRAINING ENVIRONMENT. THE TRAINEE MUST RETURN ALL PROPRIETARY MATERIALS RECEIVED DURING THE TRAINING AT THE END OF THE TRAINING. CUSTOMER ACKNOWLEDGES AND AGREES THAT NEITHER CUSTOMER NOR TRAINEE WILL RECEIVE A LICENSE TO SUCH PROPRIETARY MATERIALS AND THAT THE TRAINEE MAY NOT BE ABLE TO FULLY UTILIZE THE TRAINING WITHOUT THE USE OF SUCH PROPRIETARY MATERIALS. (CERTAIN LICENSES MAY BE OBTAINED THROUGH PURCHASE OF A PHILIPS RIGHTFIT SERVICE AGREEMENT.)** Course dates and location to be finalized by Philips. Philips shall attempt to accommodate Customer requested dates and training

IMPORTANT Notes Regarding Admission to Philips Customer Engineer Training Courses:

- | | | |
|----|--|---|
| 13 | Airfare to Cleveland for Biomed Training | 1 |
|----|--|---|

14	Food Transpt Lodging for Cleveland Biomed Training	3
-----------	---	----------

15	1st SVC Manual for Gov	1
----	------------------------	---

Product: GE VIVID S6
Serial Number: VS60733
Manufacturer: GE MEDICAL SYSTEMS