

## **SALIENT CHARACTERISTICS**

Offerors will be evaluated to the extent they can provide the items (or equal). The Government will award to the lowest priced, responsible, technically acceptable offeror. For those quoting equal items, please provide specifications sheet and other applicable documents that demonstrate *equality* and *satisfaction* for each of the salient characteristics listed.

Items must meet the following salient characteristics:

1. Must be compatible with Stryker endoscopic surgical camera systems
2. Must include all adapters and/or accessories for the proper function of all equipment.
3. High Definition Camera Control Unit
4. High Definition integrated camera heads (one piece), standard eye cup, programmable
5. LED light source intensity can be controlled from buttons on camera head
6. Information management system with ability to transmit images to electronic health records
7. High Definition medical grade color printer
8. 26 inch High Definition medical grade video monitors