

Functional Requirements for **Digital Radiography System**
For **St. Louis VA Healthcare System**
PO#: 657-B50010

This document highlights the technical specifications and services being requested by this VISN for consideration toward purchase of Digital Radiography Systems. Offerors under this proposal shall provide all equipment and accessories, installation services, training, and project management support.

Objective:

It is the intent of the St. Louis VA to conduct a purchase of Digital Radiography Systems. It is anticipated that these units will be installed at the Jefferson Barracks division of the St. Louis VA. However, the St. Louis VA reserves the right to alter quantities at its discretion.

Technical Requirements:

Generator Requirements:

1. 80 kW, 3 phase generator
2. Radiographic kVp range is 40-150
3. Radiographic mA range 800 ma@ 100KVp

Radiographic/Overhead tube Crane System

1. Patient alignment system (laser alignment/positioning lights)
2. Pre-programmed exposure settings located on tube or in control room
3. Ability to change between table top, upright bucky and table bucky from tube head
4. Provide small and large focal spot information
5. Control handle
6. System motorization
7. Automatic tube positioning
8. Automatic collimation
9. Table and upright auto tracking package

Control room Radiographic control panel

1. 19" control room LCD monitor (minimum)
2. Auto HIS/RIS refresh package
3. Computer hard drive 100 GB or greater
4. Quality control tracking package
5. Operating system is Windows 7 or newer or other (Linux, Unix, etc...)
6. UPS for x-ray control/image memory
7. Automatic parameter selection
8. Ability to send images directly from the control panel. Not requiring a separate workstation.
9. Integrated DICOM structured dose reporting(if available/describe)
10. Optional dose tracking package
11. Optional bar code reader
12. Dose reduction features on system

Fixed Wall Digital Detector Requirements:

1. Detector size up to 17"x17"
2. Resolution – minimum 2.0 lp/mm
3. Tilting Bucky

4. Height minimum – low enough to complete standing knee exams while patient is standing on the floor
5. Height maximum – high enough to complete AP C-spine exams while patient is standing on the floor
6. Ability to position detector in portrait or landscape mode.
7. Removable/adjustable patient handgrips.
8. Optional wall stand auto stitching package.
9. Optional patient stand
10. Wall Bucky adjustable height and moveable

Wireless Digital Detector Requirements:

1. Detector size – minimum 14"x17"
2. Optional small (~8"x10") plate
3. Resolution – minimum 2.0 lp/mm
4. Detector will charge in Bucky/holder
5. Weight Limit (w/battery installed) – maximum 11 lbs.
6. Wireless (not tethered to the system/table)
7. Wireless – compatible with 802.11b/g/n
8. FIPS 140-2 compliant
9. Describe maximum patient weight for weight bearing examinations.
10. Provide full coverage warranty for a minimum of one year
11. Charging station (*if required*)
12. Portrait and landscape orientation capable
13. Grid attachments
14. Table holder
15. Mobile holder

Fixed Table Digital Detector Requirements:

1. Detector size up to 17"x17"
2. Resolution – minimum 2.0 lp/mm
3. Optional table auto stitching package.

Table Requirements:

1. Static – 500 lbs patient weight limit (minimum)
2. Dynamic – 300 lbs patient weight limit (minimum)
3. Width of table – 80 cm (minimum)
4. Floating tabletop
5. Adjustable height
6. Auto centering option to exact middle while the table is vertical or horizontal positions
7. Table movement controls located table side.
8. Removable/adjustable patient handgrips.
9. Hands-free operation
10. Electromagnetic locking system

Warranty and Service:

1. VPN/Remote Access – The vendor shall provide, at no additional charge, any and all equipment service programs, such as remote diagnostics, during the warranty period. The vendor shall provide post-warranty remote diagnostic service program as an "Add Option" with the offer. The system shall provide Vendor Remote Diagnostics via VPN. Vendor shall utilize the VA national Site-to-Site VPN, or the vendor shall work with the Office of Cyber and Information Security and the VAMC Information Security Officer to establish a Client-Based VPN.

2. Service and Operator Manuals – The vendor shall provide the following documentation for the proposed system:
 - Two (2) copies of operator's instruction manuals per unit purchased
 - Two (2) copies of complete technical service manuals including detailed troubleshooting guides, necessary diagnostic software, service keys, schematic diagrams, and parts lists
 - Two (2) copies of a system manager's manual outlining back-up procedures, managing privilege group limits, routine tasks, etc.
3. Warranty – The system and accessories shall be covered under the manufacturer's warranty, and shall include all parts and labor for one year following acceptance by the VAMC. The manufacturer's factory-trained field service personnel shall perform installation and maintenance during the warranty period. The warranty period shall not begin until after VAMC authorized personnel have accepted the products delivery, installation, and functionality.

Training

1. On-site – Clinical applications during GO LIVE - **32 hours at a minimum**
2. Follow up applications training to be provided after technologists have hands-on experience with the system - within one year of installation to be determined by the station - **32 hours at a minimum**
3. Biomedical training package - Include all prerequisite classes, equivalent to what your OEM field service representatives receive. This training is to be provided within 18 months of installation. All service manuals, schematics, diagrams, diagnostic software, other special tools and hardware keys equivalent to what their OEM field service reps have available to diagnose, troubleshoot, repair and maintain the equipment. Travel including airfare, hotel and government rate per diems shall be included.
4. Quick notes – or reference guides to be provided during on-site training
5. Describe on-line or telephonic applications support and availability
6. Describe post-warranty technologist and physician training packages that are offered.

Information to be provided by each vendor:

1. Provide detail information about the curriculum and length of the Biomedical Technical Training.
2. Provide brochures and spec sheets
3. Provide the following information about the wireless detector:
 - a. maintenance requirements
 - b. battery life/charging requirements/run time
 - c. drop policy (during warranty and post warranty)
 - d. cleaning instructions for the wireless detector
4. Typical drawings (to include minimum room size requirements)
5. DICOM conformance statement
6. Completed pre-procurement assessment form (6550) and MDS² document
7. Information about your companies support structure during the warranty period (i.e. a listing of Field Service Engineer locations and availability, support 800 phone number(s), remote support, etc.)

Other Features:

Other value added features included in the vendor's proposal will be considered but not limited to:

1. Extended warranty.
2. hardware/upgrades,
3. additional training,
4. performance improvement and
5. process optimization consulting,

6. Price improvements based on volume purchase will be considered.
7. Describe any other value added items offered

Trade-in – Please provide trade in and de-installation information

1. Option 1 – VA will retain any hard drives containing electronic personal health information (ePHI).
Manufacturer: Picker International
Model: RadView
S/N: 01172S07
Acq. Year: 2000
EE# 45460
Station: St. Louis VAMC