

AMENDMENT OF SOLICITATION/MODIFICATION OF CONTRACT			BPA NO.	1. CONTRACT ID CODE 00244	PAGE 1	OF PAGES
2. AMENDMENT/MODIFICATION NO. A00001		3. EFFECTIVE DATE 06/18/2015	4. REQUISITION/PURCHASE REQ. NO. 562-15-2-711-0194		5. PROJECT NO.(If applicable) none	
6. ISSUED BY Department of Veterans Affairs Network Contracting Office 4 1010 Delafield Road Pittsburgh PA 15215		CODE 00244	7. ADMINISTERED BY (If other than Item 6) Department of Veterans Affairs Network Contracting Office 4 1010 Delafield Road Pittsburgh PA 15215		CODE 00244	
8. NAME AND ADDRESS OF CONTRACTOR (No., street, county, State and ZIP Code) To all Offerors/Bidders			(X)	9A. AMENDMENT OF SOLICITATION NO. VA244-15-Q-0918		
				9B. DATED (SEE ITEM 11) 06-18-2015		
			X	10A. MODIFICATION OF CONTRACT/ORDER NO.		
				10B. DATED (SEE ITEM 13)		
CODE		FACILITY CODE				
11. THIS ITEM ONLY APPLIES TO AMENDMENTS OF SOLICITATIONS						
<input checked="" type="checkbox"/> The above numbered solicitation is amended as set forth in Item 14. The hour and date specified for receipt of Offers <input checked="" type="checkbox"/> is extended, <input type="checkbox"/> is not extended. Offers must acknowledge receipt of this amendment prior to the hour and date specified in the solicitation or as amended, by one of the following methods: (a) By completing Items 8 and 15, and returning <u>1</u> copies of the amendment; (b) By acknowledging receipt of this amendment on each copy of the offer submitted; or (c) By separate letter or telegram which includes a reference to the solicitation and amendment numbers. FAILURE OF YOUR ACKNOWLEDGMENT TO BE RECEIVED AT THE PLACE DESIGNATED FOR THE RECEIPT OF OFFERS PRIOR TO THE HOUR AND DATE SPECIFIED MAY RESULT IN REJECTION OF YOUR OFFER. If by virtue of this amendment you desire to change an offer already submitted, such change may be made by telegram or letter, provided each telegram or letter makes reference to the solicitation and this amendment, and is received prior to the opening hour and date specified. 06/25/2015 10:00a EST						
12. ACCOUNTING AND APPROPRIATION DATA (If required) See CONTINUATION Page						
13. THIS ITEM APPLIES ONLY TO MODIFICATIONS OF CONTRACTS/ORDERS, IT MODIFIES THE CONTRACT/ORDER NO. AS DESCRIBED IN ITEM 14.						
(X)	A. THIS CHANGE ORDER IS ISSUED PURSUANT TO: (Specify authority) THE CHANGES SET FORTH IN ITEM 14 ARE MADE IN THE CONTRACT ORDER NO. IN ITEM 10A.					
	B. THE ABOVE NUMBERED CONTRACT/ORDER IS MODIFIED TO REFLECT THE ADMINISTRATIVE CHANGES (such as changes in paying office, appropriation date, etc.) SET FORTH IN ITEM 14, PURSUANT TO THE AUTHORITY OF FAR 43.103(b).					
	C. THIS SUPPLEMENTAL AGREEMENT IS ENTERED INTO PURSUANT TO AUTHORITY OF:					
	D. OTHER (Specify type of modification and authority)					
E. IMPORTANT: Contractor <input type="checkbox"/> is not, <input type="checkbox"/> is required to sign this document and return _____ copies to the issuing office.						
14. DESCRIPTION OF AMENDMENT/MODIFICATION (Organized by UCF section headings, including solicitation/contract subject matter where feasible.) The purpose of this amendment is to modify solicitation VA244-15-Q-0918 by the following. This solicitation is now set aside 100% for small business. Please see the attached revised Section B.4 Price/Cost Schedule.						
Except as provided herein, all terms and conditions of the document referenced in Item 9A or 10A, as heretofore changed, remains unchanged and in full force and effect.						
15A. NAME AND TITLE OF SIGNER (Type or print)			16A. NAME AND TITLE OF CONTRACTING OFFICER (Type or print)			
15B. CONTRACTOR/OFFEROR (Signature of person authorized to sign)		15C. DATE SIGNED	16B. UNITED STATES OF AMERICA BY (Signature of Contracting Officer)		16C. DATE SIGNED	

CONTINUATION PAGE

A.1 Price/Cost Schedule

Item Information

ITEM NUMBER	DESCRIPTION OF SUPPLIES/SERVIC ES	QUANTIT Y	UNI T	UNIT PRICE	AMOUNT
0001	CSN: 6650-433129 EIL: 430 ZEISS OPMI LUMERA I SURGICAL MICROSCOPE Surgical microscope: Apochromatic Optics; Motorized zoom system: zoom ratio 1 : 6; Magnification factors, Y = 0.4- 2.4; Focusing Range: motorized, 50 mm; Objective lens: f = 200 mm (optionally also f = 175 mm); Binocular tube: 0-180* invertertube; Eyepieces: 10 x; DeepView: depth of field management system Illumination: Stereo Coaxil Illumination(SCI): adjustable red reflex illumination and full field illumination, both are patented Fiber- optic illumination; Blue blocking filter; Fluorescence filter; Integrated 408 NM UV Barrier filter	1.00	EA		

Light Source
12 V, 100 W
Halogen light source
with fully automatic
lamp change in case of
failure;

XY coupling
61 MM x 61
MM adjustment range;
Programmable
button for starting
positions of X-Y
coupling, focus and
Zoom

Weight:
8.5 kg (with
180* inclinable
binocular tube,
objective lens and
eyepieces)

Suspension System;
Maximum Load
Capacity 14Kg
(complete microscope
equipment and all
accessories)

Foot Switch: wireless
foot control panel
(FCP WL) and
automatic inversion

HD Video Camera
with microscope
imaging and operation
system; prime function
is recording of
operation scene and
capturing snapshot.

Floor Stand w
electromagnetic
braking system for
safety

Column tray

GRAND TOTAL _____

A.2 Delivery Schedule

ITEM NUMBER		QUANTITY	DELIVERY DATE
0001	SHIP TO:	DEPARTMENT OF VETERANS AFFAIRS VA MEDICAL CENTER ERIE (562) 135 EAST 38TH STREET ERIE, PA 16504	1.00
	MARK FOR:	ATTN: WAREHOUSE	

End of Document