

SCOPE OF WORK FUME HOOD PREVENTATIVE MAINTENANCE

SCOPE

Contractor shall calibrate and provide certification for all fume hoods at the Central Texas Veteran Health Care System. Contractor shall furnish all personnel, transportation, equipment, parts and supplies, facilities, supervision, and other items and non-personal services necessary to perform fume hood preventative maintenance services.

LOCATION

Central Texas Veterans Health Care System
Olin E. Teague Veterans Medical Center
1901 Veterans Memorial Drive
Temple, TX 76504-7451

SERVICES

1. Certification – Contractor shall verify all hoods tested meet CAP standards and will also provide a bacterial sample and particle counting of the pharmacy cleanroom.
2. Schedule - Units are to be checked either annually (1) or semi-annually (2).

Manufacturer	Model Number	Serial Number	EE Number	Location	Quantity
AIRCLEAN Systems	AC710C	AC710C890	86311	1D105-163-T	1
Hamilton Ind.	54L179	19903	41342	1D104A-163	1
Hamilton Ind.	54L179	990201	41400	BA121H-163-T	1
Hamilton Ind.	54L179	990202	41401	BA119-163-T	1
Hamilton Ind.	54L179	990526	41722	1D102-163-T	1
Hamilton Ind.	54L180	29905	41359	1D108A-163	1
Hamilton Ind.	54L181	000603	47301	1A158-163-T	1
Olympic Medical	44	5397	60290	BD108C-163-T	1
Olympic Medical	44	5424	60291	BD103D-163-T	1
Thermo Forma	1284	18565375	50764	3R27-205-T	1
LABCONCO	48822	020853454G	54376	3R32-205-T	1
LABCONCO	48822	020853568G	54377	3R32-205-T	1
LABCONCO	48822	040520483	60323	2R33-205-T	1
LABCONCO	72804	020953722G	54394	3R18-205-T	1
LABCONCO	72804	020853310G	54395	3R11-205-T	1
LABCONCO	72804	020953724G	54396	3R05-205-T	1
LABCONCO	72804	020953726G	54397	3R08-205-T	1
LABCONCO	72804	020953723G	54398	3R21-205-T	1
LABCONCO	72804	020953728G	54399	3R03-205-T	1
LABCONCO	72804	020953730G	54400	3R29-205-T	1
LABCONCO	72804	020953725G	54401	3R15-205-T	1
LABCONCO	72804	020953716G	54402	3R34-205-T	1

LABCONCO	72804	040416920	60319	2R15-205-T	1
LABCONCO	72804	040419624	60320	2R05-205-T	1
LABCON CO	72804	040520348	60321	2R21-205-T	1
LABCONCO	72804	040520480	60322	2R14-205-T	1
LABCONCO	7280400	060961239	65003	2R24-205-T	1
LABCONCO	7280400	060961248	65004	2R18-205-T	1
LABCONCO	7280400	060961259	65005	2R03-205-T	1
LABCONCO	7280400	060961261	65006	2R08-205-T	1
LABCONCO	7280400	100221195	87839	2R34-205-T	1
LABCONCO	7280400	100221196	87840	2R29-205-T	1
LABCONCO	3460001	100322539	88049	2R27-205-T	2
LABCONCO	3460001	100322544	88050	2R30-205-T	2
LABCONCO	3460001	100322541	88051	2R37-205-T	2
LABCONCO	3460001	100322213	88052	2R09-205-T	2
LABCONCO	3460001	100322538	88053	2R20-205-T	2
LABCONCO	3620804	020953715G	54403	3R25-205-T	2
LABCONCO	3620804	020823624A	54404	3R36-205-T	2
LABCONCO	3620804	020923815A	54405	3R09-205-T	2
LABCONCO	3620804	020823293A	54406	3R04-205-T	2
LABCONCO	3620804	020923814A	54407	3R20-205-T	2
LABCONCO	3620804	020823625A	54408	3R24-205-T	2
LABCONCO	3620804	020823629A	54409	3R17-205-T	2
LABCONCO	Purifier Logic T	140592824	113465	2R12-205-T	2
Baker Co.	315	BM25184V	4627	1A158-163-T	2
Baker Co.	SG403A-HE	108536	104737	1B206-7901-A	2
Containment Tech.	A1000B	1006MPD552	65235	3G15-204-T	2
Containment Tech.	A1000R	1006MPA550	65236	3G15-204-T	2
Containment Tech.	MICSINGLE	1008EMPD690	80311	BB115F-163-T	2
Containment Tech.	MICTPN	0708ETPN677	80312	BB115F-163-T	2
NUAIRE	NU-201-630	80198101002	54454	3R17-205-T	2
NUAIRE	NU-201-630	80214101002	54457	3R27-205-T	2
NUAIRE	NU-201-630	80205101002	54458	3R27-205-T	2
NUAIRE	NU-201-630	80215101002	54459	3R30-205-T	2
NUAIRE	NU-201-630	80206101002	54460	3R24-205-T	2
NUAIRE	NU-201-630	80223101102	54461	3R04-205-T	2
NUAIRE	NU-201-630	80208101002	54462	3R09-205-T	2
NUAIRE	NU-201-630	80209101002	54464	3R20-205-T	2
NUAIRE	NU-425	160404112513	111463	1D105-163-T	2
NUAIRE	NU-425	160444112613	111464	1D101A-163-T	2
NUAIRE	NU-425	160405112513	111465	1D101B-163-T	2
NUAIRE	NU-425-400	18262030201	50935	3R30-205-T	1
NUAIRE	NU-425-400	163886061814	115783	6K04-204-T	2

Summary

Fume Hoods, Total: 64

Fume Hoods, Annual: 33
Fume Hoods, Semi-annual: 31

Service Description	Unit	Quantity
Bacteria Sampling of Pharmacy Cleanroom	Month (MO)	12
Particle Counting of Pharmacy Cleanroom	JB	4