


RFI FORM**Contractor Name:** SMF Constructors, Inc.**Address:** 29970 Technology Dr. #212, Murrieta CA 92563**Phone/Fax:** (951) 225-1343**Send to:** cecil.nichols@va.gov

RFI (REQUEST FOR INFORMATION)			
PROJECT NO.:	600-CSI-102	RFI NO.:	9
PROJECT NAME:	UPGRADE Utility for Mobile MRI Unit	DATE REQUESTED:	06/08/2016
SOLICITATION NO.:	VA262-16-Q-0838	REFERENCE:	
DRAWING:		SPECIFICATION SECTION:	
DESCRIPTION OF PROBLEM OR INFORMATION BEING REQUESTED			
Please be specific as possible: <ol style="list-style-type: none">1. Confirm that the MRI trailer will be provided by the VA.2. Does the contractor move the MRI trailer into position.3. What trailer accessories are provided by the VA? Jack stands, etc.?4. Identify the POC for telephone and data service and water, if applicable.			
PROJECT MANAGER'S RESPONSE			
VA Response: <ol style="list-style-type: none">1. It is not in the SOW for the contractor to provide the MRI trailer. It will be provided by others.2. Contractor will not move the MRI trailer into position. It will be done by other.3. It is not in the contractor's SOW to provide any trailer accessories. It will be provided by others.4. Telephone, data service, and water, are not in the SOW for the contractor. The only utility the contractor has to provide is 480 V, 3 phase, 5 wire electrical power close to the trailer. Follow the MRI manufacturer's Installation Guide.			
TRACKING NO.:		AMENDMENT NO.:	
VA PROJECT ENGINEER/MANAGER: 		DATE: 6/10/2016	