

Solicitation VA246-16-Q-1102
Solicitation Attachment B – Salient Characteristics
Requirement Title: Cardiac Ablation Clamps
Requiring Facility: Richmond VA Medical Center

SALIENT CHARACTERISTICS
VA Solicitation Number VA246-16-Q-1102

Need – Cardiac Ablation Clamps

Market – Open Market

(1) Functions to be performed;

- Cardiac ablation clamps shall be intended for ablation of cardiac tissues during surgery.

(2) Performance required;

- The clamps shall be available with right and left curved jaws.
- The clamps shall utilize radiofrequency (RF) technology.
- The clamps shall be compatible with existing AtriCure generators in use at the facility.
- The clamps shall include a method for guiding placement during surgery.