

THIS REQUEST FOR INFORMATION (RFI) IS ISSUED SOLELY FOR INFORMATION AND PLANNING PURPOSES ONLY AND DOES NOT CONSTITUTE A SOLICITATION.

THE SUBMISSION OF PRICING, CAPABILITIES FOR PLANNING PURPOSES, AND OTHER MARKET INFORMATION IS HIGHLY ENCOURAGED AND ALLOWED UNDER THIS RFI IN ACCORDANCE WITH (IAW) FAR 15.201(e).

DISCLAIMER

This RFI is issued solely for information and planning purposes only and does not constitute a solicitation. All information received in response to this RFI that is marked as proprietary will be handled accordingly. IAW FAR 15.201(e), responses to this notice are not offers and cannot be accepted by the Government to form a binding contract. Responders are solely responsible for all expenses associated with responding to this RFI.

SOURCES SOUGHT DESCRIPTION

This is NOT a solicitation announcement. This is a sources sought/RFI only. The purpose of this sources sought/RFI is to gain knowledge of potential qualified sources and their size classification relative to NAICS 339112 (size standard of 1,000 Employees). Responses to this sources sought will be used by the Government to make appropriate acquisition decisions. After review of the responses to this sources sought, a solicitation announcement may be published. Responses to this sources sought synopsis are not considered adequate responses for a solicitation announcement.

The Department of Veterans Affairs (VA), VISN 22 Network Contracting Office, is seeking sources that can provide an *On-Board Imager Imaging System Software Upgrade* that at a minimum meets the following salient characteristics with a brand name or equal product IAW FAR 11.104 and VAAR 811.104-70 for the VA Long Beach Healthcare System.

Brand name information:

- 1. Varian WES7 Upg H291225**
 - 2. Varian WES7 Upg H294418**
-
- A. OBI Software Updates/Upgrades must be compatible with the Windows 7 OS.**
 - B. OBI and CBCT Computer Upgrades must deliver a Windows 7 compatible On-Board Imager kV Imaging system (OBI) workstation computer that also supports Cone Beam Computed Tomography (CBCT) functionality; must support Varian OBI and CBCT products.**
 - C. 4D Integrated Treatment Console (4DITC) software must be Windows 7 OS compatible; must support 4-slot accessories for complex treatment techniques; must provide image only sessions keeping Treatment sessions in sync with fractionation pattern; photos, activity and patient note must be displayed on Treatment queue.**

- D. Multileaf Collimator (MLC) Controller Low Profile Controller (LPC) Hardware Upgrade must provide enhanced performance; must be compatible with the latest MLC software version (v)8.1; and must reduce footprint design.**
- E. Millenium Multileaf Collimator (MLC) software upgrade must be to latest version and must be Windows 7 compatible.**
- F. Training must be included with the purchase of the OBI Upgrade. Training plan details must be provided by the training management team as part of the project implementation process. Offer must be valid for 18 months after installation of product.**
- G. Upgrade for Varian Verification System (VVS for 4DIITC) must provide the capability to verify patient and accessory identification and conical collimator verification through use of a barcode scanning system. Must include two (2) bar code scanners, one (1) barcode label printer and one (1) package of labels for the bar code printer; must include Patient and Accessory Verification license; and must include Conical Collimator Verification license.**
- H. 4D Integrated Treatment Console (4DITC) workstation refresh must provide the latest computer model with Integrated Linac Verification Interface (LVI). Must be Windows 7 OS compatible.**

The information identified above is intended to be descriptive, not restrictive and to indicate the quality of the supplies/services that will be satisfactory. It is the responsibility of the interested source to demonstrate to the government that the interested parties can provide the supplies/services that fulfill the required specifications.

If you are interested, and are capable of providing the sought out supplies/services, please provide the requested information as well as the information indicated below. Response to this notice should include company name, address, point of contact, size of business pursuant to the following questions:

- (1) Please indicate the size status and representations of your business, such as but not limited to: Service Disabled Veteran Owned Small Business (SDVOSB), Veteran Owned Small Business (VOSB), Hubzone, Woman Owned Small Business (WOSB), Large Business, etc.)?**
- (2) Is your company considered small under the NAICS code identified under this RFI?**
- (3) Are you the manufacturer or distributor of the items being referenced above (or equivalent product/solution)? What is the manufacturing country of origin of these items?**
- (4) If you're a small business and you are an authorized distributor/reseller for the items identified above (or equivalent product/solution), do you alter; assemble; modify; the items requested in any way? If you do, state how and what is altered; assembled; modified?**
- (5) Does your company have an FSS contract with GSA or the NAC or are you a contract holder with NASA SEWP or any other federal contract? If so, please provide the contract type and number.**

(6) If you are an FSS GSA/NAC or NASA SEWP contract holder or other federal contract holder, are the referenced items/solutions available on your schedule/contract?

(7) Please provide general pricing for your products/solutions for market research purposes.

(8) Please submit your capabilities in regards to the salient characteristics being provided and any information pertaining to “equal to items” to establish capabilities for planning purposes?

***** Submissions addressing Section (8) should show clear, compelling and convincing***
evidence that all “equal to items” meet all required salient characteristics.**

Responses to this notice shall be submitted via email to **raphael.fontela@va.gov**. Telephone responses shall not be accepted. Responses must be received no later than Friday, August 19, 2016 at 3:00 p.m. PST. If a solicitation is issued it shall be announced at a later date, and all interested parties must respond to that solicitation announcement separately from the responses to this request for information. Responses to this notice are not a request to be added to a prospective bidders list or to receive a copy of the solicitation.