

VENDOR Q & A Set #3
VISN 23 DR ROOM SPECIAL CONSOLIDATION
VA797R-16-Q-0018

Q32: VISN23 requested dose reporting software in the specifications, however there is no requirement for a DAP meter. Without the DAP Meter there is little value in the dose reporting software. Should the DAP Meter be included in the quoted price, or included as an option for VISN23 to include later?

A32: **VISN 23 does not have dose reporting software in the specifications. VISN 23 does have a requirement for “Integrated DICOM structured dose reporting”. The intent of this requirement is to ensure the DR equipment includes DICOM structured reporting of dose and if a DAP meter is required to meet this specification please include that in the quote price.**

Q33: In several instances the specifications call for an LLI patient stand that is commonly used with the auto-stitching software. The specifications either do not call for the auto-stitching software, or it was removed via Amendment 0001. Again, the stand without the auto-stitching software is of little use. Should the auto-stitching software be included/optioned? Should the LLI stand be removed in those instance where the auto-stitching software was removed?

A33: **The specifications for the following orders were revised and the requirement for of the “Mobile patient stand to be used with auto stitching feature” was removed because stitching is not to be included on these orders and therefore no mobile patient stand is required.**

| Facility Name | Requirement Number | Delivery Order Number | Stitching |
|--------------------------|--------------------|-----------------------|-----------|
| Sioux Falls SD (438) | 618-B69008 | VA797R-16-J-0373 | No |
| Fort Meade SD (568) | 618-B69009 | VA797R-16-J-0374 | No |
| Hot Springs SD (568A4) | 618-B69011 | VA797R-16-J-0376 | No |
| Lincoln, NE (636A5) | 618-B69016 | VA797R-16-J-0381 | No |
| Grand Island, NE (636A4) | 618-B69017 | VA797R-16-J-0382 | No |
| Omaha, NE (636) | 618-B69018 | VA797R-16-J-0383 | No |
| Omaha, NE (636) | 618-B69020 | VA797R-16-J-0385 | No |
| St. Cloud, MN (656) | 618-B69026 | VA797R-16-J-0390 | No |
| St. Cloud, MN (656) | 618-B69027 | VA797R-16-J-0391 | No |
| St. Cloud, MN (656) | 618-B69028 | VA797R-16-J-0393 | No |
| St. Cloud, MN (656) | 618-B69029 | VA797R-16-J-0394 | No |