

Harris, Craig R.

Subject: VA Providence ICU project VA241-17-R-0336
Attachments: VAMC_201367_ProVRepICU --Sheet - A1-2-2 - EQUIPMENT PLAN.DWG; VAMC_201367_ProVRepICU --Sheet - A1-1-1 - OVERALL FIRST FLOOR PLAN.DWG; VAMC_201367_ProVRepICU --Sheet - A1-1-2 - OVERALL SECOND FLOOR PLAN.DWG; VAMC_201367_ProVRepICU --Sheet - A1-1-3 - ROOF PLAN.DWG; VAMC_201367_ProVRepICU --Sheet - A1-2-1 - ENLARGED SECOND FLOOR PLAN.DWG; VAMC_201367_ProVRepICU --Sheet - A1-3-0 - FIRST FLOOR REFLECTED CEILINGdwg; VAMC_201367_ProVRepICU --Sheet - A1-3-1 - SECOND FLOOR REFLECTED CEILING....dwg

Please see attached Project CAD drawings.

Q&A are open until 5/10 per the RFP.

No site visit will be done until award-

Only electronic RFP responses will be accepted-No hard copies.

1) In each ICU room, will the VA/Providence be mounting Patient Lifts? **Lifts will be provided via a separate equipment contract.**

* if so, which vendor? **VA Providence is currently standardized to Hill-Rom Liko Lifts**

* if so, how will the Patient Lift Tracks be configured (XY or J)? **Configured as shown in the drawings provided.**

Coordination between lift and boom vendor is required as part of our SOW.

2) Is the VA/Providence interested in a IV Transfer system for each room? **Potentially, we would like to review this as an optional item.**

3) What type or specialty will these ICU rooms be? Cardiac? General? **General**

4) Since the hospital ICU Boom standard is fixed height will the VA/Providence accept fixed height Booms or only Booms with vertical height adjustments? **SOW calls for vertical height adjustments where possible. Proposals will be evaluated based on the ability to meet product requirements in addition to implementation services, past performance, and price.**

5) Is the VA/Providence RFP looking for us to quote a complete mounting structure which includes the mounting plate as well as the structure above it which attaches to the deck and/or steel or just the mounting plate itself? **Construction contractor will provide the structural support. Selected boom vendor will provide the mounting plate.**