

STATEMENT OF NEED

GENERAL:

The Anesthesia Department, located within the Wilmington VAMC, Wilmington, DE proposes purchasing and installation of new anesthesia machines. These machines are used to monitor patients in the OR.

DESCRIPTION	QTY
Perseus A500	4
Basic Unit with Insul. Transf Mechanic gas mixer	4
Me.3-gas mixer, Adv. Cyl. Supp Drager Auto Exclusion, 2 vap	4
Cylinder holder PinIndex Press. Red. O2, for small cyl.	4
Pressure reducer O2, Pin Index Pres. Red. N2O, for small cyl.	4
Pres. Reducer N2O, Pin Index Pres. Red. AIR, for small cyl.	4
Pres. Reducer N2O, Pin Index Central gas supply hoses	4
Additional power outlet strip	4
Mounting Pole right	4
SW option Pressur Support	4
SW option Breating Sound EM. SW opt. Low Flow Wizard	4
Flexible breathing bag arm	4
Inf.ID CLIC Absorber 800+, 6 pcs	24
Control Valve NPT	4
Endotrach. Suction DISS, VAC	4
Mounting kit VAC suction	4
Suction acc. (hoop, tip holder)	4
Pressuer guage	4

Drawer module standard	4
Box with writing tray	4
Adapter O2 DISS-Ohmeda	4
Adapter N2O Diss-Ohmeda	4
Adapter Air Diss-Ohmeda	4
Adapter VAC Diss-Ohmeda	4
Adapter EVAC Diss-Ohmeda Length of CS hoses 10ft	4
O2 CS hose 10 ft DISSN/DISSN	4
N2O CS hose 10 ft DISSN/DISSN	4
AIR CS hose 10 ft DISSN/DISSN	4
VAC CS hose 10 ft DISSN/DISSN	4
EVAC CS hose 10 ft DISSN/DISSN	4
Perseus A500 OrevCare Opt. 1 2 year	4
3rd Party Hardware	1
Vapor 3000 Specif. Nation property	4
Vapor 3000 Sevoflurane filling system Drager Fill Drager Standard	4
D-Vapor 3000	4
D-Vapor 3000 Front Cap Drager Countryspecific power cable	4
Clinical Application Supp Anesthesia .8hsegm.	5
Seminar Persus A500	1
Trade in Perseus (Aestiva 7900)	4

1. SPECIFICATIONS:

Turbo-Vent 2: accommodates infant to adult patients.

Heated breathing system optimized for low and minimal flow.

Sampled gas is returned from the analyzer to the breathing system; to help reduce anesthetic waste.

CPAP in MAN/Spont

Optional mode of ventilation: APRV

Configure Daylight Savings Time for automatic time change.

Vapor Communication: Low agent alarms (visual and audio).

Many options for mounting for peripheral devices which allows and ergonomic configuration.

Integrated cable channels to help maintain a less cluttered workspace and promotes a more hygienic environment.

Mechanical or Electronic Gas Mixer

2. DELIVERY:

The contractor shall deliver all items at no additional cost. All deliveries shall be made Monday- Friday, 8:00 a.m. – 3:30 p.m.

3. PRICING SCHEDULE: VA anticipates award of a contract on a firm fixed unit price basis.

4. PERFORMANCE, DELIVERY, INSPECTION AND ACCEPTANCE:

4.1 Backorders: If there is a backorder on any of the requested items, the contractor must contact the VA no later than 24 hours. The contractor will offer the VA substitute of equal or higher quality. If acceptable to the VA, the substitute may be made part of the contract with a bilateral modification.

4.2 Warranty: The contractor warrants/guarantees all items against defective material and/or workmanship for a minimum of one year from the date of receipt. The contractor agrees to furnish without additional cost to the government, all labor and materials necessary to correct defects that were detected during the guarantee period.

4.3 The Contractor shall immediately notify all affected VA's and contracting activities in the event of a product recall, removal, required adjustments, or required modification that is suggested or mandated by the contractor, distributor, manufacturer, or any regulatory or official agency. The information that shall be included in this notification include:

1. A complete item description, identification.
2. Administrative identification data including contract number, order numbers, and order date.
3. Reasons for recall, removal, adjustment, or modification.
4. Instructions for appropriate corrective action.

5. EDUCATION/ TRAINING:

5.1 The contractor shall provide any and all technical training required in order for VA staff to fully utilize the equipment. The contractor will provide installation and face to face training on the equipment.

DISCLAIMER

This RFI is issued solely for information and planning purposes only and does not constitute a solicitation. All information received in response to this RFI that is marked as proprietary will be handled accordingly. In accordance with FAR 15.201(e), responses to this notice are not offers and cannot be accepted by the Government to form a binding contract. Responders are solely responsible for all expenses associated with responding to this RFI.

End of Document