

ENDO MOBILE CARTS
FOR VA SIERRA NEVADA HEALTH CARE SYSTEM (VASNHCS)
QUESTIONS AND ANSWERS
VA261-17-Q-0607 / PR# 654-17-2-5135-0128

Question 1: Are these only metallic cabinets or is wood acceptable? I noticed the solicitation contained both.

Answer 1: *These cabinets are metal and have no wood except for the substrate for the thermafoil top. They must be LEED compliant and that was what the wood referred to the substrate for the thermafoil top and meeting LEED requirements. The cabinets cannot be wood, they do not hold up with the intensive cleaning and the chemicals that we use*