

AMENDMENT OF SOLICITATION/MODIFICATION OF CONTRACT

BPA NO.

1. CONTRACT ID CODE

PAGE
1OF PAGES
22. AMENDMENT/MODIFICATION NUMBER
A000013. EFFECTIVE DATE
08-21-20174. REQUISITION/PURCHASE REQ. NUMBER
502-17-2-170-01145. PROJECT NUMBER (if applicable)
NONE

6. ISSUED BY CODE 36C256

Department of Veterans Affairs
Veterans Health Administration
Network Contracting Office 16
2575 N. Keystone Crossing
Fayetteville AR 72703

7. ADMINISTERED BY (If other than Item 6) CODE 36C256

Department of Veterans Affairs
Veterans Health Administration
Network Contracting Office 16
2575 N. Keystone Crossing
Fayetteville AR 72703

8. NAME AND ADDRESS OF CONTRACTOR (Number, street, county, State and ZIP Code)

To all Offerors/Bidders

(X)

9A. AMENDMENT OF SOLICITATION NUMBER

VA256-17-Q-1086

X

9B. DATED (SEE ITEM 11)
08-01-2017

10A. MODIFICATION OF CONTRACT/ORDER NUMBER

10B. DATED (SEE ITEM 13)

CODE

FACILITY CODE

11. THIS ITEM ONLY APPLIES TO AMENDMENTS OF SOLICITATIONS

☒ The above numbered solicitation is amended as set forth in Item 14. The hour and date specified for receipt of Offers ☒ is extended, ☐ is not extended.
Offers must acknowledge receipt of this amendment prior to the hour and date specified in the solicitation or as amended, by one of the following methods:

(a) By completing Items 8 and 15, and returning 1 copies of the amendment; (b) By acknowledging receipt of this amendment on each copy of the offer submitted; or (c) By separate letter or electronic communication which includes a reference to the solicitation and amendment numbers. FAILURE OF YOUR ACKNOWLEDGMENT TO BE RECEIVED AT THE PLACE DESIGNATED FOR THE RECEIPT OF OFFERS PRIOR TO THE HOUR AND DATE SPECIFIED MAY RESULT IN REJECTION OF YOUR OFFER. If by virtue of this amendment you desire to change an offer already submitted, such change may be made by letter or electronic communication, provided each letter or electronic communication makes reference to the solicitation and this amendment, and is received prior to the opening hour and date specified.

12:00 PM CT, 8/25/2017

12. ACCOUNTING AND APPROPRIATION DATA (If required)

13. THIS ITEM APPLIES ONLY TO MODIFICATIONS OF CONTRACTS/ORDERS,
IT MODIFIES THE CONTRACT/ORDER NO. AS DESCRIBED IN ITEM 14.CHECK
ONE

A. THIS CHANGE ORDER IS ISSUED PURSUANT TO: (Specify authority) THE CHANGES SET FORTH IN ITEM 14 ARE MADE IN THE CONTRACT ORDER NO. IN ITEM 10A.

B. THE ABOVE NUMBERED CONTRACT/ORDER IS MODIFIED TO REFLECT THE ADMINISTRATIVE CHANGES (such as changes in paying office, appropriation date, etc.) SET FORTH IN ITEM 14, PURSUANT TO THE AUTHORITY OF FAR 43.103(b).

C. THIS SUPPLEMENTAL AGREEMENT IS ENTERED INTO PURSUANT TO AUTHORITY OF:

D. OTHER (Specify type of modification and authority)

E. IMPORTANT: Contractor ☐ is not, ☐ is required to sign this document and return _____ copies to the issuing office.

14. DESCRIPTION OF AMENDMENT/MODIFICATION (Organized by UCF section headings, including solicitation/contract subject matter where feasible.)

The solicitation identified in Block 9A above is hereby amended as described below:

a. Clarification, as indicated on page 2 of this amendment, is provided in response to questions submitted in regard to the current status of equipment to be covered by this contract. And,

b. The deadline for submitting quotes is extended to August 25, 2017 at 12:00 PM Central Time, as indicated in Block 11 above.

(Continued on Page 2)

Except as provided herein, all terms and conditions of the document referenced in Item 9A or 10A, as heretofore changed, remains unchanged and in full force and effect.

15A. NAME AND TITLE OF SIGNER (Type or print)

16A. NAME AND TITLE OF CONTRACTING OFFICER (Type or print)

James Hunt
Contracting Officer

15B. CONTRACTOR/OFFEROR

15C. DATE SIGNED

16B. UNITED STATES OF AMERICA

16C. DATE SIGNED

(Signature of person authorized to sign)

BY (Signature of Contracting Officer)

No.	Question	Answer
1	When was the current x-ray tube installed?	AQUILION ONE SN: 2BA1632025 Date: 9/28/2012
2	How many scan seconds does the current x-ray tube have on it?	105,284.1
3	What is the current exposure count?	124,587