

The OVAHCS NCO 08 - 675 – Orlando Veterans Affairs Health Care System in FL requires the following: **ARM RETRACTOR.**

1. Background: The Orlando Veterans Affairs Health Care System has a requirement to purchase 2 new Martin Arm Retractor.

2. Performance Period:

- One Time.

3. Minimum Government Needs:

Straight holding system:

- Intake clamp accepts 0mm to 18mm instrumentation.
- Height of 30cm.
- Reach of 37 cm.

Extension rod:

- Length of 50cm.

ITEM INFORMATION

ITEM NUMBER	DESCRIPTION OF SUPPLIES/SERVICES	QUANTITY	UNIT	UNIT PRICE	AMOUNT
0001	PRODUCT NUMBER 28272KLA STRAIGHT HOLDING SYSTEM W/28272UL CLAMP UNIVERSAL INTAKE 0MM-18MM, H:30CM, R:37CM OR EQUAL	2.00	EA		
0002	PRODUCT NUMBER 28172HM EXTENSION ROD FOR HEIGHT ADJUSTMENT OF THE ARTICULATED STAND, 50CM Manufacturer's Warranty - 1 year OR EQUAL	2.00	EA		

DELIVERY SCHEDULE

13800 Veterans Way

Orlando FL, 32827