

## STATEMENT OF WORK

This uses a **Brand Name or Equal Description** of the product required. This permits prospective contractors to offer products other than those specifically referenced by brand name. Offerors of equal products shall include **equipment brochures which clearly define the equal product, or offer will not be considered.** Any information submitted by an offeror for consideration will not be returned and will remain the property of the VA. All costs associated with submitting the above are the responsibility of the Contractor and all information provided must be submitted to the Contracting Officer.

### Real Ear Hearing aid Analyzer

#### Background

This piece of equipment is used to measure hearing aid output in the patients' ears using probe microphones. This item also contains a hearing aid analyzer test box to analyze whether the hearing aid output is meeting the technical specifications required. These probe microphone measurements are a required part of our hearing aid fitting protocol.

The Audiology Department has existing Audioscan Verifit Model VF-2, but will need to procure additional piece of this equipment in order to increase the efficiency of our staff and scheduling.

#### Selection Criteria

Vendor must meet or exceed all requirements as specified below.

#### Specifications

1 EA - Audioscan Verifit Model VF-2 or Equal

- Unit must be able to work as a stand-alone without a PC
- Dual probe real ear module for on-ear measures
- Test box coupler for performing binaural test box measures with expanded frequency range up to 16000Hz
- NOAH database module compatible
- Installation and Operations Manual
- Service Manual provided with no charge
- 12 month manufacturer warranty on parts and labor

#### Electrical

- 115 VAC, 50/60 Hz

#### Information Security:

The vendor will not have any access to sensitive VA information. The C&A requirements do not apply, and that a Security Accreditation Package is not required.