

Attachment B: Contract Line Item Number (CLIN) pricing schedule

Line Item:	Description:	QTY:	Unit:	Period of Performance:	Price:
0001	<p>Peripheral Pressure, Oxygen and Carbon Dioxide Measuring Device:</p> <ul style="list-style-type: none"> • Must be able to simultaneously assess limb circulation and wound healing and use transcutaneous oximetry with a visual and audible alarm • Automatic calibration: In air (tcpO2)/with TC600 (tcpO2) 8 electrodes simultaneously • Must be HL7 and DICOM compatible • The unit shall be on a mobile workstation and contain: <ul style="list-style-type: none"> -Touchscreen color display -Keyboard & Printer -USB ports -2 GB storage capacity -(6) pressure outlets -Combined pO2/pCO2 electrode sensor cpO2 electrode unit with a temperature range of 37°C to 45°C set in steps of 0.5°C -drawer • Accessories & Consumables shall include: <ul style="list-style-type: none"> Fixation rings, contact liquid, membranizing kits, unit & gas calibrations kits with calibration device, camera, range of cuff sizes, range of Doppler probes, system cart, foot pedal, probe holder mounted on workstation, two accessory baskets 	1	EA	4/27/2018 – 5/27/2018	
0002	<p>Training:</p> <ul style="list-style-type: none"> • See Attachment D 	1	JB	4/27/2018 – 5/27/2018	