

Attachment B  
RFQ: 36C24818Q0851

<b>Line Item</b>	<b>Description</b>	<b>Qty</b>	<b>Unit</b>	<b>Unit Price</b>	<b>Total Price</b>
0001	<p><b>Middle Ear Analyzer – the product must include the following:</b></p> <p>Offer a comprehensive range of middle ear testing capabilities that support the needs of hearing healthcare professionals. Providing reliable and efficient clinical middle ear solutions.</p> <p>Power Consumption: 60 Watts Max</p> <p>Test Types: Tympanometry, Acoustic Reflex Threshold, Reflex Decay, Eustachian Tube Function (Intact and Perforated)</p> <p>Protocols: Diagnostic, Screening, User-defined, Frequency Tympanometry, Acoustic Reflex Latency Test, Reflex Sensitization</p> <p>Interface: USB, Keyboard, Mouse , Flash Drive, PC Communication.</p> <p>Printout: external Deskjet Printer</p> <p>Probe Tone: 226 Hz (85 dB SPL +/- 1.5 dB)</p> <p>Sensitivity Scale: Auto Scales to appropriate range, Manual selection also possible in reflex modes only</p> <p>Accuracy (226 hz)</p> <ul style="list-style-type: none"> <li>- Tympanometry Mode: +/- 5% of reading or +/- 0.1 ml, whichever is greater.</li> <li>- Reflex Mode: +/- 5% of reading or +/- 0.2 ml, whichever is greater.</li> </ul> <p>Pressure Measurements: (load volume of 0.2 to 7.0 ml).</p> <p>Range: Normal = +200 to -400 daPa. Wide = +400 to -600 daPa.</p> <p>Accuracy: +/- 10% of reading or +/- 10 daPa, whichever is greater.</p> <p>Sweep Rate: 12.5, 50, and 600/200 daPa/sec.</p> <p>Sweep Accuracy: 10% of nominal rate Maximum limits (in 0.5cc cavity): -800 daPa and +600 daPa.</p>	EA	5		

Attachment B  
RFQ: 36C24818Q0851

	<p>Frequency Accuracy: <math>\pm 3\%</math>.  Harmonic Distortion (THD): Less than 5% (measured acoustically).  Noise Signals: (3 dB bandwidths).  Low Band: 125 -1,600 Hz.  High Band: 1,600 -4,000 Hz  Broad Band: 125 -4,000 Hz  Intensity Range: 35 to 120 dB HL  Step Size: 5 dB, 1 dB and 2 dB Calibration  Accuracy: <math>\pm 3</math> dB  Step Accuracy: <math>\pm 0.5</math> dB  ON/OFF Ratio: 70 dB minimum  Temperature:  - Storage: -32° F to +122° F  - Operating: +59° F to +95° F  - Humidity: 90% at 95° F (non-condensing)</p>				
0002	<p><b>Audio Meter – The Audio Meter must include all the following:</b>  TDH 50 Phones  ER-3A Insert Phones  B81 Bone Oscillator  HDA-300 High Frequency Phones  Patient Response Button  Patch Cords  Talk Back Mic  Mic-Monitor Headset  Asst. Listening Headset w/ extension cord  Instruct. Manual  GSI- Suite Software  GSI Audio/Tymp module for NOAH  USB Cable  Calibration Certificate  Dust Cover  CD- User Manuals  Quick Guide</p>	EA	2		
0003	<p><b>On Site Installation all equipment. Middle Ear analyzer will be installed at the Malcom Randall VAMC and Compensation and pension Clinic. The Audio Meters will be installed at the compensation and pension Clinic.</b></p>	JB	2		
<b>Total Cost</b>					