

<i>Line Item</i>	<i>Description</i>	<i>Qty</i>	<i>Unit</i>	<i>Unit Price</i>	<i>Total Price</i>
0001	Ultrasound System Manufacturer: Fuji Film Make: Sonosite Model: SonoSite S II Part Number: L22406 Or Equal	2	EA		
0002	13-6 MHZ Transducer Transverse Biopsy Compatible Manufacturer: Fuji Film Make: Sonosite Model: L25X Part Number: P07691 Or Equal	2	EA		
0003	Stand Manufacturer: Fuji Film Make: Sonosite Model: S II Stand Part Number: L21432 Or Equal	2	EA		
0004	Dual Transducer Cover Manufacturer: Fuji Film Make: Sonosite Model: DTC Part Number: P22523 Or Equal	2	EA		
0005	Service Manuals and User Guides	2	EA		
Total Cost					

Salient Characteristics

- Must be able to operate in the below modes:
 - B-Mode
 - M-Mode
 - Color Doppler
 - Color Power Doppler (CPD)
 - Pulsed Wave
 - Continuous Wave
- Must support transducers with linear, phased array and curved array formats.
- Must include a full array of measurement and calculation packages.