

Statement of Work
Thermo Scientific Excelsior AS Tissue Processor
06/01/2018

A. GENERAL INFORMATION

1. Title of Project: Thermo Scientific Excelsior AS Tissue Processor
2. Type of Contract Contemplated: *Equipment Purchase*

B. SCOPE

Needed to process biopsy specimens, in order to diagnose cancer.

C. REQUIREMENT/SALIENT CHARACTERISTICS/SPECIFICATIONS

Common Nomenclature (commercial description): Tissue Processor

Kind of material (i.e., type, grade, alternatives, etc.): equipment

Electrical Data (i.e., 110 v, 60 Hz, 3 Ph, etc.): **100-240V**

Dimensions, size, capacity: 28"W X 54"H X 23"D
Weight – 363 lbs.

Principles of operation: Tissue Processor

Restrictive environmental conditions: None

Intended use: process biopsy specimens

Equipment with which the item is used

(N/A if not applicable): N/A

End item application: Tissue samples

Original equipment manufacturer part number: A82300001

Other pertinent information that describes the item, material or service required:

D. OTHER UNIQUE REQUIREMENTS

N/A

Compatibility with Existing Equipment: N/A

Applicable DICOM compliance, Directives or IT Policies: N/A

Coordination with On-going Construction Project (provide the project number): N/A

Warranty: 1 Year

Training: N/A

Service and/or Maintenance: M-F, 8AM – 5 PM

Safety/ISO/IT/BioMed/SPS & Other Departmental Compliance: N/A

E. DELIVERY

This section should clearly state:

- What the Contractor must deliver – Excelsior AS Tissue Processor
- When the Contractor must deliver – 30 days after purchase
- Where the Contractor must deliver the service – Syracuse VAMC Histology Lab.
COR – John Zavaski
- What documentation (if any) the Contractor must obtain –completed invoice
- Where inspection and acceptance will occur – Syracuse VAMC Histology Lab
- Other items to consider: Include any associated review periods or incorporation of Government comments as applicable. Deliverables are different than task completion milestones. Milestones may not require a specific submittal. Include here the schedule for when final products such as data, reports or other items are required to be furnished to the primary requestor.

F. Government-Furnished Property, Material, Equipment, or Information (GFP, GFM, GFE, or GFI) NONE

G. SECURITY

This section should identify any unique security requirements associated with contract performance (when applicable). NONE, unit will not be interfaced.

H. SPECIAL MATERIAL REQUIREMENT

None

I. PLACE OF PERFORMANCE

Syracuse VAMC, Room C301

J. PERIOD OF PERFORMANCE

2/1/18 – 2/1/28

K. BILLING/INVOICING - Syracuse VAMC

L. CUSTOMER INFORMATION

John Zavaski, Laboratory Supervisor, Syracuse VAMC, 800 Irving Avenue, Syracuse, NY 13210. Phone – (315)425-4800

M. DRAWINGS/ASBUILTS/CONSTRUCTION FLOOR PLANS

N/A