

# FedBizOpps Sources Sought Notice

<b>CLASSIFICATION CODE *</b>	65
<b>SUBJECT *</b>	PATIENT MONITORING SYSTEM
<b>GENERAL INFORMATION</b>	
<b>CONTRACTING OFFICE'S * ZIP-CODE</b>	90815
<b>SOLICITATION NUMBER *</b>	36C26218Q9452
<b>RESPONSE DATE (MM-DD-YYYY)</b>	06-19-2018
<b>ARCHIVE</b>	30 <b>DAYS AFTER THE RESPONSE DATE</b>
<b>RECOVERY ACT FUNDS</b>	N
<b>SET-ASIDE</b>	
<b>NAICS CODE *</b>	334510
<b>CONTRACTING OFFICE ADDRESS</b>	Department of Veterans Affairs Network Contracting Office 22 4811 Airport Plaza Drive Suite 600 Long Beach CA 90815
<b>DESCRIPTION *</b>	<b>See Attachment</b>
<b>POINT OF CONTACT *</b>  (POC Information Automatically Filled from User Profile Unless Entered)	FREDY A ARTEAGA CONTRACTING OFFICER
<b>PLACE OF PERFORMANCE</b>	
<b>ADDRESS</b>	VAMC GREATER LOS ANGELES 11301 WILSHIRE BLVD  LOS ANGELES CA
<b>POSTAL CODE</b>	90073
<b>COUNTRY</b>	USA
<b>ADDITIONAL INFORMATION</b>	
<b>AGENCY'S URL</b>	WWW.VA.GOV
<b>URL DESCRIPTION</b>	AGENCY WEBSITE
<b>AGENCY CONTACT'S EMAIL ADDRESS</b>	FREDY.ARTEAGA@VA.GOV
<b>EMAIL DESCRIPTION</b>	WORK EMAIL

\* = Required Field

**THIS REQUEST FOR INFORMATION (RFI)/SOURCES SOUGHT IS ISSUED SOLELY FOR INFORMATION AND PLANNING PURPOSES ONLY AND DOES NOT CONSTITUTE A SOLICITATION.**

**THE SUBMISSION OF PRICING, CAPABILITIES FOR PLANNING PURPOSES, AND OTHER MARKET INFORMATION IS HIGHLY ENCOURAGED AND ALLOWED UNDER THIS RFI IN ACCORDANCE WITH (IAW) FAR 15.201(e).**

## **DISCLAIMER**

This RFI is issued solely for information and planning purposes only and does not constitute a solicitation. All information received in response to this RFI that is marked as proprietary will be handled accordingly. IAW FAR 15.201(e), responses to this notice are not offers and cannot be accepted by the Government to form a binding contract. Responders are solely responsible for all expenses associated with responding to this RFI.

## **SOURCES SOUGHT DESCRIPTION**

This is NOT a solicitation announcement. This is a sources sought/RFI only. The purpose of this sources sought/RFI is to gain knowledge of potential qualified sources and their size classification relative to NAICS 334510 (size standard of 1250 Employees). Responses to this sources sought will be used by the Government to make appropriate acquisition decisions. After review of the responses to this sources sought, a solicitation announcement may be published. Responses to this sources sought synopsis are not considered adequate responses for a solicitation announcement.

The Department of Veterans Affairs (VA), VISN 22 Network Contracting Office, is seeking sources that can provide All in One Mobile Nurse Workstations that at a minimum meets the following salient characteristics with a brand name or equal product for the VA Greater Los Angeles Healthcare System:

### **Salient Characteristics**

**Description of Need:** Advanced hemodynamic, central venous access, patient monitoring platform for continuous cardiac output compatible with the Swan-Ganz pulmonary artery catheters and oximetry central venous catheter.

### **Scope: Purchase and Installation must include:**

- Continuous Cardiac output monitor
- Cable cco catheter interface
- Oximetry cable
- Roll stand
- Interface cable - (intellispace critical care or intellibridge)

### **Must have the capability to monitor:**

- Central Venous Oxygen Saturation (ScvO2)
- Continuous Cardiac Output (CCO, CCI)
- Continuous Right Ventricular Ejection Fraction (RVEF)
- Continuous Right Ventricular End Diastolic Volume (RVEDV, RVEDVI)
- Mixed Venous Oxygen Saturation (SvO2)
- Pulmonary Vascular Resistance (PVR)
- Right Ventricular Stroke Work Index (RVSWI)
- Stroke Volume (SV, SVI)
- Systemic Vascular Resistance (SVR, SVRI)

### **Must Include:**

- Wired and wireless communication
- Must interface with our current Critical Care Information System (Philips)
  - Capable through Capsule Tech, EC10 module, or otherwise
- Warranty
  - Covers Physical and Liquid Damage to the Equipment
  - Associated Oximetry Cable Failures Replaced at No Charge
  - Guaranteed Next Day Replacement of Covered Equipment
  - All Parts, Labor and Shipping Costs Covered
  - 24/7/365 Technical & Product Phone Support
  - Between 12-48 months of Total Protection
- End User In-Service Training
  - One day of training per 3 devices purchased
  - Service and Operators Manual

### **Brand Name Information:**

1. Item Number HEMSGOX2-US, HEMOSPHERE ADVANCED MONITORING PLATFORM
2. Item Number HEMBAT10, HEMOSPHERE BATTERY PACK
3. Item Number HEMOXSC100, HEMOSPHERE OXIMETRY CABLE
4. Item Number HEMRLSTD1000, HEMOSPHERE ROLL STAND

The information identified above is intended to be descriptive, not restrictive and to indicate the quality of the supplies/services that will be satisfactory. It is the responsibility of the interested source to demonstrate to the government that the interested parties can provide the supplies/services that fulfill the required specifications.

If you are interested, and are capable of providing the sought out supplies/services, please provide the requested information as well as the information indicated below. Response to this

notice should include company name, address, point of contact, size of business pursuant to the following questions:

- (1) Please indicate the size status and representations of your business, such as but not limited to: Service Disabled Veteran Owned Small Business (SDVOSB), Veteran Owned Small Business (VOSB), Hubzone, Woman Owned Small Business (WOSB), Large Business, etc.)?**
- (2) Is your company considered small under the NAICS code identified under this RFI?**
- (3) Are you the manufacturer, distributor, or an equivalent solution to the items being referenced above?**
- (4) If you are a large business, do you have any designated distributors? If so, please provide their company name, telephone, point of Contact and size status (if available).**
- (5) If you're a small business and you are an authorized distributor/reseller for the items identified above or an equivalent solution, do you alter; assemble; modify; the items requested in any way? If you do, state how and what is altered; assembled; modified?**
- (6) If you intend to sub contract any work on this contract, what portion of the total cost will be self-performed/will be performed by your organization? Please provide estimated detailed percentage breakdowns related to sub contracted work.**
- (7) Does your company have an FSS contract with GSA or the NAC or are you a contract holder with NASA SEWP or any other federal contract? If so, please provide the contract number.**
- (8) If you are an FSS GSA/NAC or NASA SEWP contract holder or other federal contract holder, are the items/solution you are providing information for available on your schedule/contract?**
- (9) Please provide general pricing of your products/solution for market research purposes.**
- (10) Please submit your capabilities in regards to the salient characteristics being provided and any information pertaining to "equal to items" to establish capabilities for planning purposes?**
- (11) Please review salient characteristics/statement of work (if applicable) and provide feedback or suggestions. If none, please reply as N/A.**
- (12) Please provide your DUNS number.**

**\*\*\* Submissions addressing number (10) should show clear, compelling and convincing\*\*\* evidence that all "equal to items" meet all the salient characteristics.**

Responses to this notice shall be submitted via email to [Fredy.arteaga@va.gov](mailto:Fredy.arteaga@va.gov). Telephone responses shall not be accepted. Responses must be received no later than Tuesday, June 19,

2018 at 1:00 p.m. EST. If a solicitation is issued it shall be announced at a later date, and all interested parties must respond to that solicitation announcement separately from the responses to this sources sought. Responses to this sources sought notice are not a request to be added to a prospective bidders list or to receive a copy of the solicitation