

### **Question and Answers for 36C24818Q0571**

**Question No. 1:** Is there a current incumbent?

**Answer:** Yes

**Question No. 2:** If so, what ESR brand and model is currently being used?

**Answer:** Streck

**Question No. 3:** If current equipment is VA owned equipment will you need a trade-in value?

**Answer:** No.

**Question No. 4:** Are both systems in Orlando close enough to each other to share the same UPS?

**Answer:** Probably not.

**Question No. 5:** Will you require a separate UPS per unit?

**Answer:** Yes. Orlando (2), Daytona (1) and Viera (1).

**Question No. 6:** What are the minimum UPS back up hours required by VA?

**Answer:** 20-30 minutes.

**Question No. 7:** 2.1.2 What is the maximum footprint allowable? L x W x H (Is there access to electrical for UPS)?

**Answer:** Standard or smaller footprint, table top. Yes, there is electrical for the UPS.

**Question No. 8:** 2.7.7.11 C. Please clarify number of days.

**Answer:** 2.5.7.11 C. is 90 days.

**Question No. 9:** Cost per Test (CPT) or Cost per Reportable Report (CPRR)?

**Answer:** Cost per Test.

**Question No. 10:** Should shipping be included?

**Answers:** Yes. Shipping shall be FBO Destination.