

Salient Characteristics Nuclear Medicine Dose Calibrators

The Department of Veterans Affairs Network Contracting Office (NCO) 17 has a requirement for Nuclear Medicine Dose Calibrators. The North American Industry Classification System (NAICS) code for this procurement is 334519 Other Measuring and Controlling Device Manufacturing.

All items must be covered by manufacturer's warranty and procured through a manufacturer approved distribution channel. Sellers must be able to document they are manufacturer-authorized distributors.

The Government has identified the below items as meeting the requirement:

Line Item	Manufacturer	Part No.	Description	Qty	Unit
1	Capintec, Inc.	5130-3215	CRC-25R Dose Calibrator complete with Chamber	2	EA
2	Capintec, Inc.	5130-3234	CRC-55TR Dose Calibrator complete with Chamber	1	EA

This is a brand name or equal to purchase description. Items comparable to the brand name will be acceptable as long as they meet the below salient characteristics.

Salient Characteristics Item 1 (5130-3215):

- Nuclide keys: 28 Programmable keys
- System Memory: Over 80 Nuclides
- Chamber plug and play
- Interface: RC-232 & USB
- Dose Calibrator Chamber welded at the base
- Includes: Well Insert
- Vial/Syringe Dipper
- Printer Epson, Included
- Pkg Paper rolls 12 count.
- Mounting Plate
- Instructional Manual

Salient Characteristics Item 2 (5130-3234):

- Touch Screen Interface
- Type 8" VGA LCD Color Touch screen display
- Direct reading in Bq or Ci
- Nuclide keys: 28 Programmable keys
- System Memory: Over 80 Nuclides
- Chamber plug and play
- Interface: RC-232 & USB
- Dose Calibrator Chamber welded at the base
- Includes: Well Insert
- Vial/Syringe Dipper
- Printer Epson, Included
- Pkg Paper rolls 12 count.
- Mounting Plate
- Instructional Manual

User and Service Manuals

The vendor shall provide a complete printed set as well as electronic version of operator manuals, service manuals, electronic schematics, troubleshooting guides and parts list for all items. These manuals will include all components and subassemblies, including those not manufactured by the vendor.

Delivery Address

Audie L. Murphy VA Hospital
Supply Warehouse
7400 Merton Minter
San Antonio, TX 78229