

DEPARTMENT OF VETERANS AFFAIRS

Justification for Single Source Awards IAW [FAR 13.106-1](#)

For

Over Micro-Purchase Threshold but Not Exceeding the SAT (\$250K)

Acquisition Plan Action ID: VA244-18-AP-6035

1. **Contracting Activity:** Department of Veterans Affairs, VISN 04, *Lebanon* Medical Center
1700 South Lincoln Ave., Lebanon, PA 17067, Associate Director, 595-18-3-455-0159,

2. **Brief Description of Supplies/ Services required and the intended use/Estimated Amount:**

This is a Brand Name requirement. ZEISS OS | HFA3 MODEL 860 W/ LIQUID TRIAL LENS, RELEYE, KINETIC, GPA, HEAD TRK, GAZE TRK, PRINTER & TABLE / 1 each, \$35,517.00

Fixation Control, Video eye monitor, Heijl-Krakau blind spot monitor, Gaze Tracking Head Tracking, Vertex Monitoring

Visual Field Index™: Summary measurement of the patient's visual field status, expressed as a percent of a normal age-adjusted visual field.

Liquid Trial Lens™: Using Liquid pressure, delivers each patient's refractive correction. Available correction range is -8 to +8 diopters. Spherical correction only. Liquid Trial Lens available in model 860 only.

SmartTouch interface: Touch screen interface allows for a more efficient Workflow

RelEye: The RelEYE fixation monitoring system allows for systematic and pin point fixation inspection. The HFA 3 captures and keeps the image of the eye during the testing and just after, prior to the Result Review Screen.

RelEYE cannot be viewed on the HFA 3 after terminating the test. Using the HFA 3, RelEYE is also available in FORUM Glaucoma Workplace (minimum version 2.0.3). Only FORUM can keep the RelEYE data stored for future usage for exporting exams.

3. **Unique characteristics that limit availability to only one source, with the reason no other supplies or services can be used:**

The LVAMC providers presently utilize this product in the Lebanon Facility, and the Lancaster Facility to service patients. The exact equipment is needed to maintain continuity in care, as well as allow for transmission of data between facilities with the use of Zeiss Forum which connects all equipment for immediate access of data regardless of where patient is cared for.

The Zeiss Visual Field Machine is compatible with all other components requiring comprehensive Glaucoma and other eye disease testing used in the Lebanon VA Medical Center, such as Forum, and the Cirrus OCT Machine. This device offers superior functionality in diagnosing and ultimately treating debilitating ocular diseases.

4. Description of market research conducted and results or statement why it was not conducted:
FPDS, VETBIZ, Small Business were searched – results showed that items are available.

5. Contracting Officer's Certification: *Purchase is approved in accordance with FAR13.106-1(b). I certify that the foregoing justification is accurate and complete to the best of my knowledge and belief.*

Name: Donna Cooper
Title: Contracting Officer
Date: 07/05/2018

Name: Dianne Smith
Title: Division Chief, Division 2
Date: 07/05/2018