

Attachment B
 RFQ: 36C24818Q9328

<i>Line Item</i>	<i>Description</i>	<i>Qty</i>	<i>Unit</i>	<i>Unit Price</i>	<i>Total Price</i>
0001	Panoramic X-Ray System Must meet all the specifications listed below.	EA	1		
0002	Full Day Onsite Romexis Training	JB	1		
0003	Installation	JB	1		
Total Cost					

All equipment must meet the following specifications:

- Capable of taking dental 2D Digital and 3D Digital Images
- Power Requirement: 110V/120A
- Dimensions move to 93.8" high for patients that are 7 feet tall
- Programable Bitewings
- Compatible with MiPacs and Romexis software
- Able to be bolted to the wall
- Serializable bite stick
- Able to be wipe down with hydrogen peroxide disinfectant
- USB Cable
- Onsite Romexis Training
- Installation