

PRODUCT DESCRIPTION

For

NON-EXPENDABLE EQUIPMENT:

Stereotactic Ear Nose Throat (ENT) Navigation System



Prepared By:

Strategic Acquisition Center
Department of Veterans Affairs
10300 Spotsylvania Avenue, Suite 400
Fredericksburg, VA 22408

1. SCOPE OF WORK:

The Veterans Health Administration (VHA) Non-Expendable Equipment National Program has identified the Stereotactic ENT (Ears Nose Throat) Navigation System product line as a candidate for a VA-wide (otherwise referred to as “national”) single Requirements contract award. The Stereotactic ENT Navigation System is an image-guided computer-aided stereotactic surgical system designed to aid in the performance of minimally invasive otorhinolaryngology procedures by combining medical imaging, dedicated computer software, and surgical instrument tracking systems for guidance. The systems provide surgeons with real-time, three-dimensional (3-D) positioning and orientation of surgical instruments (e.g., trephines, drills) relative to the patient's anatomy on a video display during surgery without the use of an endoscope. Stereotactic ENT Navigation Systems are used in several procedures, including skull base procedures, nasal sinus surgery, and other intranasal and middle ear procedures.

The Contractor must demonstrate the ability to meet all requirements for the solicitation. The objective is to provide ENT Navigation Systems to be used by clinicians throughout the VA medical centers and facilities. The period of performance is for one (1) 12 month base period with two (2) twelve (12) month option periods.

2. REQUIREMENT

This requirement will be IAW FAR clause 52.211-6, “Brand Name or Equal” which requires that the offeror indicate each product that is being offered as an “equal” product. For each equal product, the offeror must include a description reflecting the characteristics and level of quality that will satisfy the salient physical, functional, or performance characteristics of the equal product(s) specified in the solicitation. The offeror also must clearly identify the item by brand name (if any) and make/model number. Finally, the offeror must include descriptive literature, such as illustrations, drawings, or a clear reference to previously furnished descriptive data or information available to the contracting officer, and clearly describe any modifications it plans to make to a product to make it conform to the solicitation requirements.

CLIN	Brand	Part Number	Item Description	Base Year	Opt 1	Opt 2
0001	Medtronic	9733700XOM	Bundle Fusion Software/Instrument Set	10 EA	10 EA	10 EA
0002	Medtronic	9733560XOM	Cart Fusion EM System	10 EA	10 EA	10 EA
0003	Medtronic	9733560XOM1	Fusion Software Support Program Renewal	10 EA	10 EA	10 EA
0004	Medtronic	9733933XOM	SFTWR Kit DICOM Software Q/R	10 EA	10 EA	10 EA
0005	Medtronic	9734400XOM	Software STEALTHMERGE ENT PG	10 EA	10 EA	10 EA

The Department of Veterans Affairs (VA) is seeking vendors who can provide Covidien/Medtronic brand of ENT Navigation System as listed above or equal commodities which meet all the following salient characteristics:

SC	APPLICABLE CLINs	SALIENT CHARACTERISTICS	METHOD OF EVALUATION
SC 1	0001 - 0005	Electromagnetic Tracking	Literature Review
SC 2	0001 - 0005	Ability to use stainless steel instruments.	Literature Review
SC 3	0001 - 0005	Instrument verification that requires brief or no calibration sequence.	Literature Review
SC 4	0001 - 0005	Ability to track three different instruments at once.	Literature Review