

Department of  
Veterans Affairs

# Memorandum

**Date:** August 16, 2012

**From:** Business Manager , SSSL, VAPHS

**Subj:** SALIENT CHARACTERISTICS / FUNCTIONAL / PERFORMANCE  
SPECIFICATIONS/ STATEMENT OF NEED   JedMed CSC Cabinet

**To:** Acquisitions / Contracting

**Thru:** Business Resource Specialist

1. The following are the required characteristics, functional and performance specifications for the requested item. **Note: use generalities and NOT trade or proprietary names in your memo.**

- a. ENT Treatment Cabinet
- b. Sealed Membrane Switches
- c. Control Box:
  - i. Input: 115Vac, 60 Hz
  - ii. Output: 115Vac (controls up to 10 Amps)
  - iii. Power: 1150W maximum
  - iv. Class: Class 1 (3 prong, grounded)
  - v. Dimensions: 7.5 W x 5.0 D x 8.0 H inches
  - vi. Weight: 14 lbs. (6.4 kg)
- d. Suction Pump:
  - i. Electrical Input: 115Vac 4.2amps 60Hz
  - ii. Vacuum: 25.5 in Hg (max.)
  - iii. Flowrate: 1.1cfm @ 0 in. Hg
  - iv. Dimensions: 7.5 W x 5.0 D x 8.0 H inches
  - v. Weight: 14 lbs. (6.4 kg)
- e. Pressure Pump:
  - i. Electrical Input: 115Vac 2.5amps 60 Hz
  - ii. Pressure: 35psi
  - iii. Flowrate: 8 lpm
  - iv. Dimensions: 8.0 W x 4.0 D x 7.0 H inches
  - v. Weight: 5.0 lbs.
- f. Trap Canister:
  - i. Dimensions: 5.0 D x 6.0 H inches
  - ii. Capacity: 800cc

2. Statement of Need: The Surgical Service Line is requesting the purchase of additional ENTJedmed Cabinets to accommodate the patient load in the new ENT patient clinic. The

ENT service line increased their clinic space by 4 patient exam rooms and requires additional cabinets for patient care. At this time 4 brand new patient suites are not being utilized for what they were designed for. This has been approved on the 2012 5yr equipment plan.

3. Equipment Transaction Number: **646-431-12-00005**
4. Vista Transaction Number: **646-12-3-167-0100**
5. If you have any questions please feel free to contact me 412-360-3798.

**JOYCE CAHILL, LCSW**

**Business Manager, Surgical Specialty Service Line**

**Note: Please attach brochures or equipment specifications provided by companies do NOT include them in the body of your memo as references, use only the specifications they have provided. Be as thorough as you can to assist in the purchase of your equipment.**

Firms with interest and capability to respond to this requirement should develop a short narrative including relevant experience and other information indicating and evidencing capability to successfully meet the aforementioned requirements. Interested parties should send their company information via E-mail (preferred) or fax to Mr. Pedro Catinchi, Contracting Officer, Service Area East Office. E-mail is [pedro.catinchi@va.gov](mailto:pedro.catinchi@va.gov) and fax number is 412-822-3440.

Company information must include business type e.g. Service Disabled Veteran Owned Business, Woman Owned Small Business, etc. Information should be provided no later than (12:00 PM) AUG 29, 2012.