

Brand Name or Equal Worksheet (“Salient Characteristics”) v 2.0

Brand Name Item: Stryker System 8 Orthopedic Power Tools

1) Essential/significant physical, functional, or performance characteristics.

- Must be able to use the power tools for orthopedic, vascular, and podiatry procedures that are or will be done at SMC VAMC
- Must include a cordless driver, sabo sagittal saw, dual trigger rotary drill, sternum saw, sagittal saw, and all necessary components
- The power tool package must include saws/drills that are able to cut through small, medium and large sized bones
- Power tools must be designed for both threaded and fluted pins, as well as straight
- Must be able to temporarily submerge the power tool handpieces during the cleaning process
- Must be able to actively wash the power tools
- Must be battery powered tools
- Batteries must be able to sit on the shelf for 30 days without losing power

2) Complete generic identification – Orthopedic Power Tools

3) Applicable model/make/catalog number – Stryker System 8 Power Equip

4) Manufacturer name – Stryker

Per VAAR 811.104 and 811.105, I certify the following regarding the above information:

- The salient characteristics listed above are essential to the Government’s requirements.
- To my knowledge, the restrictions listed above will not limit the acquisition to an item peculiar to one manufacture.

Name, Date:

Madison Pallante

6/13/18