

SHIP TO:
WAREHOUSE B90007
V.A. Medical Center
JERRY L PETTIS
VETERANS HOSPITAL
11201 BENTON ST
LOMA LINDA, CA 92357

605-B90007

Qty	Item Description
1	<p>Biograph Vision 600</p> <p>The Biograph Vision 600 is a whole-body PETCT tomograph designed for the purposes of oncological, neurological and cardiac imaging and aid in diagnosis. With a single noninvasive procedure, the Biograph produces remarkable CT and PETCT images that reveal highly-detailed anatomy and biological processes at the molecular level.</p> <p>The Biograph provides:</p> <ul style="list-style-type: none">- high performance spiral computed tomography (CT) imaging and applications.- high-resolution, high-count rate, positron emission tomography (PET) imaging of metabolic and physiologic processes.- high quality anatomic and metabolic image registration for optimal lesion detection and identification within the body.- high quality attenuation correction and scatter correction for PET imaging.
1	<p>FlowMotion Option (AWP)</p> <p>Bring FlowMotion Technology to the Biograph large bore family.</p>
1	<p>RTP Pallet</p> <p>RTP Flat pallet for Biograph mCT. The carbon fiber table top utilizes a quick release latch for easy on/off.</p> <p>Varian Exact(tm) compatible indexing for accessories.</p> <p>Includes quantity 1 two-pin locator bar.</p>
1	<p>PET Cardiac Gating Opt (AWP)</p> <p>Provides PET cardiac gated list mode acquisition, offline histogramming, and reconstruction for improved accuracy in quantitation as well as visualization of cardiac motion. Supports a maximum of 24 gate bins from the list mode PET acquisition. Requires the optional UPMM for ECG signal capture.</p>
1	<p>PET Dynamic Option (AWP)</p> <p>Support for list mode acquisition, offline histogramming and reconstruction. Support for retrospective histogramming in any arbitrary frame durations of 3 second or greater, maximum of 100 frames defined by available disk space. Whole body (multi-bed) dynamic support of up to 25 passes. Dynamic Speed feature supports online processing capabilities for list mode imaging allowing reconstruction of dynamic frames from list mode data while acquisition is ongoing.</p>
1	<p>ECG module (UPMM) - Vision</p> <p>Universal Physiological Monitoring Module (UPMM) provides patient cardiac ECG information</p>

Qty	Item Description
1	patient connection. Includes patient cable.
1	<p data-bbox="397 411 747 441">PET/CT Resp Option Vision</p> <p data-bbox="397 449 1292 520">Provides both CT Respiratory and Triggering option as well as PET respiratory gated acquisition/reconstruction. Allows for the ability to automatically match gate definition between CT and PET during reconstruction for phase match attenuation correction and visualization.</p>
1	<p data-bbox="397 537 641 567">OncoFreeze (AWP)</p> <p data-bbox="397 575 1292 667">Adaptive respiratory gating for automated optimal, motion-freeze, designed to provide improved image quality by reducing respiratory motion artifacts while providing optimized count statistics. Mass preservation optical flow algorithm designed to utilize 100% of the counts.</p>
1	<p data-bbox="397 684 747 714">Anzai Respiratory Interface</p> <p data-bbox="397 722 1128 756">Configuration for connecting Anzai respiratory trigger system to the Biograph.</p>
1	<p data-bbox="397 772 738 802">ultraHD-PET Option (AWP)</p> <p data-bbox="397 810 1292 903">Utilizing timing information (time-of-flight) between the two PET coincidence events, coupled with resolution recovery of HDPET, ultraHDPET option provides improved image signal-to-noise which can be used to either enhance image quality and/or reduce patient acquisition time. The Biograph ultraHDPET option takes PET imaging to the pinnacle of performance.</p>
1	<p data-bbox="397 919 625 949">CT SAFIRE (AWP)</p> <p data-bbox="397 957 1292 1029">The Sinogram Affirmed Iterative Reconstruction (SAFIRE) enhances spatial resolution, reduces image noise and increases sharpness by introducing multiple iteration steps in the reconstruction process.</p>
1	<p data-bbox="397 1045 560 1075">iMAR (AWP)</p> <p data-bbox="397 1083 1292 1184">The iMAR metal artifact reduction algorithm combines three successful approaches (beam hardening correction, normalized sinogram inpainting and frequency split). This allows to reduce metal artifacts caused by metal implants such as coils, metal screws and plates, dental fillings or implants.</p> <p data-bbox="397 1222 1292 1268">iMAR is compatible with extended FoV, the extended CT scale as well as the newest dose reduction feature.</p> <p data-bbox="397 1306 1292 1381">Along with the new algorithm comes the simple user interface of iMAR enabling easy reconstruction of clinical images with reduced metal artifacts.</p>
1	<p data-bbox="397 1398 690 1428">PET Gantry/MARS UPS</p> <p data-bbox="397 1436 1292 1507">Uninterruptible Power Supply (UPS) option providing 5 minutes of backup power to the PET gantry and PET acquisition/reconstruction computer, enabling proper shutdown of the PET system in the event of power loss. Specifications: 6.0 KVA, 230 Volts, 50/60 Hz.</p>
1	<p data-bbox="397 1524 633 1554">Keyboard, English</p> <p data-bbox="397 1562 820 1596">Keyboard in the above-mentioned language.</p>
1	<p data-bbox="397 1612 860 1642">Biograph On-site NEMA 2018 testing</p> <p data-bbox="397 1650 1096 1684">Class standard testing NU2-2018 tests are offered as follows for PET/CT.</p>
1	<p data-bbox="397 1701 803 1730">Biograph Ge-68 Sources - Large</p> <p data-bbox="397 1738 1292 1780">Calibration sources for the Biograph PET/CT. These sources are to be purchased with a new Biograph PET/CT scanner.</p>
1	<p data-bbox="397 1797 779 1827">Uniform Source Shield - Large</p> <p data-bbox="397 1835 1063 1869">Contains shield for the large Uniform Source for the Biograph PET/CT.</p>

Qty	Item Description
1	<p>Seismic Install Kit w/PDU - Vision</p> <p>Includes installation kit items and additional items for seismic region installation. Some items still require local specification, due to site specific requirements. Includes power distribution unit for connecting entire system to a single 3-phase power drop.</p>
1	<p>Installation US</p>
1	<p>Cooling System Water/Air</p> <p>Water-to-air heat exchanger for the dissipation of heat loss generated in the gantry to the outside air. System operating temperature:20 - 26 degrees C, 20 - 75 % rel. humidity (not condensing). Ideal for installation far from the scan room. Cooling system contains to units, water/water exchanger close to the scan room and an additional remote water/air exchanger. Maximum distance between water/water unit and remote water/air exchanger up to 40 meters enabled by thin diameter of water transferring pipes.</p>
1	<p>Cooling System Install Kit</p> <p>Kit for installation of the Cooling System Water/Air in US</p> <p>Includes:</p> <ul style="list-style-type: none"> - Transformer for powering the Cooling System Water/Air - Service switch to shut off the outdoor cooling unit for maintenance or in case of emergency
1	<p>teamplay Basic</p> <p>Healthcare professionals, come together in teamplay's rich Digital Marketplace to access both the metrics from their own imaging fleet and a vast shared pool of imaging data. As a community, you connect and collaborate in a secure environment with high data privacy and security standards. teamplay BASIC applications include Dose, Usage and Image data management functionalities free-of-charge.</p> <p>Its easy onboarding allows you to register on teamplay's digital platform and download the needed software to set it up in your institution: Click > Try for free at www.siemens.com/teamplay</p>
1	<p>Additional System Manuals</p> <p>Additional user manual for the above selected MI system.</p>
1	<p>Initial onsite training 32 hrs</p> <p>Up to (32) hours of on-site clinical education training, scheduled consecutively (Monday - Friday) during standard business hours for a maximum of (4) imaging professionals. Training will cover agenda items on the ASRT approved checklist. Uptime Clinical Education phone support is provided during the warranty period for specified posted hours. This educational offering must be completed (12) months from install end date. If training is not completed within the applicable time period, Siemens obligation to provide the training will expire without refund.</p>
1	<p>CT Cross Trainer</p> <p>The CT CrossTrainer Program is a printed self-study course provided and administered by MIC and is intended for (1) imaging professional. StudyModules(tm) present CT in an easy to learn, interactive format with a comprehensive scope including physical principles, hardware and software, image parameters selection, image formation, safety, contrast administration, artifact prevention, patient handling, routine and advanced imaging techniques, and an abundance of images and illustrations.</p> <ul style="list-style-type: none"> - Target audience: Any technologist seeking to learn CT imaging. Prior CT experience is not required. - Specs: 6 StudyModules, 15 hours of study time, 17 Cat A CE credits - Time limited: A 6 month term of enrollment provides an opportunity to earn the associated CE credits starting on the date the course materials are shipped. Upon expiration of enrollment, unearned credits are forfeited without refund.
1	<p>Follow-up training 32 hrs</p> <p>Up to (32) hours of follow-up on-site clinical education training, scheduled consecutively (Monday - Friday) during standard business hours for a maximum of (4) imaging</p>

Qty	Item Description
	professionals. Uptime Clinical Education phone support is provided during the warranty period for specified posted hours. This educational offering must be completed (12) months from install end date. If training is not completed within the applicable time period, Siemens obligation to provide the training will expire without refund.
1	<p data-bbox="397 493 868 520">GOV'T ONLY - MI PET Training Class</p> <p data-bbox="397 525 1294 619">Tuition for (1) government attendee to attend a classroom course of choice at one of the Siemens training centers. This educational offering must be completed (12) months from install end date. If training is not completed within the applicable time period, Siemens obligation to provide the training will expire without refund.</p>
1	<p data-bbox="397 646 885 674">Additional CT onsite training 32 hours</p> <p data-bbox="397 678 1294 825">MI_CTPET_ADD_32 Up to (32) hours of on-site clinical education training, scheduled consecutively (Monday - Friday) during standard business hours for a maximum of (4) imaging professionals. Training will cover agenda items on the ASRT approved checklist if applicable. This educational offering must be completed (12) months from install end date. If training is not completed within the applicable time period, Siemens obligation to provide the training will expire without refund.</p>
1	<p data-bbox="397 852 868 879">GOV'T ONLY - MI PET Training Class</p> <p data-bbox="397 884 1294 978">Tuition for (1) government attendee to attend a classroom course of choice at one of the Siemens training centers. This educational offering must be completed (12) months from install end date. If training is not completed within the applicable time period, Siemens obligation to provide the training will expire without refund.</p>
1	<p data-bbox="397 1005 755 1033">MI PET Project Management</p> <p data-bbox="397 1037 1294 1184">A Siemens Project Manager (PM) will be the single point of contact for the implementation of your Siemen's equipment. The assigned PM will work with the customer's facilities management, architect or building contractor to assist you in ensuring that your site is ready for installation. Your PM will provide initial and final drawings and will coordinate the scheduling of the equipment, installation, and rigging, as well as the initiation of on-site clinical education.</p>
1	<p data-bbox="397 1211 836 1239">PET Gating Onsite Training 16 Hrs</p> <p data-bbox="397 1243 1294 1390">Up to (16) hours of on-site gating clinical education training, scheduled consecutively (Monday - Friday) during standard business hours for a maximum of (4) imaging professionals. Training will cover agenda items on the ASRT approved checklist if applicable. This educational offering must be completed (12) months from install end date. If training is not completed within the applicable time period, Siemens obligation to provide the training will expire without refund.</p>
1	<p data-bbox="397 1417 836 1444">Additonal onsite training 32 hours</p> <p data-bbox="397 1449 1294 1596">Up to (32) hours of on-site clinical education training, scheduled consecutively (Monday - Friday) during standard business hours for a maximum of (4) imaging professionals. Training will cover agenda items on the ASRT approved checklist if applicable. This educational offering must be completed (12) months from install end date. If training is not completed within the applicable time period, Siemens obligation to provide the training will expire without refund.</p>
1	<p data-bbox="397 1623 868 1650">Project Mgmt/Site Planning (US only)</p>
1	<p data-bbox="397 1677 868 1705">MI PET Riedel Chiller Start-up by SBT</p>
1	<p data-bbox="397 1738 690 1766">NEMA_XR-29 Standard</p> <p data-bbox="397 1770 1294 1791">This system is in compliance with NEMA XR-29 Standard Attributes on CT Equipment Related To Dose Optimization and Management, also know as Smart Dose</p>
1	<p data-bbox="397 1831 909 1858">Anzai Respiratory Gating (VI) - no pedes</p>
1	<p data-bbox="397 1887 852 1915">Low Contrast CT Phantom & Holder</p>
1	<p data-bbox="397 1950 876 1978">Stellant D Dual Ceiling w/Certegra WS</p> <p data-bbox="397 1982 1294 2003">Stellant D Dual Ceiling mounted with Certegra Workstation NO Informatics. Short ceiling post</p>

Qty	Part No.	Item Description
		- 580 mm.
		Other ceiling post lengths are available (different part numbers): 850 mm and 1000 mm.
		Includes Stellant D, Dual Head, ceiling mounted injector; Certegra workstation; installation and warranty through Medrad.

1		<p>Biograph PET Gantry Trinity Complimentary Biomed Training</p> <p>This educational offering includes system training tuition for 1 clinical engineering professional on the PET Biograph mCT (Definition Based), and the syngo multimodality workstation as applicable. The training curriculum depends on and is limited to the system purchased and may include multiple courses including classroom training in USA or an international site, and/or virtual and web-based training. Additional modality basics training may be required as a prerequisite to these courses and must be purchased separately. This system training includes a 15% discount. Travel and lodging are not included. This educational offering must be completed by the later of (12) months from purchase or install end date; if training is not completed within the applicable time period, Siemens obligation to provide the training will expire without refund. This forfeiture does not apply to Federal government agencies.</p>
---	--	---

Offset Part 14422287 Additional System Manuals

Offset MIP_MCTD_BIOMD Complimentary Biomed Training

Initial onsite training 32 hrsGov Offset

Detailed Technical Specifications

Description

Biograph Vision 600 consists of the following.

Scanning Unit (Integrated PET•CT Gantry):

The fully integrated PET•CT gantry incorporates CT and PET detector assemblies and electronics in an efficient, compact design. The large gantry opening, continuous patient port and short tunnel length provide ease of positioning for up to 500 lb (227 kg) patients and help to minimize patient claustrophobia. Quad operator controls are featured on the gantry for positioning from either side of patient and from either the front or rear. Dual gantry displays (front and rear) provide system status.

CT System:

The CT imaging capability of the Biograph Vision 600 consists of a 64-slice CT featuring a full range of spiral CT clinical applications with high performance.

Gantry:

Aperture: 78 cm; power supplied via low-voltage slipring.

Rotational speed of the gantry: 182 rpm with a rotation time of 330 ms.

Scanning system:

Adaptive Array Detector (AAD) system based on UFC™ (ultrafast ceramics) with up to 23,552 elements depending on configuration, and 1472 measuring channels per slice (the measuring system can contain replacement components).

STRATON tube high-performance X-ray system:

The STRATON tube provides direct oil cooling of the anode with the ball bearings located outside the vacuum. The direct anode cooling and the small and compact design of the anode eliminates the need for heat storage capacity (equivalent of 50 MHU) and enables an unprecedented cooling rate of 7.3 MHU/min. Therefore cooling delays between multiple long range scans are eliminated, even for large patients. Tube current range: 20-800 mA. Focal spot size according to IEC 60336: 0.7 x 0.7mm/7°, 0.9 x 1.1mm/7°. Computer controlled monitoring of anode temperature, multifan principle with flying focal spot.

Z-Sharp technology:

The unique STRATON X-ray tube utilizes an electron beam that is accurately and rapidly deflected, creating two precise focal spots alternating 4,608 times per second. This doubles the X-ray projections reaching each detector element. The two overlapping projections result in an oversampling in the z-direction, known as Double Z-Sampling. The resulting measurements interleave half a detector slice width, doubling the scan information without a corresponding increase in dose. Siemens' proprietary, high-speed Ultra Fast Ceramic (UFC) detector enables a virtually simultaneous readout of two projections for each detector element - 2 x 32 slices for every viewing angle - resulting in a full 64-slice acquisition.

80 kW X-ray generator:

Microprocessor-controlled, low-noise high-frequency generator with integrated, automatic self-testing system for continuous monitoring of operation. Settings: High-voltage range 70, 80, 100, 120 and 140 kV; power max. 80 kW, adjustable in fine steps (100 kW optional).

PET System:

The PET imaging capability of the Biograph Vision 600 is supported by a multi-ring LSO/SiPM (lutetium oxyorthosilicate silicon photomultiplier) PET detector system with 3D acquisition and reconstruction, featuring 159 image planes comprising a 263 mm axial field of view.

The Vision detection system provides:

Description

- High spatial slice resolution in trans-axial and axial dimensions.
- Slice spacing (1.65 mm) optimized for speed and resolution.
- LSO crystal for fast response and high light output.
- High-performance SiPM-based detectors.
- High-bandwidth electronics for low deadtime and high signal-to-noise.
- MARS combined acquisition and reconstruction computer system for high count rate capability and fast reconstruction of PET data.
- Three-dimensional display of organs with a large axial view.
- Excellent volume sensitivity.
- Fast acquisition and reconstruction at any available PET matrix size.
- Simultaneous data acquisition and image reconstruction for high patient throughput.
- Simultaneous reconstruction of two images in parallel (e.g. AC and NAC)
- Static and whole body acquisition capability.
- 820 mm detector ring diameter.
- 78 cm gantry aperture.
- 70 cm transverse field of view.
- Dedicated PET/CT Patient Handling System.
- TrueC advanced scatter correction technique

Patient Handling System:

The Biograph Vision 600 patient handling system (PHS) has a unique reinforced cantilever design that provides reliable patient support with high weight capacity and minimal pallet deflection. As one of the pillars of SMART (Siemens Molecular & Anatomical Registration Technologies), the PHS provides:

- Reinforced cantilever design for maximum patient support and absolute positioning between PET and CT scan.
- Integrated patient table design for easy patient positioning.
- Low attenuation carbon fiber pallet.
- 43 cm vertical motion range.
- Maximum 2 m PET•CT co-scan range.
- Low attenuation head holder, table extensions, head-arm support, knee-leg support.
- Maximum patient weight of 227 kg (500 lbs.).

Control and evaluation unit:

CT control box with intercom system with user-programmable patient instruction system. Dual monitors (19 inch (48 cm) LCD flat panel displays), keyboard and mouse for *syngo* Acquisition Workplace.

Computer system:

The computer system of the Biograph Vision 600 consists of three components.

- *syngo* Acquisition Workplace console for the planning and execution of the PET/CT examination, including evaluation and management of the PET/CT images.
- Reconstruction computer for the preprocessing and reconstruction of the CT data.
- PET acquisition and reconstruction system (MARS) with supported image reconstruction of 128x128, 220x220, 256x256, 440x440, 512x512, and 880x880.

The *syngo* Acquisition Workplace console consists of a high performance Windows 7 based computer with 8 core Xeon processor, 16 GB RAM, 300 GB storage capacity, and DVD DICOM with 4.7 GB storage. External USB 2.0/3.0 devices for data storage are supported.

Description

The CT reconstruction computer contains a cluster of high performance processors performing the preprocessing and reconstruction of the CT data at up to 40 images/sec (512x512). Raw data memory is 900 GB.

The PET acquisition and reconstruction system (MARS) provides high performance 3D acquisition and reconstruction. It contains dual Xeon 10 core processors with a total of 256 GB RAM. Disk storage of 3 TB for PET raw data and images is provided.

syngo User Software:

syngo features an intuitive and easy-to-learn user interface. *syngo* visualizes the examination in individual process steps on so-called *task cards*, such as patient registration or examination card. A large number of functions and input parameters as well as the language used can be selected according to individual requirements. Frequently repeated processes can be automated and saved.

Patient registration - The system can accept patient data in different ways. These include entering the data via keyboard or transfer of a worklist via network.

DICOM Worklist - Software module for accepting lists of patient data and exam requirements from a Radiology Information Systems (RIS) via DICOM "Get Worklist" functionality. The module enables efficient workflow and helps ensure consistent patient data.

Examination card - The scanner is supplied with a large number of predefined CT and fully integrated PET•CT examination protocols, making examination planning a fast and efficient procedure.

Viewing card - On the viewing card it is possible to move interactively with the mouse through the image volume of the ongoing examination. The images of different examinations can be displayed in parallel for comparison. A large number of functions are available for evaluation, documentation and archiving.

Filming card - A virtual film sheet shows a 1:1 display of the film sheets to be printed out, thus permitting an effective preview of the filming job and re-windowing the images, as well as providing a large number of evaluation functions. Layout changes are possible interactively with up to 64 images. The printout parameters for the ongoing auto-filming running parallel to acquisition or reconstruction are also defined with the filming card.

3D card - The 3D task card contains the User Interface for the operation of the MIP (Maximum Intensity Projection), SSD (Surface Shaded Display), and MPR (Multi-planar Reconstruction) three-dimensional post-processing. The 3D card also features an intuitive and fast bone removal function for CTA post processing and presentation.

3D VRT - Advanced 3D functionality as an extension to the basic 3D viewer, containing volume rendering technique (VRT) and advanced editing functions. Advanced 3D application package for the optimal display and differentiation of different organs through independent control of color, opacity, and shading in up to 4 tissue classes.

CT Angio - Software for the reconstruction of angular projections from the images of a spiral data record for the display and diagnosis, e.g. of aneurysms, plaques, stenoses, vascular anomalies or vascular origins. MIP: Maximum Intensity Projection, MinIP: Minimum Intensity Projection, and Thin MIP are available. Interfering or irrelevant parts of the image can be eliminated with the integrated volume editor. The angular projections are reconstructed around a definable axis, whereby the maximum CT values in this direction are selected for each angular projection. The resulting images can be viewed with the CINE function as a series of images with a 3D image effect.

Workstream - Planning and reconstruction of diagnostic CT coronal, sagittal, oblique and MIP images can take place directly after scanning.

DynEva card - Software for dynamic evaluation of the contrast enhancement in organs and types of tissues, enabling the reconstruction of

- Time-density curves (up to 5 ROIs)
- Peak-enhancement images
- Time-to-peak images.

Video Capture and Editing Tool - Integrated solution for imaging and visualization of 4D information, allowing the generation and editing of video files for improved diagnoses, recording and teaching. A wide range of multimedia formats is supported, e.g. AVI, Flash (SWF), GIF, QuickTime (MOV), streaming video.

HD FoV Pro - Extended Field of View option which allows visualization of objects with a CT FOV up to 78 cm, and improved CT image quality beyond the traditional 50 cm CT FOV for improved PET attenuation correction. HD FoV

Description

Pro can show reconstructed CT attenuation values outside the conventional FoV with an accuracy of +/- 50 HU within a 65 cm FoV.

TrueD Basic - Single-mode, single timepoint layout for displaying the PET and CT either fused or side-by-side comparison with viewer formats and color map tables. Support for 3D spherical regions-of-interest with units of Bq/ml or Standard Uptake Value (SUV). Allows re-registration of PET to CT data for correction of misregistration as a result of patient motion.

Media Viewer - Provides basic viewing capabilities in a portable Windows-based application that can be burned to media (CD, DVD) along with patient images. Not intended for diagnostic use.

- Review volume datasets from CT and PET
- Supports viewing single-modality or fused images
- View linked axial, coronal, and sagittal views
- Navigate in three dimensions
- View MIP images correlated to axial, coronal, and sagittal views
- Blend fused images
- Quantify Hounsfield units, SUV

CARE Solutions:

UFC Detector - Up to 30% dose reduction compared to conventional CT detectors. High efficiency for low mAs requirements enable best possible image quality with low patient dose.

CARE Filter - Specially designed X-ray exposure filter installed at the tube collimator. Up to 25% dose reduction with increased image quality.

CARE kV - Automated, organ-sensitive voltage setting to optimize contrast-to-noise-ratio and reduce dose by up to 60%.

CARE Profile - Visualization of the dose distribution along the topogram prior to the scan.

CARE Dashboard - Visualization of activated dose reduction features and technologies for each scan range of an examination.

CARE Child - Pediatric protocols, special examination protocols with 70 or 80 kV and a large range of adjustable mAs values for optimum adaptation of the radiation exposure to the age and weight of the child to be examined.

CARE Topo - Real-time topogram, Manual interruption possible once desired anatomy has been imaged.

CARE Bolus - Operating mode for CM-enhancement triggered data acquisition. The objective is optimum utilization of the contrast medium bolus in its "plateau" phase in the target organ. This option has been especially adapted to the increased speed and timing requirements resulting from the multirow capability and faster rotation. The CM enhancement is observed via monitoring scans in a user-defined ROI with a trigger threshold. As soon as the enhancement reaches its predefined threshold, the spiral scan is triggered as quickly as possible. License for software use on one modality.

CARE Dose4D - This software feature provides automatic, real-time x-ray dose management for all scan modes. The minimal x-ray dose level needed to obtain optimal image quality is determined from extensive computer analysis of the topogram image and also from the data collected during every slice scanned, on a real time basis. This automatic approach ensures optimal image quality at the lowest possible x-ray dose. CARE Dose4D uses at first a automated adjustment of the dose level depending on patient size based on the attenuation values obtained from the standard topogram along the patient axis. In addition CARE Dose4D uses a real-time adaptation of the tube current during the scan based on the actual attenuation of the X-ray beam measured around the patient. Up to 2,320 projections are evaluated per second to optimize the mA level instantaneously. In combination with the extreme adjustment speed of the tube current, CARE Dose4D ensures consistent high quality images in every anatomical position. And that's at anytime with the minimal possible X-ray dose.

Several clinical benefits are achieved with CARE Dose4D:

- Significant x-ray dose reduction (up to 68 %) possible for all body regions scanned compared with standard sequence or spiral scanning;
- Consistent, optimal image quality with the x-ray dose level unique for every patient and for every anatomical

Description

region;

- Thinner axial slices and/or longer scan ranges possible because of reduced tube loading;

- Ultra-low dose examinations for pediatric patients.

CARE Dose Configurator - Enhancement of Siemens' renowned realtime dose modulation CARE Dose4D, introducing new reference curves for each body region and for each body habitus allowing to adjust the configuration even more precisely to the patient's anatomy.

Dose Notification - The Biograph Vision provides the ability to set dose reference values (CTDIvol, DLP) for each scan range. If these reference values are exceeded the Dose Notification window informs the user.

Dose Alert - The Biograph Vision automatically adds up CTDIvol and DLP depending on z-position (scan axis). The Dose Alert window appears, if either of these cumulative values exceeds a user-defined threshold.

Adaptive Dose Shield - eliminates clinically irrelevant radiation in every spiral scan, adding to the lowest possible dose that CARE Solutions provide.

Ultra-low CT dose protocols for PET attenuation correction - non-diagnostic CT imaging protocols enabling sub 1 mSv CT acquisitions for PET attenuation correction

FAST CARE:

With Siemens' unique FAST CARE platform, the Biograph Vision is set to raise the standard of patient-centric productivity. Utilizing FAST – Fully Assisting Scanner Technologies, typically time-consuming and complex procedures during the scan process are extremely simplified and automated, not only improving workflow efficiency, but optimizing the overall clinical outcome by facilitating reproducible results, making diagnosis more reliable and reducing patient burden through streamlined examinations.

FAST Scan Assistant - An intuitive user interface for solving conflicts by changing the scan time, resp. the pitch and/or the maximum tube current manually.

FAST Adjust - assists the user to handle system settings in a fast and easy way by automatically solving of conflicts within user defined limits by one single click on the FAST Adjust button. The limits for scan time and tube current per scan are defined via the Scan Protocol Assistant. FAST Adjust offers an undo functionality to return to previously set values.

Examination and Evaluation Functions:

Topogram - Scanning perspectives: a.p., p.a., lat.; length of scan field: 128 - 2200 mm, width of scan field: 512 mm, 1.5 - 20s. The topogram can be switched off manually when the desired examination length is reached.

Tomogram - Scan field size: 50 cm. Standard scan times: 0.33, 0.5, and 1 seconds. Slice thickness in sequence: 0.6, 0.75, 1, 1.2, 1.5, 2.0, 2.4, 3, 4.0, 4.8, 5, 6, 7, 7.2, 8, 10, 12, 14.4, and 15 mm. Slice thickness in spiral: 0.4**, 0.5**, 0.6, 0.75, 1.0, 1.5, 2, 3, 4, 5, 6, 7, 8, and 10 mm (**optional with z-UHR).

Real-time image display - Immediate image reconstruction and display without time delay simultaneously to data acquisition in 512x512 matrix size.

Spiral - Scanning technique for continuous volume scans with continuous table feed in multirotation mode. Max. scan time 100 seconds and up to 2M scan length with full low-contrast resolution. Selection of the pitch factor between 0.35 and 1.5 depending on scan mode. Selection of up to 33 separately parameterizable examination ranges in a patient protocol. In addition individual anatomic sections can be successively combined and then scanned automatically. Storage of up to 10,000 examination protocols. Rotation times/cycle: 0.33 sec, 0.5 sec, and 1 sec.

Dynamic - Program for functional dynamic examinations. Serial scanning technique in one slice position with variable scan cycle times.

Serio-sequential examination without table feed - Up to 100 scans in uninterrupted, continuous sequence without table feed. Scan cycle time: 0.75 - 60 seconds.

Multiscan spiral examination without table feed – Continuous multirotational data acquisition in one slice position. Quantitative evaluation and graphical display of time-density curves.

WorkStream4D with Asynchronous Recon - 4D workflow with direct generation of axial, sagittal, coronal, or double-oblique images from standard scanning protocols. Elimination of manual reconstruction steps.

Asynchronous Recon allows for multiple image reconstructions and reformats, parallel to scanning. With this

Description

feature, up to eight reconstruction job requests can be loaded into a scan protocol. Immediately upon completion of the scan acquisition, these reconstruction jobs are automatically executed in the background without delaying the start of next patient examination.

CT Image reconstruction and storage - Image reconstruction in full resolution (512x512 matrix) takes place during the examination with up to 40 images per second, with full cone beam reconstruction, z-Sharp Technology and full image quality. Reconstruction fields of 5 cm to 50 cm through raw data zoom with the possibility of freely selecting the image center either prospectively before each scan or retrospectively. Reconstructions of different slice thicknesses from a single raw data record, e.g. lung soft tissue and lung high-contrast with CombiScan, with simultaneous suppression of partial volume artifacts. Up to 8 reconstructions per scan range can be predefined with the examination protocol. Patient-related storage of the image and raw data.

Image display - 1024x1024 display matrix; screen splitting configurable up to 64 image segments; CT value scale from -1024 to +3071 HU. For very dense objects, the CT value scale can be extended from -10240 to +30710 HU (extended CT scale) e.g. for suppressing metal artifacts.

Image evaluation - Complete software-controlled image evaluation program for all diagnostic requirements.

CINE Display - Dynamic display technique for the visualization of time or volume series. A series of up to 1024 images can be displayed at a frame rate of at least 30 f/s. Automatic or interactive mouse-operated control.

Multitasking functions - Simultaneous processing during operation of the scanner.

Real-time Display - Image reconstruction in pace with the examination in full image quality (512x512 matrix) with up to 40 images/second (with full cone beam reconstruction and z-Sharp Technology).

Metro Display - Simultaneous display, processing and evaluation of images from other patients while the current patient is being scanned.

Metro Documentation - Simultaneous documentation of images from any previously examined patient while the current patient is being scanned.

Metro Copy - Automatic transfer of image data to the syngo CT Workplace (optional) or a DICOM network node.

Networking and Documentation:

For the connection to a local Ethernet (10, 100 Mbit or 1-Gigabit) in order to communicate with networked printers, diagnostic and therapy workstations, RIS or HIS systems and teleradiology routers.

Scope of functions:

- Configurable network stations.

- Unlimited selection of stations.

- DICOM Standard (Digital Imaging and Communications in Medicine) for the transfer of information between DICOM-compatible units from different manufacturers. The scope of functions is described in detail in the DICOM Conformance Statement, and the standard version comprises the functions Send/Receive, Query/Retrieve and BasicPrint, Worklist, Storage Commitment, and MPPS (Modality Performed Procedure Step).

System Documentation (1 set)

Smart Remote Services:

Smart Remote Services (SRS) offers a wide range of medical equipment-related remote services resulting in increased system availability and efficiency. SRS employs sophisticated authentication and authorization procedures, state-of-the-art encryption technologies and logging routines together with strictly enforced organizational measures that provide optimal patient data security and access protection. The following SRS services are included for all service agreement customers and during warranty period:

Remote Diagnosis & Repair: In case of an unforeseen system malfunction, Siemens competent experts may directly connect with the PET/CT system in order to identify the problem quickly. Moreover the remote repair function enables Siemens to often correct software errors immediately. Should an engineer on site be required, Remote Diagnosis & Repair allows Siemens to identify defective parts efficiently and accelerate their delivery, thereby keeping repair times to a minimum.

Event Monitoring: Event Monitoring screens the performance of the system. If a parameter deviates from a predefined value, a status message is automatically sent to the Siemens UPTIME Service Center. Service Engineers may evaluate the status message at periodic intervals and may initiate appropriate action within the scope of the service agreement.

Description

FlowMotion Technology provides:

Improved planning workflow, acquisition and reconstruction with continuous table movement PET acquisition, a simple CT-like planning workflow and high resolution optimized image processing. Precise CT-like planning for PET provides flexibility for scan area allowing reduced radiation dose. Simple customization of protocols provides the ability to incorporate features for motion management, varied speed for focused image quality and improved throughput, and maximum reconstructed image resolution for the most demanding clinical and research applications. Additional benefits include:

- PET acquisition without the limitation of bed positions
- Ability to customize scan protocols to the specific physician preference for each patient
- Enables easy setup for up to 4 scan regions for optimized scan parameters in each body area
- Designed to improve axial noise uniformity, end-plane sensitivity, and increase reproducibility
- Designed for CT dose reduction for ALARA dose
- One-click CT-like range planning to scan the precise area needed
- Improved patient satisfaction due to a sense of progress being made with constant bed motion during acquisition

The CT Respiratory Gating and Triggering option is comprised of software components that allow for the capture and storage of a signal representing a patient's respiratory cycle during a spiral or sequence CT acquisition. With the Respiratory Gating feature, the respiratory data is synchronized with the CT acquisition data so that a user can freely select the point at which images are retrospectively reconstructed based on the corresponding respiration amplitude. With the Respiration Triggering feature, the user prospectively selects a point in the respiratory cycle at which sequence images will be acquired.

Through the selection and reconstruction processes, organ motion artifacts caused by respiration are minimized or eliminated and a better visualization and localization is possible resulting in more accurate assessment of tumor and organ motion, their position, size, and volume during respiration.

These applications generate 4D CT datasets that can be used to create more accurate treatment plans and also for the delivery of respiratory-triggered radiation therapy.

Provides PET respiratory gated list mode acquisition, offline histogramming, and reconstruction for improved accuracy in quantitation as well as visualization of organ motion. Supports a maximum of 24 gate bins from the list mode PET acquisition.

Includes single-click gate definition matching between PET and CT or vice versa during retrospective processing.

Requires the optional Respiratory Trigger System.

HD•PET Package provides improved PET image quality compared to conventionally reconstructed images. HD•PET Package contains TrueX, an innovative image processing technique and 3D iterative reconstruction.

TrueX is an innovative image processing technology that is the key to achieving HD•PET performance levels. Conventional PET does not take into account the detector geometry and incorrect positioning of the LORs. HD•PET incorporates measured point spread functions (PSF) into the iterative reconstruction algorithm. Through modeling of the PSF, HD•PET more precisely accounts for the positioning of the LOR yielding visually sharper clinical images, visual improvements in contrast and in resolution.

3D Iterative reconstruction (OSEM) provides improved image quality in the most demanding low statistics acquisitions.

Dose reduction with CT has been limited by the currently used filtered back projection (FBP) reconstruction algorithm. When using this conventional reconstruction of acquired raw data into image data, a trade-off between spatial resolution and image noise has to be considered. Higher spatial resolution increases the ability to see the smallest detail; however, it is directly correlated with increased image noise in standard filtered back projection reconstructions as they are used in CT scanners today.

Description

Iterative reconstruction approaches allow decoupling of spatial resolution and image noise. With the Sinogram Affirmed Iterative Reconstruction (SAFIRE), correction loops are introduced into the image generation process. These iteration loops utilize raw-data information to significantly improve image quality. Additionally, image noise is reduced in the iterative corrections without degrading image sharpness. The noise texture of the images is comparable to standard well-established convolution kernels. This technique can improve image quality and be translated to dose savings for a wide range of clinical applications.

The iMAR CT metal artifact reduction algorithm combines three successful approaches to reduce metal artifacts: beam hardening correction (in sinogram regions of less severe metal attenuation), normalized sinogram inpainting (in sinogram regions of high metal attenuation), and frequency split (to mix back noise texture and sharp details that are potentially lost during inpainting).

The correction process is then iteratively refined by repeating the normalized sinogram inpainting and the mixing steps thanks to the Adaptive Sinogram Mixing.

Along with the new algorithm comes the simple user interface of iMAR. Besides the typical reconstruction parameters it only requires to select the desired protocol from a drop down menu which contains the following type of implants: dental fillings, neuro coil, thoracic coil, hip implants, extremity implants, pacemakers, spine implants and shoulder implants.

Including 3D Spatial Resolution (NU 2-2018 Ch. 3); 3D scatter fraction, Count losses, and randoms measurement (NU 2-2018 Ch 4); 3D Sensitivity (NU 2-2012 Ch 5); Time-of-Flight resolution (NU2-2018 Ch 8); PET-CT coregistration accuracy (NU2-2018 Ch 9) Modifications, changes or additions to these tests subject to factory approval and quotation.

All radioisotopes (e.g. Fluorine-18, Sodium-22) to be provided by the site.

Loaner phantoms available from factory.

90 day lead time on scheduling. Estimated 1 week to complete measurements on site.

Purchase of a dedicated 22-Na source is required for the NEMA NU2 spatial resolution test for Vision. The 0.25 mm solid 22-Na point source is available in pre-calibrated activity levels (10-100 uCi). Sources should be ordered directly from the source vendor:

Eckert & Ziegler Isotope Products.
Address: Eckert & Ziegler Isotope Products
Dba Isotope Products Laboratories
24937 Avenue Tibbitts
Valencia, CA 91355
Phone: +1 661 309-1082

Vendor Part # MMS05-022

Description: Multimodal spot marker, Na-22 nominal, 10 microCuries

Sources consist of the following:

2 LS-ART Set-up rod sources (Max. 1.25 mCi (46.25 MBq) per rod source)
Uniform QC Phantom, Ge-68, specific for the Biograph Vision. Max 3.0 mCi (111 MBq)

Disposal of sources is not included in sale price.

teampay BASIC applications provide a locally installed teampay Receiver software with web-based data analytics capabilities in the area of Dose, Usage and Image data management.

teampay Dose BASIC application and teampay Usage BASIC application are restricted to datasets produced by Siemens modalities.

Its easy onboarding allows you to register on teampay's digital platform and download the needed software to set it up in your institution: Click > Try for free at www.siemens.com/teampay

teampay Receiver software

teampay Receiver software is a DICOM gateway that is installed on the local institution network to manage

Description

communication of data between hospital systems and teamplay servers and manage conformity with local data privacy regulations.

Functionalities of the teamplay Receiver software:

- DICOM fetch and receive (Query/Retrieve and C-STORE)
- Choice of three data privacy levels
- Automatic software updates

The teamplay Receiver software can be installed on hardware or virtual machines provided by the customer meeting the minimum requirements listed below.

- Windows 64-bit server or client operating system
(at least Windows server 2012 R2 or Windows server 2016 is recommended for a production environment although Windows 10 may be used):
min. dual core CPU system
min. 4 GB memory
at least 200 GB free disk space (< 500 GB recommended)
- Recommendation for small sites (up to 50.000 procedures per year):
at least 4 core system
at least 8 GB memory
- Recommendation for medium sites (between 50.000 and 250.000 procedures per year):
at least 6 core system
at least 8 GB memory
- Recommendation for large sites (more than 250.000 procedures per year):
at least 8 core system
at least 16 GB memory
- At least 6 Mbit/s upload bandwidth to the Internet

The minimum hardware requirements are applicable when fetching data for teamplay Dose and Usage using DICOM Q/R. For scenarios where data will be actively sent to teamplay using DICOM C-Store the hardware requirements depend on the actual load and needs to be adjusted accordingly. Please contact the Siemens teamplay support for help

teamplay Dose:

teamplay Dose provides easy access to dose data to support the quality assurance process for monitoring imaging radiation doses across the fleet of scanners in the institution.

teamplay Usage:

teamplay Usage provides an intuitive way to display an overview of the fleet utilization of all institutional diagnostic imaging scanners.

teamplay Images Research:

teamplay Images Research provides the functionality to receive imaging data from other teamplay members for research and education.

Caution: teamplay Images Research is not intended for clinical use.

teamplay Images:

teamplay Images provides the functionality to receive imaging data from other teamplay members.

This system is in compliance with NEMA XR-29 Standard Attributes on CT Equipment Related To Dose Optimization and Management, also known as Smart Dose