

Equipment Specifications

Ultrasound

VISN 15, Colmery O'Neil VA Medical Center

589-B92017

A. REQUIREMENT OVERVIEW

The Colmery O'Neil VA Medical Center is in the process of constructing a new Emergency Department. As part of the construction process two (2) Point of Care Ultrasound Systems are needed for specialty providers to perform more diagnostic testing at bedside.

Facility	Quantity
Topeka, Kansas Colmery O'Neil VA Medical Center	2

B. TECHNICAL REQUIREMENTS

1. Unit physical specifications

a. Minimum screen size [in]	18 inches
b. Monitor height range from floor [in]	47 in to 57 inches
c. Minimum number of transducer ports	4
d. Minimum image storage [days or TB]	60 GB
e. Minimum battery life [min]	200 (scan time)
f. Maximum equipment weight [lb]	150 lbs. (without probes)

2. Scanning modes

<input checked="" type="checkbox"/>	a. Power Doppler Imaging
<input checked="" type="checkbox"/>	b. M-Mode
<input checked="" type="checkbox"/>	c. B-Mode
<input checked="" type="checkbox"/>	d. Pulsed Wave Doppler
<input checked="" type="checkbox"/>	e. Continuous Wave Doppler
<input checked="" type="checkbox"/>	f. Color Flow Mode
<input checked="" type="checkbox"/>	g. Tissue Doppler Imaging
<input checked="" type="checkbox"/>	h. Pulsed Wave Doppler
<input checked="" type="checkbox"/>	i. Anatomical M-Mode

3. Doppler displays

<input checked="" type="checkbox"/>	a. Frequency
<input checked="" type="checkbox"/>	b. Velocity



<input checked="" type="checkbox"/>	c. Power (microvascular/directorial)
<input checked="" type="checkbox"/>	d. Duplex
<input checked="" type="checkbox"/>	e. Triplex
<input checked="" type="checkbox"/>	f. Steering for color Doppler is required to be 3° increments or less

4. Control panel specifications

<input checked="" type="checkbox"/>	a. Touchscreen monitor
<input checked="" type="checkbox"/>	b. Keyboard on touchscreen
<input checked="" type="checkbox"/>	c. Articulating support arm to allow for vertical and horizontal adjustment of the monitors for viewing from anywhere in the room

5. Additional specifications

<input checked="" type="checkbox"/>	a. Multi-modality display
-------------------------------------	---------------------------

6. Security/Connectivity requirements

<input checked="" type="checkbox"/>	a. OEM-supported operating system
<input checked="" type="checkbox"/>	b. Latest DICOM print, store, commit, and modality worklist
<input checked="" type="checkbox"/>	c. Wireless connectivity to VA network – Compatible with 802.11b/g/n and FIPS 140-2 compliant
<input checked="" type="checkbox"/>	d. Encrypted hard drive

7. Analysis packages

<input checked="" type="checkbox"/>	a. Abdominal
<input checked="" type="checkbox"/>	b. Musculoskeletal Clinical
<input checked="" type="checkbox"/>	c. Small Parts
<input checked="" type="checkbox"/>	d. Vascular – measurement and analysis of vessels
<input checked="" type="checkbox"/>	e. Thoracic
<input checked="" type="checkbox"/>	f. Transcranial
<input checked="" type="checkbox"/>	g. Nerve
<input checked="" type="checkbox"/>	h. Needle Guidance
<input checked="" type="checkbox"/>	i. Urology
<input checked="" type="checkbox"/>	j. Obstetric
<input checked="" type="checkbox"/>	k. Breast
<input checked="" type="checkbox"/>	l. Cardiovascular Clinical

Vendors must include in their offers analysis packages that meet the criteria above. Please include all other analysis packages offered by your company in the optional section on the quotes.

8. Transducers



Description	Qty	Frequency Range [MHz]
a. Wide Band Phased Array	2	2.5 – 4.5
b. Wide Band Linear with Buttons	2	7 – 12 MHz
c. Wide Band Microconvex	2	6 – 10

Vendors must include in their offers transducers that meet the criteria above. Please include all other transducers offered by your company in the optional section on the quotes.

9. Advanced features

10. Added Value

Specifications listed below are not required, but preferred. Vendors who do not include the below specifications in the submitted offer will not be docked or excluded from consideration. Specifications listed below will be evaluated based on added value.

<input checked="" type="checkbox"/>	a. Additional year(s) of warranty
<input checked="" type="checkbox"/>	b. Version/platform long-range plan

C. TRAINING REQUIREMENTS

1. Clinical Training

<input checked="" type="checkbox"/>	a. On-site clinical applications training for [10] providers total, during go-live at both campuses
<input checked="" type="checkbox"/>	b. On-site follow-up clinical applications training for [6] technologists once technologists have hands-on experience with the system
<input checked="" type="checkbox"/>	c. Technologists who complete the clinical applications training shall receive continuing education credits (CMEs).
<input checked="" type="checkbox"/>	d. Vendors shall be responsible for accommodating different personnel shifts for clinical applications training during go-live.

2. Biomedical Technician Training

Please reference the “Instructions to Offers” section 2.8.g for further information about the type of information to provide by equipment type not by specific request. Please also reference the “Instructions to Offers” section 7.3.3. for response format.

Technical training information to include detailed information about the curriculum and length of the biomedical technical training required for each equipment type.

Although the NAC will not award this training along with the equipment, it is imperative that the customer is informed that this training is available. Vendors must demonstrate that they can provide any required



off-site training, therefore off-site training should be quoted as an optional item. Off-site training will be purchased at the time of need via a modification (if the original order remains open) or via a separate order. No travel expenses for any VA employees will be included in any HTME equipment or training order.

D. SERVICE REQUIREMENTS

1. VPN/Remote Access – The vendor shall provide any and all equipment service programs, such as remote diagnostics, during the warranty period. The vendor shall provide post-warranty remote diagnostic service program as an “Add Option” with the offer. The system shall provide vendor remote diagnostics via VPN. The vendor shall either utilize the VA national site-to-site VPN or work with the Office of Cyber and Information Security and the VAMC Information Systems Security Officer to establish a client-based VPN.
2. Service and Operator Manuals – The vendor shall provide the following documentation for the proposed systems:
 - a. Two (2) copies of operator instruction manuals (one (1) electronic and one (1) physical copy)
 - b. Two (2) copies of a service manuals (one (1) electronic and one (1) physical copy)*Vendors can include the physical copy as a priced line item in their quote as applicable.
3. Minimum Warranty – The system and accessories shall be covered under the manufacturer’s warranty and shall include all parts and labor for one year following acceptance by the VAMC. This warranty must include PMs as required by the manufacturer. The manufacturer’s factory-trained field service representatives shall perform installation and maintenance during the warranty period.

Vendors are encouraged to include any offerings for service, warranty, and training that may exceed the minimum requirements, to include information on their service support structure during and after the warranty period. Vendors who do not include any added value offerings for service, warranty, and training will not be docked or excluded from consideration. However, any such offerings will be evaluated based on added value.

E. OTHER INFORMATION/DOCUMENTATION REQUESTED

Please reference the “Instructions to Offers” section 2.8a-h for further information about the type of information to provide by equipment type not by specific request. Please also reference the “Instructions to Offers” section 7.3.3. for response format.

1. Completed pre-procurement assessment form (6550 Appendix A)
2. Completed Manufacture Disclosure Statement for Medical Device Security (MDS2) form
3. Federal Information Processing Standard (FIPS) 140-2 certification
4. Product brochures
5. Technical specification sheets, to include dimensions and weight of the system
6. Typical drawings (pdf version of the CAD drawings)
7. Technical training- Biomedical: information to include detailed information about the curriculum and length of the biomedical technical training required for each equipment type.
 - Although the NAC will not award this training along with the equipment, it is imperative that the customer is informed that this training is available. Vendors must demonstrate that they can provide any required off-site training, therefore off-site training should be quoted as an optional item. Off-site training will be purchased at the time of need via a modification (if the original order remains open) or via a separate order. No travel expenses for any VA employees will be included in any HTME equipment or training order.



8. Support information to include your company's support structure during and after the warranty period
- On-line or telephonic applications support and availability (include third party coverage)
 - A listing of field service engineer locations and availability
 - A listing of part depots

F. TRADE-IN

- | | |
|--|---------------------------------------|
| <input checked="checked" type="checkbox"/> | a. VA has no trade-in units to offer. |
|--|---------------------------------------|

