

Equipment Specifications

Radiation Treatment Planning System

VISN 23/Minneapolis VA Healthcare System

618-B99013

A. REQUIREMENT OVERVIEW

Radiation treatment planning systems allow physicians, physicists, dosimetrists, and therapists to plan and optimize the delivery of radiation therapy plans used in the treatment of cancer and related diseases. These systems use image data and dosimetric data to help clinicians determine the optimum treatment parameters to match the prescribed dose and constraints. Various computer algorithms are used to model the interactions between the radiation beam and the patient's anatomy to determine the spatial distribution of the radiation dose. Different algorithms are necessary to account for the different types of radiation and computational complexity. Planning systems are available for all types of radiation treatment delivery. The Minneapolis VA Healthcare System requires a radiation treatment planning system, with the required hardware and software components, for treatment of veterans.

Facility	Quantity
Minneapolis VA Healthcare System	1

B. TECHNICAL REQUIREMENTS

1. Treatment Planning

<input checked="" type="checkbox"/>	a. Photon External Beam Planning
<input checked="" type="checkbox"/>	b. Electron External Beam Planning
<input checked="" type="checkbox"/>	c. Frame Stereotactic External Beam Planning
<input checked="" type="checkbox"/>	d. Frameless Stereotactic External Beam Planning
<input checked="" type="checkbox"/>	e. 3-D Conformal
<input checked="" type="checkbox"/>	f. 4-D Conformal
<input checked="" type="checkbox"/>	g. Step and Shoot intensity modulated radiation therapy (IMRT)
<input checked="" type="checkbox"/>	h. Dynamic intensity modulated radiation therapy (IMRT)
<input checked="" type="checkbox"/>	i. Multileaf Collimator (MLC) External Beam Planning
<input checked="" type="checkbox"/>	j. Solid Block External Beam Planning
<input checked="" type="checkbox"/>	k. Virtual Wedge
<input checked="" type="checkbox"/>	l. Image-Guided Radiation Therapy (IGRT)
<input checked="" type="checkbox"/>	m. Adaptive Therapy
<input checked="" type="checkbox"/>	n. Volumetric Arc Therapy

2. Compatible Treatment Delivery

<input checked="" type="checkbox"/>	a. Linear Accelerator – Varian TrueBeam
<input checked="" type="checkbox"/>	b. Multileaf Collimator (MLC) – 120 leaf MLC



<input checked="" type="checkbox"/>	c. Stereotactic Frames – SRS Encompass Immobilization Package from Qfix
-------------------------------------	---

3. Image Data

<input checked="" type="checkbox"/>	a. DICOM 3.0
<input checked="" type="checkbox"/>	b. DICOM RT
<input checked="" type="checkbox"/>	c. Conventional Simulator
<input checked="" type="checkbox"/>	d. CT Image Data
<input checked="" type="checkbox"/>	e. MR Image Data
<input checked="" type="checkbox"/>	f. PET Image Data
<input checked="" type="checkbox"/>	g. Meets requirements for IHE RT Profile
<input checked="" type="checkbox"/>	h. Manual Registration Image Fusion
<input checked="" type="checkbox"/>	i. Fiducial-Based Image Fusion
<input checked="" type="checkbox"/>	j. Anatomy-Based Image Fusion
<input checked="" type="checkbox"/>	k. Deformable Fusion

4. Planning Methodology and Tools

<input checked="" type="checkbox"/>	a. Module-based Software
<input checked="" type="checkbox"/>	b. Inverse Planning
<input checked="" type="checkbox"/>	c. Multiple Dosimetric Algorithms
<input checked="" type="checkbox"/>	d. Plan Resolution – 1 mm ³
<input checked="" type="checkbox"/>	e. Max Number of Beamlets - 100
<input checked="" type="checkbox"/>	f. Max Number of Beam Angles - Unlimited
<input checked="" type="checkbox"/>	g. Template Library
<input checked="" type="checkbox"/>	h. Automatic Organ Contouring
<input checked="" type="checkbox"/>	i. Semiautomatic Organ Contouring
<input checked="" type="checkbox"/>	j. Input Prescription Limitations
<input checked="" type="checkbox"/>	k. Real-time Plan Adjustment and Optimization
<input checked="" type="checkbox"/>	l. Composite Modality Planning

5. Workstation Requirements

<input checked="" type="checkbox"/>	a. Minimum number of dosimetry workstations	4
<input checked="" type="checkbox"/>	b. Minimum dosimetry workstation monitor size [in]	30
<input checked="" type="checkbox"/>	c. Minimum dosimetry workstation hard drive size [GB]	250
<input checked="" type="checkbox"/>	d. Minimum dosimetry workstation RAM size [GB]	8
<input checked="" type="checkbox"/>	e. Minimum number of physician workstations	2
<input checked="" type="checkbox"/>	f. Minimum physician workstation monitor size [in]	30
<input checked="" type="checkbox"/>	g. Minimum physician workstation hard drive size [GB]	250
<input checked="" type="checkbox"/>	h. Minimum physician workstation RAM size [GB]	8



<input checked="" type="checkbox"/>	i. Minimum number of physicist workstations	2
<input checked="" type="checkbox"/>	j. Minimum physicist workstation monitor size [in]	30
<input checked="" type="checkbox"/>	k. Minimum physicist workstation hard drive size [GB]	250
<input checked="" type="checkbox"/>	l. Minimum physicist workstation RAM size [GB]	8
<input checked="" type="checkbox"/>	m. Minimum number of resident workstations	1
<input checked="" type="checkbox"/>	n. Minimum resident workstation monitor size [in]	30
<input checked="" type="checkbox"/>	o. Minimum resident workstation hard drive size [GB]	250
<input checked="" type="checkbox"/>	p. Minimum resident workstation RAM size [GB]	8
<input checked="" type="checkbox"/>	q. LCD monitor(s)	
<input checked="" type="checkbox"/>	r. Solid state hard drive solution on all workstations	
<input checked="" type="checkbox"/>	s. Dual video output on all workstations	

6. *Security/Connectivity Requirements*

<input checked="" type="checkbox"/>	m. OEM-supported operating system
<input checked="" type="checkbox"/>	n. Latest DICOM print, store, commit, and modality worklist
<input checked="" type="checkbox"/>	o. Encrypted hard drive
<input checked="" type="checkbox"/>	p. Oncology Information System compatibility – Varian Aria
<input checked="" type="checkbox"/>	q. Remote diagnostic service availability

7. *Added Value*

Specifications listed below are not required, but preferred. Vendors who do not include the below specifications in the submitted offer will not be docked or excluded from consideration. Specifications listed below will be evaluated based on added value.

<input checked="" type="checkbox"/>	a. Additional year(s) of warranty
<input checked="" type="checkbox"/>	b. Post-warranty remote diagnostic service program
<input checked="" type="checkbox"/>	c. Version/platform long-range plan
<input checked="" type="checkbox"/>	d. OEM-supported operating system through 2022
<input checked="" type="checkbox"/>	e. Compatibility with Cerner EMR
<input checked="" type="checkbox"/>	f. Ability to complete Secondary Calculation check
<input checked="" type="checkbox"/>	g. Printer for workstations

C. TRAINING REQUIREMENTS

1. *Clinical Training*

<input checked="" type="checkbox"/>	a. On-site clinical applications training for 2 dosimetrists and 2 physicists during go-live
<input checked="" type="checkbox"/>	b. On-site clinical applications training for 2 physicians during go-live
<input checked="" type="checkbox"/>	c. On-site clinical applications training for 6 total physicians, dosimetrists, and physicists post go-live, not to exceed 1 year.



<input checked="" type="checkbox"/>	d. Off-site advanced planning training for 2 dosimetrists
<input checked="" type="checkbox"/>	e. Off-site physics modeling training for 2 physicists
<input checked="" type="checkbox"/>	f. Technologists who complete the clinical applications training shall receive continuing education credits (CMEs).
<input checked="" type="checkbox"/>	g. Vendors shall be responsible for accommodating different personnel shifts for clinical applications training during go-live.

2. Biomedical Technician Training

Please reference the “Instructions to Offers” section 2.8.g for further information about the type of information to provide by equipment type not by specific request. Please also reference the “Instructions to Offers” section 7.3.3. for response format.

Technical training information to include detailed information about the curriculum and length of the biomedical technical training required for each equipment type.

Although the NAC will not award this training along with the equipment, it is imperative that the customer is informed that this training is available. Vendors must demonstrate that they can provide any required off-site training, therefore off-site training should be quoted as an optional item. Off-site training will be purchased at the time of need via a modification (if the original order remains open) or via a separate order. No travel expenses for any VA employees will be included in any HTME equipment or training order.

D. SERVICE REQUIREMENTS

1. VPN/Remote Access – The vendor shall provide any and all equipment service programs, such as remote diagnostics, during the warranty period. The vendor shall provide post-warranty remote diagnostic service program as an “Add Option” with the offer. The system shall provide vendor remote diagnostics via VPN. The vendor shall either utilize the VA national site-to-site VPN or work with the Office of Cyber and Information Security and the VAMC Information Systems Security Officer to establish a client-based VPN.
2. Service and Operator Manuals – The vendor shall provide the following documentation for the proposed systems:
 - a. Two (2) copies of operator instruction manuals (one (1) electronic and one (1) physical copy)
 - b. Two (2) copies of a service manuals (one (1) electronic and one (1) physical copy)

*Vendors can include the physical copy as a priced line item in their quote as applicable.
3. Minimum Warranty – The system and accessories shall be covered under the manufacturer’s warranty and shall include all parts and labor for one year following acceptance by the VAMC. This warranty must include PMs as required by the manufacturer. The manufacturer’s factory-trained field service representatives shall perform installation and maintenance during the warranty period.

Vendors are encouraged to include any offerings for service, warranty, and training that may exceed the minimum requirements, to include information on their service support structure during and after the warranty period. Vendors who do not include any added value offerings for service, warranty, and training will not be docked or excluded from consideration. However, any such offerings will be evaluated based on added value.



E. OTHER INFORMATION/DOCUMENTATION REQUESTED

1. Please reference the “Instructions to Offers” section 2.8a-h for further information about the type of information to provide by equipment type not by specific request. Please also reference the “Instructions to Offers” section 7.3.3. for response format.
 1. Completed pre-procurement assessment form (6550 Appendix A)
 2. Completed Manufacture Disclosure Statement for Medical Device Security (MDS2) form
 3. Federal Information Processing Standard (FIPS) 140-2 certification
 4. Product brochures
 5. Technical specification sheets, to include dimensions and weight of the system
 6. Typical drawings (pdf version of the CAD drawings)
 7. Technical training- Biomedical: information to include detailed information about the curriculum and length of the biomedical technical training required for each equipment type.
 - Although the NAC will not award this training along with the equipment, it is imperative that the customer is informed that this training is available. Vendors must demonstrate that they can provide any required off-site training, therefore off-site training should be quoted as an optional item. Off-site training will be purchased at the time of need via a modification (if the original order remains open) or via a separate order. No travel expenses for any VA employees will be included in any HTME equipment or training order.
 8. Support information to include your company’s support structure during and after the warranty period
 - On-line or telephonic applications support and availability (include third party coverage)
 - A listing of field service engineer locations and availability
 - A listing of part depots

F. TRADE-IN

- | | |
|-------------------------------------|---------------------------------------|
| <input checked="" type="checkbox"/> | a. VA has no trade-in units to offer. |
|-------------------------------------|---------------------------------------|

