Digital Radiographic Upright Unit Specifications
(Radiographic Room 8)

System with two detectors

- Fixed detector in the wall stand (Fixed detector preferred)
- At least one wireless digital detector is required
- Wireless detector on a dolly
- Detector sharing capability

Minimum 80kW generator

UPS to maintain total system functionality for 10 minutes without facility power

Vertical Wall Stand

- Motorized Tilting – define tilt range
- Motorized tracking
- Removable patient grip
- Fixed or wireless digital detector

Ceiling suspended tube carrier

Control Room Workstation

- DICOM store, storage commit, query/retrieve, worklist management
- MPPS
- Dose Reporting
- Runs on Windows 7 or latest version

Wireless detector accessories

- Detector protector for weight bearing studies
- Detachable grids (landscape and portrait)

Training

- On-site training for go-live
- On-site follow up training (64 hours)
- One (1) Biomedical Engineering Training (tuition and travel)

Desired Options

- Extended 6 month warranty