Digital Radiographic Specifications
(Radiographic Rooms 6 and 7 with Pivot Tables)

System with two detectors

- Fixed detector in the wall stand and wireless digital detector in table preferred
- At least one wireless digital detector is required
- Detector sharing capability

Minimum 80kW generator

UPS to maintain total system functionality for 10 minutes without facility power

Table

- Height adjustable – define range of movement
- Four way table float (positioning)
- Minimum patient weight in all positions, 400 lbs
- Lock position to disengage footswitch
- Fixed or wireless digital detector
- Pivoting table

Ceiling suspended tube carrier

Control Room Workstation

- DICOM store, storage commit, query/retrieve, worklist management
- MPPS
- Dose Reporting
- Runs on Windows 7 or latest version

Vertical Wall Stand

- Motorized tilting – define tilt range
- Motorized tracking
- Vertical travel
- Removable patient grip
- Fixed or wireless digital detector

Wireless detector accessories

- Detector protector for weight bearing studies
- Table fixation support for cross table lateral exams
- Detachable grids (landscape and portrait)

Training

- On-site training for go-live
- On-site follow up training (64 hours)
- One (1) Biomedical Engineering Training (tuition and travel)

Desired Options

- Extended 6 month warranty